

PART 3000

METALS

Part Coordinators

Randy A. Gottler and Marvin D. Piwoni

JOINT TASK GROUP CHAIRS

3010 Introduction	Brian J. Condike
3020 Quality Assurance/Quality Control	David W. Tucker, Cindy A. Ziernicki
3030 Preliminary Treatment of Samples	Jonathan Talbott
3113 Metals by Electrothermal Atomic Absorption Spectrometry	Raymond J. Lovett
3125 Metals by Inductively Coupled Plasma/Mass Spectrometry	William R. Kammin
3130 Metals by Anodic Stripping Voltammetry	Malgorzata Ciszewska
3500-(all)	Brian J. Condike
3500-Cr Chromium	Rock J. Vitale

SUMMARY OF MAJOR CHANGES SINCE 1998

Chromium (3500-Cr) contains updated information on the stability of hexavalent chromium. Quality Assurance/Quality Control (3020) was revised to make it consistent with current regulatory requirements.

3010 INTRODUCTION

3010 A. General Discussion

1. Significance

The effects of metals in water and wastewater range from beneficial through troublesome to dangerously toxic. Some metals are essential to plant and animal growth while others may adversely affect water consumers, wastewater treatment systems, and receiving waters. The benefits versus toxicity of some metals depend on their concentrations in waters.

2. Types of Methods

Preliminary treatment is often required to present the metals to the analytical methodology in an appropriate form. Alternative methods for pretreatment of samples are presented in Section 3030.

Metals may be determined satisfactorily by a variety of methods, with the choice often depending on the precision and sensitivity required. Part 3000 describes colorimetric methods as well as instrumental methods, i.e., atomic absorption spectrometry, including flame, electrothermal (furnace), hydride, and cold vapor techniques; flame photometry; inductively coupled plasma emission spectrometry; inductively coupled plasma mass spectrometry, and anodic stripping voltammetry. Flame atomic absorption methods generally are applicable at moderate (0.1- to 10-mg/L) concentrations in clean and complex-matrix samples. Electrothermal methods generally can increase sensitivity if matrix problems do not interfere. Inductively coupled plasma emis-

sion techniques are applicable over a broad linear range and are especially sensitive for refractory elements. Inductively coupled plasma mass spectrometry offers significantly increased sensitivity for some elements (as low as 0.01 $\mu\text{g/L}$) in a variety of environmental matrices. Flame photometry gives good results at higher concentrations for several Group I and II elements. Anodic stripping offers high sensitivity for several elements in relatively clean matrices. Colorimetric methods are applicable to specific metal determinations where interferences are known not to compromise method accuracy; these methods may provide speciation information for some metals. Table 3010:I lists the methods available in Part 3000 for each metal.

3. Definition of Terms

a. Dissolved metals: Those metals in an unacidified sample that pass through a 0.45- μm membrane filter.

b. Suspended metals: Those metals in an unacidified sample that are retained by a 0.45- μm membrane filter.

c. Total metals: The concentration of metals determined in an unfiltered sample after vigorous digestion, or the sum of the concentrations of metals in the dissolved and suspended fractions. Note that total metals are defined operationally by the digestion procedure.

d. Acid-extractable metals: The concentration of metals in solution after treatment of an unfiltered sample with hot dilute mineral acid. To determine either dissolved or suspended metals, filter sample immediately after collection. Do not preserve with acid until after filtration.

Joint Task Group: 20th Edition—Brian J. Condit (chair).

TABLE 3010:I. APPLICABLE METHODS FOR ELEMENTAL ANALYSIS

Element	Flame Atomic Absorption (Direct)	Flame Atomic Absorption (Extracted)	Flame Photometry	Electrothermal Atomic Absorption	Hydride/Cold Vapor Atomic Absorption	Inductively Coupled Plasma (ICP)	ICP/Mass Spectrometry (ICP/MS)	Anodic Stripping Voltammetry	Alternative Methods†
Aluminum	3111D	3111E		3113B		3120A	3125		3500-Al,B
Antimony	3111B			3113B		3120A	3125		
Arsenic				3113B	3114B	3120A	3125		3500-As,B
Barium	3111D	3111E		3113B		3120A	3125		
Beryllium	3111D	3111E		3113B		3120A	3125		
Bismuth	3111B			3113B			3125*		
Boron						3120A	3125*		4500-B,B,C
Cadmium	3111B	3111C		3113B		3120A	3125	3130B	
Calcium	3111B,D	3111E				3120A	3125*		3500-Ca,B
Cesium	3111B						3125*		
Chromium	3111B	3111C		3113B		3120A	3125		3500-Cr,B,C
Cobalt	3111B	3111C		3113B		3120A	3125		
Copper	3111B	3111C		3113B		3120A	3125		3500-Cu,B,C
Gallium				3113B			3125*		
Germanium				3113B			3125*		
Gold	3111B			3113B			3125*		
Indium				3113B			3125*		
Iridium	3111B						3125*		

TABLE 3010:I. CONT.

Element	Flame Atomic Absorption (Direct)	Flame Atomic Absorption (Extracted)	Flame Photometry	Electrothermal Atomic Absorption	Hydride/Cold Vapor Atomic Absorption	Inductively Coupled Plasma (ICP)	ICP/Mass Spectrometry (ICP/MS)	Anodic Stripping Voltammetry	Alternative Methods†
Iron	3111B	3111C		3113B		3120A	3125*		3500-Fe.B
Lead	3111B	3111C		3113B		3120A	3125	3130B	3500-Pb.B
Lithium	3111B		3500-Li.B			3120A	3125*		
Magnesium	3111B					3120A	3125*		3500-Mg.B,C
Manganese	3111B	3111C		3113B		3120A	3125		3500-Mn.B
Mercury					3112B		3125*		
Molybdenum	3111D	3111E		3113B		3120A	3125		
Nickel	3111B	3111C		3113B		3120A	3125		
Osmium	3111D	3111E					3125*		
Palladium	3111B						3125*		
Platinum	3111B						3125*		
Potassium	3111B		3500-K.B			3120A	3125*		3500-K.C
Rhenium	3111D	3111E					3125*		
Rhodium	3111B						3125*		
Ruthenium	3111B						3125*		
Selenium				3113B	3114B.C	3120A	3125		3500-Se.C,D,E
Silicon	3111D	3111E				3120A	3125*		
Silver	3111B	3111C		3113B		3120A	3125		
Sodium	3111B		3500-Na.B			3120A	3125*		
Strontium	3111B		3500-Sr.B			3120A	3125		
Tellurium				3113B			3125*		
Thallium	3111B			3113B		3120A	3125		
Thorium	3111D	3111E					3125*		
Tin	3111B			3113B			3125*		
Titanium	3111D	3111E					3125*		
Uranium							3125		
Vanadium	3111D	3111E		3113B		3120A	3125		3500-V.B
Zinc	3111B	3111C		3113B		3120A	3125	3130B	3500-Zn.B

*Metal is not specifically mentioned in the method, but 3125 may be used successfully in most cases.

† Additional alternative methods for aluminum, beryllium, cadmium, mercury, selenium, silver, and zinc may be found in the 19th Edition of *Standard Methods*.

3010 B. Sampling and Sample Preservation

Before collecting a sample, decide what fraction is to be analyzed (dissolved, suspended, total, or acid-extractable). This decision will determine in part whether the sample is acidified with or without filtration and the type of digestion required.

Serious errors may be introduced during sampling and storage because of contamination from sampling device, failure to remove residues of previous samples from sample container, and loss of metals by adsorption on and/or precipitation in sample container caused by failure to acidify the sample properly.

1. Sample Containers

The best sample containers are made of quartz or TFE. Because these containers are expensive, the preferred sample container is made of polypropylene or linear polyethylene with a polyethylene cap. Borosilicate glass containers also may be used, but avoid soft glass containers for samples containing metals in the microgram-per-liter range. Store samples for determination of silver in light-absorbing containers. Use only containers and filters that have been acid rinsed.

2. Preservation

Preserve samples immediately after sampling by acidifying with concentrated nitric acid (HNO₃) to pH < 2. Filter samples for dissolved metals before preserving (see Section 3030). Usually 1.5 mL conc HNO₃/L sample (or 3 mL 1 + 1 HNO₃/L sample) is sufficient for short-term preservation. For samples with high buffer capacity, increase amount of acid (5 mL may be required for some alkaline or highly buffered samples). Use commercially available high-purity acid* or prepare high-purity acid by sub-boiling distillation of acid.

After acidifying sample, preferably store it in a refrigerator at approximately 4°C to prevent change in volume due to evaporation. Under these conditions, samples with metal concentrations of several milligrams per liter are stable for up to 6 months (except mercury, for which the limit is 5 weeks). For microgram-per-liter metal levels, analyze samples as soon as possible after sample collection.

* Ultrex, J.T. Baker, or equivalent.

Alternatively, preserve samples for mercury analysis by adding 2 mL/L 20% (w/v) $K_2Cr_2O_7$ solution (prepared in 1 + 1 HNO_3). Store in a refrigerator not contaminated with mercury. (CAUTION: Mercury concentrations may increase in samples stored in plastic bottles in mercury-contaminated laboratories.)

3. Bibliography

- STRUEMPLER, A.W. 1973. Adsorption characteristics of silver, lead, calcium, zinc and nickel on borosilicate glass, polyethylene and polypropylene container surfaces. *Anal. Chem.* 45:2251.
- FELDMAN, C. 1974. Preservation of dilute mercury solutions. *Anal. Chem.* 46:99.

- KING, W.G., J.M. RODRIGUEZ & C.M. WAI. 1974. Losses of trace concentrations of cadmium from aqueous solution during storage in glass containers. *Anal. Chem.* 46:771.
- BATLEY, G.E. & D. GARDNER. 1977. Sampling and storage of natural waters for trace metal analysis. *Water Res.* 11:745.
- SUBRAMANIAN, K.S., C.L. CHAKRABARTI, J.E. SUETIAS & I.S. MAINES. 1978. Preservation of some trace metals in samples of natural waters. *Anal. Chem.* 50:444.
- BERMAN, S. & P. YEATS. 1985. Sampling of seawater for trace metals. *Crit. Rev. Anal. Chem.* 16:1.
- WENDLANDT, E. 1986. Sample containers and analytical accessories made of modern plastics for trace analysis. *Gewaess. Wass. Abwass.* 86:79.

3010 C. General Precautions

1. Sources of Contamination

Avoid introducing contaminating metals from containers, distilled water, or membrane filters. Some plastic caps or cap liners may introduce metal contamination; for example, zinc has been found in black bakelite-type screw caps as well as in many rubber and plastic products, and cadmium has been found in plastic pipet tips. Lead is a ubiquitous contaminant in urban air and dust.

2. Contaminant Removal

Thoroughly clean sample containers with a metal-free non-ionic detergent solution, rinse with tap water, soak in acid, and then rinse with metal-free water. For quartz, TFE, or glass materials, use 1 + 1 HNO_3 , 1 + 1 HCl, or aqua regia (3 parts conc HCl + 1 part conc HNO_3) for soaking. For plastic material, use 1 + 1 HNO_3 or 1 + 1 HCl. Reliable soaking conditions are 24 h at 70°C. Chromic acid or chromium-free substitutes* may be used to remove organic deposits from containers, but rinse

* Nochromix, Godax Laboratories, or equivalent.

containers thoroughly with water to remove traces of chromium. Do not use chromic acid for plastic containers or if chromium is to be determined. Always use metal-free water in analysis and reagent preparation (see 3111B.3c). In these methods, the word "water" means metal-free water.

3. Airborne Contaminants

For analysis of microgram-per-liter concentrations of metals, airborne contaminants in the form of volatile compounds, dust, soot, and aerosols present in laboratory air may become significant. To avoid contamination use "clean laboratory" facilities such as commercially available laminar-flow clean-air benches or custom-designed work stations and analyze blanks that reflect the complete procedure.

4. Bibliography

- MITCHELL, J.W. 1973. Ultrapurity in trace analysis. *Anal. Chem.* 45:492A.
- GARDNER, M., D. HUNT & G. TOPPING. 1986. Analytical quality control (AQC) for monitoring trace metals in the coastal and marine environment. *Water Sci. Technol.* 18:35.

3020 QUALITY ASSURANCE/QUALITY CONTROL

3020 A. Introduction

General information and recommendations for quality assurance (QA) and quality control (QC) are provided in Sections 1020 Quality Assurance, 1030 Data Quality, and 1040 Method Development and Evaluation. This section discusses QA/QC requirements that are common to the analytical methods presented in Part 3000.

The requirements described in this section are recommended minimum QA/QC activities; refer to individual methods and regulatory program requirements for method-specific QA/QC requirements. NOTE: If an individual method in Part 3000 has requirements more stringent than those shown here, follow re-

quirements of that method. If analyses are to be conducted for regulatory purposes and QA/QC requirements are set in regulations or in a reference method, follow those requirements.

Always consider the overall purpose of analyses. QA/QC measures and substantiation for operational-control determinations may differ significantly from those for determinations of trace metals at water quality criteria levels. Levels of trace metals in environmental samples may be orders of magnitude lower than in potential sources of contamination.

Use replicates of measurable concentration to establish precision and known-additions recovery to determine bias. Use blanks, calibrations, control charts, known additions, standards, and other ancillary measurement tools as appropriate. Provide adequate documentation and record keeping to satisfy client requirements and performance criteria established by the laboratory.

Joint Task Group: Cindy A. Ziernicki (chair), Katherine B. Adams, Myriam E. Cardenas, David Eugene Kimbrough; 20th Edition—David W. Tucker (chair), Nilda B. Cox, David W. Haddaway, Daniel A. McLean, Jonalea V. Ostlund, John T. Pivinski.

3020 B. Quality Control Practices

1. Initial Quality Control

a. Initial demonstration of capability: Verify analyst capability before analyzing any samples and repeat periodically to demonstrate proficiency with the analytical method. Verify that the method being used provides sufficient sensitivity for the purpose of the measurement. Test analyst capability by analyzing at least four reagent water portions containing known additions of the analyte of interest. Confirm proficiency by generating analytical results that demonstrate precision and bias within acceptable limits representative of the analytical method.

b. Method detection level (MDL): Before samples are analyzed, determine the MDL for each analyte by the procedures of Section 1030, or other applicable procedure.¹ Determine MDL at least annually for each method and major matrix category. Verify MDL for a new analyst or whenever instrument hardware or method operating conditions are modified. Analyze samples for MDL determinations over a 3- to 5-d period to generate a realistic value. Preferably use pooled data from several analysts rather than data from a single analyst.

c. Dynamic range (DR): Before using a new method, determine the dynamic range, i.e., the concentration range over which a method has an increasing response (linear or second-order), for each analyte by analyzing several standard solutions that bracket the range of interest. Each standard measurement should be within 10% of the true value for acceptance into the DR determination. Take measurements at both the low and high end of the calibration range to determine method suitability. Analytical instrumentation with curve-fitting features may allow utilization of nonlinear instrument response.

2. Calibration

a. Initial calibration: Calibrate initially with a minimum of a blank and three calibration standards of the analyte(s) of interest. Select calibration standards that bracket the expected concentration of the sample and that are within the method's dynamic range. The number of calibration points depends on the width of the dynamic range and the shape of the calibration curve. One calibration standard should be at or below the reporting limit for the method. As a general rule, differences between calibration standard concentrations should not be greater than one order of magnitude (i.e., 1, 10, 100, 1000). Apply linear or polynomial curve-fitting statistics, as appropriate, for analysis of the concentration-instrument response relationship. The appropriate linear or nonlinear correlation coefficient for standard concentration to instrument response should be ≥ 0.995 . Use initial calibration for quantitation of analyte concentration in samples. Use calibration verification, ¶ *b* below, only for checks on the initial calibration and not for sample quantitation. Repeat initial calibration daily and whenever calibration verification acceptance criteria are not satisfied.

b. Calibration verification: Calibration verification is the periodic confirmation that instrument response has not changed significantly from the initial calibration. Verify calibration by analyzing a midpoint calibration standard (check standard) and calibration blank at the beginning and end of a sample run, periodically during a run (normally after each set of ten samples). A check standard determination outside 90 to 110% of the expected concentration indicates a potential problem. If a check standard determination is outside 80 to 120% of the expected concentration, immediately cease sample analyses and initiate cor-

rective action. For instrumental methods (3111, 3113, 3120, and 3125), cease analysis and initiate corrective action if check standards exceed 90 to 110%. Repeat initial calibration and sample determinations since the last acceptable calibration verification. Use calculated control limits (Section 1020B) to provide better indications of system performance and to provide tighter limits.

c. Quality control sample: Analyze an externally generated quality control sample of known concentration at least quarterly and whenever new calibration stock solutions are prepared. Obtain this sample from a source external to the laboratory or prepare it from a source different from those used to prepare working standards. Use to validate the laboratory's working standards both qualitatively and quantitatively.

3. Batch Quality Control

a. Method blank (MB): A method blank (also known as reagent blank) is a portion of reagent water treated exactly as a sample, including exposure to all equipment, glassware, procedures, and reagents. The MB is used to assess whether analytes or interference are present within the analytical process or system. No analyte of interest should be present in the MB at a warning level based on the end user's requirements. Undertake immediate corrective action for MB measurements above the MDL. Include a minimum of one MB with each set of 20 or fewer samples.

b. Laboratory-fortified blank (LFB): The laboratory-fortified blank (also known as blank spike) is a method blank that has been fortified with a known concentration of analyte. It is used to evaluate ongoing laboratory performance and analyte recovery in a clean matrix. Prepare fortified concentrations approximating the midpoint of the calibration curve or lower with stock solutions prepared from a source different from those used to develop working standards. Calculate percent recovery, plot control charts, and determine control limits (Section 1020B) for these measurements. Ensure that the LFB meets performance criteria for the method when such criteria are specified. Establish

corrective actions to be taken in the event that LFB does not satisfy acceptance criteria. Include a minimum of one LFB with each set of 20 or fewer samples.

c. Duplicates: Use duplicate samples of measurable concentration to measure precision of the analytical process. Randomly select routine samples to be analyzed twice. Process duplicate sample independently through entire sample preparation and analytical process. Include a minimum of one duplicate for each matrix type with each set of 20 or fewer samples.

d. Laboratory-fortified matrix (LFM)/Laboratory-fortified matrix duplicate: Use LFM (also known as matrix spike) and LFM duplicate to evaluate the bias and precision, respectively, of the method as influenced by a specific matrix. Prepare by adding a known concentration of analytes to a randomly selected routine sample. Prepare addition concentrations to approximately double the concentration present in the original sample. If necessary, dilute sample to bring the measurement within the established calibration curve. Limit addition volume to 5% or less of sample volume. Calculate percent recovery and relative percent difference, plot control charts, and determine control limits (Section 1020B). Ensure the performance criteria for the method are satisfied. Process fortified samples independently through entire sample preparation and analytical process. Include a minimum of one LFM/LFM duplicate with each set of 20 or fewer samples.

e. Method of known additions: To analyze a new or unfamiliar matrix, use the method of known additions (Section 1020B) to demonstrate freedom from interference before reporting concentration data for the analyte. Verify absence of interferences by analyzing such samples undiluted and in a 1:10 dilution; results should be within 10% of each other. Limit known-addition volume to 10% or less of the sample volume.

4. Reference

1. U.S. ENVIRONMENTAL PROTECTION AGENCY. 1995. Definition and procedure for the determination of the method detection limit, revision 1.11. 40 CFR Part 136, Appendix B. *Federal Register* 5:23703.

3030 PRELIMINARY TREATMENT OF SAMPLES*

3030 A. Introduction

Samples containing particulates or organic material generally require pretreatment before spectroscopic analysis. "Total metals" includes all metals, inorganically and organically bound, both dissolved and particulate. Colorless, transparent samples (primarily drinking water) having a turbidity of <1 NTU, no odor, and single phase may be analyzed directly by atomic absorption spectroscopy (flame or electrothermal vaporization) or inductively coupled

plasma spectroscopy (atomic emission or mass spectrometry) for total metals without digestion. For further verification or if changes in existing matrices are encountered, compare digested and undigested samples to ensure comparable results. On collection, acidify such samples to pH <2 with conc nitric acid (1.5 mL HNO₃/L is usually adequate for drinking water) and analyze directly. Digest all other samples before determining total metals. To analyze for dissolved metals, filter sample, acidify filtrate, and store until analyses can be performed. To determine suspended metals, filter sample, digest filter and the material on it, and analyze. To determine acid-extractable metals, extract metals as indicated in Sections 3030E through K and analyze extract.

* Approved by Standard Methods Committee, 1997.

Joint Task Group: 20th Edition—Jonathan Talbott (chair), Paul R. Fritschel, Elly M. Gabrielian, David Eugene Kimbrough, H.M. Kingston, Nimi Kocherlakota, Dennis Neuin, Mark E. Tatro, Mark M. Ultis, Melissa A. Weekley, Aaron D. Weiss, Ruth E. Wolf.

This section describes general pretreatment for samples in which metals are to be determined according to Sections 3110 through 3500-Zn with several exceptions. The special digestion techniques for mercury are given in Sections 3112B.4b and c, and those for arsenic and selenium in Sections 3114 and 3500-Se.

Take care not to introduce metals into samples during preliminary treatment. During pretreatment avoid contact with rubber, metal-based paints, cigarette smoke, paper tissues, and all metal products including those made of stainless steel, galvanized metal, and brass. Conventional fume hoods can contribute significantly to sample contamination, particularly during acid digestion in open containers. Keep vessels covered with watch glasses and turn spouts away from incoming air to reduce airborne contamination. Plastic pipet tips often are contaminated

with copper, iron, zinc, and cadmium; before use soak in 2N HCl or HNO₃ for several days and rinse with deionized water. Avoid using colored plastics, which can contain metals. Use certified metal-free plastic containers and pipet tips when possible. Avoid using glass if analyzing for aluminum or silica.

Use metal-free water (see 3111B.3c) for all operations. Check reagent-grade acids used for preservation, extraction, and digestion for purity. If excessive metal concentrations are found, purify the acids by distillation or use ultra-pure acids. Inductively coupled plasma mass spectrometry (ICP-MS) may require use of ultra-pure acids and reagents to avoid measurable contamination. Process blanks through all digestion and filtration steps and evaluate blank results relative to corresponding sample results. Either apply corrections to sample results or take other corrective actions as necessary or appropriate.

3030 B. Filtration for Dissolved and Suspended Metals

1. Filtration Procedures

If dissolved or suspended metals (see Section 3010A) are to be determined, filter sample at time of collection using a preconditioned plastic filtering device with either vacuum or pressure, containing a filter support of plastic or fluorocarbon, through a prewashed ungridded 0.4- to 0.45- μ m-pore-diam membrane filter (polycarbonate or cellulose esters). Before use filter a blank consisting of metal-free (deionized) water to insure freedom from contamination. Precondition filter and filter device by rinsing with 50 mL deionized water. If the filter blank contains significant metals concentrations, soak membrane filters in approximately 0.5N HCl or 1N HNO₃ (recommended for electrothermal and ICP-MS analyses) and rinse with deionized water before use.

Before filtering, centrifuge highly turbid samples in acid-washed fluorocarbon or high-density plastic tubes to reduce loading on filters. Stirred, pressure filter units foul less readily than vacuum filters; filter at a pressure of 70 to 130 kPa. After filtration acidify filtrate to pH 2 with conc HNO₃ and store until analyses can be performed. If a precipitate forms on acidification, digest acidified filtrate before analysis as directed (see Section 3030E). Retain filter and digest it for direct determination of suspended metals.

If it is not possible to field-filter the sample without contaminating it, obtain sample in an "unpreserved" bottle as above and

promptly cool to 4°C. Do not acid-preserve the sample. Then, without delay, filter sample under cleaner conditions in the laboratory.

Test pH of a portion of aqueous sample upon receipt in the laboratory to ensure that the sample has been properly filtered and acid-preserved.¹

NOTE: Different filters display different sorption and filtration characteristics²; for trace analysis, test filter and filtration system to verify complete recovery of metals.

If suspended metals (see Section 3010A) are to be determined, filter sample as above for dissolved metals, but do not centrifuge before filtration. Retain filter and digest it for direct determination of suspended metals. Record sample volume filtered and include a filter in determination of the blank.

CAUTION: Do not use perchloric acid to digest membrane filters. (See 3030H for more information on handling HClO₄).

2. References

1. U.S. ENVIRONMENTAL PROTECTION AGENCY. 1994. Sample Preparation Procedure for Spectrochemical Determination of Total Recoverable Elements, Method 200.2. Environmental Monitoring Systems Lab., Cincinnati, Ohio.
2. HOROWITZ, A.J., K.R. LUM, J.R. GARBARINO, G.E.M. HALL, C. LEMIEUX & C.R. DEMAS. 1996. Problems with using filtration to define dissolved trace element concentrations in natural water samples. *Environ. Sci. Technol.* 30: 954.

3030 C. Treatment for Acid-Extractable Metals

Extractable metals (see Section 3010A) are lightly adsorbed on particulate material. Because some sample digestion may be unavoidable use rigidly controlled conditions to obtain meaningful and reproducible results. Maintain constant sample volume,

acid volume, and contact time. Express results as extractable metals and specify extraction conditions.

At collection, acidify entire sample with 5 mL conc HNO₃/L sample. To prepare sample, mix well, transfer 100 mL to a

beaker or flask, and add 5 mL 1 + 1 high-purity HCl. Heat 15 min on a steam bath. Filter through a membrane filter (preconditioned as in Section 3030B) and carefully transfer filtrate to a tared volumetric flask. Adjust volume to 100 mL with metal-free

water, mix, and analyze. If volume is greater than 100 mL, determine volume to nearest 0.1 mL by weight, analyze, and correct final concentration measurement by multiplying by the dilution factor (final volume ÷ 100).

3030 D. Digestion for Metals

To reduce interference by organic matter and to convert metals associated with particulates to a form (usually the free metal) that can be determined by atomic absorption spectrometry or inductively-coupled plasma spectroscopy, use one of the digestion techniques presented below. Use the least rigorous digestion method required to provide acceptable and consistent recovery compatible with the analytical method and the metal being analyzed.¹⁻³

1. Selection of Acid

Nitric acid will digest most samples adequately (Section 3030E). Nitrate is an acceptable matrix for both flame and electrothermal atomic absorption and the preferred matrix for ICP-MS.⁴ Some samples may require addition of perchloric, hydrochloric, hydrofluoric, or sulfuric acid for complete digestion. These acids may interfere in the analysis of some metals and all provide a poorer matrix for both electrothermal and ICP-MS analysis. Confirm metal recovery for each digestion and analytical procedure used. Use Table 3030:I as a guide in determining which acids (in addition to HNO₃) to use for complete digestion. As a general rule, HNO₃ alone is adequate for clean samples or easily oxidized materials; HNO₃-H₂SO₄ or HNO₃-HCl digestion is adequate for readily oxidizable organic matter; HNO₃-HClO₄ or HNO₃-HClO₄-HF digestion is necessary for difficult-to-oxidize organic matter or minerals containing silicates. Although dry ashing is not generally recommended because of the loss of many volatile elements, it may be helpful if large amounts of organic matter are present.

2. Digestion Procedures

Dilute samples with Ag concentrations greater than 1 mg/L to contain less than 1 mg Ag/L for flame atomic absorption meth-

ods and 25 µg/L or less for electrothermal analysis.^{2,5,6} To address problems with silver halide solubility in HNO₃, digest using method 3030F.3b.

Report digestion technique used.

Acid digestion techniques (Sections 3030E through I) generally yield comparable precision and bias for most sample types that are totally digested by the technique. Because acids used in digestion will add metals to the samples and blanks, minimize the volume of acids used.

Because the acid digestion techniques (3030E and F) normally are not total digestions, the microwave digestion procedure (3030K) may be used as an alternative. The microwave method is a closed-vessel procedure and thus is expected to provide improved precision when compared with hot-plate techniques. Microwave digestion is recommended for samples being analyzed by ICP-MS. The microwave digestion method is recommended for the analysis of Ag, Al, As, Ba, Be, Ca, Cd, Co, Cr, Cu, Fe, K, Mg, Mn, Mo, Na, Ni, Pb, Sb, Se, Tl, V, and Zn. Microwave digestion may be acceptable for additional analytes provided its performance for those elements is validated.

Suggested sample volumes are indicated below for flame atomic absorption spectrometry. Lesser volumes, to a minimum of 5 mL, are appropriate for graphite furnace, ICP, and ICP-MS. Do not subsample volumes less than 5 mL, especially when particulates are present. Instead dilute samples with elevated analyte concentrations after digestion. If the recommended volume exceeds digestion vessel capacity, add sample as evaporation proceeds. For samples containing particulates, wide-bore pipets may be useful for volume measurement and transfer.

When samples are concentrated during digestion (e.g., >100 mL sample used) determine metal recovery for each matrix digested, to verify method validity. Using larger samples will require additional acid, which also would increase the concentration of impurities.

TABLE 3030:I. ACIDS USED IN CONJUNCTION WITH HNO₃ FOR SAMPLE PREPARATION

Acid	Recommended for	May Be Helpful for	Not Recommended for
HCl	Ag	Sb, Ru, Sn	Th, Pb
H ₂ SO ₄	Ti	—	Ag, Pb, Ba
HClO ₄	—	Organic materials	—
HF	—	Siliceous materials	—

Estimated Metal Concentration mg/L	Sample Volume* mL
<0.1	1000
0.1-10	100
10-100+	10

*For flame atomic absorption spectrometry.

Report results as follows:

$$\text{Metal concentration, mg/L} = A \times \frac{B}{C}$$

where:

A = concentration of metal in digested solution, mg/L,

B = final volume of digested solution, mL, and

C = sample size, mL.

Prepare solid samples or liquid sludges with high solids contents on a weight basis. Mix sample and transfer a suitable amount (typically 1 g of a sludge with 15% total solids) directly into a preweighed digestion vessel. Reweigh and calculate weight of sample. Proceed with one of the digestion techniques presented below. However, as these digestion methods are predominantly for dissolved and extractable metals in aqueous samples, other approaches may be more appropriate for solid samples. For complete mineralization of solid samples, consult methods available elsewhere.^{1,4,6,7} Report results on wet- or dry-weight basis as follows:

$$\text{Metal concentration, mg/kg (wet-weight basis)} = \frac{A \times B}{\text{g sample}}$$

$$\text{Metal concentration, mg/kg (dry-weight basis)} = \frac{A \times B}{\text{g sample}} \times \frac{100}{D}$$

where:

A = concentration of metal in digested solution, mg/L,

B = final volume of digested solution, mL, and

D = total solids, % (see Section 2540G).

Always prepare acid blanks for each type of digestion performed. Although it is always best to eliminate all relevant

sources of contamination, a reagent blank prepared with the same acids and subjected to the same digestion procedure as the sample can correct for impurities present in acids and reagent water. However, blank correction is not recommended for any other sources of contamination such as impurities adsorbed on glassware.

3. References

1. BOUMANS, P.W.J.M., ed. 1987. Inductively Coupled Plasma Emission Spectroscopy, Part II: Applications and Fundamentals. John Wiley & Sons, New York, N.Y.
2. U.S. ENVIRONMENTAL PROTECTION AGENCY. 1992. Test Methods for Evaluating Solid Waste, Physical/Chemical Methods, SW-846, 3rd ed. Update 1, Methods 3005A, 3010A, 3020A & 3050A. Off. Solid Waste & Emergency Response, Washington, D.C.
3. HOENIG, M. & A.M. DE KERSABIEC. 1996. Sample preparation steps for analysis by atomic spectroscopy methods: Present status. *Spectrochim. Acta.* B51:1297.
4. JARVIS, K.E., A.L. GRAY & R.S. HOUK, eds. 1992. Sample preparation for ICP-MS. Chapter 7 in Handbook of Inductively Coupled Plasma Mass Spectrometry. Blackie, Glasgow & London, U.K.
5. U.S. ENVIRONMENTAL PROTECTION AGENCY. 1994. Sample Preparation Procedure for Spectrochemical Determination of Total Recoverable Elements, Method 200.2. Environmental Monitoring Systems Lab., Cincinnati, Ohio.
6. KINGSTON, H.M. & S. HASWELL, eds. 1997. Microwave Enhanced Chemistry: Fundamentals, Sample Preparation and Applications. American Chemical Soc., Washington, D.C.
7. BOCK, R. 1979. A Handbook of Decomposition Methods in Analytical Chemistry. Blackie, Glasgow, U.K.

3030 E. Nitric Acid Digestion

Because of the wide variation in concentration levels detected by various instrumental techniques and the need to deal adequately with sources of contamination at trace levels, this method presents one approach for high-level analytes (>0.1 mg/L) and another for trace levels (≤0.1 mg/L).

1. Digestion for Flame Atomic Absorption and High-Level Concentrations

a. Apparatus:

- 1) Hot plate.
- 2) Conical (erlenmeyer) flasks, 125-mL, or Griffin beakers, 150-mL, acid-washed and rinsed with water.
- 3) Volumetric flasks, 100-mL.
- 4) Watch glasses, ribbed and unribbed.

b. Reagent:

Nitric acid, HNO₃, conc, analytical or trace-metals grade.

c. Procedure: Transfer a measured volume (100 mL recommended) of well-mixed, acid-preserved sample appropriate for the expected metals concentrations to a flask or beaker (see 3030D for sample volume). In a hood, add 5 mL conc HNO₃. If a beaker is used, cover with a ribbed watch glass to minimize contamination. Boiling chips, glass beads, or Hengar granules may be added to aid boiling and minimize spatter when high

concentration levels (>10 mg/L) are being determined. Bring to a slow boil and evaporate on a hot plate to the lowest volume possible (about 10 to 20 mL) before precipitation occurs. Continue heating and adding conc HNO₃ as necessary until digestion is complete as shown by a light-colored, clear solution. Do not let sample dry during digestion.

Wash down flask or beaker walls and watch glass cover (if used) with metal-free water and then filter if necessary (see Section 3030B). Transfer filtrate to a 100-mL volumetric flask with two 5-mL portions of water, adding these rinsings to the volumetric flask. Cool, dilute to mark, and mix thoroughly. Take portions of this solution for required metal determinations.

2. Digestion for Trace-Level (≤0.1 mg/L) Concentrations for ICP and ICP-MS¹

a. Apparatus:

- 1) Block heater, dry, with temperature control.
- 2) Polypropylene tubes*, graduated, round-bottom tubes with caps, 17 × 100 mm, acid-washed and rinsed with metal-free water. Preferably use tubes that simultaneously match the anal-

* Falcon tubes or equivalent.

ysis instrument autosampler and the block digester. A fit with the centrifuge is secondary but also desirable.

3) *Pipettors*, assorted sizes or adjustable.

4) *Pipet tips*.

5) *Centrifuge*.

b. *Reagent*:

Nitric acid, HNO_3 , conc, double distilled.†

c. *Procedure*: Soak new polypropylene tubes and caps overnight or for several days in 2N HNO_3 . Triple rinse with metal-free water, and preferably dry in poly racks or baskets in a low-temperature oven overnight. Store cleaned tubes in plastic bags before use. Pipet tips also may need to be cleaned; evaluate before use.

Pipet 10 mL well-mixed, acid-preserved sample into a pre-cleaned, labeled tube with a macropipet. With a minimum volume change (<0.5 mL), add appropriate amount of analyte for matrix fortified samples. With a pipet, add 0.5 mL conc HNO_3

(or 1.0 mL 1 + 1 HNO_3) to all samples, blanks, standards, and quality control samples.

Place tubes in block heater in a hood and adjust temperature to 105°C. Drape caps over each tube to allow escape of acid vapors while preventing contamination. NOTE: Do not screw on caps at this time. Digest samples for a minimum of 2 h. Do not let samples boil. Add more conc nitric acid as necessary until digestion is complete by observation of a clear solution.

Remove tubes from heat and cool. Dilute back to original 10 mL volume with metal-free water. Adjust over-volume samples to next convenient gradation for calculations and note volume. (Apply concentration correction from Section 3030D.) If tubes contain particulates, centrifuge and decant clear portion into another pre-cleaned tube. Tighten screw caps and store at 4°C until ready for analysis.

3. Reference

- JARVIS, K.E., A.L. GRAY & R.S. HOUK, eds. 1992. Sample preparation for ICP-MS. Chapter 7 in Handbook of Inductively Coupled Plasma Mass Spectrometry. Blackie & Son, Ltd., Glasgow & London, U.K.

† Ultrex, Optima grade or equivalent.

3030 F. Nitric Acid-Hydrochloric Acid Digestion

1. Apparatus

See 3030E.1a. The following also may be needed:

Steam bath.

2. Reagents

a. *Nitric acid*, HNO_3 , conc, analytical grade or better (see Section 3030E).

b. *Hydrochloric acid*, HCl, 1 + 1.

c. *Nitric acid*, HNO_3 , 1 + 1.

3. Procedure

a. *Total HNO_3/HCl* : Transfer a measured volume of well-mixed, acid-preserved sample appropriate for the expected metals concentrations to a flask or beaker (see 3030D for sample volume). In a hood add 3 mL conc HNO_3 and cover with a ribbed watch glass. Place flask or beaker on a hot plate and cautiously evaporate to less than 5 mL, making certain that sample does not boil and that no area of the bottom of the container is allowed to go dry. Cool. Rinse down walls of beaker and watch glass with a minimum of metal-free water and add 5

mL conc HNO_3 . Cover container with a nonribbed watch glass and return to hot plate. Increase temperature of hot plate so that a gentle reflux action occurs. Continue heating, adding additional acid as necessary, until digestion is complete (generally indicated when the digestate is light in color or does not change in appearance with continued refluxing). Cool. Add 10 mL 1 + 1 HCl and 15 mL water per 100 mL anticipated final volume. Heat for an additional 15 min to dissolve any precipitate or residue. Cool, wash down beaker walls and watch glass with water, filter to remove insoluble material that could clog the nebulizer (see Section 3030B), and transfer filtrate to a 100-mL volumetric flask with rinsings. Alternatively centrifuge or let settle overnight. Adjust to volume and mix thoroughly.

b. *Recoverable HNO_3/HCl* : For this less rigorous digestion procedure, transfer a measured volume of well-mixed, acid-preserved sample to a flask or beaker. Add 2 mL 1 + 1 HNO_3 and 10 mL 1 + 1 HCl and cover with a ribbed watch glass. Heat on a steam bath or hot plate until volume has been reduced to near 25 mL, making certain sample does not boil. Cool and filter to remove insoluble material or alternatively centrifuge or let settle overnight. Quantitatively transfer sample to volumetric flask, adjust volume to 100 mL, and mix.

For trace-level digestion, use precautionary measures similar to those detailed in Section 3030E.

3030 G. Nitric Acid-Sulfuric Acid Digestion

1. Apparatus

See 3030E.1a.

2. Reagents

- Nitric acid, HNO_3 , conc. (See 3030E for acid grades.)
- Sulfuric acid, H_2SO_4 , conc.

3. Procedure

Transfer a measured volume of well-mixed, acid-preserved sample appropriate for the expected metals concentrations to a flask or beaker (see 3030D for sample volume). Add 5 mL

conc HNO_3 and cover with a ribbed watch glass. Bring to slow boil on hot plate and evaporate to 15 to 20 mL. Add 5 mL conc HNO_3 and 10 mL conc H_2SO_4 , cooling flask or beaker between additions. Evaporate on a hot plate until dense white fumes of SO_3 just appear. If solution does not clear, add 10 mL conc HNO_3 and repeat evaporation to fumes of SO_3 . Heat to remove all HNO_3 before continuing treatment. All HNO_3 will be removed when the solution is clear and no brownish fumes are evident. Do not let sample dry during digestion.

Cool and dilute to about 50 mL with water. Heat to almost boiling to dissolve slowly soluble salts. Filter if necessary, then complete procedure as directed in Section 3030E.1c beginning with, "Transfer filtrate . . ."

3030 H. Nitric Acid-Perchloric Acid Digestion

1. Apparatus

See 3030E.1a. The following also are needed:

- Safety shield.
- Safety goggles.
- Watch glasses.

2. Reagents

- Nitric acid, HNO_3 , conc.
- Perchloric acid, HClO_4 .
- Ammonium acetate solution: Dissolve 500 g $\text{NH}_4\text{C}_2\text{H}_3\text{O}_2$ in 600 mL water.

3. Procedure

CAUTION: Heated mixtures of HClO_4 and organic matter may explode violently. Avoid this hazard by taking the following precautions: (a) do not add HClO_4 to a hot solution containing organic matter; (b) always pretreat samples containing organic matter with HNO_3 before adding HClO_4 ; (c) avoid repeated fuming with HClO_4 in ordinary hoods (For routine operations, use a water pump attached to a glass fume eradicator. Stainless steel fume hoods with adequate water washdown facilities are available commercially and are acceptable for use with HClO_4);

and (d) never let samples being digested with HClO_4 evaporate to dryness.

Transfer a measured volume of well-mixed, acid-preserved sample appropriate for the expected metals concentrations to a flask or beaker (see 3030D for sample volume). In a hood add 5 mL conc HNO_3 and cover with a ribbed watch glass. Evaporate sample to 15 to 20 mL on a hot plate. Add 10 mL each of conc HNO_3 and HClO_4 , cooling flask or beaker between additions. Evaporate gently on a hot plate until dense white fumes of HClO_4 just appear. If solution is not clear, keep solution just boiling until it clears. If necessary, add 10 mL conc HNO_3 to complete digestion. Cool, dilute to about 50 mL with water, and boil to expel any chlorine or oxides of nitrogen. Filter, then complete procedure as directed in 3030E.1c beginning with, "Transfer filtrate . . ."

If lead is to be determined in the presence of high amounts of sulfate (e.g., determination of Pb in power plant fly ash samples), dissolve PbSO_4 precipitate as follows: Add 50 mL ammonium acetate solution to flask or beaker in which digestion was carried out and heat to incipient boiling. Rotate container occasionally to wet all interior surfaces and dissolve any deposited residue. Reconnect filter and slowly draw solution through it. Transfer filtrate to a 100-mL volumetric flask, cool, dilute to mark, mix thoroughly, and set aside for determination of lead.

3030 I. Nitric Acid-Perchloric Acid-Hydrofluoric Acid Digestion

1. Apparatus

- Hot plate.
- TFE beakers, 250-mL, acid-washed and rinsed with water.
- Volumetric flasks, 100-mL, polypropylene or other suitable plastic.

2. Reagents

- Nitric acid, HNO_3 , conc and 1 + 1.
- Perchloric acid, HClO_4 .
- Hydrofluoric acid, HF, 48 to 51%.

3. Procedure

CAUTION: See precautions for using HClO_4 in 3030H; handle HF with extreme care and provide adequate ventilation, especially for the heated solution. Avoid all contact with exposed skin. Provide medical attention for HF burns.

Transfer a measured volume of well-mixed, acid-preserved sample appropriate for the expected metals concentrations into a

250-mL TFE beaker (see 3030D for sample volume). Evaporate on a hot plate to 15 to 20 mL. Add 12 mL conc HNO_3 and evaporate to near dryness. Repeat HNO_3 addition and evaporation. Let solution cool, add 20 mL HClO_4 and 1 mL HF, and boil until solution is clear and white fumes of HClO_4 have appeared. Cool, add about 50 mL water, filter, and proceed as directed in 3030E.1c beginning with, "Transfer filtrate . . ."

3030 J. Dry Ashing

The procedure appears in the Eighteenth Edition of *Standard Methods*. It has not been included in subsequent versions of this publication.

3030 K. Microwave-Assisted Digestion

1. Apparatus

a. **Microwave unit** with programmable power (minimum 545 W) to within ± 10 W of required power, having a corrosion-resistant, well-ventilated cavity and having all electronics protected against corrosion for safe operation. Use a unit having a rotating turntable with a minimum speed of 3 rpm to insure homogeneous distribution of microwave radiation. Use only laboratory-grade microwave equipment and closed digestion containers with pressure relief that are specifically designed for hot acid.¹

b. **Vessels:** Construction requires an inner liner of perfluoroalkoxy (PFA) Teflon™,* other TFE, or composite fluorinated polymers,† capable of withstanding pressures of at least 760 ± 70 kPa (110 ± 10 psi), and capable of controlled pressure relief at the manufacturer's maximum pressure rating.

Acid wash all digestion vessels and rinse with water (§ 2a). For new vessels or when changing between high- and low-concentration samples, clean by leaching with hot‡ hydrochloric acid (1:1) for a minimum of 2 h and then with hot nitric acid (1:1) for a minimum of 2 h; rinse with water and dry in a clean environment. Use this procedure whenever the previous use of digestion vessels is unknown or cross-contamination from vessels is suspected.

c. **Temperature feedback control system**, using shielded thermocouple, fiber-optic probe, or infrared detector.

d. **Bottles**, polyethylene, 125-mL, with caps.

e. **Thermometer**, accurate to $\pm 0.1^\circ\text{C}$.

f. **Balance**, large-capacity (1500 g), accurate to 0.1 g.

g. **Filtration or centrifuge equipment** (optional).

h. **Plastic container** with cover, 1-L, preferably made of PFA Teflon™§.

2. Reagents

a. **Metal-free water:** See Section 3111B.3c.

b. **Nitric acid**, HNO_3 , conc, sub-boiling distilled. Non-sub-boiling acids can be used if they are shown not to contribute blanks.

3. Calibration of Microwave Unit

NOTE: For microwave units equipped with temperature feedback electronic controls, calibration of the microwave unit is not required provided performance specifications can be duplicated.

For cavity-type microwave equipment, evaluate absolute power (watts) by measuring the temperature rise in 1 kg water exposed to microwave radiation for a fixed time. With this measurement, the relationship between available power (W) and the partial power setting (%) of the unit can be estimated, and any absolute power in watts may be transferred from one unit to another. The calibration format required depends on type of electronic system used by manufacturer to provide partial microwave power. Few units have an accurate and precise linear relationship between percent power settings and absorbed power. Where linear circuits have been used, determine calibration curve by a three-point calibration method; otherwise, use the multiple-point calibration method.

a. **Three-point calibration method:** Measure power at 100% and 50% power using the procedure described in § 3c and calculate power setting corresponding to required power in watts as specified in the procedure from the two-point line. Measure absorbed power at the calculated partial power setting. If the measured absorbed power does not correspond to the calculated power within ± 10 W, use the multiple-point calibration method, § 3b. Use this point periodically to verify integrity of calibration.

b. **Multiple-point calibration method:** For each microwave unit, measure the following power settings: 100, 99, 98, 97, 95, 90, 80, 70, 60, 50, and 40% using the procedure described in § 3c. These data are clustered about the customary working power ranges. Nonlinearity commonly is encountered at the upper end of the calibration curve. If the unit's electronics are known to have nonlinear deviations in any region of proportional power control, make a set of measurements that bracket the power to be used. The final calibration point should be at the partial power setting that will be used in the test. Check this setting periodi-

* Or equivalent.

† Such as TFM™ or equivalent.

‡ At temperatures greater than 80°C , but not boiling.

§ Or equivalent.

cally to evaluate the integrity of the calibration. If a significant change (± 10 W) is detected, re-evaluate entire calibration.

c. Equilibrate a large volume of water to room temperature ($23 \pm 2^\circ\text{C}$). Weigh 1 kg water ($1000 \text{ g} \pm 1 \text{ g}$) or measure ($1000 \text{ mL} \pm 1 \text{ mL}$) into a plastic, not glass, container, and measure the temperature to $\pm 0.1^\circ\text{C}$. Condition microwave unit by heating a glass beaker with 500 to 1000 mL tap water at full power for 5 min with the exhaust fan on. Loosely cover plastic container to reduce heat loss and place in normal sample path (at outer edge of rotating turntable); circulate continuously through the microwave field for 120 s at desired power setting with exhaust fan on as it will be during normal operation. Remove plastic container and stir water vigorously. Use a magnetic stirring bar inserted immediately after microwave irradiation; record maximum temperature within the first 30 s to $\pm 0.1^\circ\text{C}$. Use a new sample for each additional measurement. If the water is reused, return both water and beaker to $23 \pm 2^\circ\text{C}$. Make three measurements at each power setting. When any part of the high-voltage circuit, power source, or control components in the unit have been serviced or replaced, recheck calibration power. If power output has changed by more than ± 10 W, re-evaluate entire calibration.

Compute absorbed power by the following relationship:

$$P = \frac{(K)(Cp)(m)(\Delta T)}{t}$$

where:

P = apparent power absorbed by sample, W,

K = conversion factor for thermochemical calories sec^{-1} to watts (4.184),

Cp = heat capacity, thermal capacity, or specific heat ($\text{cal g}^{-1} \text{ } ^\circ\text{C}^{-1}$) of water,

m = mass of water sample, g,

ΔT = final temperature minus initial temperature, $^\circ\text{C}$, and

t = time, s.

For the experimental conditions of 120 s and 1 kg water (Cp at $25^\circ\text{C} = 0.9997$), the calibration equation simplifies to:

$$P = (\Delta T)(34.85)$$

Stable line voltage within the manufacturer's specification is necessary for accurate and reproducible calibration and operation. During measurement and operation it must not vary by more than ± 2 V. A constant power supply may be necessary if line voltage is unstable.

4. Procedure

CAUTION: *This method is designed for microwave digestion of waters only. It is not intended for the digestion of solids, for which high concentrations of organic compounds may result in high pressures and possibly unsafe conditions.*

CAUTION: *As a safety measure, never mix different manufacturers' vessels in the same procedure. Vessels constructed differently will retain heat at different rates; control of heating conditions assumes that all vessels have the same heat-transfer characteristics. Inspect casements for cracks and chemical corrosion. Failure to maintain the vessels' integrity may result in catastrophic failure.*

Both prescription controls and performance controls are provided for this procedure. Performance controls are the most general and most accurate. When equipment capability permits, use the performance criterion.

a. Performance criterion: The following procedure is based on heating acidified samples in two stages where the first stage is to reach $160 \pm 4^\circ$ in 10 min and the second stage is to permit a slow rise to 165 to 170°C during the second 10 min. This performance criterion is based on temperature feedback control system capability that is implemented in various ways by different manufacturers. Because the temperature of the acid controls the reaction, this is the essential condition that will reproduce results in this preparation method. Verification of temperature conditions inside the vessel at these specific times is sufficient to verify the critical procedural requirements.

b. Prescription criterion: For all PFA vessels without liners, a verified program that meets the performance-based temperature-time profile is 545 W for 10 min followed by 344 W for 10 min using five single-wall PFA Teflon™ digestion vessels.² Any verified program for a given microwave unit depends on unit power and operational power settings, heating times, number, type, and placement of digestion vessels within the unit, and sample and acid volumes. The change in power, time, and temperature profile is not directly proportional to the change in the number of sample vessels. Any deviations from the verified program conditions will require verification of the time-temperature profile to conform to the given two-stage profile. This may be done by laboratory personnel if suitable test equipment is available, or by the manufacturer of the microwave equipment.

c. General conditions: Weigh entire digestion vessel assembly to 0.1 g before use and record (A). Accurately transfer 45 mL of well-shaken sample into the digestion vessel. Pipet 5 mL conc HNO_3 into each vessel. Attach all safety equipment required for appropriate and safe vessel operation following manufacturer's specifications. Tighten cap to manufacturer's specifications. Weigh each capped vessel to the nearest 0.1 g (B).

Place appropriate number of vessels evenly distributed in the carousel. Treat sample blanks, known additions, and duplicates in the same manner as samples. For prescription control only, when fewer samples than the appropriate number are digested, fill remaining vessels with 45 mL water and 5 mL conc HNO_3 to obtain full complement of vessels for the particular program being used.

Place carousel in unit and seat it carefully on turntable. Program microwave unit to heat samples to $160 \pm 4^\circ\text{C}$ in 10 min and then, for the second stage, to permit a slow rise to 165 to 170°C for 10 min. Start microwave generator, making sure that turntable is turning and that exhaust fan is on.

At completion of the microwave program, let vessels cool for at least 5 min in the unit before removal. Cool samples further outside the unit by removing the carousel and letting them cool on a bench or in a water bath. When cooled to room temperature, weigh each vessel (to 0.1 g) and record weight (C).

If the net weight of sample plus acid decreased by more than 10%, discard sample.

Complete sample preparation by carefully uncapping and venting each vessel in a fume hood. Follow individual manufac-

|| Or equivalent.

turer's specifications for relieving pressure in individual vessel types. Transfer to acid-cleaned noncontaminating plastic bottles. If the digested sample contains particulates, filter, centrifuge, or settle overnight and decant.

5. Calculations

a. Dilution correction: Multiply results by 50/45 or 1.11 to account for the dilution caused by the addition of 5 mL acid to 45 mL sample.

b. Discarding of sample: To determine if the net weight of sample plus acid decreased by more than 10% during the digestion process, use the following calculation

$$\frac{[(B - A) - (C - A)]}{(B - A)} \times 100 > 10\% \text{ (1\% for multilayer vessels)}$$

6. Quality Control

NOTE: When nitric acid digestion is used, recoveries of silver and antimony in some matrices may be unacceptably low. Verify recoveries using appropriate known additions.

Preferably include a quality-control sample in each loaded carousel. Prepare samples in batches including preparation blanks, sample duplicates, and pre-digestion known additions. Determine size of batch and frequency of quality-control samples by method of analysis and laboratory practice. The power of the microwave unit and batch size may prevent including one or more of the quality-control samples in each carousel. Do not group quality-control samples together but distribute them throughout the various carousels to give the best monitoring of digestion.

7. References

1. KINGSTON, H.M. & S. HASWELL, eds. 1997. *Microwave Enhanced Chemistry: Fundamentals, Sample Preparation and Applications*. American Chemical Soc., Washington, D.C.
2. U.S. ENVIRONMENTAL PROTECTION AGENCY. 1990. *Microwave assisted acid digestion of aqueous samples and extracts*. SW-846 Method 3015, Test Methods for Evaluating Solid Waste. U.S. Environmental Protection Agency, Washington, D.C.

3110 METALS BY ATOMIC ABSORPTION SPECTROMETRY

Because requirements for determining metals by atomic absorption spectrometry vary with metal and/or concentration to be determined, the method is presented as follows:

Section 3111, Metals by Flame Atomic Absorption Spectrometry, encompasses:

- Determination of antimony, bismuth, cadmium, calcium, cesium, chromium, cobalt, copper, gold, iridium, iron, lead, lithium, magnesium, manganese, nickel, palladium, platinum, potassium, rhodium, ruthenium, silver, sodium, strontium, thallium, tin, and zinc by direct aspiration into an air-acetylene flame (3111B),

- Determination of low concentrations of cadmium, chromium, cobalt, copper, iron, lead, manganese, nickel, silver, and zinc by chelation with ammonium pyrrolidine dithiocarbamate (APDC), extraction into methyl isobutyl ketone (MIBK), and aspiration into an air-acetylene flame (3111C),

- Determination of aluminum, barium, beryllium, calcium, molybdenum, osmium, rhenium, silicon, thorium, titanium, and

vanadium by direct aspiration into a nitrous oxide-acetylene flame (3111D), and

- Determination of low concentrations of aluminum and beryllium by chelation with 8-hydroxyquinoline, extraction into MIBK, and aspiration into a nitrous oxide-acetylene flame (3111E).

Section 3112 covers determination of mercury by the cold vapor technique.

Section 3113 concerns determination of micro quantities of aluminum, antimony, arsenic, barium, beryllium, cadmium, chromium, cobalt, copper, iron, lead, manganese, molybdenum, nickel, selenium, silver, and tin by electrothermal atomic absorption spectrometry.

Section 3114 covers determination of arsenic and selenium by conversion to their hydrides and aspiration into an argon-hydrogen or nitrogen-hydrogen flame.

3111 METALS BY FLAME ATOMIC ABSORPTION SPECTROMETRY*

3111 A. Introduction

1. Principle

In flame atomic absorption spectrometry, a sample is aspirated into a flame and atomized. A light beam is directed through the flame, into a monochromator, and onto a detector that measures

the amount of light absorbed by the atomized element in the flame. For some metals, atomic absorption exhibits superior sensitivity over flame emission. Because each metal has its own characteristic absorption wavelength, a source lamp composed of that element is used; this makes the method relatively free from spectral or radiation interferences. The amount of energy at the characteristic wavelength absorbed in the flame is proportional to the concentration of the element in the sample over a limited

* Approved by Standard Methods Committee, 1999.

concentration range. Most atomic absorption instruments also are equipped for operation in an emission mode, which may provide better linearity for some elements.

2. Selection of Method

See Section 3110.

3. Interferences

a. Chemical interference: Many metals can be determined by direct aspiration of sample into an air-acetylene flame. The most troublesome type of interference is termed "chemical" and results from the lack of absorption by atoms bound in molecular combination in the flame. This can occur when the flame is not hot enough to dissociate the molecules or when the dissociated atom is oxidized immediately to a compound that will not dissociate further at the flame temperature. Such interferences may be reduced or eliminated by adding specific elements or compounds to the sample solution. For example, the interference of phosphate in the magnesium determination can be overcome by adding lanthanum. Similarly, introduction of calcium eliminates silica interference in the determination of manganese. However, silicon and metals such as aluminum, barium, beryllium, and vanadium require the higher-temperature, nitrous oxide-acetylene flame to dissociate their molecules. The nitrous oxide-acetylene flame also can be useful in minimizing certain types of chemical interferences encountered in the air-acetylene flame. For example, the interference caused by high concentrations of phosphate in the determination of calcium in the air-acetylene flame is reduced in the nitrous oxide-acetylene flame.

MIBK extractions with APDC (see 3111C) are particularly useful where a salt matrix interferes, for example, in seawater. This procedure also concentrates the sample so that the detection limits are lowered.

Brines and seawater can be analyzed by direct aspiration but sample dilution is recommended. Aspiration of solutions containing high concentrations of dissolved solids often results in solids buildup on the burner head. This requires frequent shut-down of the flame and cleaning of the burner head. Preferably use background correction when analyzing waters that contain in excess of 1% solids, especially when the primary resonance line of the element of interest is below 240 nm. Make more frequent recovery checks when analyzing brines and seawaters to insure accurate results in these concentrated and complex matrices.

Barium and other metals ionize in the flame, thereby reducing the ground state (potentially absorbing) population. The addition of an excess of a cation (sodium, potassium, or lithium) having a similar or lower ionization potential will overcome this problem. The wavelength of maximum absorption for arsenic is 193.7 nm and for selenium 196.0 nm—wavelengths at which the air-acetylene flame absorbs intensely. The sensitivity for arsenic and selenium can be improved by conversion to their gaseous hydrides and analyzing them in either a nitrogen-hydrogen or an argon-hydrogen flame with a quartz tube (see Section 3114).

b. Background correction: Molecular absorption and light scattering caused by solid particles in the flame can cause erroneously high absorption values resulting in positive errors. When such phenomena occur, use background correction to obtain accurate values. Use any one of three types of background

correction: continuum-source, Zeeman, or Smith-Hieftje correction.

1) Continuum-source background correction—A continuum-source background corrector utilizes either a hydrogen-filled hollow cathode lamp with a metal cathode or a deuterium arc lamp. When both the line source hollow-cathode lamp and the continuum source are placed in the same optical path and are time-shared, the broadband background from the elemental signal is subtracted electronically, and the resultant signal will be background-compensated.

Both the hydrogen-filled hollow-cathode lamp and deuterium arc lamp have lower intensities than either the line source hollow-cathode lamp or electrodeless discharge lamps. To obtain a valid correction, match the intensities of the continuum source with the line source hollow-cathode or electrodeless discharge lamp. The matching may result in lowering the intensity of the line source or increasing the slit width; these measures have the disadvantage of raising the detection limit and possibly causing nonlinearity of the calibration curve. Background correction using a continuum source corrector is susceptible to interference from other absorbing lines in the spectral bandwidth. Mis-correction occurs from significant atomic absorption of the continuum source radiation by elements other than that being determined. When a line source hollow-cathode lamp is used without background correction, the presence of an absorbing line from another element in the spectral bandwidth will not cause an interference unless it overlaps the line of interest.

Continuum-source background correction will not remove direct absorption spectral overlap, where an element other than that being determined is capable of absorbing the line radiation of the element under study.

2) Zeeman background correction—This correction is based on the principle that a magnetic field splits the spectral line into two linearly polarized light beams parallel and perpendicular to the magnetic field. One is called the pi (π) component and the other the sigma (σ) component. These two light beams have exactly the same wavelength and differ only in the plane of polarization. The π line will be absorbed by both the atoms of the element of interest and by the background caused by broadband absorption and light scattering of the sample matrix. The σ line will be absorbed only by the background.

Zeeman background correction provides accurate background correction at much higher absorption levels than is possible with continuum source background correction systems. It also virtually eliminates the possibility of error from structured background. Because no additional light sources are required, the alignment and intensity limitations encountered using continuum sources are eliminated.

Disadvantages of the Zeeman method include reduced sensitivity for some elements, reduced linear range, and a "rollover" effect whereby the absorbance of some elements begins to decrease at high concentrations, resulting in a two-sided calibration curve.

3) Smith-Hieftje background correction—This correction is based on the principle that absorbance measured for a specific element is reduced as the current to the hollow cathode lamp is increased while absorption of nonspecific absorbing substances remains identical at all current levels. When this method is applied, the absorbance at a high-current mode is subtracted from the absorbance at a low-current mode. Under these condi-

TABLE 3111:I. ATOMIC ABSORPTION CONCENTRATION RANGES WITH DIRECT ASPIRATION ATOMIC ABSORPTION

Element	Wave-length nm	Flame Gases*	Instrument Detection		Optimum Concentration Range mg/L
			Level mg/L	Sensitivity mg/L	
Ag	328.1	A-Ac	0.01	0.06	0.1-4
Al	309.3	N-Ac	0.1	1	5-100
Au	242.8	A-Ac	0.01	0.25	0.5-20
Ba	553.6	N-Ac	0.03	0.4	1-20
Be	234.9	N-Ac	0.005	0.03	0.05-2
Bi	223.1	A-Ac	0.06	0.4	1-50
Ca	422.7	A-Ac	0.003	0.08	0.2-20
Cd	228.8	A-Ac	0.002	0.025	0.05-2
Co	240.7	A-Ac	0.03	0.2	0.5-10
Cr	357.9	A-Ac	0.02	0.1	0.2-10
Cs	852.1	A-Ac	0.02	0.3	0.5-15
Cu	324.7	A-Ac	0.01	0.1	0.2-10
Fe	248.3	A-Ac	0.02	0.12	0.3-10
Ir	264.0	A-Ac	0.6	8	—
K	766.5	A-Ac	0.005	0.04	0.1-2
Li	670.8	A-Ac	0.002	0.04	0.1-2
Mg	285.2	A-Ac	0.0005	0.007	0.02-2
Mn	279.5	A-Ac	0.01	0.05	0.1-10
Mo	313.3	N-Ac	0.1	0.5	1-20
Na	589.0	A-Ac	0.002	0.015	0.03-1
Ni	232.0	A-Ac	0.02	0.15	0.3-10
Os	290.9	N-Ac	0.08	1	—
Pb†	283.3	A-Ac	0.05	0.5	1-20
Pt	265.9	A-Ac	0.1	2	5-75
Rh	343.5	A-Ac	0.5	0.3	—
Ru	349.9	A-Ac	0.07	0.5	—
Sb	217.6	A-Ac	0.07	0.5	1-40
Si	251.6	N-Ac	0.3	2	5-150
Sn	224.6	A-Ac	0.8	4	10-200
Sr	460.7	A-Ac	0.03	0.15	0.3-5
Ti	365.3	N-Ac	0.3	2	5-100
V	318.4	N-Ac	0.2	1.5	2-100
Zn	213.9	A-Ac	0.005	0.02	0.05-2

* A-Ac = air-acetylene; N-Ac = nitrous oxide-acetylene..

† The more sensitive 217.0 nm wavelength is recommended for instruments with background correction capabilities.

Copyright © ASTM. Reprinted with permission.

tions, any absorbance due to nonspecific background is subtracted out and corrected for.

Smith-Hieftje background correction provides a number of advantages over continuum-source correction. Accurate correction at higher absorbance levels is possible and error from structured background is virtually eliminated. In some cases, spectral interferences also can be eliminated. The usefulness of Smith-Hieftje background correction with electrodeless discharge lamps has not yet been established.

4. Sensitivity, Detection Levels, and Optimum Concentration Ranges

The sensitivity of flame atomic absorption spectrometry is defined as the metal concentration that produces an absorption of 1% (an absorbance of approximately 0.0044). The instrument

detection level is defined here as the concentration that produces absorption equivalent to twice the magnitude of the background fluctuation. Sensitivity and detection levels vary with the instrument, the element determined, the complexity of the matrix, and the technique selected. The optimum concentration range usually starts from the concentration of several times the detection level and extends to the concentration at which the calibration curve starts to flatten. To achieve best results, use concentrations of samples and standards within the optimum concentration range of the spectrometer. See Table 3111:I for indication of concentration ranges measurable with conventional atomization. In many instances the concentration range shown in Table 3111:I

TABLE 3111:II. INTERLABORATORY PRECISION AND BIAS DATA FOR ATOMIC ABSORPTION METHODS—DIRECT ASPIRATION AND EXTRACTED METALS

Metal	Conc.*	SD*	Relative		No. of Participants
			SD %	Error %	
Direct determination:					
Aluminum ¹	4.50	0.19	4.2	8.4	5
Barium ²	1.00	0.089	8.9	2.7	11
Beryllium ¹	0.46	0.0213	4.6	23.0	11
Cadmium ³	0.05	0.0108	21.6	8.2	26
Cadmium ¹	1.60	0.11	6.9	5.1	16
Calcium ¹	5.00	0.21	4.2	0.4	8
Chromium ¹	3.00	0.301	10.0	3.7	9
Cobalt ¹	4.00	0.243	6.1	0.5	14
Copper ³	1.00	0.112	11.2	3.4	53
Copper ¹	4.00	0.331	8.3	2.8	15
Iron ¹	4.40	0.260	5.8	2.3	16
Iron ³	0.30	0.0495	16.5	0.6	43
Lead ¹	6.00	0.28	4.7	0.2	14
Magnesium ³	0.20	0.021	10.5	6.3	42
Magnesium ¹	1.10	0.116	10.5	10.0	8
Manganese ¹	4.05	0.317	7.8	1.3	16
Manganese ³	0.05	0.0068	13.5	6.0	14
Nickel ¹	3.93	0.383	9.8	2.0	14
Silver ³	0.05	0.0088	17.5	10.6	7
Silver ¹	2.00	0.07	3.5	1.0	10
Sodium ¹	2.70	0.122	4.5	4.1	12
Strontium ¹	1.00	0.05	5.0	0.2	12
Zinc ³	0.50	0.041	8.2	0.4	48
Extracted determination:					
Aluminum ²	300	32	10.7	0.7	15
Beryllium ²	5	1.7	34.0	20.0	9
Cadmium ³	50	21.9	43.8	13.3	12
Cobalt ¹	300	28.5	9.5	1.0	6
Copper ¹	100	71.7	71.7	12.0	8
Iron ¹	250	19.0	7.6	3.6	4
Manganese ¹	21.5	2.4	11.2	7.4	8
Molybdenum ¹	9.5	1.1	11.6	1.3	5
Nickel ¹	56.8	15.2	26.8	13.6	14
Lead ³	50	11.8	23.5	19.0	8
Silver ¹	5.2	1.4	26.9	3.0	7

* For direct determinations, mg/L; for extracted determinations, µg/L.

Superscripts refer to reference numbers.

Source: AMERICAN SOCIETY FOR TESTING AND MATERIALS, 1986. Annual Book of ASTM Standards. Volume 11.01, Water and Environmental Technology. American Soc. Testing & Materials, Philadelphia, Pa. Copyright © ASTM. Reprinted with permission.

TABLE 3111:III. SINGLE-OPERATOR PRECISION AND RECOMMENDED CONTROL RANGES FOR ATOMIC ABSORPTION METHODS—DIRECT ASPIRATION AND EXTRACTED METALS

Metal	Conc.*	SD*	Relative SD %	No. of Participants	QC Std.*	Acceptable Range*
Direct determination:						
Aluminum ¹	4.50	0.23	5.1	15	5.00	4.3–5.7
Beryllium ¹	0.46	0.012	2.6	10	0.50	0.46–0.54
Calcium ¹	5.00	0.05	1.0	8	5.00	4.8–5.2
Chromium ¹	7.00	0.69	9.9	9	5.00	3.3–6.7
Cobalt ¹	4.00	0.21	5.3	14	4.00	3.4–4.6
Copper ¹	4.00	0.115	2.9	15	4.00	3.7–4.3
Iron ¹	5.00	0.19	3.8	16	5.00	4.4–5.6
Magnesium ¹	1.00	0.009	0.9	8	1.00	0.97–1.03
Nickel ⁴	5.00	0.04	0.8	—	5.00	4.9–5.1
Silver ¹	2.00	0.25	12.5	10	2.00	1.2–2.8
Sodium ⁴	8.2	0.1	1.2	—	5.00	4.8–5.2
Strontium ¹	1.00	0.04	4.0	12	1.00	0.87–1.13
Potassium ⁴	1.6	0.2	12.5	—	1.6	1.0–2.2
Molybdenum ⁴	7.5	0.07	0.9	—	10.0	9.7–10.3
Tin ⁴	20.0	0.5	2.5	—	20.0	18.5–21.5
Titanium ⁴	50.0	0.4	0.8	—	50.0	48.8–51.2
Vanadium	50.0	0.2	0.4	—	50.0	49.4–50.6
Extracted determination:						
Aluminum ¹	300	12	4.0	15	300	264–336
Cobalt ¹	300	20	6.7	6	300	220–380
Copper ¹	100	21	21	8	100	22–178
Iron ¹	250	12	4.8	4	250	180–320
Manganese ¹	21.5	202	10.2	8	25	17–23
Molybdenum ¹	9.5	1.0	10.5	5	10	5.5–14.5
Nickel ¹	56.8	9.2	16.2	14	50	22–78
Silver ¹	5.2	1.2	23.1	7	5.0	0.5–9.5

* For direct determinations, mg/L; for extracted determinations, $\mu\text{g/L}$.

Superscripts refer to reference numbers.

Source: AMERICAN SOCIETY FOR TESTING AND MATERIALS, 1986. Annual Book of ASTM Standards, Volume 11.01, Water and Environmental Technology. American Soc. Testing & Materials, Philadelphia, Pa. Copyright © ASTM. Reprinted with permission.

may be extended downward either by scale expansion or by integrating the absorption signal over a long time. The range may be extended upward by dilution, using a less sensitive wavelength, rotating the burner head, or utilizing a microprocessor to linearize the calibration curve at high concentrations.

5. Preparation of Standards

Prepare standard solutions of known metal concentrations in water with a matrix similar to the sample. Use standards that bracket expected sample concentration and are within the method's working range. Very dilute standards should be prepared daily from stock solutions in concentrations greater than 500 mg/L. Stock standard solutions can be obtained from several commercial sources. They also can be prepared from National Institute of Standards and Technology (NIST) reference materials or by procedures outlined in the following sections.

For samples containing high and variable concentrations of matrix materials, make the major ions in the sample and the dilute standard similar. If the sample matrix is complex and components cannot be matched accurately with standards, use the method of standard additions, 3113B.4d2), to correct for matrix effects. If digestion is used, carry standards through the same digestion procedure used for samples.

6. Apparatus

a. Atomic absorption spectrometer, consisting of a light source emitting the line spectrum of an element (hollow-cathode lamp or electrodeless discharge lamp), a device for vaporizing the sample (usually a flame), a means of isolating an absorption line (monochromator or filter and adjustable slit), and a photoelectric detector with its associated electronic amplifying and measuring equipment.

b. Burner: The most common type of burner is a premix, which introduces the spray into a condensing chamber for removal of large droplets. The burner may be fitted with a conventional head containing a single slot; a three-slot Boling head, which may be preferred for direct aspiration with an air-acetylene flame; or a special head for use with nitrous oxide and acetylene.

c. Readout: Most instruments are equipped with either a digital or null meter readout mechanism. Most modern instruments are equipped with microprocessors or stand-alone control computers capable of integrating absorption signals over time and linearizing the calibration curve at high concentrations.

d. Lamps: Use either a hollow-cathode lamp or an electrodeless discharge lamp (EDL). Use one lamp for each element being measured. Multi-element hollow-cathode lamps generally pro-

vide lower sensitivity than single-element lamps. EDLs take a longer time to warm up and stabilize.

e. Pressure-reducing valves: Maintain supplies of fuel and oxidant at pressures somewhat higher than the controlled operating pressure of the instrument by using suitable reducing valves. Use a separate reducing valve for each gas.

f. Vent: Place a vent about 15 to 30 cm above the burner to remove fumes and vapors from the flame. This precaution protects laboratory personnel from toxic vapors, protects the instrument from corrosive vapors, and prevents flame stability from being affected by room drafts. A damper or variable-speed blower is desirable for modulating air flow and preventing flame disturbance. Select blower size to provide the air flow recommended by the instrument manufacturer. In laboratory locations with heavy particulate air pollution, use clean laboratory facilities (Section 3010C).

7. Quality Assurance/Quality Control

Some data typical of the precision and bias obtainable with the methods discussed are presented in Tables 3111:II and III.

Analyze a blank between sample or standard readings to verify baseline stability. Rezero when necessary.

To one sample out of every ten (or one sample from each group of samples if less than ten are being analyzed) add a known amount of the metal of interest and reanalyze to confirm recovery. The amount of metal added should be approximately equal to the amount found. If little metal is present add an amount close to the middle of the linear range of the test. Recovery of added metal should be between 85 and 115%.

Analyze an additional standard solution after every ten samples or with each batch of samples, whichever is less, to confirm that the test is in control. Recommended concentrations of standards to be run, limits of acceptability, and

reported single-operator precision data are listed in Table 3111:III.

See Section 3020 for additional recommended quality control procedures.

8. References

1. AMERICAN SOCIETY FOR TESTING AND MATERIALS. 1986. Annual Book of ASTM Standards, Volume 11.01, Water and Environmental Technology. American Soc. Testing & Materials, Philadelphia, Pa.
2. U.S. DEPARTMENT HEALTH, EDUCATION AND WELFARE. 1970. Water Metals No. 6, Study No. 37. U.S. Public Health Serv. Publ. No. 2029, Cincinnati, Ohio.
3. U.S. DEPARTMENT HEALTH, EDUCATION AND WELFARE. 1968. Water Metals No. 4, Study No. 30. U.S. Public Health Serv. Publ. No. 999-UTH-8, Cincinnati, Ohio.
4. U.S. ENVIRONMENTAL PROTECTION AGENCY. 1983. Methods for Chemical Analysis of Water and Wastes. Cincinnati, Ohio.

9. Bibliography

- KAHN, H.L. 1968. Principles and Practice of Atomic Absorption. Advan. Chem. Ser. No. 73, Div. Water, Air & Waste Chemistry, American Chemical Soc., Washington, D.C.
- RAMIRIZ-MUNOZ, J. 1968. Atomic Absorption Spectroscopy and Analysis by Atomic Absorption Flame Photometry. American Elsevier Publishing Co., New York, N.Y.
- SLAVIN, W. 1968. Atomic Absorption Spectroscopy. John Wiley & Sons, New York, N.Y.
- PAUS, P.E. 1971. The application of atomic absorption spectroscopy to the analysis of natural waters. *Atomic Absorption Newsletter* 10:69.
- EDIGER, R.D. 1973. A review of water analysis by atomic absorption. *Atomic Absorption Newsletter* 12:151.
- PAUS, P.E. 1973. Determination of some heavy metals in seawater by atomic absorption spectroscopy. *Fresenius Zeitschr. Anal. Chem.* 264:118.
- BURRELL, D.C. 1975. Atomic Spectrometric Analysis of Heavy-Metal Pollutants in Water. Ann Arbor Science Publishers, Inc., Ann Arbor, Mich.

3111 B. Direct Air-Acetylene Flame Method

1. General Discussion

This method is applicable to the determination of antimony, bismuth, cadmium, calcium, cesium, chromium, cobalt, copper, gold, iridium, iron, lead, lithium, magnesium, manganese, nickel, palladium, platinum, potassium, rhodium, ruthenium, silver, sodium, strontium, thallium, tin, and zinc.

2. Apparatus

Atomic absorption spectrometer and associated equipment: See Section 3111A.6. Use burner head recommended by the manufacturer.

3. Reagents

a. Air, cleaned and dried through a suitable filter to remove oil, water, and other foreign substances. The source may be a compressor or commercially bottled gas.

b. Acetylene, standard commercial grade. Acetone, which always is present in acetylene cylinders, can be prevented from entering and damaging the burner head by replacing a cylinder when its pressure has fallen to 689 kPa (100 psi) acetylene.

CAUTION: *Acetylene gas represents an explosive hazard in the laboratory. Follow instrument manufacturer's directions in plumbing and using this gas. Do not allow gas contact with copper, brass with >65% copper, silver, or liquid mercury; do not use copper or brass tubing, regulators, or fittings.*

c. Metal-free water: Use metal-free water for preparing all reagents and calibration standards and as dilution water. Prepare

metal-free water by deionizing tap water and/or by using one of the following processes, depending on the metal concentration in the sample: single distillation, redistillation, or sub-boiling. Always check deionized or distilled water to determine whether the element of interest is present in trace amounts. (NOTE: *If the source water contains Hg or other volatile metals, single- or redistilled water may not be suitable for trace analysis because these metals distill over with the distilled water. In such cases, use sub-boiling to prepare metal-free water.*)

d. *Calcium solution:* Dissolve 630 mg calcium carbonate, CaCO_3 , in 50 mL of 1 + 5 HCl. If necessary, boil gently to obtain complete solution. Cool and dilute to 1000 mL with water.

e. *Hydrochloric acid,* HCl, 1%, 10%, 20% (all v/v), 1 + 5, 1 + 1, and conc.

f. *Lanthanum solution:* Dissolve 58.65 g lanthanum oxide, La_2O_3 , in 250 mL conc HCl. Add acid slowly until the material is dissolved and dilute to 1000 mL with water.

g. *Hydrogen peroxide,* 30%.

h. *Nitric acid,* HNO_3 , 2% (v/v), 1 + 1, and conc.

i. *Aqua regia:* Add 3 volumes conc HCl to 1 volume conc HNO_3 .

j. *Standard metal solutions:* Prepare a series of standard metal solutions in the optimum concentration range by appropriate dilution of the following stock metal solutions with water containing 1.5 mL conc HNO_3/L . Stock standard solutions are available from a number of commercial suppliers. Alternatively, prepare as described below. Thoroughly dry reagents before use. In general, use reagents of the highest purity. For hydrates, use fresh reagents.

1) *Antimony:* Dissolve 0.2669 g $\text{K}(\text{SbO})\text{C}_4\text{H}_4\text{O}_6$ in water, add 10 mL 1 + 1 HCl and dilute to 1000 mL with water; 1.00 mL = 100 μg Sb.

2) *Bismuth:* Dissolve 0.100 g bismuth metal in a minimum volume of 1 + 1 HNO_3 . Dilute to 1000 mL with 2% (v/v) HNO_3 ; 1.00 mL = 100 μg Bi.

3) *Cadmium:* Dissolve 0.100 g cadmium metal in 4 mL conc HNO_3 . Add 8.0 mL conc HNO_3 and dilute to 1000 mL with water; 1.00 mL = 100 μg Cd.

4) *Calcium:* Suspend 0.2497 g CaCO_3 (dried at 180° for 1 h before weighing) in water and dissolve cautiously with a minimum amount of 1 + 1 HNO_3 . Add 10.0 mL conc HNO_3 and dilute to 1000 mL with water; 1.00 mL = 100 μg Ca.

5) *Cesium:* Dissolve 0.1267 g cesium chloride, CsCl , in 1000 mL water; 1.00 mL = 100 μg Cs.

6) *Chromium:* Dissolve 0.1923 g CrO_3 in water. When solution is complete, acidify with 10 mL conc HNO_3 and dilute to 1000 mL with water; 1.00 mL = 100 μg Cr.

7) *Cobalt:* Dissolve 0.1000 g cobalt metal in a minimum amount of 1 + 1 HNO_3 . Add 10.0 mL 1 + 1 HCl and dilute to 1000 mL with water; 1.00 mL = 100 μg Co.

8) *Copper:* Dissolve 0.100 g copper metal in 2 mL conc HNO_3 , add 10.0 mL conc HNO_3 and dilute to 1000 mL with water; 1.00 mL = 100 μg Cu.

9) *Gold:* Dissolve 0.100 g gold metal in a minimum volume of aqua regia. Evaporate to dryness, dissolve residue in 5 mL conc HCl, cool, and dilute to 1000 mL with water; 1.00 mL = 100 μg Au.

10) *Iridium:* Dissolve 0.1147 g ammonium chloroiridate, $(\text{NH}_4)_2\text{IrCl}_6$, in a minimum volume of 1% (v/v) HCl and dilute to 100 mL with 1% (v/v) HCl; 1.00 mL = 500 μg Ir.

11) *Iron:* Dissolve 0.100 g iron wire in a mixture of 10 mL 1 + 1 HCl and 3 mL conc HNO_3 . Add 5 mL conc HNO_3 and dilute to 1000 mL with water; 1.00 mL = 100 μg Fe.

12) *Lead:* Dissolve 0.1598 g lead nitrate, $\text{Pb}(\text{NO}_3)_2$, in a minimum amount of 1 + 1 HNO_3 , add 10 mL conc HNO_3 , and dilute to 1000 mL with water; 1.00 mL = 100 μg Pb.

13) *Lithium:* Dissolve 0.5323 g lithium carbonate, Li_2CO_3 , in a minimum volume of 1 + 1 HNO_3 . Add 10.0 mL conc HNO_3 and dilute to 1000 mL with water; 1.00 mL = 100 μg Li.

14) *Magnesium:* Dissolve 0.1658 g MgO in a minimum amount of 1 + 1 HNO_3 . Add 10.0 mL conc HNO_3 and dilute to 1000 mL with water; 1.00 mL = 100 μg Mg.

15) *Manganese:* Dissolve 0.1000 g manganese metal in 10 mL conc HCl mixed with 1 mL conc HNO_3 . Dilute to 1000 mL with water; 1.00 mL = 100 μg Mn.

16) *Nickel:* Dissolve 0.1000 g nickel metal in 10 mL hot conc HNO_3 , cool, and dilute to 1000 mL with water; 1.00 mL = 100 μg Ni.

17) *Palladium:* Dissolve 0.100 g palladium wire in a minimum volume of aqua regia and evaporate just to dryness. Add 5 mL conc HCl and 25 mL water and warm until dissolution is complete. Dilute to 1000 mL with water; 1.00 mL = 100 μg Pd.

18) *Platinum:* Dissolve 0.100 g platinum metal in a minimum volume of aqua regia and evaporate just to dryness. Add 5 mL conc HCl and 0.1 g NaCl and again evaporate just to dryness. Dissolve residue in 20 mL of 1 + 1 HCl and dilute to 1000 mL with water; 1.00 mL = 100 μg Pt.

19) *Potassium:* Dissolve 0.1907 g potassium chloride, KCl , (dried at 110°C) in water and make up to 1000 mL; 1.00 mL = 100 μg K.

20) *Rhodium:* Dissolve 0.386 g ammonium hexachlororhodate, $(\text{NH}_4)_3\text{RhCl}_6 \cdot 1.5\text{H}_2\text{O}$, in a minimum volume of 10% (v/v) HCl and dilute to 1000 mL with 10% (v/v) HCl; 1.00 mL = 100 μg Rh.

21) *Ruthenium:* Dissolve 0.205 g ruthenium chloride, RuCl_3 , in a minimum volume of 20% (v/v) HCl and dilute to 1000 mL with 20% (v/v) HCl; 1.00 mL = 100 μg Ru.

22) *Silver:* Dissolve 0.1575 g silver nitrate, AgNO_3 , in 100 mL water, add 10 mL conc HNO_3 , and make up to 1000 mL; 1.00 mL = 100 μg Ag.

23) *Sodium:* Dissolve 0.2542 g sodium chloride, NaCl , dried at 140°C , in water, add 10 mL conc HNO_3 and make up to 1000 mL; 1.00 mL = 100 μg Na.

24) *Strontium:* Suspend 0.1685 g SrCO_3 in water and dissolve cautiously with a minimum amount of 1 + 1 HNO_3 . Add 10.0 mL conc HNO_3 and dilute to 1000 mL with water; 1 mL = 100 μg Sr.

25) *Thallium:* Dissolve 0.1303 g thallium nitrate, TlNO_3 , in water. Add 10 mL conc HNO_3 and dilute to 1000 mL with water; 1.00 mL = 100 μg Tl.

26) *Tin:* Dissolve 1.000 g tin metal in 100 mL conc HCl and dilute to 1000 mL with water; 1.00 mL = 1.00 mg Sn.

27) *Zinc:* Dissolve 0.100 g zinc metal in 20 mL 1 + 1 HCl and dilute to 1000 mL with water; 1.00 mL = 100 μg Zn.

4. Procedure

a. Sample preparation: Required sample preparation depends on the metal form being measured.

If dissolved metals are to be determined, see Section 3030B for sample preparation. If total or acid-extractable metals are to be determined, see Sections 3030C through K. For all samples, make certain that the concentrations of acid and matrix modifiers are the same in both samples and standards.

When determining *Ca* or *Mg*, dilute and mix 100 mL sample or standard with 10 mL lanthanum solution (§ 3f) before aspirating. When determining *Fe* or *Mn*, mix 100 mL with 25 mL of *Ca* solution (§ 3d) before aspirating. When determining *Cr*, mix 1 mL 30% H_2O_2 with each 100 mL before aspirating. Alternatively use proportionally smaller volumes.

b. Instrument operation: Because of differences between makes and models of atomic absorption spectrometers, it is not possible to formulate instructions applicable to every instrument. See manufacturer's operating manual. In general, proceed according to the following: Install a hollow-cathode lamp for the desired metal in the instrument and roughly set the wavelength dial according to Table 3111:I. Set slit width according to manufacturer's suggested setting for the element being measured. Turn on instrument, apply to the hollow-cathode lamp the current suggested by the manufacturer, and let instrument warm up until energy source stabilizes, generally about 10 to 20 min. Readjust current as necessary after warmup. Optimize wavelength by adjusting wavelength dial until optimum energy gain is obtained. Align lamp in accordance with manufacturer's instructions.

Install suitable burner head and adjust burner head position. Turn on air and adjust flow rate to that specified by manufacturer to give maximum sensitivity for the metal being measured. Turn on acetylene, adjust flow rate to value specified, and ignite flame. Let flame stabilize for a few minutes. Aspirate a blank consisting of deionized water containing the same concentration of acid in standards and samples. Zero the instrument. Aspirate a standard solution and adjust aspiration rate of the nebulizer to obtain maximum sensitivity. Adjust burner both vertically and horizontally to obtain maximum response. Aspirate blank again and zero the instrument. Aspirate a standard near the middle of the

linear range. Record absorbance of this standard when freshly prepared and with a new hollow-cathode lamp. Refer to these data on subsequent determinations of the same element to check consistency of instrument setup and aging of hollow-cathode lamp and standard.

The instrument now is ready to operate. When analyses are finished, extinguish flame by turning off first acetylene and then air.

c. Standardization: Select at least three concentrations of each standard metal solution (prepared as in § 3j above) to bracket the expected metal concentration of a sample. Aspirate blank and zero the instrument. Then aspirate each standard in turn into flame and record absorbance.

Prepare a calibration curve by plotting on linear graph paper absorbance of standards versus their concentrations. For instruments equipped with direct concentration readout, this step is unnecessary. With some instruments it may be necessary to convert percent absorption to absorbance by using a table generally provided by the manufacturer. Plot calibration curves for *Ca* and *Mg* based on original concentration of standards before dilution with lanthanum solution. Plot calibration curves for *Fe* and *Mn* based on original concentration of standards before dilution with *Ca* solution. Plot calibration curve for *Cr* based on original concentration of standard before addition of H_2O_2 .

d. Analysis of samples: Rinse nebulizer by aspirating water containing 1.5 mL conc HNO_3/L . Aspirate blank and zero instrument. Aspirate sample and determine its absorbance.

5. Calculations

Calculate concentration of each metal ion, in micrograms per liter for trace elements, and in milligrams per liter for more common metals, by referring to the appropriate calibration curve prepared according to § 4c. Alternatively, read concentration directly from the instrument readout if the instrument is so equipped. If the sample has been diluted, multiply by the appropriate dilution factor.

6. Bibliography

WILLIS, J.B. 1962. Determination of lead and other heavy metals in urine by atomic absorption spectrophotometry. *Anal. Chem.* 34:614. Also see Section 3111A.8 and 9.

3111 C. Extraction/Air-Acetylene Flame Method

1. General Discussion

This method is suitable for the determination of low concentrations of cadmium, chromium, cobalt, copper, iron, lead, manganese, nickel, silver, and zinc. The method consists of chelation with ammonium pyrrolidine dithiocarbamate (APDC) and extraction into methyl isobutyl ketone (MIBK), followed by aspiration into an air-acetylene flame.

2. Apparatus

a. Atomic absorption spectrometer and associated equipment: See Section 3111A.6.

b. Burner head, conventional. Consult manufacturer's operating manual for suggested burner head.

3. Reagents

a. Air: See 3111B.3a.

b. Acetylene: See 3111B.3b.

c. Metal-free water: See 3111B.3c.

d. Methyl isobutyl ketone (MIBK), reagent grade. For trace analysis, purify MIBK by redistillation or by sub-boiling distillation.

e. Ammonium pyrrolidine dithiocarbamate (APDC) solution: Dissolve 4 g APDC in 100 mL water. If necessary, purify APDC

with an equal volume of MIBK. Shake 30 s in a separatory funnel, let separate, and withdraw lower portion. Discard MIBK layer.

f. *Nitric acid*, HNO₃, conc, ultrapure.

g. *Standard metal solutions*: See 3111B.3j.

h. *Potassium permanganate solution*, KMnO₄, 5% (w/v) aqueous.

i. *Sodium sulfate*, Na₂SO₄, anhydrous.

j. *Water-saturated MIBK*: Mix one part purified MIBK with one part water in a separatory funnel. Shake 30 s and let separate. Discard aqueous layer. Save MIBK layer.

k. *Hydroxylamine hydrochloride solution*, 10% (w/v). This solution can be purchased commercially.

4. Procedure

a. *Instrument operation*: See Section 3111B.4b. After final adjusting of burner position, aspirate water-saturated MIBK into flame and gradually reduce fuel flow until flame is similar to that before aspiration of solvent.

b. *Standardization*: Select at least three concentrations of standard metal solutions (prepared as in 3111B.3j) to bracket expected sample metal concentration and to be, after extraction, in the optimum concentration range of the instrument. Adjust 100 mL of each standard and 100 mL of a metal-free water blank to pH 3 by adding 1N HNO₃ or 1N NaOH. For individual element extraction, use the following pH ranges to obtain optimum extraction efficiency:

Element	pH Range for Optimum Extraction
Ag	2-5 (complex unstable)
Cd	1-6
Co	2-10
Cr	3-9
Cu	0.1-8
Fe	2-5
Mn	2-4 (complex unstable)
Ni	2-4
Pb	0.1-6
Zn	2-6

NOTE: For Ag and Pb extraction the optimum pH value is 2.3 ± 0.2. The Mn complex deteriorates rapidly at room temperature, resulting in decreased instrument response. Chilling the extract to 0°C may preserve the complex for a few hours. If this is not possible and Mn cannot be analyzed immediately after extraction, use another analytical procedure.

Transfer each standard solution and blank to individual 200-mL volumetric flasks, add 1 mL APDC solution, and shake to mix. Add 10 mL MIBK and shake vigorously for 30 s. (The maximum volume ratio of sample to MIBK is 40.) Let contents of each flask separate into aqueous and organic layers, then carefully add water (adjusted to the same pH at which the extraction was carried out) down the side of each flask to bring the organic layer into the neck and accessible to the aspirating tube.

Aspirate organic extracts directly into the flame (zeroing instrument on a water-saturated MIBK blank) and record absorbance.

Prepare a calibration curve by plotting on linear graph paper absorbances of extracted standards against their concentrations before extraction.

c. *Analysis of samples*: Prepare samples in the same manner as the standards. Rinse atomizer by aspirating water-saturated MIBK. Aspirate organic extracts treated as above directly into the flame and record absorbances.

With the above extraction procedure only hexavalent chromium is measured. To determine total chromium, oxidize trivalent chromium to hexavalent chromium by bringing sample to a boil and adding sufficient KMnO₄ solution dropwise to give a persistent pink color while the solution is boiled for 10 min. Destroy excess KMnO₄ by adding 1 to 2 drops hydroxylamine hydrochloride solution to the boiling solution, allowing 2 min for the reaction to proceed. If pink color persists, add 1 to 2 more drops hydroxylamine hydrochloride solution and wait 2 min. Heat an additional 5 min. Cool, extract with MIBK, and aspirate.

During extraction, if an emulsion forms at the water-MIBK interface, add anhydrous Na₂SO₄ to obtain a homogeneous organic phase. In that case, also add Na₂SO₄ to all standards and blanks.

To avoid problems associated with instability of extracted metal complexes, determine metals immediately after extraction.

5. Calculations

Calculate the concentration of each metal ion in micrograms per liter by referring to the appropriate calibration curve.

6. Bibliography

- ALLAN, J.E. 1961. The use of organic solvents in atomic absorption spectrophotometry. *Spectrochim. Acta* 17:467.
- SACHDEV, S.L. & P.W. WEST. 1970. Concentration of trace metals by solvent extraction and their determination by atomic absorption spectrophotometry. *Environ. Sci. Technol.* 4:749.

3111 D. Direct Nitrous Oxide-Acetylene Flame Method

1. General Discussion

This method is applicable to the determination of aluminum, barium, beryllium, calcium, molybdenum, osmium, rhenium, silicon, thorium, titanium, and vanadium.

2. Apparatus

a. *Atomic absorption spectrometer and associated equipment*: See Section 3111A.6.

b. *Nitrous oxide burner head*: Use special burner head as

suggested in manufacturer's manual. At roughly 20-min intervals of operation it may be necessary to dislodge the carbon crust that forms along the slit surface with a carbon rod or appropriate alternative.

c. *T-junction valve* or other switching valve for rapidly changing from nitrous oxide to air, so that flame can be turned on or off with air as oxidant to prevent flashbacks.

3. Reagents

- a. *Air*: See 3111B.3a.
- b. *Acetylene*: See 3111B.3b.
- c. *Metal-free water*: See 3111B.3c.
- d. *Hydrochloric acid*, HCl, 1N, 1+1, and conc.
- e. *Nitric acid*, HNO₃, conc.
- f. *Sulfuric acid*, H₂SO₄, 1% (v/v).
- g. *Hydrofluoric acid*, HF, 1N.
- h. *Nitrous oxide*, commercially available cylinders. Fit nitrous oxide cylinder with a special nonfreezable regulator or wrap a heating coil around an ordinary regulator to prevent flashback at the burner caused by reduction in nitrous oxide flow through a frozen regulator. (Most modern atomic absorption instruments have automatic gas control systems that will shut down a nitrous oxide-acetylene flame safely in the event of a reduction in nitrous oxide flow rate.)

CAUTION: Use nitrous oxide with strict adherence to manufacturer's directions. Improper sequencing of gas flows at startup and shutdown of instrument can produce explosions from flashback.

- i. *Potassium chloride solution*: Dissolve 250 g KCl in water and dilute to 1000 mL.
- j. *Aluminum nitrate solution*: Dissolve 139 g Al(NO₃)₃ · 9H₂O in 150 mL water. Acidify slightly with conc HNO₃ to preclude possible hydrolysis and precipitation. Warm to dissolve completely. Cool and dilute to 200 mL.
- k. *Standard metal solutions*: Prepare a series of standard metal solutions in the optimum concentration ranges by appropriate dilution of stock metal solutions with water containing 1.5 mL conc HNO₃/L. Stock standard solutions are available from a number of commercial suppliers. Alternatively, prepare as described below.

1) *Aluminum*: Dissolve 0.100 g aluminum metal in an acid mixture of 4 mL 1 + 1 HCl and 1 mL conc HNO₃ in a beaker. Warm gently to effect solution. Transfer to a 1-L flask, add 10 mL 1 + 1 HCl, and dilute to 1000 mL with water; 1.00 mL = 100 μg Al.

2) *Barium*: Dissolve 0.1516 g BaCl₂ (dried at 250° for 2 h), in about 10 mL water with 1 mL 1 + 1 HCl. Add 10.0 mL 1 + 1 HCl and dilute to 1000 mL with water; 1.00 mL = 100 μg Ba.

3) *Beryllium*: Do not dry. Dissolve 1.966 g BeSO₄ · 4H₂O in water, add 10.0 mL conc HNO₃, and dilute to 1000 mL with water; 1.00 mL = 100 μg Be.

4) *Calcium*: See 3111B.3j4).

5) *Molybdenum*: Dissolve 0.2043 g (NH₄)₂ MoO₄ in water and dilute to 1000 mL; 1.00 mL = 100 μg Mo.

6) *Osmium*: Obtain standard 0.1M osmium tetroxide solution* and store in glass bottle; 1.00 mL = 19.02 mg Os. Make

dilutions daily as needed using 1% (v/v) H₂SO₄. CAUTION: OsO₄ is extremely toxic and highly volatile.

7) *Rhenium*: Dissolve 0.1554 g potassium perrhenate, KReO₄, in 200 mL water. Dilute to 1000 mL with 1% (v/v) H₂SO₄; 1.00 mL = 100 μg Re.

8) *Silica*: Do not dry. Dissolve 0.4730 g Na₂SiO₃ · 9H₂O in water. Add 10.0 mL conc HNO₃ and dilute to 1000 mL with water. 1.00 mL = 100 μg SiO₂. Store in polyethylene.

9) *Thorium*: Dissolve 0.238 g thorium nitrate, Th(NO₃)₄ · 4H₂O in 1000 mL water; 1.00 mL = 100 μg Th.

10) *Titanium*: Dissolve 0.3960 g pure (99.8 or 99.9%) titanium chloride, TiCl₄,† in a mixture of equal volumes of 1N HCl and 1N HF. Make up to 1000 mL with this acid mixture; 1.00 mL = 100 μg Ti.

11) *Vanadium*: Dissolve 0.2297 g ammonium metavanadate, NH₄VO₃, in a minimum amount of conc HNO₃. Heat to dissolve. Add 10 mL conc HNO₃, and dilute to 1000 mL with water; 1.00 mL = 100 μg V.

4. Procedure

a. *Sample preparation*: See Section 3111B.4a.

When determining Al, Ba, or Ti, mix 2 mL KCl solution into 100 mL sample and standards before aspiration. When determining Mo and V, mix 2 mL Al(NO₃)₃ · 9H₂O into 100 mL sample and standards before aspiration.

b. *Instrument operation*: See Section 3111B.4b. After adjusting wavelength, install a nitrous oxide burner head. Turn on acetylene (without igniting flame) and adjust flow rate to value specified by manufacturer for a nitrous oxide-acetylene flame. Turn off acetylene. With both air and nitrous oxide supplies turned on, set T-junction valve to nitrous oxide and adjust flow rate according to manufacturer's specifications. Turn switching valve to the air position and verify that flow rate is the same. Turn acetylene on and ignite to a bright yellow flame. With a rapid motion, turn switching valve to nitrous oxide. The flame should have a red cone above the burner. If it does not, adjust fuel flow to obtain red cone. After nitrous oxide flame has been ignited, let burner come to thermal equilibrium before beginning analysis.

Aspirate a blank consisting of deionized water containing 1.5 mL conc HNO₃/L and check aspiration rate. Adjust if necessary to a rate between 3 and 5 mL/min. Zero the instrument. Aspirate a standard of the desired metal with a concentration near the midpoint of the optimum concentration range and adjust burner (both horizontally and vertically) in the light path to obtain maximum response. Aspirate blank again and re-zero the instrument. The instrument now is ready to run standards and samples.

To extinguish flame, turn switching valve from nitrous oxide to air and turn off acetylene. This procedure eliminates the danger of flashback that may occur on direct ignition or shutdown of nitrous oxide and acetylene. (See also discussion in 3111B.4b.)

c. *Standardization*: Select at least three concentrations of standard metal solutions (prepared as in ¶ 3k) to bracket the expected metal concentration of a sample. Aspirate each in turn into the flame and record absorbances.

Most modern instruments are equipped with microprocessors and digital readout which permit calibration in direct concentra-

* GFS Chemicals, Inc., Columbus, OH, Cat. No. 64, or equivalent.

† Alpha Ventron, P.O. Box 299, 152 Andover St., Danvers, MA 01923, or equivalent.

tion terms. If instrument is not so equipped, prepare a calibration curve by plotting on linear graph paper absorbance of standards versus concentration. Plot calibration curves for Al, Ba, and Ti based on original concentration of standard before adding KCl solution. Plot calibration curves for Mo and V based on original concentration of standard before adding $\text{Al}(\text{NO}_3)_3$ solution.

d. Analysis of samples: Rinse atomizer by aspirating water containing 1.5 mL conc HNO_3/L and zero instrument. Aspirate a sample and determine its absorbance.

5. Calculations

Calculate concentration of each metal ion in micrograms per liter by referring to the appropriate calibration curve prepared according to ¶ 4c.

3111 E. Extraction/Nitrous Oxide-Acetylene Flame Method

1. General Discussion

a. Application: This method is suitable for the determination of aluminum at concentrations less than 900 $\mu\text{g}/\text{L}$ and beryllium at concentrations less than 30 $\mu\text{g}/\text{L}$. The method consists of chelation with 8-hydroxyquinoline, extraction with methyl isobutyl ketone (MIBK), and aspiration into a nitrous oxide-acetylene flame.

b. Interferences: Concentrations of Fe greater than 10 mg/L interfere by suppressing Al absorption. Iron interference can be masked by addition of hydroxylamine hydrochloride/1,10-phenanthroline. Mn concentrations up to 80 mg/L do not interfere if turbidity in the extract is allowed to settle. Mg forms an insoluble chelate with 8-hydroxyquinoline at pH 8.0 and tends to remove Al complex as a coprecipitate. However, the Mg complex forms slowly over 4 to 6 min; its interference can be avoided if the solution is extracted immediately after adding buffer.

2. Apparatus

Atomic absorption spectrometer and associated equipment: See Section 3111A.6.

3. Reagents

- Air:* See 3111B.3a.
- Acetylene:* See 3111B.3b.
- Ammonium hydroxide,* NH_4OH , conc.
- Buffer:* Dissolve 300 g ammonium acetate, $\text{NH}_4\text{C}_2\text{H}_3\text{O}_2$, in water, add 105 mL conc NH_4OH , and dilute to 1 L.
- Metal-free water:* See 3111B.3c.
- Hydrochloric acid,* HCl, conc.
- 8-Hydroxyquinoline solution:* Dissolve 20 g 8-hydroxyquinoline in about 200 mL water, add 60 mL glacial acetic acid, and dilute to 1 L with water.
- Methyl isobutyl ketone:* See 3111C.3d.
- Nitric acid,* HNO_3 , conc.

Alternatively, read the concentration directly from the instrument readout if the instrument is so equipped. If sample has been diluted, multiply by the appropriate dilution factor.

6. Bibliography

WILLIS, J.B. 1965. Nitrous oxide-acetylene flame in atomic absorption spectroscopy. *Nature* 207:715.

Also see Section 3111A.8 and 9.

j. Nitrous oxide: See 3111D.3h.

k. Standard metal solutions: Prepare a series of standard metal solutions containing 5 to 1000 $\mu\text{g}/\text{L}$ by appropriate dilution of the stock metal solutions prepared according to 3111D.3k.

l. Iron masking solution: Dissolve 1.3 g hydroxylamine hydrochloride and 6.58 g 1,10-phenanthroline monohydrate in about 500 mL water and dilute to 1 L with water.

4. Procedure

a. Instrument operation: See Sections 3111B.4b, C.4a, and D.4b. After final adjusting of burner position, aspirate MIBK into flame and gradually reduce fuel flow until flame is similar to that before aspiration of solvent. Adjust wavelength setting according to Table 3111:I.

b. Standardization: Select at least three concentrations of standard metal solutions (prepared as in ¶ 3k) to bracket the expected metal concentration of a sample and transfer 100 mL of each (and 100 mL water blank) to four different 200-mL volumetric flasks. Add 2 mL 8-hydroxyquinoline solution, 2 mL masking solution (if required), and 10 mL buffer to one flask, immediately add 10 mL MIBK, and shake vigorously. The duration of shaking affects the forms of aluminum complexed. A fast, 10-s shaking time favors monomeric Al, whereas 5 to 10 min of shaking also will complex polymeric species. Adjustment of the 8-hydroxyquinoline to sample ratio can improve recoveries of extremely high or low concentrations of aluminum. Treat each blank, standard, and sample in similar fashion. Continue as in Section 3111C.4b.

c. Analysis of samples: Rinse atomizer by aspirating water-saturated MIBK. Aspirate extracts of samples treated as above, and record absorbances.

5. Calculations

Calculate concentration of each metal in micrograms per liter by referring to the appropriate calibration curve prepared according to ¶ 4b.

3112 METALS BY COLD-VAPOR ATOMIC ABSORPTION SPECTROMETRY*

3112 A. Introduction

For general introductory material on atomic absorption spectrometric methods, see Section 3111A.

* Approved by Standard Methods Committee, 1999.

3112 B. Cold-Vapor Atomic Absorption Spectrometric Method

1. General Discussion

This method is applicable to the determination of mercury.

2. Apparatus

When possible, dedicate glassware for use in Hg analysis. Avoid using glassware previously exposed to high levels of Hg, such as those used in COD, TKN, or Cl^- analysis.

a. Atomic absorption spectrometer and associated equipment: See Section 3111A.6. Instruments and accessories specifically designed for measurement of mercury by the cold vapor technique are available commercially and may be substituted.

b. Absorption cell, a glass or plastic tube approximately 2.5 cm in diameter. An 11.4-cm-long tube has been found satisfactory but a 15-cm-long tube is preferred. Grind tube ends perpendicular to the longitudinal axis and cement quartz windows in place. Attach gas inlet and outlet ports (6.4 mm diam) 1.3 cm from each end.

c. Cell support: Strap cell to the flat nitrous-oxide burner head or other suitable support and align in light beam to give maximum transmittance.

d. Air pumps: Use any peristaltic pump with electronic speed control capable of delivering 2 L air/min. Any other regulated compressed air system or air cylinder also is satisfactory.

e. Flowmeter, capable of measuring an air flow of 2 L/min.

f. Aeration tubing, a straight glass frit having a coarse porosity for use in reaction flask.

g. Reaction flask, 250-mL erlenmeyer flask or a BOD bottle, fitted with a rubber stopper to hold aeration tube.

h. Drying tube, 150-mm \times 18-mm-diam, containing 20 g $\text{Mg}(\text{ClO}_4)_2$. A 60-W light bulb with a suitable shade may be substituted to prevent condensation of moisture inside the absorption cell. Position bulb to maintain cell temperature at 10°C above ambient.

i. Connecting tubing, glass tubing to pass mercury vapor from reaction flask to absorption cell and to interconnect all other components. Clear vinyl plastic* tubing may be substituted for glass.

3. Reagents†

a. Metal-free water: See 3111B.3c.

* Tygon or equivalent.

† Use specially prepared reagents low in mercury.

b. Stock mercury solution: Dissolve 0.1354 g mercuric chloride, HgCl_2 , in about 70 mL water, add 1 mL conc HNO_3 , and dilute to 100 mL with water; 1.00 mL = 1.00 mg Hg.

c. Standard mercury solutions: Prepare a series of standard mercury solutions containing 0 to 5 $\mu\text{g/L}$ by appropriate dilution of stock mercury solution with water containing 10 mL conc HNO_3/L . Prepare standards daily.

d. Nitric acid, HNO_3 , conc.

e. Potassium permanganate solution: Dissolve 50 g KMnO_4 in water and dilute to 1 L.

f. Potassium persulfate solution: Dissolve 50 g $\text{K}_2\text{S}_2\text{O}_8$ in water and dilute to 1 L.

g. Sodium chloride-hydroxylamine sulfate solution: Dissolve 120 g NaCl and 120 g $(\text{NH}_2\text{OH})_2 \cdot \text{H}_2\text{SO}_4$ in water and dilute to 1 L. A 10% hydroxylamine hydrochloride solution may be substituted for the hydroxylamine sulfate.

h. Stannous ion (Sn^{2+}) solution: Use either stannous chloride, ¶ 1), or stannous sulfate, ¶ 2), to prepare this solution containing about 7.0 g $\text{Sn}^{2+}/100$ mL.

1) Dissolve 10 g SnCl_2 in water containing 20 mL conc HCl and dilute to 100 mL.

2) Dissolve 11 g SnSO_4 in water containing 7 mL conc H_2SO_4 and dilute to 100 mL.

Both solutions decompose with aging. If a suspension forms, stir reagent continuously during use. Reagent volume is sufficient to process about 20 samples; adjust volumes prepared to accommodate number of samples processed.

i. Sulfuric acid, H_2SO_4 , conc.

4. Procedure

a. Instrument operation: See Section 3111B.4b. Set wavelength to 253.7 nm. Install absorption cell and align in light path to give maximum transmission. Connect associated equipment to absorption cell with glass or vinyl plastic tubing as indicated in Figure 3112:1. Turn on air and adjust flow rate to 2 L/min. Allow air to flow continuously. Alternatively, follow manufacturer's directions for operation. NOTE: Fluorescent lighting may increase baseline noise.

b. Standardization: Transfer 100 mL of each of the 1.0, 2.0, and 5.0 $\mu\text{g/L}$ Hg standard solutions and a blank of 100 mL water to 250-mL erlenmeyer reaction flasks. Add 5 mL conc H_2SO_4 and 2.5 mL conc HNO_3 to each flask. Add 15 mL KMnO_4 solution to each flask and let stand at least 15 min. Add 8 mL




Figure 3112:1. Schematic arrangement of equipment for measurement of mercury by cold-vapor atomic absorption technique.

$K_2S_2O_8$ solution to each flask and heat for 2 h in a water bath at $95^\circ C$. Cool to room temperature.

Treating each flask individually, add enough NaCl-hydroxylamine solution to reduce excess $KMnO_4$, then add 5 mL $SnCl_2$ or $SnSO_4$ solution and immediately attach flask to aeration apparatus. As Hg is volatilized and carried into the absorption cell, absorbance will increase to a maximum within a few seconds. As soon as recorder returns approximately to the base line, remove stopper holding the frit from reaction flask, and replace with a flask containing water. Flush system for a few seconds and run the next standard in the same manner. Construct a standard curve by plotting peak height versus micrograms Hg.

c. Analysis of samples: Transfer 100 mL sample or portion diluted to 100 mL containing not more than $5.0 \mu g$ Hg/L to a reaction flask. Treat as in ¶ 4b. Seawaters, brines, and effluents high in chlorides require as much as an additional 25 mL $KMnO_4$ solution. During oxidation step, chlorides are converted to free chlorine, which absorbs at 253 nm. Remove all free chlorine before the Hg is reduced and swept into the cell by using an excess (25 mL) of hydroxylamine reagent.

Remove free chlorine by sparging sample gently with air or nitrogen after adding hydroxylamine reducing solution. Use a separate tube and frit to avoid carryover of residual stannous chloride, which could cause reduction and loss of mercury.

TABLE 3112:I. INTERLABORATORY PRECISION AND BIAS OF COLD-VAPOR ATOMIC ABSORPTION SPECTROMETRIC METHOD FOR MERCURY¹

Form	Conc. $\mu g/L$	SD $\mu g/L$	Relative SD %	Relative Error %	No. of Participants
Inorganic	0.34	0.077	22.6	21.0	23
Inorganic	4.2	0.56	13.3	14.4	21
Organic	4.2	0.36	8.6	8.4	21

5. Calculation

Determine peak height of sample from recorder chart and read mercury value from standard curve prepared according to ¶ 4b.

6. Precision and Bias

Data on interlaboratory precision and bias for this method are given in Table 3112:I.

7. Reference

1. KOPP, J.F., M.C. LONGBOTTOM & L.B. LOBRING. 1972. "Cold vapor" method for determining mercury. *J. Amer. Water Works Assoc.* 64:20.

8. Bibliography

- HATCH, W.R. & W.L. OTT. 1968. Determination of submicrogram quantities of mercury by atomic absorption spectrophotometry. *Anal. Chem.* 40:2085.
- UTHE, J.F., F.A.J. ARMSTRONG & M.P. STANTON. 1970. Mercury determination in fish samples by wet digestion and flameless atomic absorption spectrophotometry. *J. Fish. Res. Board Can.* 27:805.
- FELDMAN, C. 1974. Preservation of dilute mercury solutions. *Anal. Chem.* 46:99.
- BOTHNER, M.H. & D.E. ROBERTSON. 1975. Mercury contamination of sea water samples stored in polyethylene containers. *Anal. Chem.* 47:592.
- HAWLEY, J.E. & J.D. INGLE, JR. 1975. Improvements in cold vapor atomic absorption determination of mercury. *Anal. Chem.* 47:719.
- LO, J.M. & C.M. WAL. 1975. Mercury loss from water during storage: Mechanisms and prevention. *Anal. Chem.* 47:1869.
- EL-AWADY, A.A., R.B. MILLER & M.J. CARTER. 1976. Automated method for the determination of total and inorganic mercury in water and wastewater samples. *Anal. Chem.* 48:110.
- ODA, C.E. & J.D. INGLE, JR. 1981. Speciation of mercury by cold vapor atomic absorption spectrometry with selective reduction. *Anal. Chem.* 53:2305.
- SUDDENDORF, R.F. 1981. Interference by selenium or tellurium in the determination of mercury by cold vapor generation atomic absorption spectrometry. *Anal. Chem.* 53:2234.
- HEIDEN, R.W. & D.A. AIKENS. 1983. Humic acid as a preservative for trace mercury (II) solutions stored in polyolefin containers. *Anal. Chem.* 55:2327.
- CHOU, H.N. & C.A. NALEWAY. 1984. Determination of mercury by cold vapor atomic absorption spectrometry. *Anal. Chem.* 56:1737.

3113 METALS BY ELECTROTHERMAL ATOMIC ABSORPTION SPECTROMETRY*

3113 A. Introduction

1. Applications

Electrothermal atomic absorption permits determination of most metallic elements with sensitivities and detection limits from 20 to 1000 times better than those of conventional flame techniques without extraction or sample concentration. This increase in sensitivity results from an increase in atom density within the furnace as compared to flame atomic absorption. Many elements can be determined at concentrations as low as 1.0 $\mu\text{g/L}$. An additional advantage of electrothermal atomic absorption is that only a very small volume of sample is required.

The electrothermal technique is used only at concentration levels below the optimum range of direct flame atomic absorption because it is subject to more interferences than the flame procedure and requires increased analysis time. The method of standard additions may be required to insure validity of data. Because of the high sensitivity of this technique, it is extremely susceptible to contamination; extra care in sample handling and analysis may be required.

2. Principle

Electrothermal atomic absorption spectroscopy is based on the same principle as direct flame atomization but an electrically heated atomizer or graphite furnace replaces the standard burner head. A discrete sample volume is dispensed into the graphite sample tube (or cup). Typically, determinations are made by heating the sample in three or more stages. First, a low current heats the tube to dry the sample. The second, or charring, stage destroys organic matter and volatilizes other matrix components at an intermediate temperature. Finally, a high current heats the tube to incandescence and, in an inert atmosphere, atomizes the element being determined. Additional stages frequently are added to aid in drying and charring, and to clean and cool the tube between samples. The resultant ground-state atomic vapor absorbs monochromatic radiation from the source. A photoelectric detector measures the intensity of transmitted radiation. The inverse of the transmittance is related logarithmically to the absorbance, which is directly proportional to the number density of vaporized ground-state atoms (the Beer-Lambert law) over a limited concentration range.

3. Interferences

Electrothermal atomization determinations may be subject to significant interferences from molecular absorption as well as chemical and matrix effects. Molecular absorption may occur when components of the sample matrix volatilize during atom-

ization, resulting in broadband absorption. Several background correction techniques are available commercially to compensate for this interference. A continuum source such as a deuterium arc can correct for background up to absorbance levels of about 0.8. Continuum lamp intensity diminishes at long wavelengths and use of continuum background correction is limited to analytical wavelengths below 350 nm. Zeeman effect background correctors can handle background absorbance up to 1.5 to 2.0. The Smith-Hieftje correction technique can accommodate background absorbance levels as large as 2.5 to 3.0 (see Section 3111A.3). Both Zeeman and Smith-Hieftje background corrections are susceptible to rollover (development of a negative absorbance-concentration relationship) at high absorbances. The rollover absorbance for each element should be available in the manufacturer's literature. Curvature due to rollover should become apparent during calibration; dilution produces a more linear calibration plot. Use background correction when analyzing samples containing high concentrations of acid or dissolved solids and in determining elements for which an absorption line below 350 nm is used.

Matrix modification can be useful in minimizing interference and increasing analytical sensitivity. Determine need for a modifier by evaluating recovery of a sample with a known addition. Recovery near 100% indicates that sample matrix does not affect analysis. Chemical modifiers generally modify relative volatilities of matrix and metal. Some modifiers enhance matrix removal, isolating the metal, while other modifiers inhibit metal volatilization, allowing use of higher ashing/charring temperatures and increasing efficiency of matrix removal. Chemical modifiers are added at high concentration (percent level) and can lead to sample contamination from impurities in the modifier solution. Heavy use of chemical modifiers may reduce the useful life (normally 50 to 100 firings) of the graphite tube. Some specific chemical modifiers and approximate concentrations are listed in Table 3113:I.

Addition of a chemical modifier directly to the sample before analysis is restricted to inexpensive additives (e.g. phosphoric acid). Use of palladium salts for matrix modification normally requires methods of co-addition, in which sample and modifier are added consecutively to the furnace either manually or, preferably, with an automatic sampler. Palladium salts (nitrate is preferred, chloride is acceptable) are listed in Table 3113:I as a modifier for many metals. The palladium solution (50 to 2000 mg/L) generally includes citric or ascorbic acid, which aids reduction of palladium in the furnace. Citric acid levels of 1 to 2% are typical. Use of hydrogen (5%) in the coolant gas (available commercially as a mixture) also reduces palladium, eliminating need for organic reducing acids. *CAUTION: Do not mix hydrogen and other gases in the laboratory; hydrogen gas is very flammable—handle with caution.* Use low levels of palladium (50 to 250 mg/L) for normal samples and higher levels for complex samples. Addition of excess palladium modifier may widen atomization peaks; in such cases peak area measurements

* Approved by Standard Methods Committee, 1999.
Joint Task Group: 20th Edition—Raymond J. Lovett (chair), David J. Kaptain.

TABLE 3113:I. POTENTIAL MATRIX MODIFIERS FOR ELECTROTHERMAL ATOMIC ABSORPTION SPECTROMETRY*

Modifier	Analyses for Which Modifier May Be Useful
1500 mg Pd/L + 1000 mg Mg(NO ₃) ₂ /L ¹	Ag, As, Au, Bi, Cu, Ge, Mn, Hg, In, Sb, Se, Sn, Te, Tl
500–2000 mg Pd/L + reducing agent ^{2†}	Ag, As, Bi, Cd, Co, Cr, Cu, Fe, Hg, Mn, Ni, Pb, Sb
5000 mg Mg(NO ₃) ₂ /L ¹	Be, Co, Cr, Fe, Mn, V
100–500 mg Pd/L ²	As, Ga, Ge, Sn
50 mg Ni/L ²	As, Se, Sb
2% PO ₄ ³⁻ + 1000 mg Mg (NO ₃) ₂ /L ¹	Cd, Pb

*Assumes 10 µL modifier/10 µL sample.

†Citric acid (1–2%) preferred; ascorbic acid or H₂ acceptable.

may provide higher quality results. The recommended mode of modifier use is through co-addition to the furnace of about 10 µL of the palladium (or other) modifier solution. Palladium may not be the best modifier in all cases and cannot be recommended unconditionally. Test samples requiring a modifier first with palladium; test other modifiers only if palladium is unsuccessful or to minimize modifier cost. See Section 3113B.3 for preparation of modifier solution.

Temperature ramping, i.e., gradual heating, can be used to decrease background interferences and permits analysis of samples with complex matrices. Ramping permits a controlled, continuous increase of furnace temperature in any of the various steps of the temperature sequence. Ramp drying is used for samples containing mixtures of solvents or for samples with a high salt content (to avoid spattering). If spattering is suspected, develop drying ramp by visual inspection of the drying stage, using a mirror. Samples that contain a complex mixture of matrix components sometimes require ramp charring to effect controlled, complete thermal decomposition. Ramp atomization may minimize background absorption by permitting volatilization of the element being determined before the matrix. This is especially applicable in the determination of such volatile elements as cadmium and lead. Use of time-resolved absorbance profiles (available on most modern instruments) greatly aids method development. Changes in atomization, notably the element peak appearance time and magnitude of background and metal absorbances, can be monitored directly.

Improve analysis by using a graphite platform, inserted into the graphite tube, as the atomization site. The platform is not heated as directly by the current flowing through the graphite tube; thus the metal atomizes later and under more uniform conditions.

Use standard additions to compensate for matrix interferences. When making standard additions, determine whether the added metal and that in the sample behave similarly under the specified conditions. [See Section 3113B.4d2)]. In the extreme, test every sample for recovery (85 to 115% recovery desired) to determine if standard addition is needed. Test every new sample type for recovery. Recovery of only 40 to 85% generally indicates that standard addition is required. Often, as long as the samples are from sources of consistent properties, a representative recovery can be used to characterize the analysis and determine the

TABLE 3113:II. DETECTION LEVELS AND CONCENTRATION RANGES FOR ELECTROTHERMAL ATOMIZATION ATOMIC ABSORPTION SPECTROMETRY

Element	Wavelength nm	Estimated Detection Level µg/L	Optimum Concentration Range µg/L
Al	309.3	3	20–200
Sb	217.6	3	20–300
As	193.7	1	5–100
Ba	553.6	2	10–200
Be	234.9	0.2	1–30
Cd	228.8	0.1	0.5–10
Cr	357.9	2	5–100
Co	240.7	1	5–100
Cu	324.7	1	5–100
Fe	248.3	1	5–100
Pb*	283.3	1	5–100
Mn	279.5	0.2	1–30
Mo	313.3	1	3–60
Ni	232.0	1	5–100
Se	196.0	2	5–100
Ag	328.1	0.2	1–25
Sn	224.6	5	20–300

*The more sensitive 217.0-nm wavelength is recommended for instruments with background correction capabilities.

necessity of standard addition. Test samples of unknown origin or of complex composition (digestates, for example) individually for metal recovery. Ideally, chemical modifiers and graphite platforms render the sample fit to be analyzed using a standard analytical calibration curve. Always verify this assumption; however, a properly developed method with judicious use of chemical modifiers should eliminate the necessity for standard addition in all but the most extreme samples.

Chemical interaction of the graphite tube with various elements to form refractory carbides occurs at high charring and atomization temperatures. Elements that form carbides are barium, molybdenum, nickel, titanium, vanadium, and silicon. Carbide formation is characterized by broad, tailing atomization peaks and reduced sensitivity. Using pyrolytically coated tubes for these metals minimizes the problem.

4. Sensitivity, Detection Levels, and Optimum Concentration Range

Estimated detection levels and optimum concentration ranges are listed in Table 3113:II. These values may vary with the chemical form of the element being determined, sample composition, or instrumental conditions.

For a given sample, increased sensitivity may be achieved by using a larger sample volume or by reducing flow rate of the purge gas or by using gas interrupt during atomization. Note, however, that these techniques also will increase the effects of any interferences present. Sensitivity can be decreased by diluting the sample, reducing sample volume, increasing purge-gas flow, or using a less sensitive wavelength. Use of argon, rather than nitrogen, as the purge gas generally improves sensitivity and reproducibility. Hydrogen mixed with the inert gas may suppress chemical interference and increase sensitivity by acting

as a reducing agent, thereby aiding in producing more ground-state atoms. Pyrolytically coated graphite tubes can increase sensitivity for the more refractory elements and are recommended. The optical pyrometer/maximum power accessory available on some instruments also offers increased sensitivity with lower atomization temperatures for many elements.

Using the Stabilized Temperature Platform Furnace (STPF) technique, which is a combination of individual techniques, also offers significant interference reduction with improved sensitivity. Sensitivity changes with sample tube age. Discard graphite tubes when significant variations in sensitivity or poor reproducibility are observed. The use of high acid concentrations, brine samples, and matrix modifiers often drastically reduces tube life. Preferably use the graphite platform in such situations.

5. References

1. PERKIN-ELMER CORP. 1991. Summary of Standard Conditions for Graphite Furnace. Perkin-Elmer Corp., Norwalk, Conn.
2. ROTHERY, E., ed. 1988. Analytical Methods for Graphite Tube Atomizers. Varian Techtron Pty, Ltd., Mulgrave, Victoria, Australia.

6. Bibliography

- FERNANDEZ, F.J. & D.C. MANNING. 1971. Atomic absorption analyses of metal pollutants in water using a heated graphite atomizer. *Atomic Absorption Newsletter* 10:65.
- SEGAR, D.A. & J.G. GONZALEZ. 1972. Evaluation of atomic absorption with a heated graphite atomizer for the direct determination of trace transition metals in sea water. *Anal. Chim. Acta* 58:7.

- BARNARD, W.M. & M.J. FISHMAN. 1973. Evaluation of the use of heated graphite atomizer for the routine determination of trace metals in water. *Atomic Absorption Newsletter* 12:118.
- KAHN, H.L. 1973. The detection of metallic elements in wastes and waters with the graphite furnace. *Int. J. Environ. Anal. Chem.* 3:121.
- WALSH, P.R., J.L. FASCHING & R.A. DUCE. 1976. Matrix effects and their control during the flameless atomic absorption determination of arsenic. *Anal. Chem.* 48:1014.
- HENN, E.L. 1977. Use of Molybdenum in Eliminating Matrix Interferences in Flameless Atomic Absorption. Spec. Tech. Publ. 618, American Soc. Testing & Materials, Philadelphia, Pa.
- MARTIN, T.D. & J.F. KOPP. 1978. Methods for Metals in Drinking Water. U.S. Environmental Protection Agency, Environmental Monitoring and Support Lab., Cincinnati, Ohio.
- HYDES, D.J. 1980. Reduction of matrix effects with a soluble organic acid in the carbon furnace atomic absorption spectrometric determination of cobalt, copper, and manganese in seawater. *Anal. Chem.* 52:289.
- SOTERA, J.J. & H.L. KAHN. 1982. Background correction in AAS. *Amer. Lab.* 14:100.
- SMITH, S.B. & G.M. HIEFTJE. 1983. A new background-correction method for atomic absorption spectrometry. *Appl. Spectrosc.* 37:419.
- GROSSER, Z. 1985. Techniques in Graphite Furnace Atomic Absorption Spectrophotometry. Perkin-Elmer Corp., Ridgefield, Conn.
- SLAVIN, W. & G.R. CARNICK. 1985. A survey of applications of the stabilized temperature platform furnace and Zeeman correction. *Atomic Spectrosc.* 6:157.
- BRUEGGEMEYER, T. & F. FRICKE. 1986. Comparison of furnace & atomization behavior of aluminum from standard & thorium-treated L'vov platforms. *Anal. Chem.* 58:1143.

3113 B. Electrothermal Atomic Absorption Spectrometric Method

1. General Discussion

This method is suitable for determination of micro quantities of aluminum, antimony, arsenic, barium, beryllium, cadmium, chromium, cobalt, copper, iron, lead, manganese, molybdenum, nickel, selenium, silver, and tin. It is also applicable to analysis of bismuth, gallium, germanium, gold, indium, mercury, tellurium, thallium, and vanadium, but precision and accuracy data are not yet available.

2. Apparatus

a. Atomic absorption spectrometer: See Section 3111A.6a. The instrument must have background correction capability.

b. Source lamps: See Section 3111A.6d.

c. Graphite furnace: Use an electrically heated device with electronic control circuitry designed to carry a graphite tube or cup through a heating program that provides sufficient thermal energy to atomize the elements of interest. Furnace heat controllers with only three heating steps are adequate only for fresh waters with low dissolved solids content. For salt waters, brines, and other complex matrices, use a furnace controller with up to seven individually programmed heating steps. Fit the furnace into the sample compartment of the spectrometer in place of the

conventional burner assembly. Use argon as a purge gas to minimize oxidation of the furnace tube and to prevent the formation of metallic oxides. Use graphite tubes with platforms to minimize interferences and to improve sensitivity.

d. Readout: See Section 3111A.6c.

e. Sample dispensers: Use microliter pipets (5 to 100 μ L) or an automatic sampling device designed for the specific instrument.

f. Vent: See Section 3111A.6f.

g. Cooling water supply: Cool with tap water flowing at 1 to 4 L/min or use a recirculating cooling device.

h. Membrane filter apparatus: Use an all-glass filtering device and 0.45- μ m or smaller-pore-diameter membrane filters. For trace analysis of aluminum, use polypropylene or TFE devices.

3. Reagents

a. Metal-free water: See Section 3111B.3c.

b. Hydrochloric acid, HCl, 1 + 1 and conc.

c. Nitric acid, HNO₃, 1 + 1 and conc.

d. Matrix modifier stock solutions:

1) *Magnesium nitrate*, 10 000 mg Mg/L: Dissolve 10.5 g Mg(NO₃)₂ · 6H₂O in water. Dilute to 100 mL.

2) *Nickel nitrate*, 10 000 mg Ni/L: Dissolve 4.96 g $\text{Ni}(\text{NO}_3)_2 \cdot 6\text{H}_2\text{O}$ in water. Dilute to 100 mL.

3) *Phosphoric acid*, 10% (v/v): Add 10 mL conc H_3PO_4 to water. Dilute to 100 mL.

4) *Palladium nitrate*, 4000 mg Pd/L: Dissolve 8.89 g $\text{Pd}(\text{NO}_3)_2 \cdot \text{H}_2\text{O}$ in water. Dilute to 1 L.

5) *Citric acid*, 4%: Dissolve 40 g citric acid in water. Dilute to 1 L.

NOTE: All of the modifier solutions recommended in Table 3113:I can be prepared with volumetric combination of the above solutions and water. For preparation of other matrix modifiers, see references or follow manufacturers' instructions.

e. Stock metal solutions: Refer to Sections 3111B and 3114.

f. Chelating resin: 100 to 200 mesh* purified by heating at 60°C in 10N NaOH for 24 h. Cool resin and rinse 10 times each with alternating portions of 1N HCl, metal-free water, 1N NaOH, and metal-free water.

g. Metal-free seawater (or brine): Fill a 1.4-cm-ID \times 20-cm-long borosilicate glass column to within 2 cm of the top with purified chelating resin. Elute resin with successive 50-mL portions of 1N HCl, metal-free water, 1N NaOH, and metal-free water at the rate of 5 mL/min just before use. Pass salt water or brine through the column at a rate of 5 mL/min to extract trace metals present. Discard the first 10 bed volumes (300 mL) of eluate.

4. Procedures

a. Sample pretreatment: Before analysis, pretreat all samples as indicated below. Rinse all glassware with 1 + 1 HNO_3 and water. Carry out digestion procedures in a clean, dust-free laboratory area to avoid sample contamination. For digestion of trace aluminum, use polypropylene or TFE utensils to avoid leachable aluminum from glassware.

1) Dissolved metals—See Section 3030B. For samples requiring arsenic and/or selenium analysis add 3 mL 30% hydrogen peroxide/100 mL sample and an appropriate volume of nickel nitrate solution (see Table 3113:I) before analysis. For all other metals no further pretreatment is required except for adding an optional matrix modifier.

2) Total recoverable metals (Al, Sb, Ba, Be, Cd, Cr, Co, Cu, Fe, Pb, Mn, Mo, Ni, Ag, and Sn)—NOTE: Sb and Sn are not recovered unless HCl is used in the digestion. See Section 3030D. Quantitatively transfer digested sample to a 100-mL volumetric flask, add an appropriate amount of matrix modifier (see Table 3113:I), and dilute to volume with water.

3) Total recoverable metals (As, Se)—Transfer 100 mL of shaken sample, 1 mL conc HNO_3 , and 2 mL 30% H_2O_2 to a clean, acid-washed 250-mL beaker. Heat on a hot plate without allowing solution to boil until volume has been reduced to about 50 mL. Remove from hot plate and let cool to room temperature. Add an appropriate concentration of nickel (see Table 3113:I), and dilute to volume in a 100-mL volumetric flask with water. Substitution of palladium is uneconomical. Nickel may be deleted if palladium is co-added during analysis. Simultaneously prepare a digested blank by substituting water for sample and proceed with digestion as described above.

b. Instrument operation: Mount and align furnace device according to manufacturer's instructions. Turn on instrument and data collection system. Select appropriate light source and adjust to recommended electrical setting. Select proper wavelength and set all conditions according to manufacturer's instructions, including background correction. Background correction is important when elements are determined at short wavelengths or when sample has a high level of dissolved solids. Background correction normally is not necessary at wavelengths longer than 350 nm. If background correction above 350 nm is needed deuterium arc background correction is not useful and other types must be used.

Select proper inert- or sheath-gas flow. In some cases, it is desirable to interrupt the inert-gas flow during atomization. Such interruption results in increased sensitivity by increasing residence time of the atomic vapor in the optical path. Gas interruption also increases background absorption and intensifies interference effects, but modern background correction methods usually eliminate these problems. Consider advantages and disadvantages of this option for each matrix when optimizing analytical conditions.

To optimize graphite furnace conditions, carefully adjust furnace temperature settings to maximize sensitivity and precision and to minimize interferences. Follow manufacturer's instructions.

Use drying temperatures slightly above the solvent boiling point and provide enough time and temperature for complete evaporation without boiling or spattering.

Select atomization temperature by determining the lowest temperature providing maximum sensitivity without significantly eroding precision. Optimize by a series of successive determinations at various atomization temperatures using a standard solution giving an absorbance of 0.2 to 0.5.

The charring temperature must be high enough to maximize volatilization of interfering matrix components yet too low to volatilize the element of interest. With the drying and atomization temperatures set to their optimum values, analyze a standard solution at a series of charring temperatures in increasing increments of 50 to 100°C. When the optimum charring temperature is exceeded, there will be a significant drop in sensitivity. Plot charring temperature versus sample absorbance; the optimum charring temperature is the highest temperature without reduced sensitivity. Verify optimization with major changes in sample matrix.

c. Instrument calibration: Prepare standard solutions for instrument calibration by diluting metal stock solutions. Prepare standard solutions fresh daily.

Prepare a blank and at least three calibration standards in the appropriate concentration range (see Table 3113:II) for correlating element concentration and instrument response. Match the matrix of the standard solutions to those of the samples as closely as possible. In most cases, this simply requires matching the acid background of the samples. For seawaters or brines, however, use the metal-free matrix (HCl 3g) as the standard solution diluent. In addition, add the same concentration of matrix modifier (if required for sample analysis) to the standard solutions.

Inject a suitable portion of each standard solution, in order of increasing concentration. Analyze each standard solution in triplicate to verify method precision.

*Chelex 100, or equivalent, available from Bio-Rad Laboratories, Richmond, CA.

Construct an analytical curve by plotting the average peak absorbances or peak areas of the standard solution versus concentration on linear graph paper. Alternatively, use electronic instrument calibration if the instrument has this capability.

d. *Sample analysis:* Analyze all samples except those demonstrated to be free of matrix interferences (based on recoveries of 85% to 115% for known additions) using the method of standard additions. Analyze all samples at least in duplicate or until

reproducible results are obtained. A variation of $\leq 10\%$ is considered acceptable reproducibility. Average replicate values.

1) Direct determination—Inject a measured portion of pretreated sample into the graphite furnace. Use the same volume as was used to prepare the calibration curve. Usually add modifier immediately after the sample, preferably using an automatic sampler or a micropipet. Some methods require modifier to be injected before the sample. Use the same volume and concen-

TABLE 3113:III. INTERLABORATORY SINGLE-ANALYST PRECISION DATA FOR ELECTROTHERMAL ATOMIZATION METHODS¹

Element	Concentration $\mu\text{g/L}$	Single-Analyst Precision % RSD								
		Lab Pure Water	Drinking Water	Surface Water	Effluent 1	Effluent 2	Effluent 3			
Al	28	66	108	70	—	—	66			
	125	27	35	24	—	—	34			
	11 000	11	—	—	22	—	—			
	58 300	27	—	—	19	—	—			
As	460	9	—	—	—	30	—			
	2 180	28	—	—	—	4	—			
	10.5	20	13	13	13	56	18			
	230	10	18	13	21	94	14			
Ba	9.78	40	25	15	74	23	11			
	227	10	6	8	11	15	6			
Be	56.5	36	21	29	59	23	27			
	418	14	12	20	24	24	18			
Cd	0.45	18	27	15	30	2	11			
	10.9	14	4	9	7	12	12			
Cr	0.43	72	49	1	121	35	27			
	12	11	17	22	14	11	15			
Co	9.87	24	33	10	23	15	10			
	236	16	7	11	13	16	7			
Cu	29.7	10	17	10	19	24	12			
	420	8	11	13	14	9	5			
Fe	10.1	49	47	17	17	—	30			
	234	8	15	6	21	—	11			
	300	6	—	—	—	11	—			
	1 670	11	—	—	—	6	—			
Mn	26.1	144	52	153	—	—	124			
	455	48	37	45	—	—	31			
	1 030	17	—	—	30	—	—			
	5 590	6	—	—	32	—	—			
Pb	370	14	—	—	—	19	—			
	2 610	9	—	—	—	18	—			
	10.4	6	19	17	21	19	33			
	243	17	7	17	18	12	16			
Ni	0.44	187	180	—	—	—	275			
	14.8	32	19	—	—	—	18			
	91.0	15	—	—	48	—	—			
	484.0	4	—	—	12	—	—			
Se	111.0	12	—	—	—	21	—			
	666.0	6	—	—	—	20	—			
	26.2	20	26	25	24	18	19			
	461.0	15	11	9	8	11	4			
Ag	10.0	12	27	16	35	41	13			
	235.0	6	6	15	6	13	14			
Zn	8.48	10	—	—	15	27	16			
	56.5	14	—	—	7	16	23			
	0.45	27	166	48	—	—	—			
Cd	13.6	15	4	10	—	—	—			

tration of modifier for all standards and samples. Dry, char, and atomize according to the preset program. Repeat until reproducible results are obtained.

Compare the average absorbance value or peak area to the calibration curve to determine concentration of the element of interest. Alternatively, read results directly if the instrument is equipped with this capability. If absorbance (or concentration) or peak area of the sample is greater than absorbance (concentra-

tion) or peak area of the most concentrated standard solution, dilute sample and reanalyze. If very large dilutions are required, another technique (e.g., flame AA or ICP) may be more suitable for this sample. Large dilution factors magnify small errors on final calculation. Keep acid background and concentration of matrix modifier (if present in the solutions) constant. Dilute the sample in a blank solution of acid and matrix modifiers.

TABLE 3113:IV. INTERLABORATORY OVERALL PRECISION DATA FOR ELECTROTHERMAL ATOMIZATION METHODS¹

Element	Concentration μg/L	Overall Precision % RSD					
		Lab Pure Water	Drinking Water	Surface Water	Effluent 1	Effluent 2	Effluent 3
Al	28	99	114	124	—	—	131
	125	45	47	49	—	—	40
	11 000	19	—	—	43	—	—
	58 300	31	—	—	32	—	—
	460	20	—	—	—	47	—
	2 180	30	—	—	—	15	—
As	10.5	37	19	22	50	103	39
	230	26	16	16	17	180	21
Ba	9.78	43	26	37	72	50	39
	227	18	12	13	20	15	14
Be	56.5	68	38	43	116	43	65
	418	35	35	28	38	48	16
Cd	0.45	28	31	15	67	50	35
	10.9	33	15	26	20	9	19
Cr	0.43	73	60	5	88	43	65
	12	19	25	41	26	20	27
Co	9.87	30	53	24	60	41	23
	236	18	14	24	20	14	20
Cu	29.7	13	26	17	18	21	17
	420	21	21	17	18	13	13
Fe	10.1	58	82	31	32	—	74
	234	12	33	19	21	—	26
	300	13	—	—	—	14	—
	1 670	12	—	—	—	13	—
Pb	26.1	115	93	306	—	—	204
	455	53	46	53	—	—	44
	1 030	32	—	—	25	—	—
	5 590	10	—	—	43	—	—
	370	28	—	—	—	22	—
Mn	2 610	13	—	—	—	22	—
	10.4	27	42	31	23	28	47
Ni	243	18	19	17	19	19	25
	0.44	299	272	—	—	—	248
	14.8	52	41	—	—	—	29
	91.0	16	—	—	45	—	—
	484.0	5	—	—	17	—	—
Se	111.0	15	—	—	—	17	—
	666.0	8	—	—	—	24	—
Ag	26.2	35	30	49	35	37	43
	461.0	23	22	15	12	21	17
Ag	10.0	17	48	32	30	44	51
	235.0	16	18	18	17	22	34
Ag	8.48	23	—	—	16	35	34
	56.5	15	—	—	24	32	28
	0.45	57	90	368	—	—	—
Ag	13.6	19	19	59	—	—	—

Proceed to ¶ 5a below.

2) Method of standard additions—Refer to ¶ 4c above. The method of standard additions is valid only when it falls in the linear portion of the calibration curve. Once instrument sensitivity has been optimized for the element of interest and the linear range for the element has been established, proceed with sample analyses.

Inject a measured volume of sample into furnace device. Dry, char or ash, and atomize samples according to preset program. Repeat until reproducible results are obtained. Record instrument response in absorbance or concentration as appropriate. Add a known concentration of the element of interest to a separate portion of sample so as not to change significantly the sample volume. Repeat the determination.

TABLE 3113:V. INTERLABORATORY RELATIVE ERROR DATA FOR ELECTROTHERMAL ATOMIZATION METHODS¹

Element	Concentration µg/L	Relative Error %					
		Lab Pure Water	Drinking Water	Surface Water	Effluent 1	Effluent 2	Effluent 3
Al	28.0	86	150	54	—	—	126
	125.0	4	41	39	—	—	30
	11 000.0	2	—	—	14	—	—
	58 300.0	12	—	—	7	—	—
	460.0	2	—	—	—	11	—
Sb	2 180.0	11	—	—	—	9	—
	10.5	30	32	28	24	28	36
As	230.0	35	14	19	13	73	39
	9.78	36	1	22	106	13	16
Ba	227.0	3	7	10	19	6	13
	56.5	132	54	44	116	59	40
Be	418.0	4	0	0	13	6	60
	0.45	40	16	11	16	10	15
Cd	10.9	13	2	9	7	8	8
	0.43	58	45	37	66	16	19
Cr	12.0	4	6	5	22	18	3
	9.87	10	9	4	2	5	15
Co	236.0	11	0	9	13	3	8
	29.7	7	7	1	6	5	13
Cu	420.0	12	8	8	11	5	18
	10.1	16	48	2	5	—	15
	234.0	8	7	0	4	—	19
	300.0	4	—	—	—	21	—
Fe	1 670.0	6	—	—	—	2	—
	26.1	85	60	379	—	—	158
	455.0	43	22	31	—	—	18
	1 030.0	8	—	—	8	—	—
	5 590.0	2	—	—	12	—	—
Pb	370.0	4	—	—	—	11	—
	2 610.0	35	—	—	—	2	—
	10.4	16	10	17	1	34	14
	243.0	5	15	8	18	15	29
Mn	0.44	332	304	—	—	—	556
	14.8	10	1	—	—	—	36
	91.0	31	—	—	10	—	—
	484.0	42	—	—	4	—	—
	111.0	1	—	—	—	29	—
Ni	666.0	6	—	—	—	23	—
	26.2	9	16	10	7	33	54
	461.0	15	19	18	31	16	18
Se	10.0	12	9	6	36	17	37
	235.0	7	7	0	13	10	17
Ag	8.48	12	—	—	1	51	20
	56.5	16	—	—	8	51	22
	0.45	34	162	534	—	—	—
	13.6	3	12	5	—	—	—

Add a known concentration (preferably twice that used in the first addition) to a separate sample portion. Mix well and repeat the determination.

Using linear graph paper, plot average absorbance or instrument response for the sample and the additions on the vertical axis against the concentrations of the added element on the horizontal axis, using zero as the concentration for the sample. Draw a straight line connecting the three points and extrapolate to zero absorbance. The intercept at the horizontal axis is the negative of the element concentration in the sample. The concentration axis to the left of the origin should be a mirror image of the axis to the right.

5. Calculations

a. Direct determination:

$$\mu\text{g metal/L} = C \times F$$

where:

C = metal concentration as read directly from the instrument or from the calibration curve, $\mu\text{g/L}$, and

F = dilution factor.

b. Method of additions:

$$\mu\text{g metal/L} = C \times F$$

where:

C = metal concentration as read from the method of additions plot, $\mu\text{g/L}$, and

F = dilution factor.

6. Precision and Bias

Data typical of the precision and bias obtainable are presented in Tables 3113:III, IV, and V.

7. Quality Control

See Section 3020 for specific quality control procedures to be followed during analysis. Although previous indications were that very low optimum concentration ranges were attainable for most metals (see Table 3113:II), data in Table 3113:III using variations of these protocols show that this may not be so. Exercise extreme care when applying this

method to the lower concentration ranges. Verify analyst precision at the beginning of each analytical run by making triplicate analyses. Verify autosampler precision by checking volumes (by weight) delivered by the autosampler at routinely used injection volume settings.

8. Reference

1. COPELAND, T.R. & J.P. MANEY. 1986. EPA Method Study 31: Trace Metals by Atomic Absorption (Furnace Techniques). EPA-600/S4-85-070, U.S. Environmental Protection Agency, Environmental Monitoring and Support Lab., Cincinnati, Ohio.

9. Bibliography

- RENSHAW, G.D. 1973. The determination of barium by flameless atomic absorption spectrophotometry using a modified graphite tube atomizer. *Atomic Absorption Newsletter* 12:158.
- YANAGISAWA, M., T. TAKEUCHI & M. SUZUKI. 1973. Flameless atomic absorption spectrometry of antimony. *Anal. Chim. Acta* 64:381.
- RATTONETTI, A. 1974. Determination of soluble cadmium, lead, silver and indium in rainwater and stream water with the use of flameless atomic absorption. *Anal. Chem.* 46:739.
- HENN, E.L. 1975. Determination of selenium in water and industrial effluents by flameless atomic absorption. *Anal. Chem.* 47:428.
- MARTIN, T.D. & J.F. KOPP. 1975. Determining selenium in water, wastewater, sediment and sludge by flameless atomic absorption spectrometry. *Atomic Absorption Newsletter* 14:109.
- MARUTA, T., K. MINEGISHI & G. SUDOH. 1976. The flameless atomic absorption spectrometric determination of aluminum with a carbon atomization system. *Anal. Chim. Acta* 81:313.
- CRANSTON, R.E. & J.W. MURRAY. 1978. The determination of chromium species in natural waters. *Anal. Chim. Acta* 99:275.
- HOFFMEISTER, W. 1978. Determination of iron in ultrapure water by atomic absorption spectroscopy. *Z. Anal. Chem.* 50:289.
- LAGAS, P. 1978. Determination of beryllium, barium, vanadium and some other elements in water by atomic absorption spectrometry with electrothermal atomization. *Anal. Chim. Acta* 98:261.
- CARRONDO, M.J.T., J.N. LESTER & R. PERRY. 1979. Electrothermal atomic absorption determination of total aluminum in waters and waste waters. *Anal. Chim. Acta* 111:291.
- NAKAHARA, T. & C.L. CHAKRABARTI. 1979. Direct determination of traces of molybdenum in synthetic sea water by atomic absorption spectrometry with electrothermal atomization and selective volatilization of the salt matrix. *Anal. Chim. Acta* 104:99.
- TIMINAGA, M. & Y. UMEZAKI. 1979. Determination of submicrogram amounts of tin by atomic absorption spectrometry with electrothermal atomization. *Anal. Chim. Acta* 110:55.

3114 ARSENIC AND SELENIUM BY HYDRIDE GENERATION/ATOMIC ABSORPTION SPECTROMETRY*

3114 A. Introduction

For general introductory material on atomic absorption spectrometric methods, see Section 3111A.

Two methods are presented in this section: A manual method and a continuous-flow method especially recommended for selenium. Continuous-flow automated systems are preferable to manual hydride generators because the effect of sudden hydrogen generation on light-path transparency is removed and any blank response from contamination of the HCl reagent by the elements being determined is incorporated into the background base line.

* Approved by Standard Methods Committee, 1997.

3114 B. Manual Hydride Generation/Atomic Absorption Spectrometric Method

1. General Discussion

a. Principle: This method is applicable to the determination of arsenic and selenium by conversion to their hydrides by sodium borohydride reagent and transport into an atomic absorption atomizer.

Arsenous acid and selenous acid, the As(III) and Se(IV) oxidation states of arsenic and selenium, respectively, are instantaneously converted by sodium borohydride reagent in acid solution to their volatile hydrides. The hydrides are purged continuously by argon or nitrogen into a quartz cell heated electrically or by the flame of an atomic absorption spectrometer and converted to the gas-phase atoms. The sodium borohydride reducing agent, by rapid generation of the elemental hydrides in an appropriate reaction cell, minimizes dilution of the hydrides by the carrier gas and provides rapid, sensitive determinations of arsenic and selenium.

CAUTION: Arsenic and selenium and their hydrides are toxic. Handle with care.

At room temperature and solution pH values of 1 or less, arsenic acid, the As(V) oxidation state of arsenic, is reduced relatively slowly by sodium borohydride to As(III), which is then instantaneously converted to arsine. The arsine atomic absorption peaks commonly are decreased by one-fourth to one-third for As(V) when compared to As(III). Determination of total arsenic requires that all inorganic arsenic compounds be in the As(III) state. Organic and inorganic forms of arsenic are first oxidized to As(V) by acid digestion. The As(V) then is quantitatively reduced to As(III) with sodium or potassium iodide before reaction with sodium borohydride.

Selenic acid, the Se(VI) oxidation state of selenium, is not measurably reduced by sodium borohydride. To determine total selenium by atomic absorption and sodium borohydride, first reduce Se(VI) formed during the acid digestion procedure to Se(IV), being careful to prevent reoxidation by chlorine. Efficiency of reduction depends on temperature, reduction time, and HCl concentration. For 4*N* HCl, heat 1 h at 100°C. For 6*N* HCl, boiling for 10 min is sufficient.¹⁻³ Alternatively, autoclave samples in sealed containers at 121°C for 1 h. NOTE: Autoclaving in sealed containers may result in incomplete reduction, apparently due to the buildup of chlorine gas. To obtain equal instrument responses for reduced Se(VI) and Se(IV) solutions of equal concentrations, manipulate HCl concentration and heating time. For further details, see Section 3500-Se.

b. Equipment selection: Certain atomic absorption atomizers and hydride reaction cells are available commercially for use with the sodium borohydride reagent. A functional manual system that can be constructed in the laboratory is presented in Figure 3114:1. Irrespective of the hydride reaction cell-atomizer system selected, it must meet the following quality-control considerations: (a) it must provide a precise and reproducible stan-




Figure 3114:1. Manual reaction cell for producing As and Se hydrides.

ard curve between 0 and 20 μg As or Se/L and an instrumental detection limit between 0.1 and 0.5 μg As or Se/L; (b) when carried through the entire procedure, oxidation state couples [As(III) - As(V) or Se(IV) - Se(VI)] must cause equal instrument response; and (c) sample digestion must yield 80% or greater recovery of added cacodylic acid (dimethyl arsine acid) and 90% or greater recovery of added As(III), As(V), Se(VI), or Se(IV).

Quartz atomization cells provide for the most sensitive arsenic and selenium hydride determinations. The quartz cell can be heated electrically or by an air-acetylene flame in an atomic absorption unit.

c. Digestion techniques: Waters and wastewaters may contain varying amounts of organic arsenic compounds and inorganic compounds of As(III), As(V), Se(IV), and Se(VI). To measure total arsenic and selenium in these samples requires sample digestion to solubilize particulate forms and oxidize reduced forms of arsenic and selenium and to convert any organic compounds to inorganic ones. Organic selenium compounds rarely

have been demonstrated in water. It is left to the experienced analyst's judgment whether sample digestion is required.

Various alternative digestion procedures are provided in ¶s 4c and d below. Consider sulfuric-nitric-perchloric acid digestion (¶ 4c) or sulfuric-nitric acid digestion (3030F) as providing a measure of total recoverable arsenic rather than total arsenic because they do not completely convert certain organic arsenic compounds to As(V). The sulfuric-nitric-perchloric acid digestion effectively destroys organics and most particulates in untreated wastewaters or solid samples, but does not convert all organic arsenicals to As(V). The potassium persulfate digestion (¶ 4d) is effective for converting organic arsenic and selenium compounds to As(V) and Se(VI) in potable and surface waters and in most wastewaters.⁴

The HCl-autoclave reduction of Se(VI) described above and in ¶ 4f is an effective digestion procedure for total inorganic selenium; however, it has not been found effective for converting benzene substituted selenium compounds to inorganic selenium. In all cases, verify the effectiveness of digestion methods by carrying samples with known additions of organic As or Se(IV) through the entire procedure.

d. *Interferences:* Interferences are minimized because the As and Se hydrides are removed from the solution containing most potential interfering substances. Slight response variations occur when acid matrices are varied. Control these variations by treating standards and samples in the same manner. Low concentrations of noble metals (approximately 100 µg/L of Ag, Au, Pt, Pd, etc.), concentrations of copper, lead, and nickel at or greater than 1 mg/L, and concentrations between 0.1 and 1 mg/L of hydride-forming elements (Bi, Sb, Sn, and Te) may suppress the response of As and Se hydrides. Interference by transition metals depends strongly on HCl concentration. Interferences are less pronounced at 4 to 6N HCl than at lower concentrations.⁵ The presence of As or Se in each other's matrices can cause similar suppression. Reduced nitrogen oxides resulting from HNO₃ digestion and nitrite also can suppress instrumental response for both elements. Large concentrations of iodide interfere with the Se determination by reducing Se to its elemental form. Do not use any glassware for determining Se that has been used for iodide reduction of As(V).

To prevent chlorine gas produced in the reduction of Se(VI) to Se(IV) from reoxidizing the Se(IV), generate the hydride within a few hours of the reduction steps or purge the chlorine from the samples by sparging.⁶

Interferences depend on system design and defy quantitative description because of their synergistic effects. Certain waters and wastewaters can contain interferences in sufficient concentration to suppress absorption responses of As and Se. For representative samples in a given laboratory and for initial analyses of unknown wastewaters, add appropriate inorganic forms of As or Se to digested sample portions and measure recovery. If average recoveries are less than 90%, consider using alternative analytical procedures.

e. *Detection level and optimum concentration range:* For both arsenic and selenium, analyzed by aspiration into a nitrogen-hydrogen flame after reduction, the method detection level is 2 µg/L or lower and the optimum concentration range 2 to 20 µg/L.

2. Apparatus

a. *Atomic absorption spectrometer* equipped with air-acetylene flame and quartz cell with mounting bracket or an electrically heated quartz cell, As and Se electrodeless discharge lamps with power supply, background correction at measurement wavelengths, and appropriate strip-chart recorder. A good-quality 10-mV recorder with high sensitivity and a fast response time is needed.

b. *Atomizer:* Use one of the following:

- 1) *Cylindrical quartz cell*, 10 to 20 cm long, bracket-mountable above air-acetylene burner.
- 2) *Cylindrical quartz cell*, 10 to 20 cm long, electrically heated by external nichrome wire to 800 to 900°C.⁷
- 3) *Cylindrical quartz cell* with internal fuel rich hydrogen-oxygen (air) flame.⁸

The sensitivity of quartz cells deteriorates over several months of use. Sensitivity sometimes may be restored by treatment with 40% HF. CAUTION: HF is extremely corrosive. Avoid all contact with exposed skin. Handle with care.

c. *Reaction cell for producing As or Se hydrides:* See Figure 3114:1 for an example of a manual, laboratory-made system. A commercially available system is acceptable if it utilizes liquid sodium borohydride reagents; accepts samples digested in accordance with ¶s 4c, d, and e; accepts 4 to 6N HCl; and is efficiently and precisely stirred by the purging gas and/or a magnetic stirrer.

d. *Eye dropper or syringe* capable of delivering 0.5 to 3.0 mL sodium borohydride reagent. Exact and reproducible addition is required so that production of hydrogen gas does not vary significantly between determinations.

e. *Vent:* See Section 3111A.6f.

3. Reagents

a. *Sodium borohydride reagent:* Dissolve 8 g NaBH₄ in 200 mL 0.1N NaOH. Prepare fresh daily.

b. *Sodium iodide prereductant solution:* Dissolve 50 g NaI in 500 mL water. Prepare fresh daily. Alternatively use an equivalent KI solution.

c. *Sulfuric acid, H₂SO₄, 18N.*

d. *Sulfuric acid, 2.5N:* Cautiously add 35 mL conc H₂SO₄ to about 400 mL water, let cool, and adjust volume to 500 mL.

e. *Potassium persulfate, 5% solution:* Dissolve 25 g K₂S₂O₈ in water and dilute to 500 mL. Store in glass and refrigerate. Prepare weekly.

f. *Nitric acid, HNO₃, conc.*

g. *Perchloric acid, HClO₄, conc.*

h. *Hydrochloric acid, HCl, conc.*

i. *Argon (or nitrogen), commercial grade.*

j. *Arsenic(III) solutions:*

1) *Stock As(III) solution:* Dissolve 1.320 g arsenic trioxide, As₂O₃, in water containing 4 g NaOH. Dilute to 1 L; 1.00 mL = 1.00 µg As(III).

2) *Intermediate As(III) solution:* Dilute 10 mL stock As solution to 1000 mL with water containing 5 mL conc HCl; 1.00 mL = 10.0 µg As(III).

3) *Standard As(III) solution:* Dilute 10 mL intermediate As(III) solution to 1000 mL with water containing the same concentration of acid used for sample preservation (2 to 5 mL

conc HNO₃); 1.00 mL = 0.100 µg As(III). Prepare diluted solutions daily.

k. Arsenic(V) solutions:

1) *Stock As(V) solution:* Dissolve 1.534 g arsenic pentoxide, As₂O₅, in distilled water containing 4 g NaOH. Dilute to 1 L; 1.00 mL = 1.00 mg As(V).

2) *Intermediate As(V) solution:* Prepare as for As(III) above; 1.00 mL = 10.0 µg As(V).

3) *Standard As(V) solution:* Prepare as for As(III) above; 1.00 mL = 0.100 µg As(V).

l. Organic arsenic solutions:

1) *Stock organic arsenic solution:* Dissolve 1.842 g dimethylarsinic acid (cacodylic acid), (CH₃)₂AsOOH, in water containing 4 g NaOH. Dilute to 1 L; 1.00 mL = 1.00 mg As. [NOTE: Check purity of cacodylic acid reagent against an intermediate arsenic standard (50 to 100 mg As/L) using flame atomic absorption.]

2) *Intermediate organic arsenic solution:* Prepare as for As(III) above; 1.00 mL = 10.0 µg As.

3) *Standard organic arsenic solution:* Prepare as for As(III) above; 1.00 mL = 0.100 µg As.

m. Selenium(IV) solutions:

1) *Stock Se(IV) solution:* Dissolve 2.190 g sodium selenite, Na₂SeO₃, in water containing 10 mL HCl and dilute to 1 L; 1.00 mL = 1.00 mg Se(IV).

2) *Intermediate Se(IV) solution:* Dilute 10 mL stock Se(IV) to 1000 mL with water containing 10 mL conc HCl; 1.00 mL = 10.0 µg Se(IV).

3) *Standard Se(IV) solution:* Dilute 10 mL intermediate Se(IV) solution to 1000 mL with water containing the same concentration of acid used for sample preservation (2 to 5 mL conc HNO₃). Prepare solution daily when checking the equivalency of instrument response for Se(IV) and Se(VI); 1.00 mL = 0.100 µg Se(IV).

n. Selenium(VI) solutions:

1) *Stock Se(VI) solution:* Dissolve 2.393 g sodium selenate, Na₂SeO₄, in water containing 10 mL conc HNO₃. Dilute to 1 L; 1.00 mL = 1.00 mg Se(VI).

2) *Intermediate Se(VI) solution:* Prepare as for Se(IV) above; 1.00 mL = 10.0 µg Se(VI).

3) *Standard Se(VI) solution:* Prepare as for Se(IV) above; 1.00 mL = 0.100 µg Se(VI).

4. Procedure

a. Apparatus setup: Either see Figure 3114:1 or follow manufacturer's instructions. Connect inlet of reaction cell with auxiliary purging gas controlled by flow meter. If a drying cell between the reaction cell and atomizer is necessary, use only anhydrous CaCl₂ but not CaSO₄ because it may retain SeH₂. Before using the hydride generation/analysis system, optimize operating parameters. Align quartz atomizers for maximum absorbance. Aspirate a blank until memory effects are removed. Establish purging gas flow, concentration and rate of addition of sodium borohydride reagent, solution volume, and stirring rate for optimum instrument response for the chemical species to be analyzed. Optimize quartz cell temperature. If sodium borohydride reagent is added too quickly, rapid evolution of hydrogen will unbalance the system. If the volume of solution being purged is too large, the absorption signal will be decreased.

Recommended wavelengths are 193.7 and 196.0 nm for As and Se, respectively.

b. Instrument calibration standards: Transfer 0.00, 1.00, 2.00, 5.00, 10.00, 15.00, and 20.00 mL standard solutions of As(III) or Se(IV) to 100-mL volumetric flasks and bring to volume with water containing the same acid concentration used for sample preservation (commonly 2 to 5 mL conc HNO₃/L). This yields blank and standard solutions of 0, 1, 2, 5, 10, 15, and 20 µg As or Se/L. Prepare fresh daily. In all cases, standards must be carried through the same digestion protocol as the samples to monitor digestion effectiveness.

c. Preparation of samples and standards for total recoverable arsenic and selenium: Use digestion procedure described in 3030F for samples and standards. Alternatively, add 50 mL sample, As(III), or Se(IV) standard to 200-mL Berzelius beaker or 100-mL micro-kjeldahl flask. Add 7 mL 18N H₂SO₄ and 5 mL conc HNO₃. Add a small boiling chip or glass beads if necessary. Evaporate to SO₃ fumes. Maintain oxidizing conditions at all times by adding small amounts of HNO₃ to prevent solution from darkening. Maintain an excess of HNO₃ until all organic matter is destroyed. Complete digestion usually is indicated by a light-colored solution. Cool slightly, add 25 mL water and 1 mL conc HClO₄ and again evaporate to SO₃ fumes to expel oxides of nitrogen. CAUTION: See Section 3030H for cautions on use of HClO₄. Monitor effectiveness of either digestion procedure used by adding 5 mL of standard organic arsenic solution or 5 mL of a standard selenium solution to a 50-mL sample and measuring recovery, carrying standards and the sample with known addition through entire procedure. To report total recoverable arsenic as total arsenic, average recoveries of cacodylic acid must exceed 80%. After final evaporation of SO₃ fumes, dilute to 50 mL for arsenic measurements or to 30 mL for selenium measurements. For analysis of both elements in a single sample, increase sample volume to 100 mL and double the volumes of acids used in the digestion. Adjust final digestate volume to 100 mL. Use 50 mL for As and 30 mL for Se determinations, making appropriate volume corrections in calculating results.

d. Preparation of samples and standards for total arsenic and selenium: Add 50 mL undigested sample or standard to a 200-mL Berzelius beaker or 100-mL micro-kjeldahl flask. Add 1 mL 2.5N H₂SO₄ and 5 mL 5% K₂S₂O₈. Boil gently on a pre-heated hot plate for approximately 30 to 40 min or until a final volume of 10 mL is reached. Do not let sample go to dryness. Alternatively heat in an autoclave at 121°C for 1 h in capped containers. After manual digestion, dilute to 50 mL for subsequent arsenic measurements and to 30 mL for selenium measurements. Monitor effectiveness of digestion by measuring recovery of As or Se as above. If poor recovery of arsenic added as cacodylic acid is obtained, reanalyze using double the amount of K₂S₂O₈. For analysis of both elements in a single sample, increase sample volume to 100 mL and double the volumes of acids used in the digestion. Adjust final digestate volume to 100 mL. Use 50 mL for As and 30 mL for Se determinations, making appropriate volume corrections in calculating results.

e. Determination of arsenic with sodium borohydride: To 50 mL digested standard or sample in a 200-mL Berzelius beaker (see Figure 3114:1) add 5 mL conc HCl and mix. Add 5 mL NaI prereductant solution, mix, and wait at least 30 min. (NOTE: The NaI reagent has not been found necessary for certain hydride reaction cell designs if a 20 to 30% loss in instrument sensitivity

is not important and variables of solution acid conditions, temperatures, and volumes for production of As(V) and arsine can be controlled strictly. Such control requires an automated delivery system; see Section 3114C.)

Attach one Berzelius beaker at a time to the rubber stopper containing the gas dispersion tube for the purging gas, the sodium borohydride reagent inlet, and the outlet to the quartz cell. Turn on strip-chart recorder and wait until the base line is established by the purging gas and all air is expelled from the reaction cell. Add 0.5 mL sodium borohydride reagent. After the instrument absorbance has reached a maximum and returned to the base line, remove beaker, rinse dispersion tube with water, and proceed to the next sample or standard. Periodically compare standard As(III) and As(V) curves for response consistency. Check for presence of chemical interferences that suppress instrument response for arsine by treating a digested sample with 10 $\mu\text{g/L}$ As(III) or As(V) as appropriate. Average recoveries should be not less than 90%.

f. Determination of selenium with sodium borohydride: To 30 mL digested standard or sample in a 200-mL Berzelius beaker or 100-mL micro-kjeldahl flask, add 15 mL conc HCl and mix. Heat for a predetermined period at 90 to 100°C. Alternatively autoclave at 121°C in capped containers for 60 min, or heat for a predetermined time in open test tubes using a 90 to 100°C hot water bath or an aluminum block digester. Check effectiveness of the selected heating time by demonstrating equal instrument responses for calibration curves prepared either from standard Se(IV) or from Se(VI) solutions. Effective heat exposure for converting Se(VI) to Se(IV), with no loss of Se(IV), ranges between 5 and 60 min when open beakers or test tubes are used. Establish a heating time for effective conversion and apply this time to all samples and standards. Do not digest standard Se(IV) and Se(VI) solutions used for this check of equivalency. After pre-reduction of Se(VI) to Se(IV), attach Berzelius beakers, one at a time, to the purge apparatus. For each, turn on the strip-chart recorder and wait until the base line is established. Add 0.50 mL sodium borohydride reagent. After the instrument absorbance has reached a maximum and returned to the base line, remove beaker, rinse dispersion tube with water, and proceed to the next sample or standard. Check for presence of chemical interferences that suppress selenium hydride instrument response by treating a digested sample with 10 μg Se (IV)/L. Average recoveries should be not less than 90%.

5. Calculation

Construct a standard curve by plotting peak heights or areas of standards versus concentration of standards. Measure peak heights or areas of samples and read concentrations from curve. If sample was diluted (or concentrated) before sample digestion, apply an appropriate factor. On instruments so equipped, read concentrations directly after standard calibration.

6. Precision and Bias

Single-laboratory, single-operator data were collected for As(III) and organic arsenic by both manual and automated methods, and for the manual determination of selenium. Recovery values (%) from seven replicates are given below:

Method	As(III)	Org As	Se(IV)	Se(VI)
Manual with digestion	91.8	87.3	—	—
Manual without digestion	109.4	19.4	100.6	110.8
Automated with digestion	99.8	98.4	—	—
Automated without digestion	92.5	10.4	—	—

7. References

1. VIJAN, P.N. & D. LEUNG. 1980. Reduction of chemical interference and speciation studies in the hydride generation-atomic absorption method for selenium. *Anal. Chim. Acta* 120:141.
2. VOTH-BEACH, L.M. & D.E. SHRADER. 1985. Reduction of interferences in the determination of arsenic and selenium by hydride generation. *Spectroscopy* 1:60.
3. JULSHAMN, K., O. RINGDAL, K.-E. SLINNING & O.R. BRAEKKAN. 1982. Optimization of the determination of selenium in marine samples by atomic absorption spectrometry: Comparison of a flameless graphite furnace atomic absorption system with a hydride generation atomic absorption system. *Spectrochim. Acta* 37B:473.
4. NYGAARD, D.D. & J.H. LOWRY. 1982. Sample digestion procedures for simultaneous determination of arsenic, antimony, and selenium by inductively coupled argon plasma emission spectrometry with hydride generation. *Anal. Chem.* 54:803.
5. WELZ, B. & M. MELCHER. 1984. Mechanisms of transition metal interferences in hydride generation atomic-absorption spectrometry. Part 1. Influence of cobalt, copper, iron and nickel on selenium determination. *Analyst* 109:569.
6. KRIVAN, V., K. PETRICK, B. WELZ & M. MELCHER. 1985. Radiotracer error-diagnostic investigation of selenium determination by hydride-generation atomic absorption spectrometry involving treatment with hydrogen peroxide and hydrochloric acid. *Anal. Chem.* 57:1703.
7. CHU, R.C., G.P. BARRON & P.A.W. BAUMGARNER. 1972. Arsenic determination at submicrogram levels by arsine evolution and flameless atomic absorption spectrophotometric technique. *Anal. Chem.* 44:1476.
8. SIEMER, D.D., P. KOTEEL & V. JARIWALA. 1976. Optimization of arsine generation in atomic absorption arsenic determinations. *Anal. Chem.* 48:836.

8. Bibliography

- FERNANDEZ, F.J. & D.C. MANNING. 1971. The determination of arsenic at submicrogram levels by atomic absorption spectrophotometry. *Atomic Absorption Newsletter* 10:86.
- JARREL-ASH CORPORATION. 1971. High Sensitivity Arsenic Determination by Atomic Absorption. Jarrel-Ash Atomic Absorption Applications Laboratory Bull. No. As-3.
- MANNING, D.C. 1971. A high sensitivity arsenic-selenium sampling system for atomic absorption spectroscopy. *Atomic Absorption Newsletter* 10:123.
- BRAMAN, R.S. & C.C. FOREBACK. 1973. Methylated forms of arsenic in the environment. *Science* 182:1247.
- CALDWELL, J.S., R.J. LISHKA & E.F. MCFARREN. 1973. Evaluation of a low-cost arsenic and selenium determination of microgram-per-liter levels. *J. Amer. Water Works Assoc.*, 65:731.
- AGGETT, J. & A.C. ASPELL. 1976. The determination of arsenic (III) and total arsenic by atomic-absorption spectroscopy. *Analyst* 101:341.
- FIORINO, J.A., J.W. JONES & S.G. CAPAR. 1976. Sequential determination of arsenic, selenium, antimony, and tellurium in foods via rapid hydride evolution and atomic absorption spectrometry. *Anal. Chem.* 48:120.

- CUTTER, G.A. 1978. Species determination of selenium in natural waters. *Anal. Chem. Acta* 98:59.
- GODDEN, R.G. & D.R. THOMERSON. 1980. Generation of covalent hydrides in atomic absorption spectroscopy. *Analyst* 105:1137.
- BROWN, R.M., JR., R.C. FRY, J.L. MOYERS, S.J. NORTHWAY, M.B. DENTON & G.S. WILSON. 1981. Interference by volatile nitrogen oxides and transition-metal catalysis in the preconcentration of arsenic and selenium as hydrides. *Anal. Chem.* 53:1560.

- SINEMUS, H.W., M. MELCHER & B. WELZ. 1981. Influence of valence state on the determination of antimony, bismuth, selenium, and tellurium in lake water using the hydride AA technique. *Atomic Spectrosc.* 2:81.
- DEDINA, J. 1982. Interference of volatile hydride forming elements in selenium determination by atomic absorption spectrometry with hydride generation. *Anal. Chem.* 54:2097.

3114 C. Continuous Hydride Generation/Atomic Absorption Spectrometric Method

1. General Discussion

The continuous hydride generator offers the advantages of simplicity in operation, excellent reproducibility, low detection limits, and high sample volume throughput for selenium analysis following preparations as described in 3500-Se.B or 3114B.4c and d.

a. *Principle:* See Section 3114B.

b. *Interferences:* Free chlorine in hydrochloric acid is a common but difficult-to-diagnose interference. (The amount of chlorine varies with manufacturer and with each lot from the same manufacturer.) Chlorine oxidizes the hydride and can contaminate the hydride generator to prevent recoveries under any conditions. When interference is encountered, or preferably before using each new bottle of HCl, eliminate chlorine from a 2.3-L bottle of conc HCl by bubbling with helium (commercial grade, 100 mL/min) for 3 h.

Excess oxidant (peroxide, persulfate, or permanganate) from the total selenium digestion can oxidize the hydride. Follow procedures in 3500-Se.B.2, 3, or 4 to ensure removal of all oxidizing agents before hydride generation.

Nitrite is a common trace constituent in natural and waste waters, and at levels as low as 10 $\mu\text{g/L}$ nitrite can reduce the recovery of hydrogen selenide from Se(IV) by over 50%. Moreover, during the reduction of Se(VI) to Se(IV) by digestion with HCl (3500-Se.B.5), some nitrate is converted to nitrite, which subsequently interferes. When this interference is suspected, add sulfanilamide after sample acidification (or HCl digestion). The diazotization reaction between nitrite and sulfanilamide completely removes the interferent effect (i.e., the standard addition slope is normal).

2. Apparatus

a. *Continuous hydride generator:* The basic unit is composed of two parts: a precision peristaltic pump, which is used to meter and mix reagents and sample solutions, and the gas-liquid separator. At the gas-liquid separator a constant flow of argon strips out the hydrogen and metal hydride gases formed in the reaction and carries them to the heated quartz absorption cell (3114B.1b and 2b), which is supported by a metal bracket mounted on top of the regular air acetylene burner head. The spent liquid flows




Figure 3114:2. Schematic of a continuous hydride generator.

out of the separator via a constant level side drain to a waste bucket. Schematics and operating parameters are shown in Figure 3114:2.

Check flow rates frequently to ensure a steady flow; an uneven flow in any tubing will cause an erratic signal. Remove tubings from pump rollers when not in use. Typical flow rates are: sample, 7 mL/min; acid, 1 mL/min; borohydride reagent, 1 mL/min. Argon flow usually is pre-fixed, typically at 90 mL/min.

b. *Atomic absorption spectrometric equipment:* See Section 3111A.6.

3. Reagents

a. *Hydrochloric acid, HCl, 5 + 1:* Handle conc HCl under a fume hood. If necessary, remove free Cl_2 by stripping conc HCl with helium as described above.

b. *Borohydride reagent:* Dissolve 0.6 g NaBH_4 and 0.5 g NaOH in 100 mL water. CAUTION: Sodium borohydride is toxic, flammable, and corrosive.

c. *Selenium reference standard solution, 1000 mg/L:* Use commercially available standard; verify that selenium is Se(IV).

d. *Intermediate standard solution, 1 mg/L:* Dilute 1 mL reference standard solution to 1 L in a volumetric flask with distilled water.

e. *Working standard solutions*, 5, 10, 20, 30, and 40 $\mu\text{g/L}$: Dilute 0.5, 1.0, 2.0, 3.0, and 4.0 mL intermediate standard solution to 100 mL in a volumetric flask.

f. *Sulfanilamide solution*: Prepare a 2.5% (w/v) solution daily; add several drops conc HCl per 50 mL solution to facilitate dissolution.

4. Procedure

a. *Sample preparation*: See Section 3500-Se or 3114B.4c and d for preparation steps for various Se fractions or total Se.

b. *Preconditioning hydride generator*: For newly installed tubing, turn on pump for at least 10 to 15 min before instrument calibration. Sample the highest standard for a few minutes to let volatile hydride react with the reactive sites in the transfer lines and on the quartz absorption cell surfaces.

c. *Instrument calibration*: Depending on total void volume in sample tubing, sampling time of 15 to 20 s generally is sufficient to obtain a steady signal. Between samples, submerge uptake tube in rinse water. Calibrate instrument daily after a 45-min lamp warmup time. Use either the hollow cathode or the electrodeless discharge lamp.

d. *Antifoaming agents*: Certain samples, particularly wastewater samples containing a high concentration of proteinaceous substances, can cause excessive foaming that could carry the liquid directly into the heated quartz absorption cell and cause splattering of salty deposits onto the windows of the spectrometer. Add a drop of antifoaming agent* to eliminate this problem.

e. *Nitrite removal*: After samples have been acidified, or after acid digestion, add 0.1 mL sulfanilamide solution per 10 mL sample and let react for 2 min.

f. *Analysis*: Follow manufacturer's instructions for operation of analytical equipment.

5. Calculation

Construct a calibration curve based on absorbance vs. standard concentration. Apply dilution factors on diluted samples.

6. Precision and Bias

Working standards were analyzed together with batches of water samples on a routine production basis. The standards were

* Dow Corning or equivalent.

compounded using chemically pure sodium selenite and sodium selenate. The values of Se(IV) + Se(VI) were determined by converting Se(VI) to Se(IV) by digestion with HCl. Results are tabulated below.

No. Analyses	Mean Se(IV) $\mu\text{g/L}$	Rel. Dev. %	Se(IV) + Se(VI) $\mu\text{g/L}$	Rel. Del. %
21	4.3	12	10.3	7
26	8.5	12	19.7	6
22	17.2	7	39.2	8
20	52.8	5	106.0	6

7. Bibliography

- REAMER, D.C. & C. VEILOON. 1981. Preparation of biological materials for determination of selenium by hydride generation-AAS. *Anal. Chem.* 53:1192.
- SINEMUS, H.W., M. MELCHER & B. WELZ. 1981. Influence of valence state on the determination of antimony, bismuth, selenium, and tellurium in lake water using the hydride AA technique. *Atomic Spectrosc.* 2:81.
- RODEN, D.R. & D.E. TALLMAN. 1982. Determination of inorganic selenium species in groundwaters containing organic interferences by ion chromatography and hydride generation/atomic absorption spectrometry. *Anal. Chem.* 54:307.
- CUTTER, G. 1983. Elimination of nitrite interference in the determination of selenium by hydride generation. *Anal. Chim. Acta* 149:391.
- NARASAKI, H. & M. IKEDA. 1984. Automated determination of arsenic and selenium by atomic absorption spectrometry with hydride generation. *Anal. Chem.* 56:2059.
- WELZ, B. & M. MELCHER. 1985. Decomposition of marine biological tissues for determination of arsenic, selenium, and mercury using hydride-generation and cold-vapor atomic absorption spectrometries. *Anal. Chem.* 57:427.
- EBDON, L. & S.T. SPARKES. 1987. Determination of arsenic and selenium in environmental samples by hydride generation-direct current plasma-atomic emission spectrometry. *Microchem. J.* 36:198.
- EBDON, L. & J.R. WILKINSON. 1987. The determination of arsenic and selenium in coal by continuous flow hydride-generation atomic absorption spectroscopy and atomic fluorescence spectrometry. *Anal. Chim. Acta.* 194:177.
- VOTH-BEACH, L.M. & D.E. SHRADER. 1985. Reduction of interferences in the determination of arsenic and selenium by hydride generation. *Spectroscopy* 1:60.

3120 METALS BY PLASMA EMISSION SPECTROSCOPY*

3120 A. Introduction

1. General Discussion

Emission spectroscopy using inductively coupled plasma (ICP) was developed in the mid-1960's^{1,2} as a rapid, sensitive,

and convenient method for the determination of metals in water and wastewater samples.³⁻⁶ Dissolved metals are determined in filtered and acidified samples. Total metals are determined after appropriate digestion. Care must be taken to ensure that potential interferences are dealt with, especially when dissolved solids exceed 1500 mg/L.

* Approved by Standard Methods Committee, 1999.

2. References

1. GREENFIELD, S., I.L. JONES & C.T. BERRY. 1964. High-pressure plasma-spectroscopic emission sources. *Analyst* 89: 713.
2. WENDT, R.H. & V.A. FASSEL. 1965. Induction-coupled plasma spectrometric excitation source. *Anal. Chem.* 37:920.
3. U.S. ENVIRONMENTAL PROTECTION AGENCY. 1994. Method 200.7. Inductively coupled plasma-atomic emission spectrometric method for trace element analysis of water and wastes. Methods for the Determination of Metals in Environmental Samples—Supplement I. EPA 600/R-94-111, May 1994.
4. AMERICAN SOCIETY FOR TESTING AND MATERIALS. 1987. Annual Book of ASTM Standards, Vol. 11.01. American Soc. Testing & Materials, Philadelphia, Pa.
5. FISHMAN, M.J. & W.L. BRADFORD, eds. 1982. A Supplement to Methods for the Determination of Inorganic Substances in Water and Fluvial Sediments. Rep. No. 82-272, U.S. Geological Survey, Washington, D.C.
6. GARBARINO, J.R. & H.E. TAYLOR. 1985. Trace Analysis. Recent Developments and Applications of Inductively Coupled Plasma Emission Spectroscopy to Trace Elemental Analysis of Water. Volume 4. Academic Press, New York, N.Y.

3120 B. Inductively Coupled Plasma (ICP) Method

1. General Discussion

a. Principle: An ICP source consists of a flowing stream of argon gas ionized by an applied radio frequency field typically oscillating at 27.1 MHz. This field is inductively coupled to the ionized gas by a water-cooled coil surrounding a quartz “torch” that supports and confines the plasma. A sample aerosol is generated in an appropriate nebulizer and spray chamber and is carried into the plasma through an injector tube located within the torch. The sample aerosol is injected directly into the ICP, subjecting the constituent atoms to temperatures of about 6000 to 8000°K.¹ Because this results in almost complete dissociation of molecules, significant reduction in chemical interferences is achieved. The high temperature of the plasma excites atomic emission efficiently. Ionization of a high percentage of atoms produces ionic emission spectra. The ICP provides an optically “thin” source that is not subject to self-absorption except at very high concentrations. Thus linear dynamic ranges of four to six orders of magnitude are observed for many elements.²

The efficient excitation provided by the ICP results in low detection limits for many elements. This, coupled with the extended dynamic range, permits effective multielement determination of metals.³ The light emitted from the ICP is focused onto the entrance slit of either a monochromator or a polychromator that effects dispersion. A precisely aligned exit slit is used to isolate a portion of the emission spectrum for intensity measurement using a photomultiplier tube. The monochromator uses a single exit slit/photomultiplier and may use a computer-controlled scanning mechanism to examine emission wavelengths sequentially. The polychromator uses multiple fixed exit slits and corresponding photomultiplier tubes; it simultaneously monitors all configured wavelengths using a computer-controlled readout system. The sequential approach provides greater wavelength selection while the simultaneous approach can provide greater sample throughput.

b. Applicable metals and analytical limits: Table 3120:I lists elements for which this method applies, recommended analytical wavelengths, and typical estimated instrument detection levels using conventional pneumatic nebulization. Actual working detection levels are sample-dependent. Typical upper limits for linear calibration also are included in Table 3120:I.

c. Interferences: Interferences may be categorized as follows:

1) Spectral interferences—Light emission from spectral sources other than the element of interest may contribute to apparent net signal intensity. Sources of spectral interference include direct spectral line overlaps, broadened wings of intense spectral lines, ion-atom recombination continuum emission, molecular band emission, and stray (scattered) light from the emission of elements at high concentrations.⁴ Avoid line overlaps by selecting alternate analytical wavelengths. Avoid or minimize other spectral interference by judicious choice of background correction positions. A wavelength scan of the element line region is useful for detecting potential spectral interferences and for selecting positions for background correction. Make corrections for residual spectral interference using empirically determined correction factors in conjunction with the computer software supplied by the spectrometer manufacturer or with the calculation detailed below. The empirical correction method cannot be used with scanning spectrometer systems if the analytical and interfering lines cannot be precisely and reproducibly located. In addition, if using a polychromator, verify absence of spectral interference from an element that could occur in a sample but for which there is no channel in the detector array. Do this by analyzing single-element solutions of 100 mg/L concentration and noting for each element channel the apparent concentration from the interfering substance that is greater than the element’s instrument detection limit.

2) Nonspectral interferences

a) Physical interferences are effects associated with sample nebulization and transport processes. Changes in the physical properties of samples, such as viscosity and surface tension, can cause significant error. This usually occurs when samples containing more than 10% (by volume) acid or more than 1500 mg dissolved solids/L are analyzed using calibration standards containing $\leq 5\%$ acid. Whenever a new or unusual sample matrix is encountered, use the test described in ¶ 4g. If physical interference is present, compensate for it by sample dilution, by using matrix-matched calibration standards, or by applying the method of standard addition (see ¶ 5d below).

High dissolved solids content also can contribute to instrumental drift by causing salt buildup at the tip of the nebulizer gas orifice. Using prehumidified argon for sample nebulization lessens this problem. Better control of the argon flow rate to the

TABLE 3120:I. SUGGESTED WAVELENGTHS, ESTIMATED DETECTION LEVELS, ALTERNATE WAVELENGTHS, CALIBRATION CONCENTRATIONS, AND UPPER LIMITS

Element	Suggested Wavelength nm	Estimated Detection Level µg/L	Alternate Wavelength* nm	Calibration Concentration mg/L	Upper Limit Concentration† mg/L
Aluminum	308.22	40	237.32	10.0	100
Antimony	206.83	30	217.58	10.0	100
Arsenic	193.70	50	189.04‡	10.0	100
Barium	455.40	2	493.41	1.0	50
Beryllium	313.04	0.3	234.86	1.0	10
Boron	249.77	5	249.68	1.0	50
Cadmium	226.50	4	214.44	2.0	50
Calcium	317.93	10	315.89	10.0	100
Chromium	267.72	7	206.15	5.0	50
Cobalt	228.62	7	230.79	2.0	50
Copper	324.75	6	219.96	1.0	50
Iron	259.94	7	238.20	10.0	100
Lead	220.35	40	217.00	10.0	100
Lithium	670.78	4§	—	5.0	100
Magnesium	279.08	30	279.55	10.0	100
Manganese	257.61	2	294.92	2.0	50
Molybdenum	202.03	8	203.84	10.0	100
Nickel	231.60	15	221.65	2.0	50
Potassium	766.49	100§	769.90	10.0	100
Selenium	196.03	75	203.99	5.0	100
Silica (SiO ₂)	212.41	20	251.61	21.4	100
Silver	328.07	7	338.29	2.0	50
Sodium	589.00	30§	589.59	10.0	100
Strontium	407.77	0.5	421.55	1.0	50
Thallium	190.86‡	40	377.57	10.0	100
Vanadium	292.40	8	—	1.0	50
Zinc	213.86	2	206.20	5.0	100

* Other wavelengths may be substituted if they provide the needed sensitivity and are corrected for spectral interference.

† Defines the top end of the effective calibration range. Do not extrapolate to concentrations beyond highest standard.

‡ Available with vacuum or inert gas purged optical path.

§ Sensitive to operating conditions.

nebulizer using a mass flow controller improves instrument performance.

b) Chemical interferences are caused by molecular compound formation, ionization effects, and thermochemical effects associated with sample vaporization and atomization in the plasma. Normally these effects are not pronounced and can be minimized by careful selection of operating conditions (incident power, plasma observation position, etc.). Chemical interferences are highly dependent on sample matrix and element of interest. As with physical interferences, compensate for them by using matrix matched standards or by standard addition (§ 5d). To determine the presence of chemical interference, follow instructions in ¶ 4g.

2. Apparatus

a. *ICP source:* The ICP source consists of a radio frequency (RF) generator capable of generating at least 1.1 kW of power, torch, tesla coil, load coil, impedance matching network, nebulizer, spray chamber, and drain. High-quality flow regulators are required for both the nebulizer argon and the plasma support gas flow. A peristaltic pump is recommended to regulate sample flow to the nebulizer. The type of nebulizer and spray chamber used

may depend on the samples to be analyzed as well as on the equipment manufacturer. In general, pneumatic nebulizers of the concentric or cross-flow design are used. Viscous samples and samples containing particulates or high dissolved solids content (>5000 mg/L) may require nebulizers of the Babington type.⁵

b. *Spectrometer:* The spectrometer may be of the simultaneous (polychromator) or sequential (monochromator) type with air-path, inert gas purged, or vacuum optics. A spectral bandpass of 0.05 nm or less is required. The instrument should permit examination of the spectral background surrounding the emission lines used for metals determination. It is necessary to be able to measure and correct for spectral background at one or more positions on either side of the analytical lines.

3. Reagents and Standards

Use reagents that are of ultra-high-purity grade or equivalent. Redistilled acids are acceptable. Except as noted, dry all salts at 105°C for 1 h and store in a desiccator before weighing. Use deionized water prepared by passing water through at least two stages of deionization with mixed bed cation/anion exchange resins.⁶ Use deionized water for preparing all calibration standards, reagents, and for dilution.

a. *Hydrochloric acid*, HCl, conc and 1+1.

b. *Nitric acid*, HNO₃, conc.

c. *Nitric acid*, HNO₃, 1+1: Add 500 mL conc HNO₃ to 400 mL water and dilute to 1 L.

d. *Standard stock solutions*: See 3111B, 3111D, and 3114B. CAUTION: Many metal salts are extremely toxic and may be fatal if swallowed. Wash hands thoroughly after handling.

1) *Aluminum*: See 3111D.3k1).

2) *Antimony*: See 3111B.3j1).

3) *Arsenic*: See 3114B.3k1).

4) *Barium*: See 3111D.3k2).

5) *Beryllium*: See 3111D.3k3).

6) *Boron*: Do not dry but keep bottle tightly stoppered and store in a desiccator. Dissolve 0.5716 g anhydrous H₃BO₃ in water and dilute to 1000 mL; 1 mL = 100 µg B.

7) *Cadmium*: See 3111B.3j3).

8) *Calcium*: See 3111B.3j4).

9) *Chromium*: See 3111B.3j6).

10) *Cobalt*: See 3111B.3j7).

11) *Copper*: See 3111B.3j8).

12) *Iron*: See 3111B.3j11).

13) *Lead*: See 3111B.3j12).

14) *Lithium*: See 3111B.3j13).

15) *Magnesium*: See 3111B.3j14).

16) *Manganese*: See 3111B.3j15).

17) *Molybdenum*: See 3111D.3k4).

18) *Nickel*: See 3111B.3j16).

19) *Potassium*: See 3111B.3j19).

20) *Selenium*: See 3114B.3n1).

21) *Silica*: See 3111D.3k7).

22) *Silver*: See 3111B.3j22).

23) *Sodium*: See 3111B.3j23).

24) *Strontium*: See 3111B.3j24).

25) *Thallium*: See 3111B.3j25).

26) *Vanadium*: See 3111D.3k10).

27) *Zinc*: See 3111B.3j27).

e. *Calibration standards*: Prepare mixed calibration standards containing the concentrations shown in Table 3120:I by combining appropriate volumes of the stock solutions in 100-mL volumetric flasks. Add 2 mL 1+1 HNO₃ and 10 mL 1+1 HCl and dilute to 100 mL with water. Before preparing mixed standards, analyze each stock solution separately to determine possible spectral interference or the presence of impurities. When preparing mixed standards take care that the elements are compatible and stable. Store mixed standard solutions in an FEP fluorocarbon or unused polyethylene bottle. Verify calibration standards initially using the quality control standard; monitor weekly for stability. The following are recommended combinations using the suggested analytical lines in Table 3120:I. Alternative combinations are acceptable.

1) *Mixed standard solution I*: Manganese, beryllium, cadmium, lead, selenium, and zinc.

2) *Mixed standard solution II*: Barium, copper, iron, vanadium, and cobalt.

3) *Mixed standard solution III*: Molybdenum, silica, arsenic, strontium, and lithium.

4) *Mixed standard solution IV*: Calcium, sodium, potassium, aluminum, chromium, and nickel.

5) *Mixed standard solution V*: Antimony, boron, magnesium, silver, and thallium. If addition of silver results in an

initial precipitation, add 15 mL water and warm flask until solution clears. Cool and dilute to 100 mL with water. For this acid combination limit the silver concentration to 2 mg/L. Silver under these conditions is stable in a tap water matrix for 30 d. Higher concentrations of silver require additional HCl.

f. *Calibration blank*: Dilute 2 mL 1+1 HNO₃ and 10 mL 1+1 HCl to 100 mL with water. Prepare a sufficient quantity to be used to flush the system between standards and samples.

g. *Method blank*: Carry a reagent blank through entire sample preparation procedure. Prepare method blank to contain the same acid types and concentrations as the sample solutions.

h. *Instrument check standard*: Prepare instrument check standards by combining compatible elements at a concentration of 2 mg/L.

i. *Instrument quality control sample*: Obtain a certified aqueous reference standard from an outside source and prepare according to instructions provided by the supplier. Use the same acid matrix as the calibration standards.

j. *Method quality control sample*: Carry the instrument quality control sample (¶ 3i) through the entire sample preparation procedure.

k. *Argon*: Use technical or welder's grade. If gas appears to be a source of problems, use purified grade.

4. Procedure

a. *Sample preparation*: See Section 3030F.

b. *Operating conditions*: Because of differences among makes and models of satisfactory instruments, no detailed operating instructions can be provided. Follow manufacturer's instructions. Establish instrumental detection limit, precision, optimum background correction positions, linear dynamic range, and interferences for each analytical line. Verify that the instrument configuration and operating conditions satisfy the analytical requirements and that they can be reproduced on a day-to-day basis. An atom-to-ion emission intensity ratio [Cu(I) 324.75 nm/Mn(II) 257.61 nm] can be used to reproduce optimum conditions for multielement analysis precisely. The Cu/Mn intensity ratio may be incorporated into the calibration procedure, including specifications for sensitivity and for precision.⁷ Keep daily or weekly records of the Cu and Mn intensities and/or the intensities of critical element lines. Also record settings for optical alignment of the polychromator, sample uptake rate, power readings (incident, reflected), photomultiplier tube attenuation, mass flow controller settings, and system maintenance.

c. *Instrument calibration*: Set up instrument as directed (¶ b). Warm up for 30 min. For polychromators, perform an optical alignment using the profile lamp or solution. Check alignment of plasma torch and spectrometer entrance slit, particularly if maintenance of the sample introduction system was performed. Make Cu/Mn or similar intensity ratio adjustment.

Calibrate instrument according to manufacturer's recommended procedure using calibration standards and blank. Aspirate each standard or blank for a minimum of 15 s after reaching the plasma before beginning signal integration. Rinse with calibration blank or similar solution for at least 60 s between each standard to eliminate any carryover from the previous standard. Use average intensity of multiple integrations of standards or samples to reduce random error.

Before analyzing samples, analyze instrument check standard. Concentration values obtained should not deviate from the actual values by more than $\pm 5\%$ (or the established control limits, whichever is lower).

d. Analysis of samples: Begin each sample run with an analysis of the calibration blank, then analyze the method blank. This permits a check of the sample preparation reagents and procedures for contamination. Analyze samples, alternating them with analyses of calibration blank. Rinse for at least 60 s with dilute acid between samples and blanks. After introducing each sample or blank let system equilibrate before starting signal integration. Examine each analysis of the calibration blank to verify that no carry-over memory effect has occurred. If carry-over is observed, repeat rinsing until proper blank values are obtained. Make appropriate dilutions and acidifications of the sample to determine concentrations beyond the linear calibration range.

e. Instrumental quality control: Analyze instrument check standard once per 10 samples to determine if significant instrument drift has occurred. If agreement is not within $\pm 5\%$ of the expected values (or within the established control limits, whichever is lower), terminate analysis of samples, correct problem, and recalibrate instrument. If the intensity ratio reference is used, resetting this ratio may restore calibration without the need for reanalyzing calibration standards. Analyze instrument check standard to confirm proper recalibration. Reanalyze one or more samples analyzed just before termination of the analytical run. Results should agree to within $\pm 5\%$, otherwise all samples analyzed after the last acceptable instrument check standard analysis must be reanalyzed.

Analyze instrument quality control sample within every run. Use this analysis to verify accuracy and stability of the calibration standards. If any result is not within $\pm 5\%$ of the certified value, prepare a new calibration standard and recalibrate the instrument. If this does not correct the problem, prepare a new stock solution and a new calibration standard and repeat calibration.

f. Method quality control: Analyze the method quality control sample within every run. Results should agree to within $\pm 5\%$ of the certified values. Greater discrepancies may reflect losses or contamination during sample preparation.

g. Test for matrix interference: When analyzing a new or unusual sample matrix verify that neither a positive nor negative nonlinear interference effect is operative. If the element is present at a concentration above 1 mg/L, use serial dilution with calibration blank. Results from the analyses of a dilution should be within $\pm 5\%$ of the original result. Alternately, or if the concentration is either below 1 mg/L or not detected, use a post-digestion addition equal to 1 mg/L. Recovery of the addition should be either between 95% and 105% or within established control limits of ± 2 standard deviations around the mean. If a matrix effect causes test results to fall outside the critical limits, complete the analysis after either diluting the sample to eliminate the matrix effect while maintaining a detectable concentration of at least twice the detection limit or applying the method of standard additions.

5. Calculations and Corrections

a. Blank correction: Subtract result of an adjacent calibration blank from each sample result to make a baseline drift correction. (Concentrations printed out should include negative and positive values to compensate for positive and negative baseline drift. Make certain that the calibration blank used for blank correction has not been contaminated by carry-over.) Use the result of the method blank analysis to correct for reagent contamination. Alternatively, intersperse method blanks with appropriate samples. Reagent blank and baseline drift correction are accomplished in one subtraction.

b. Dilution correction: If the sample was diluted or concentrated in preparation, multiply results by a dilution factor (*DF*) calculated as follows:

$$DF = \frac{\text{Final weight or volume}}{\text{Initial weight or volume}}$$

c. Correction for spectral interference: Correct for spectral interference by using computer software supplied by the instrument manufacturer or by using the manual method based on interference correction factors. Determine interference correction factors by analyzing single-element stock solutions of appropriate concentrations under conditions matching as closely as possible those used for sample analysis. Unless analysis conditions can be reproduced accurately from day to day, or for longer periods, redetermine interference correction factors found to affect the results significantly each time samples are analyzed.^{7,8} Calculate interference correction factors (K_{ij}) from apparent concentrations observed in the analysis of the high-purity stock solutions:

$$K_{ij} = \frac{\text{Apparent concentration of element } i}{\text{Actual concentration of interfering element } j}$$

where the apparent concentration of element *i* is the difference between the observed concentration in the stock solution and the observed concentration in the blank. Correct sample concentrations observed for element *i* (already corrected for baseline drift), for spectral interferences from elements *j*, *k*, and *l*; for example:

Concentration of element *i* corrected for spectral interference

$$= \begin{array}{r} \text{Observed} \\ \text{concentration} \\ \text{of } i \end{array} - (K_{ij}) \begin{array}{r} \text{Observed} \\ \text{concentration} \\ \text{of interfering} \\ \text{element } j \end{array} - (K_{ik}) \begin{array}{r} \text{Observed} \\ \text{concentration} \\ \text{of interfering} \\ \text{element } k \end{array} - (K_{il}) \begin{array}{r} \text{Observed} \\ \text{concentration} \\ \text{of interfering} \\ \text{element } l \end{array}$$

Interference correction factors may be negative if background correction is used for element *i*. A negative K_{ij} can result where an interfering line is encountered at the background correction wavelength rather than at the peak wavelength. Determine concentrations of interfering elements *j*, *k*, and *l* within their respective linear ranges. Mutual interferences (*i*

TABLE 3120:II. ICP PRECISION AND BIAS DATA

Element	Concentration Range µg/L	Total Digestion* µg/L	Recoverable Digestion* µg/L
Aluminum	69-4792	$X = 0.9273C + 3.6$ $S = 0.0559X + 18.6$ $SR = 0.0507X + 3.5$	$X = 0.9380C + 22.1$ $S = 0.0873X + 31.7$ $SR = 0.0481X + 18.8$
Antimony	77-1406	$X = 0.7940C - 17.0$ $S = 0.1556X - 0.6$ $SR = 0.1081X + 3.9$	$X = 0.8908C + 0.9$ $S = 0.0982X + 8.3$ $SR = 0.0682X + 2.5$
Arsenic	69-1887	$X = 1.0437C - 12.2$ $S = 0.1239X + 2.4$ $SR = 0.0874X + 6.4$	$X = 1.0175C + 3.9$ $S = 0.1288X + 6.1$ $SR = 0.0643X + 10.3$
Barium	9-377	$X = 0.7683C + 0.47$ $S = 0.1819X + 2.78$ $SR = 0.1285X + 2.55$	$X = 0.8380C + 1.68$ $S = 0.2540X + 0.30$ $SR = 0.0826X + 3.54$
Beryllium	3-1906	$X = 0.9629C + 0.05$ $S = 0.0136X + 0.95$ $SR = 0.0203X - 0.07$	$X = 1.0177C - 0.55$ $S = 0.0359X + 0.90$ $SR = 0.0445X - 0.10$
Boron	19-5189	$X = 0.8807C + 9.0$ $S = 0.1150X + 14.1$ $SR = 0.0742X + 23.2$	$X = 0.9676C + 18.7$ $S = 0.1320X + 16.0$ $SR = 0.0743X + 21.1$
Cadmium	9-1943	$X = 0.9874C - 0.18$ $S = 0.0557X + 2.02$ $SR = 0.0300X + 0.94$	$X = 1.0137C - 0.65$ $S = 0.0585X + 1.15$ $SR = 0.0332X + 0.90$
Calcium	17-47 170	$X = 0.9182C - 2.6$ $S = 0.1228X + 10.1$ $SR = 0.0189X + 3.7$	$X = 0.9658C + 0.8$ $S = 0.0917X + 6.9$ $SR = 0.0327X + 10.1$
Chromium	13-1406	$X = 0.9544C + 3.1$ $S = 0.0499X + 4.4$ $SR = 0.0009X + 7.9$	$X = 1.0049C - 1.2$ $S = 0.0698X + 2.8$ $SR = 0.0571X + 1.0$
Cobalt	17-2340	$X = 0.9209C - 4.5$ $S = 0.0436X + 3.8$ $SR = 0.0428X + 0.5$	$X = 0.9278C - 1.5$ $S = 0.0498X + 2.6$ $SR = 0.0407X + 0.4$
Copper	8-1887	$X = 0.9297C - 0.30$ $S = 0.0442X + 2.85$ $SR = 0.0128X + 2.53$	$X = 0.9647C - 3.64$ $S = 0.0497X + 2.28$ $SR = 0.0406X + 0.96$
Iron	13-9359	$X = 0.8829C + 7.0$ $S = 0.0683X + 11.5$ $SR = -0.0046X + 10.0$	$X = 0.9830C + 5.7$ $S = 0.1024X + 13.0$ $SR = 0.0790X + 11.5$
Lead	42-4717	$X = 0.9699C - 2.2$ $S = 0.0558X + 7.0$ $SR = 0.0353X + 3.6$	$X = 1.0056C + 4.1$ $S = 0.0799X + 4.6$ $SR = 0.0448X + 3.5$
Magnesium	34-13 868	$X = 0.9881C - 1.1$ $S = 0.0607X + 11.6$ $SR = 0.0298X + 0.6$	$X = 0.9879C + 2.2$ $S = 0.0564X + 13.2$ $SR = 0.0268X + 8.1$
Manganese	4-1887	$X = 0.9417C + 0.13$ $S = 0.0324X + 0.88$ $SR = 0.0153X + 0.91$	$X = 0.9725C + 0.07$ $S = 0.0557X + 0.76$ $SR = 0.0400X + 0.82$
Molybdenum	17-1830	$X = 0.9682C + 0.1$ $S = 0.0618X + 1.6$ $SR = 0.0371X + 2.2$	$X = 0.9707C - 2.3$ $S = 0.0811X + 3.8$ $SR = 0.0529X + 2.1$
Nickel	17-47 170	$X = 0.9508C + 0.4$ $S = 0.0604X + 4.4$ $SR = 0.0425X + 3.6$	$X = 0.9869C + 1.5$ $S = 0.0526X + 5.5$ $SR = 0.0393X + 2.2$
Potassium	347-14 151	$X = 0.8669C - 36.4$ $S = 0.0934X + 77.8$ $SR = -0.0099X + 144.2$	$X = 0.9355C - 183.1$ $S = 0.0481X + 177.2$ $SR = 0.0329X + 60.9$
Selenium	69-1415	$X = 0.9363C - 2.5$ $S = 0.0855X + 17.8$ $SR = 0.0284X + 9.3$	$X = 0.9737C - 1.0$ $S = 0.1523X + 7.8$ $SR = 0.0443X + 6.6$

TABLE 3120:II, CONT.

Element	Concentration Range $\mu\text{g/L}$	Total Digestion* $\mu\text{g/L}$	Recoverable Digestion* $\mu\text{g/L}$
Silicon	189-9434	$X = 0.5742C - 35.6$ $S = 0.4160X + 37.8$ $SR = 0.1987X + 8.4$	$X = 0.9737C - 60.8$ $S = 0.3288X + 46.0$ $SR = 0.2133X + 22.6$
Silver	8-189	$X = 0.4466C + 5.07$ $S = 0.5055X - 3.05$ $SR = 0.2086X - 1.74$	$X = 0.3987C + 8.25$ $S = 0.5478X - 3.93$ $SR = 0.1836X - 0.27$
Sodium	35-47 170	$X = 0.9581C + 39.6$ $S = 0.2097X + 33.0$ $SR = 0.0280X + 105.8$	$X = 1.0526C + 26.7$ $S = 0.1473X + 27.4$ $SR = 0.0884X + 50.5$
Thallium	79-1434	$X = 0.9020C - 7.3$ $S = 0.1004X + 18.3$ $SR = 0.0364X + 11.5$	$X = 0.9238C + 5.5$ $S = 0.2156X + 5.7$ $SR = -0.0106X + 48.0$
Vanadium	13-4698	$X = 0.9615C - 2.0$ $S = 0.0618X + 1.7$ $SR = 0.0220X + 0.7$	$X = 0.9551C + 0.4$ $S = 0.0927X + 1.5$ $SR = 0.0472X + 0.5$
Zinc	7-7076	$X = 0.9356C - 0.30$ $S = 0.0914X + 3.75$ $SR = -0.0130X + 10.07$	$X = 0.9500C + 1.22$ $S = 0.0597X + 6.50$ $SR = 0.0153X + 7.78$

* X = mean recovery, $\mu\text{g/L}$,

C = true value, $\mu\text{g/L}$,

S = multi-laboratory standard deviation, $\mu\text{g/L}$,

SR = single-analyst standard deviation, $\mu\text{g/L}$.

interferes with j and j interferes with i) require iterative or matrix methods for calculation.

d. Correction for nonspectral interference: If nonspectral interference correction is necessary, use the method of standard additions. It is applicable when the chemical and physical form of the element in the standard addition is the same as in the sample, or the ICP converts the metal in both sample and addition to the same form; the interference effect is independent of metal concentration over the concentration range of standard additions; and the analytical calibration curve is linear over the concentration range of standard additions.

Use an addition not less than 50% nor more than 100% of the element concentration in the sample so that measurement precision will not be degraded and interferences that depend on element/interferent ratios will not cause erroneous results. Apply the method to all elements in the sample set using background correction at carefully chosen off-line positions. Multielement standard addition can be used if it has been determined that added elements are not interferences.

e. Reporting data: Report analytical data in concentration units of milligrams per liter using up to three significant figures. Report results below the determined detection limit as not detected less than the stated detection limit corrected for sample dilution.

6. Precision and Bias

As a guide to the generally expected precision and bias, see the linear regression equations in Table 3120:II.⁹ Additional inter-laboratory information is available.¹⁰

7. References

- FAIRES, L.M., B.A. PALMER, R. ENGLEMAN, JR. & T.M. NIEMCZYK. 1984. Temperature determinations in the inductively coupled plasma using a Fourier transform spectrometer. *Spectrochim. Acta* 39B:819.
- BARNES, R.M. 1978. Recent advances in emission spectroscopy: inductively coupled plasma discharges for spectrochemical analysis. *CRC Crit. Rev. Anal. Chem.* 7:203.
- PARSONS, M.L., S. MAJOR & A.R. FORSTER. 1983. Trace element determination by atomic spectroscopic methods - State of the art. *Appl. Spectrosc.* 37:411.
- LARSON, G.F., V.A. FASSEL, R. K. WINGE & R.N. KNISELEY. 1976. Ultratrace analysis by optical emission spectroscopy: The stray light problem. *Appl. Spectrosc.* 30:384.
- GARBARINO, J.R. & H.E. TAYLOR. 1979. A Babington-type nebulizer for use in the analysis of natural water samples by inductively coupled plasma spectrometry. *Appl. Spectrosc.* 34:584.
- AMERICAN SOCIETY FOR TESTING AND MATERIALS. 1988. Standard specification for reagent water, D1193-77 (reapproved 1983). Annual Book of ASTM Standards. American Soc. Testing & Materials, Philadelphia, Pa.
- BOTTO, R.I. 1984. Quality assurance in operating a multielement ICP emission spectrometer. *Spectrochim. Acta* 39B:95.
- BOTTO, R.I. 1982. Long-term stability of spectral interference calibrations for inductively coupled plasma atomic emission spectrometry. *Anal. Chem.* 54:1654.
- MAXFIELD, R. & B. MINDAK. 1985. EPA Method Study 27, Method 200. 7 (Trace Metals by ICP). EPA-600/S4-85/05. National Technical Information Serv., Springfield, Va.
- GARBARINO, J.R., B.E. JONES, G. P. STEIN, W.T. BELSER & H.E. TAYLOR. 1985. Statistical evaluation of an inductively coupled plasma atomic emission spectrometric method for routine water quality testing. *Appl. Spectrosc.* 39:53.

3125 METALS BY INDUCTIVELY COUPLED PLASMA/MASS SPECTROMETRY*

3125 A. Introduction

1. General Discussion

This method is used for the determination of trace metals and metalloids in surface, ground, and drinking waters by inductively coupled plasma/mass spectrometry (ICP/MS). It may also be suitable for wastewater, soils, sediments, sludge, and biological samples after suitable digestion followed by dilution and/or cleanup.^{1,2} Additional sources of information on quality assurance and other aspects of ICP/MS analysis of metals are available.³⁻⁵

The method is intended to be performance-based, allowing extension of the elemental analyte list, implementation of "clean" preparation techniques as they become available, and other appropriate modifications of the base method as technology evolves. Preferably validate modifications to the base method by use of the quality control standards specified in the method.

Instrument detection levels for many analytes are between 1 and 100 ng/L. The method is best suited for the determination of metals in ambient or pristine fresh-water matrices. More complex matrices may require some type of cleanup to reduce matrix effects to a manageable level. Various cleanup techniques are available to reduce matrix interferences and/or concentrate analytes of interest.⁶⁻¹⁰

This method is ideally used by analysts experienced in the use of ICP/MS, the interpretation of spectral and matrix interference, and procedures for their correction. Preferably demonstrate analyst proficiency through analysis of a performance evaluation sample before the generation of data.

* Approved by Standard Methods Committee, 1997.

Joint Task Group; 20th Edition—William R. Kammin (chair), John R. Barnett, Isabel C. Chamberlain, Robert Henry, James O. Ross, Ruth E. Wolf, Cindy A. Ziernicki.

2. References

1. MONTASER, A. & D.W. GOLIGHTLY, eds. 1992. *Inductively Coupled Plasmas in Analytical Atomic Spectrometry*, 2nd ed. VCH Publishers, Inc., New York, N.Y.
2. DATE, A.R. & A.L. GRAY. 1989. *Applications of Inductively Coupled Plasma Mass Spectrometry*. Blackie & Son, Ltd., Glasgow, U.K.
3. U.S. ENVIRONMENTAL PROTECTION AGENCY. 1994. Determination of trace elements in waters and wastes by inductively coupled plasma-mass spectrometry, Method 200.8. U.S. Environmental Protection Agency, Environmental Monitoring Systems Lab., Cincinnati, Ohio.
4. LONGBOTTOM, J.E., T.D. MARTIN, K.W. EDGELL, S.E. LONG, M.R. PLANTZ & B.E. WARDEN. 1994. Determination of trace elements in water by inductively coupled plasma-mass spectrometry: collaborative study. *J. AOAC Internat.* 77:1004.
5. U.S. ENVIRONMENTAL PROTECTION AGENCY. 1995. Method 1638: Determination of trace elements in ambient waters by inductively coupled plasma-mass spectrometry. U.S. Environmental Protection Agency, Off. Water, Washington, D.C.
6. McLAREN, J.W., A.P. MYKTYIUK, S.N. WILLIE & S. S. BERMAN. 1985. Determination of trace metals in seawater by inductively coupled plasma mass spectrometry with preconcentration on silica-immobilized 8-hydroxyquinoline. *Anal. Chem.* 57:2907.
7. BURBA, P. & P.G. WILLMER. 1987. Multielement preconcentration for atomic spectroscopy by sorption of dithiocarbamate metal complexes (e.g., HMDC) on cellulose collectors. *Fresenius Z. Anal. Chem.* 329:539.
8. WANG, X. & R.M. BARNES. 1989. Chelating resins for on-line flow injection preconcentration with inductively coupled plasma atomic emission spectroscopy. *J. Anal. Atom. Spectrom.* 4:509.
9. SIRIRAKS, A., H.M. KINGSTON & J.M. RIVIELLO. 1990. Chelation ion chromatography as a method for trace elemental analysis in complex environmental and biological samples. *Anal. Chem.* 62:1185.
10. PUGET SOUND WATER QUALITY AUTHORITY. 1996. Recommended Guidelines for Measuring Metals in Puget Sound Marine Water, Sediment and Tissue Samples. Appendix D: Alternate Methods for the Analysis of Marine Water Samples. Puget Sound Water Quality Authority, Olympia, Wash.

3125 B. Inductively Coupled Plasma/Mass Spectrometry (ICP/MS) Method

1. General Discussion

a. Principle: Sample material is introduced into an argon-based, high-temperature radio-frequency plasma, usually by pneumatic nebulization. Energy transfer from the plasma to the sample stream causes desolvation, atomization, and ionization of

target elements. Ions generated by these energy-transfer processes are extracted from the plasma through a differential vacuum interface, and separated on the basis of their mass-to-charge ratio by a mass spectrometer. The mass spectrometer usually is of the quadrupole or magnetic sector type. The ions passing through the mass spectrometer are counted, usually by

TABLE 3125:I. RECOMMENDED ANALYTE MASSES, INSTRUMENTAL DETECTION LEVELS (IDL), AND INTERNAL STANDARDS

Element	Analytical Mass	IDL* µg/L	Recommended Internal Standard
Be	9	0.025	Li
Al	27	0.03	Sc
V	51	0.02	Sc
Cr	52	0.04	Sc
Cr	53	0.03	Sc
Mn	55	0.002	Sc
Co	59	0.002	Sc
Ni	60	0.004	Sc
Ni	62	0.025	Sc
Cu	63	0.003	Sc
Cu	65	0.004	Sc
Zn	66	0.017	Ge
Zn	68	0.020	Ge
As	75	0.025	Ge
Se	77	0.093	Ge
Se	82	0.064	Ge
Ag	107	0.003	In
Ag	109	0.002	In
Cd	111	0.006	In
Cd	114	0.003	In
Sb	121	0.07	In
Sb	123	0.07	In
Tl	203	0.03	Th
Tl	205	0.03	Th
Pb	208	0.005	Th
U	235	0.032	Th
U	238	0.001	Th
Mo	98	0.003†	In
Ba	135	0.008†	In
Sr	88	0.001‡	In

* IDLs were determined on a Perkin Elmer Elan 6000 ICP/MS using seven replicate analyses of a 1% nitric acid solution, at Manchester Environmental Laboratory, July 1996.

† From EPA Method 200.8 for the Analysis of Drinking Waters-Application Note, Order No. ENVA-300A, The Perkin Elmer Corporation, 1996.

‡ From Perkin Elmer Technical Summary TSMS-12.

an electron multiplier detector, and the resulting information processed by a computer-based data-handling system.

b. *Applicable elements and analytical limits:* This method is suitable for aluminum, antimony, arsenic, barium, beryllium, cadmium, chromium, cobalt, copper, lead, manganese, molybdenum, nickel, selenium, silver, strontium, thallium, uranium, vanadium, and zinc. The method is also acceptable for other elemental analytes as long as the same quality assurance practices are followed. The basic element suite and recommended analytical masses are given in Table 3125:I.

Typical instrument detection levels (IDL)^{1,2} for method analytes are presented in Table 3125:I. Determine the IDL and method detection level (or limit) (MDL) for all analytes before method implementation. Section 1030 contains additional information and approaches for the evaluation of detection capabilities.

The MDL is defined in Section 1010C and elsewhere.² Determination of the MDL for each element is critical for complex matrices such as seawater, brines, and industrial effluents. The MDL will typically be higher than the IDL, because of back-

ground analyte in metals preparation and analysis laboratories and matrix-based interferences. Determine both IDL and MDL upon initial implementation of this method, and then yearly or whenever the instrument configuration changes or major maintenance occurs, whichever comes first.

Determine linear dynamic ranges (LDR) for all method analytes. LDR is defined as the maximum concentration of analyte above the highest calibration point where analyte response is within ±10% of the theoretical response. When determining linear dynamic ranges, avoid using unduly high concentrations of analyte that might damage the detector. Determine LDR on multielement mixtures, to account for possible interelement effects. Determine LDR on initial implementation of this method, and then yearly.

c. *Interferences:* ICP/MS is subject to several types of interferences.

1) Isotopes of different elements that form ions of the same nominal mass-to-charge ratio are not resolved by the quadrupole mass spectrometer, and cause isobaric elemental interferences. Typically, ICP/MS instrument operating software will have all

TABLE 3125:II. ELEMENTAL ABUNDANCE EQUATIONS AND COMMON MOLECULAR ION CORRECTION EQUATIONS

Elemental and Molecular Equations*†	
Li 6	= C 6
Be 9	= C 9
Al 27	= C 27
Sc 45	= C 45
V 51	= C 51 - (3.127)[(C 53) - (0.113 × C 52)]
Cr 52	= C 52
Cr 53	= C 53
Mn 55	= C 55
Co 59	= C 59
Ni 60	= C 60
Ni 62	= C 62
Cu 63	= C 63
Cu 65	= C 65
Zn 66	= C 66
Zn 68	= C 68
As 75	= C 75 - (3.127)[(C 77) - (0.815 × C 82)]
Se 77	= C 77
Se 82	= C 82 - (1.008696 × C 83)
Sr 88	= C 88
Mo 98	= C 98 - (0.110588 × C 101)
Rh 103	= C 103
Ag 107	= C 107
Ag 109	= C 109
Cd 111	= C 111 - (1.073)[(C 108) - (0.712 × C 106)]
Cd 114	= C 114 - (0.02686 × C 118)
Sb 121	= C 121
Sb 123	= C 123 - (0.127189 × C 125)
Ba 135	= C 135
Ho 165	= C 165
Tl 203	= C 203
Tl 205	= C 205
Pb 208	= C 208 + (1 × C 206) + (1 × C 207)
Th 232	= C 232
U 238	= C 238

* C = calibration blank corrected counts at indicated mass.

† From EPA Method 200.8 for the Analysis of Drinking Waters - Application Note, Order No. ENVA-300A, The Perkin Elmer Corporation, 1996.

TABLE 3125:III. COMMON MOLECULAR ION INTERFERENCES IN ICP/MS¹

Molecular Ion	Mass	Element Measurement Affected by Interference
Background molecular ions:		
NH ⁺	15	—
OH ⁺	17	—
OH ₂ ⁺	18	—
C ₂ ⁺	24	Mg
CO ⁺	26	Mg
CO ⁺	28	Si
N ₂ ⁺	28	Si
N ₂ H ⁺	29	Si
NO ⁺	30	—
NOH ⁺	31	P
O ₂ ⁺	32	S
O ₂ H ⁺	33	—
³⁶ ArH ⁺	37	Cl
³⁸ ArH ⁺	39	K
⁴⁰ ArH ⁺	41	—
CO ₂ ⁺	44	Ca
CO ₂ ⁺ H	45	Sc
ArC ⁺ , ArO ⁺	52	Cr
ArN ⁺	54	Cr
ArNH ⁺	55	Mn
ArO ⁺	56	Fe
ArH ⁺	57	Fe
⁴⁰ Ar ³⁶ Ar ⁺	76	Se
⁴⁰ Ar ³⁸ Ar ⁺	78	Se
⁴⁰ Ar ₂ ⁺	80	Se
Matrix molecular ions:		
Bromide:		
⁸¹ BrH ⁺	82	Se
⁷⁹ BrO ⁺	95	Mo
⁸¹ BrO ⁺	97	Mo
⁸¹ BrOH ⁺	98	Mo
Ar ⁸¹ Br ⁺	121	Sb
Chloride:		
³⁵ ClO ⁺	51	V
³⁵ ClOH ⁺	52	Cr
³⁷ ClO ⁺	53	Cr
³⁷ ClOH ⁺	54	Cr
Ar ³⁵ Cl ⁺	75	As
Ar ³⁷ Cl ⁺	77	Se
Sulfate:		
³² SO ⁺	48	Ti
³² SOH ⁺	49	—
³⁴ SO ⁺	50	V, Cr
³⁴ SOH ⁺	51	V
SO ₂ ⁺ , S ₂ ⁺	64	Zn
Ar ³² S ⁺	72	Ge
Ar ³⁴ S ⁺	74	Ge
Phosphate:		
PO ⁺	47	Ti
POH ⁺	48	Ti
PO ₂ ⁺	63	Cu
ArP ⁺	71	Ga
Group I & II metals:		
ArNa ⁺	63	Cu
ArK ⁺	79	Br
ArCa ⁺	80	Se
Matrix oxides*		
TiO	62-66	Ni, Cu, Zn
ZrO	106-112	Ag, Cd
MoO	108-116	Cd
NbO	109	Ag

* Oxide interferences normally will be very small and will affect the method elements only when oxide-producing elements are present at relatively high concentrations, or when the instrument is improperly tuned or maintained. Preferably monitor Ti and Zr isotopes for soil, sediment, or solid waste samples, because these samples potentially contain high levels of these interfering elements.

known isobaric interferences entered, and will perform necessary calculations automatically. Table 3125:II shows many of the commonly used corrections. Monitor the following additional masses: ⁸³Kr, ⁹⁹Ru, ¹¹⁸Sn, and ¹²⁵Te. It is necessary to monitor these masses to correct for isobaric interference caused by ⁸²Kr on ⁸²Se, by ⁹⁸Ru on ⁹⁸Mo, by ¹¹⁴Sn on ¹¹⁴Cd, and by ¹²⁵Te on ¹²³Sb. Monitor ArCl at mass 77, to estimate chloride interferences. Verify that all elemental and molecular correction equations used in this method are correct and appropriate for the mass spectrometer used and sample matrix.

2) Abundance sensitivity is an analytical condition in which the tails of an abundant mass peak contribute to or obscure adjacent masses. Adjust spectrometer resolution to minimize these interferences.

3) Polyatomic (molecular) ion interferences are caused by ions consisting of more than one atom and having the same nominal mass-to-charge ratio as the isotope of interest. Most of the common molecular ion interferences have been identified and are listed in Table 3125:III. Because of the severity of chloride ion interference on important analytes, particularly arsenic and selenium, hydrochloric acid is not recommended for use in preparation of any samples to be analyzed by ICP/MS. The mathematical corrections for chloride interferences only correct chloride to a concentration of 0.4%. Because chloride ion is present in most environmental samples, it is critical to use chloride correction equations for affected masses. A high-resolution ICP/MS may be used to resolve interferences caused by polyatomic ions. Polyatomic interferences are strongly influenced by instrument design and plasma operating conditions, and can be reduced in some cases by careful adjustment of nebulizer gas flow and other instrument operating parameters.

4) Physical interferences include differences in viscosity, surface tension, and dissolved solids between samples and calibration standards. To minimize these effects, dissolved solid levels in analytical samples should not exceed 0.5%. Dilute water and wastewater samples containing dissolved solids at or above 0.5% before analysis. Use internal standards for correction of physical interferences. Any internal standards used should demonstrate comparable analytical behavior to the elements being determined.

5) Memory interferences occur when analytes from a previous sample or standard are measured in the current sample. Use a sufficiently long rinse or flush between samples to minimize this type of interference. If memory interferences persist, they may be indications of problems in the sample introduction system. Severe memory interferences may require disassembly and cleaning of the entire sample introduction system, including the plasma torch, and the sampler and skimmer cones.

6) Ionization interferences result when moderate (0.1 to 1%) amounts of a matrix ion change the analyte signal. This effect, which usually reduces the analyte signal, also is known as "suppression." Correct for suppression by use of internal standardization techniques.

2. Apparatus

a. *Inductively coupled plasma/mass spectrometer:* Instrumentation, available from several manufacturers, includes a mass spectrometer detector, inductively coupled plasma source, mass flow controllers for regulation of ICP gas flows, peristaltic pump for sample introduction, and a computerized data acquisition and

instrument control system. An x-y autosampler also may be used with appropriate control software.

b. Laboratory ware: Use precleaned plastic laboratory ware for standard and sample preparation. Teflon,* either tetrafluoroethylene hexafluoropropylene-copolymer (FEP), polytetrafluoroethylene (PTFE), or perfluoroalkoxy PTFE (PFA) is preferred for standard preparation and sample digestion, while high-density polyethylene (HDPE) and other dense, metal-free plastics may be acceptable for internal standards, known-addition solutions, etc. Check each new lot of autosampler tubes for suitability, and preclean autosampler tubes and pipettor tips (see Section 3010C.2).

c. Air displacement pipets, 10 to 100 μL , 100 to 1000 μL , and 1 to 10 mL size.

d. Analytical balance, accurate to 0.1 mg.

e. Sample preparation apparatus, such as hot plates, microwave digestors, and heated sand baths. Any sample preparation device has the potential to introduce trace levels of target analytes to the sample.

f. Clean hood (optional), Class 100 (certified to contain less than 100 particles/ m^3), for sample preparation and manipulation. Preferably perform all sample manipulations, digestions, dilutions, etc. in a certified Class 100 environment. Alternatively, handle samples in glove boxes, plastic fume hoods, or other environments where random contamination by trace metals can be minimized.

3. Reagents

a. Acids: Use ultra-high-purity grade (or equivalent) acids to prepare standards and to process sample. Redistilled acids are acceptable if each batch is demonstrated to be free from contamination by target analytes. Use extreme care in the handling of acids in the laboratory to avoid contamination of the acids with trace levels of metals.

1) *Nitric acid*, HNO_3 , conc (specific gravity 1.41).

2) *Nitric acid*, 1 + 1: Add 500 mL conc HNO_3 to 500 mL reagent water.

3) *Nitric acid*, 2%: Add 20 mL conc HNO_3 to 100 mL reagent water; dilute to 1000 mL.

4) *Nitric acid*, 1%: Add 10 mL conc HNO_3 to 100 mL reagent water; dilute to 1000 mL.

b. Reagent water: Use water of the highest possible purity for blank, standard, and sample preparation (see Section 1080). Alternatively, use the procedure described below to produce water of acceptable quality. Other water preparation regimes may be used, provided that the water produced is metal-free. Reagent water containing trace amounts of analyte elements will cause erroneous results.

Produce reagent water using a softener/reverse osmosis unit with subsequent UV sterilization. After the general deionization system use a dual-column strong acid/strong base ion exchange system to polish laboratory reagent water before production of metal-free water. Use a multi-stage reagent water system, with two strong acid/strong base ion exchange columns and an activated carbon filter for organics removal for final polishing of laboratory reagent water. Use only high-purity water for preparation of samples and standards.

* Or equivalent.

c. Stock, standard, and other required solutions: See 3120B.3d for preparation of standard stock solutions from elemental materials (pure metals, salts). Preferably, purchase high-purity commercially prepared stock solutions and dilute to required concentrations. Single- or multi-element stock solutions (1000 mg/L) of the following elements are required: aluminum, antimony, arsenic, barium, beryllium, cerium, cadmium, chromium, cobalt, copper, germanium, indium, lead, magnesium, manganese, molybdenum, nickel, rhodium, scandium, selenium, silver, strontium, terbium, thallium, thorium, uranium, vanadium, and zinc. Prepare internal standard stock separately from target element stock solution. The potential for incompatibility between target elements and/or internal standards exists, and could cause precipitation or other solution instability.

1) *Internal standard stock solution:* Lithium, scandium, germanium, indium, and thorium are suggested as internal standards. The following masses are monitored: ^6Li , ^{45}Sc , ^{72}Ge , ^{115}In , and ^{232}Th . Add to all samples, standards, and quality control (QC) samples a level of internal standard that will give a suitable counts/second (cps) signal (for most internal standards, 200 000 to 500 000 cps; for lithium, 20 000 to 70 000 cps). Minimize error introduced by dilution during this addition by using an appropriately high concentration of internal standard mix solution. Maintain volume ratio for all internal standard additions.

Prepare internal standard mix as follows: Prepare a nominal 50-mg/L solution of ^6Li by dissolving 0.15 g $^6\text{Li}_2\text{CO}_3$ (isotopically pure, i.e., 95% or greater purity†) in a minimal amount of 1:1 HNO_3 . Pipet 5.0 mL 1000-mg/L scandium, germanium, indium, and thorium standards into the lithium solution, dilute resulting solution to 500.0 mL, and mix thoroughly. The resultant concentration of Sc, Ge, In, and Th will be 10 mg/L. Older instruments may require higher levels of internal standard to achieve acceptable levels of precision.

Other internal standards, such as rhodium, yttrium, terbium, holmium, and bismuth may also be used in this method. Ensure that internal standard mix used is stable and that there are no undesired interactions between elements.

Screen all samples for internal standard elements before analysis. The analysis of a few representative samples for internal standards should be sufficient. Analyze samples "as received" or "as digested" (before addition of internal standard), then add internal standard mix and reanalyze. Monitor counts at the internal standard masses. If the "as received" or "as digested" samples show appreciable detector counts (10% or higher of samples with added internal standard), dilute sample or use an alternate internal standard. If the internal standard response of the sample with the addition is not within 70 to 125% of the response for a calibration blank with the internal standard added, either dilute the sample before analysis, or use an alternate internal standard. During actual analysis, monitor internal standard masses and note all internal standard recoveries over 125% of internal standard response in calibration blank. Interpret results for these samples with caution.

The internal standard mix may be added to blanks, standards, and samples by pumping the solution so it is mixed with the sample stream in the sample introduction process.

† Cambridge Isotope Laboratories or equivalent.

TABLE 3125:IV. SUGGESTED ANALYTICAL RUN SEQUENCE

Sample Type	Comments
Tuning/optimization standard	Check mass calibration and resolution
Tuning/optimization standard	Optimize instrument for maximum rhodium counts while keeping oxides, double charged ions, and background within instrument specifications
Rinse	—
Reagent blank	Check for contamination
Reagent blank	Calibration standard blank
5- $\mu\text{g/L}$ standard	—
10- $\mu\text{g/L}$ standard	—
20- $\mu\text{g/L}$ standard	—
50- $\mu\text{g/L}$ standard	—
100- $\mu\text{g/L}$ standard	—
Rinse	—
Initial calibration verification, 50 $\mu\text{g/L}$	—
Initial calibration blank	—
0.30- $\mu\text{g/L}$ standard	Low-level calibration verification
1.0- $\mu\text{g/L}$ standard	Low-level calibration verification
External reference material	NIST 1643c or equivalent
Continuing calibration verification	—
Continuing blank calibration	—
Project sample method blank	—
Project sample laboratory-fortified blank	—
Project sample 1-4	—
Project sample 5	—
Project sample 5 with known addition	—
Project sample 5 duplicate with known addition	—
Continuing calibration verification	—
Continuing calibration blank	—

TABLE 3125:V. SUMMARY OF PERFORMANCE CRITERIA

Performance Characteristic	Criteria
Mass resolution	Manufacturer's specification
Mass calibration	Manufacturer's specification
Ba ²⁺ /Ba ⁺	Manufacturer's specification
CeO/Ce	Manufacturer's specification
Background counts at mass 220	Manufacturer's specification
Correlation coefficient	≥ 0.995
Calibration blanks	< Reporting limit
Calibration verification standards	$\pm 10\%$ of true value
Laboratory fortified blank (control sample)	$\pm 30\%$ of true value
Precision	$\pm 20\%$ relative percent difference for lab duplicates
Known-addition recovery	75-125%
0.3 and 1.0 $\mu\text{g/L}$ standards	Dependent on data quality objectives
Reference materials	Dependent on data quality objectives
Internal standard response	70-125% of response in calibration blank with known addition

2) *Instrument optimization/tuning solution*, containing the following elements: barium, beryllium, cadmium, cerium, cobalt, copper, germanium, indium, magnesium, rhodium, scandium, terbium, thallium, and lead. Prepare this solution in 2% HNO₃. This mix includes all common elements used in optimization and tuning of the various ICP/MS operational parameters. It may be possible to use fewer elements in this solution, depending on the instrument manufacturer's recommendations.

3) *Calibration standards*, 0, 5, 10, 20, 50, and 100 $\mu\text{g/L}$.[‡] Other calibration regimes are acceptable, provided the full suite of quality assurance samples and standards is run to validate these method changes. Fewer standards may be used, and a two-point blank/mid-range calibration technique commonly used in ICP optical methods should also produce acceptable results. Calibrate all analytes using the selected concentrations. Prepare all calibration standards and blanks in a matrix of 2% nitric acid. Add internal standard mix to all calibration standards to provide appropriate count rates for interference correction. NOTE: All standards and blanks used in this method have the internal standard mix added at the same ratio.

4) *Method blank*, consisting of reagent water (¶ 3b) taken through entire sample preparation process. For dissolved samples, take reagent water through same filtration and preservation processes used for samples. For samples requiring digestion, process reagent water with the same digestion techniques as samples. Add internal standard mix to method blank.

[‡] Performance data for the method were obtained with these concentrations.

TABLE 3125:VI. QUALITY CONTROL ANALYSES FOR ICP/MS METHOD

Analysis	Frequency	Acceptance Criteria
Reference material [¶ 3c9]	Greater of: once per sample batch, or 5%	Dependent on data quality objectives
Preparatory/method blank [¶ 3c4]	Greater of: once per sample batch, or 5%	\pm Absolute value of instrument detection limit; \pm absolute value of laboratory reporting limit or MDL is acceptable
Laboratory fortified blank [¶ 3c7]	Greater of: once per sample batch, or 5%	$\pm 30\%$ of true value
Duplicate known-addition samples	Greater of: once per sample batch, or 5%	$\pm 20\%$ relative percent difference
Continuing calibration verification standards [¶ 3c5]	10%	$\pm 10\%$ of known concentration
Continuing calibration verification blank [¶ 3c6]	10%	\pm Absolute value of instrument detection limit; \pm absolute value of laboratory reporting limit or MDL is acceptable

TABLE 3125:VII. METHOD PERFORMANCE WITH CALIBRATION VERIFICATION STANDARDS*

Element	Mass	Continuing Calibration Verification Standard (N = 44)				Initial Calibration Verification Standard (N = 12)			
		Mean Recovery %	Mean $\mu\text{g/L}$	Standard Deviation $\mu\text{g/L}$	Relative Standard Deviation %	Mean Recovery %	Mean $\mu\text{g/L}$	Standard Deviation $\mu\text{g/L}$	Relative Standard Deviation %
Be	9	98.71	49.35	3.43	6.94	100.06	50.03	1.90	3.80
Al	27	99.62	49.81	2.99	6.01	98.42	49.21	1.69	3.44
V	51	100.97	50.48	1.36	2.68	99.91	49.96	1.23	2.47
Cr	52	101.39	50.70	1.86	3.66	99.94	49.97	1.47	2.95
Cr	53	100.68	50.34	1.91	3.79	99.13	49.56	1.44	2.90
Mn	55	101.20	50.60	1.98	3.91	99.48	49.74	1.40	2.82
Co	59	101.67	50.83	2.44	4.79	99.44	49.72	1.61	3.24
Ni	60	99.97	49.99	2.14	4.28	97.98	48.99	1.70	3.47
Ni	62	99.79	49.89	2.09	4.18	97.57	48.79	1.32	2.71
Cu	63	100.51	50.25	2.19	4.36	97.87	48.93	1.63	3.33
Cu	65	100.39	50.19	2.26	4.51	98.34	49.17	1.58	3.20
Zn	66	101.07	50.53	1.93	3.82	98.75	49.38	0.87	1.76
Zn	68	100.42	50.21	1.89	3.77	97.75	48.87	0.50	1.02
As	75	100.76	50.38	1.15	2.28	98.83	49.41	0.89	1.80
Se	77	101.71	50.85	1.43	2.81	99.54	49.77	1.01	2.03
Se	82	101.97	50.98	1.50	2.95	99.76	49.88	0.94	1.89
Ag	107	101.50	50.75	1.68	3.30	99.27	49.63	1.17	2.36
Ag	109	101.65	50.83	1.68	3.31	99.66	49.83	1.54	3.08
Cd	111	100.92	50.46	1.94	3.84	98.61	49.30	1.36	2.77
Cd	114	100.90	50.45	2.07	4.10	99.20	49.60	1.41	2.84
Sb	121	100.14	50.07	2.39	4.77	99.38	49.69	1.38	2.78
Sb	123	99.98	49.99	2.48	4.97	99.09	49.54	1.34	2.71
Tl	203	101.36	50.68	1.64	3.23	100.05	50.02	1.01	2.01
Tl	205	102.40	51.20	1.93	3.78	101.23	50.62	1.45	2.87
Pb	208	101.21	50.61	1.65	3.25	99.33	49.67	0.84	1.69
U	238	101.54	50.77	1.93	3.80	99.80	49.90	1.36	2.72

* Single-laboratory, single-operator, single-instrument data, determined using a 50- $\mu\text{g/L}$ standard prepared from sources independent of calibration standard source. Data acquired January-November 1996 during actual sample determinations. Performance of continuing calibration verification standards at different levels may vary. Perkin-Elmer Elan 6000 ICP/MS used for determination.

5) *Calibration verification standard*: Prepare a mid-range standard, from a source different from the source of the calibration standards, in 2% HNO_3 , with equivalent addition of internal standard.

6) *Calibration verification blank*: Use 2% HNO_3 .

7) *Laboratory fortified blank* (optional): Prepare solution with 2% nitric acid and method analytes added at about 50 $\mu\text{g/L}$. This standard, sometimes called a laboratory control sample (LCS), is used to validate digestion techniques and known-addition levels.

8) *Reference materials*: Externally prepared reference material, preferably from National Institute of Standards and Technology (NIST) 1643 series or equivalent.

9) *Known-addition solution for samples*: Add stock standard to sample in such a way that volume change is less than 5%. In the absence of information on analyte levels in the sample, prepare known additions at around 50 $\mu\text{g/L}$. If analyte concentration levels are known, add at 50 to 200% of the sample levels. For samples undergoing digestion, make additions before digestion. For the determination of dissolved metals, make additions after filtration, preferably immediately before analysis.

10) *Low-level standards*: Use both a 0.3- and a 1.0- $\mu\text{g/L}$ standard when expected analyte concentration is below 5 $\mu\text{g/L}$. Prepare both these standards in 2% nitric acid.

Prepare volumetrically a mixed standard containing the method analytes at desired concentration(s) (0.30 $\mu\text{g/L}$, 1.0 $\mu\text{g/L}$, or both). Prepare weekly in 100-mL quantities.

d. *Argon*: Use a prepurified grade of argon unless it can be demonstrated that other grades can be used successfully. The use of prepurified argon is usually necessary because of the presence of krypton as an impurity in technical argon. ^{82}Kr interferes with the determination of ^{82}Se . Monitor ^{83}Kr at all times.

4. Procedures

a. *Sample preparation*: See Sections 3010 and 3020 for general guidance regarding sampling and quality control. See Section 3030E for recommended sample digestion technique for all analytes except silver and antimony. If silver and antimony are target analytes, use method given in 3030F, paying special attention to interferences caused by chloride ion, and using all applicable elemental corrections. Alternative digestion techniques and additional guidance on sample preparation are available.^{3,4}

Ideally use a "clean" environment for any sample handling, manipulation, or preparation. Preferably perform all sample manipulations in a Class 100 clean hood or room to minimize potential contamination artifacts in digested or filtered samples.

b. *Instrument operating conditions*: Follow manufacturer's standard operating procedures for initialization, mass calibration, gas flow optimization, and other instrument operating conditions. Maintain complete and detailed information on the operational status of the instrument whenever it is used.

c. *Analytical run sequence:* A suggested analytical run sequence, including instrument tuning/optimization, checking of reagent blanks, instrument calibration and calibration verification, analysis of samples, and analysis of quality control samples and blanks, is given in Table 3125:IV.

d. *Instrument tuning and optimization:* Follow manufacturer's instructions for optimizing instrument performance. The most important optimization criteria include nebulizer gas flows, detector and lens voltages, radio-frequency forward power, and mass calibration. Periodically check mass calibration and instrument resolution. Ideally, optimize the instrument to minimize oxide formation and doubly-charged species formation. Measure the CeO/Ce ratio to monitor oxide formation, and measure doubly-charged species by determination of the Ba²⁺/Ba⁺ ratio. Both these ratios should meet the manufacturer's criteria before instrument calibration. Monitor background counts at mass 220 after optimization and compare with manufacturer's criteria. A summary of performance criteria related to optimization and tuning, calibration, and analytical performance for this method is given in Table 3125:V.

e. *Instrument calibration:* After optimization and tuning, calibrate instrument using an appropriate range of calibration standards. Use appropriate regression techniques to determine calibration lines or curves for each analyte. For acceptable calibrations, correlation coefficients for regression curves are ideally 0.995 or greater.

Immediately after calibration, run initial calibration verification standard, ¶ 3c5; acceptance criteria are ±10% of known analyte concentration. Next run initial calibration verification blank, ¶ 3c6; acceptance criteria are ideally ± the absolute value of the instrument detection limit for each analyte, but in practice, ± the absolute value of the laboratory reporting limit or the laboratory method detection limit for each analyte is acceptable. Verify low-level calibration by running 0.3- and/or 1.0-µg/L standards, if analyte concentrations are less than 5 µg/L.

f. *Sample analysis:* Ensure that all vessels and reagents are free from contamination. During analytical run (see Table 3125:IV), include quality control analyses according to schedule of Table 3125:VI, or follow project-specific QA/QC protocols.

Internal standard recoveries must be between 70% and 125% of internal standard response in the laboratory-fortified blank; otherwise, dilute sample, add internal standard mix, and reanalyze.

Make known-addition analyses for each separate matrix in a digestion or filtration batch.

5. Calculations and Corrections

Configure instrument software to report internal standard corrected results. For water samples, preferably report results in micrograms per liter. Report appropriate number of significant figures.

TABLE 3125:VIII. METHOD PERFORMANCE FOR RECOVERY OF KNOWN ADDITION IN NATURAL WATERS*

Element	Mass	Total Recoverable Metals†		Dissolved Metals‡	
		Mean Recovery %	Relative Standard Deviation %	Mean Recovery %	Relative Standard Deviation %
Be	9	89.09	5.77	—	—
V	51	87.00	8.82	—	—
Cr	52	87.33	8.42	88.38	6.43
Cr	53	86.93	7.90	88.52	5.95
Mn	55	91.81	10.12	—	—
Co	59	87.67	8.92	—	—
Ni	60	85.07	8.42	89.31	5.70
Ni	62	84.67	8.21	89.00	5.82
Cu	63	84.13	8.46	88.55	8.33
Cu	65	84.37	8.05	88.26	7.80
Zn	66	86.14	23.01	95.59	13.81
Zn	68	81.95	20.31	91.94	13.27
As	75	90.43	4.46	97.30	8.84
Se	77	83.09	4.76	105.36	10.80
Se	82	83.42	4.73	105.36	10.75
Ag	107	—	—	91.98	5.06
Ag	109	—	—	92.25	4.96
Cd	111	91.37	5.47	96.91	6.03
Cd	114	91.47	6.04	97.03	5.42
Sb	121	94.40	5.24	—	—
Sb	123	94.56	5.36	—	—
Tl	203	97.24	5.42	—	—
Tl	205	98.14	6.21	—	—
Pb	208	96.09	7.08	100.69	7.28

* Single-laboratory, single-operator, single-instrument data. Samples were Washington State surface waters from various locations. Data acquired January-November 1996 during actual sample determinations. Performance of known additions at different levels may vary. Perkin-Elmer Elan 6000 ICP/MS used for determination.

† Known-addition level 20 µg/L. Additions made before preparation according to Section 3030E (modified by cleanhood digestion in TFE beakers). *N* = 20.

‡ Known-addition level for Cd and Pb 1 µg/L; for other analytes 10 µg/L. Additions made after filtration through 1:1 HNO₃ precleaned 0.45-µm filters. *N* = 28.

TABLE 3125:IX. METHOD PERFORMANCE WITH LOW-LEVEL CHECK STANDARDS*

Element	Mass	1.0- $\mu\text{g/L}$ Standard			0.3- $\mu\text{g/L}$ Standard				
		Mean Recovery %	Mean $\mu\text{g/L}$	Standard Deviation $\mu\text{g/L}$	Relative Standard Deviation %	Mean Recovery %	Mean $\mu\text{g/L}$	Standard Deviation $\mu\text{g/L}$	Relative Standard Deviation %
Be	9	97	0.97	0.06	6.24	95	0.284	0.03	12.11
Al	27	121	1.21	0.32	26.49	196	0.588	0.44	74.30
V	51	104	1.04	0.06	5.83	111	0.332	0.10	28.96
Cr	52	119	1.19	0.34	28.62	163	0.490	0.37	75.90
Cr	53	102	1.02	0.36	35.54	113	0.338	0.32	93.70
Mn	55	103	1.03	0.07	6.55	110	0.329	0.08	25.64
Co	59	103	1.03	0.07	6.42	102	0.307	0.04	12.53
Ni	60	101	1.01	0.05	5.24	107	0.321	0.05	14.14
Ni	62	102	1.02	0.06	5.42	109	0.326	0.05	15.94
Cu	63	107	1.07	0.09	8.78	118	0.355	0.06	18.29
Cu	65	107	1.07	0.10	9.05	117	0.352	0.06	17.69
Zn	66	117	1.17	0.51	43.52	182	0.547	0.68	124.13
Zn	68	116	1.16	0.50	42.90	179	0.537	0.66	122.12
As	75	97	0.97	0.05	5.23	101	0.302	0.06	18.29
Se	77	89	0.89	0.08	8.72	88	0.265	0.08	29.07
Se	82	92	0.92	0.14	15.50	106	0.317	0.14	43.91
Ag	107	101	1.01	0.05	4.53	94	0.282	0.04	15.74
Ag	109	103	1.03	0.07	6.57	92	0.277	0.04	13.68
Cd	111	98	0.98	0.04	3.80	96	0.288	0.03	8.74
Cd	114	100	1.00	0.03	3.39	98	0.293	0.03	8.70
Sb	121	94	0.94	0.05	5.28	93	0.280	0.06	21.89
Sb	123	94	0.94	0.05	5.36	93	0.278	0.06	22.39
Tl	203	101	1.01	0.04	3.57	98	0.294	0.03	11.89
Tl	205	104	1.04	0.05	5.15	100	0.300	0.03	10.43
Pb	208	104	1.04	0.04	3.65	104	0.312	0.03	11.13
U	238	106	1.06	0.05	4.64	102	0.307	0.03	9.92

* Single-laboratory, single-operator, single-instrument data. $N = 24$ for both standards.

a. Correction for dilutions and solids: Correct all results for dilutions, and raise reporting limit for all analytes reported from the diluted sample by a corresponding amount. Similarly, if results for solid samples are to be determined, use Method 2540B to determine total solids. Report results for solid samples as micrograms per kilogram, dry weight. Correct all results for solids content of solid samples. Use the following equation to correct solid or sediment sample results for dilution during digestion and moisture content:

$$R_{corr} = \frac{R_{uncorr} \times V}{W \times \% TS/100}$$

where:

R_{corr} = corrected result, $\mu\text{g/kg}$,

R_{uncorr} = uncorrected elemental result, $\mu\text{g/L}$,

V = volume of digestate (after digestion), L,

W = mass of the wet sample, kg, and

$\% TS$ = percent total solids determined in the solid sample.

b. Compensation for interferences: Use instrument software to correct for interferences listed previously for this method. See Table 3125:III for a listing of the most common molecular ion interferences.

c. Data reporting: Establish appropriate reporting limits for method analytes based on instrument detection limits and the

laboratory blank. For regulatory programs, ensure that reporting limits for method analytes are a factor of three below relevant regulatory criteria.

If method blank contamination is typically random, sporadic, or otherwise not in statistical control, do not correct results for the method blank. Consider the correction of results for laboratory method blanks only if it can be demonstrated that the concentration of analytes in the method blank is within statistical control over a period of months. Report all method blank data explicitly in a manner identical to sample reporting procedures.

d. Documentation: Maintain documentation for the following (where applicable): instrument tuning, mass calibration, calibration verification, analyses of blanks (method, field, calibration, and equipment blanks), IDL and MDL studies, analyses of samples and duplicates with known additions, laboratory and field duplicate information, serial dilutions, internal standard recoveries, and any relevant quality control charts.

Also maintain, and keep available for review, all raw data generated in support of the method.⁵

6. Method Performance

Table 3125:I presents instrument detection limit (IDL) data generated by this method; this represents optimal state-of-the-art instrument detection capabilities, not recommended

method detection or reporting limits. Tables 3125:VII through IX contain single-laboratory, single-operator, single-instrument performance data generated by this method for calibration verification standards, low-level standards, and known-addition recoveries for fresh-water matrices. Performance data for this method for some analytes are not currently available. However, performance data for similar ICP/MS methods are available in the literature.^{1,4}

7. References

1. U.S. ENVIRONMENTAL PROTECTION AGENCY. 1994. Determination of trace elements in waters and wastes by inductively coupled plasma-mass spectrometry, Method 200.8. U.S. Environmental Protection Agency, Environmental Monitoring Systems Lab., Cincinnati, Ohio.
2. U.S. ENVIRONMENTAL PROTECTION AGENCY. 1984. Definition and procedure for the determination of the method detection limit, revision 1.11. 40 CFR 136, Appendix B.
3. U.S. ENVIRONMENTAL PROTECTION AGENCY. 1991. Methods for the determination of metals in environmental samples. U.S. Environmental Protection Agency, Off. Research & Development, Washington D.C.
4. U.S. ENVIRONMENTAL PROTECTION AGENCY. 1995. Method 1638: Determination of trace elements in ambient waters by inductively coupled plasma mass spectrometry. U.S. Environmental Protection Agency, Off. Water, Washington, D.C.
5. U.S. ENVIRONMENTAL PROTECTION AGENCY. 1995. Guidance on the Documentation and Evaluation of Trace Metals Data Collected for Clean Water Act Compliance Monitoring. U.S. Environmental Protection Agency, Off. Water, Washington, D.C.

8. Bibliography

- GRAY, A.L. 1974. A plasma source for mass analysis. *Proc. Soc. Anal. Chem.* 11:182.
- HAYHURST, A.N. & N.R. TELFORD. 1977. Mass spectrometric sampling of ions from atmospheric pressure flames. I. Characteristics and calibration of the sampling system. *Combust. Flame.* 67.
- HOUK, R.S., V.A. FASSEL, G.D. FLESCH, H.J. SVEC, A.L. GRAY & C.E. TAYLOR. 1980. Inductively coupled argon plasma as an ion source

- for mass spectrometric determination of trace elements. *Anal. Chem.* 52:2283.
- DOUGLAS, D.J. & J.B. FRENCH. 1981. Elemental analysis with a microwave-induced plasma/quadrupole mass spectrometer system. *Anal. Chem.* 53:37.
- HOUK, R.S., V.A. FASSEL & H.J. SVEC. 1981. Inductively coupled plasma-mass spectrometry: Sample introduction, ionization, ion extraction and analytical results. *Dyn. Mass Spectrom.* 6:234.
- OLIVARES, J.A. & R.S. HOUK. 1985. Ion sampling for inductively coupled plasma mass spectrometry. *Anal. Chem.* 57:2674.
- HOUK, R.S. 1986. Mass spectrometry of inductively coupled plasmas. *Anal. Chem.* 58:97.
- THOMPSON, J.J. & R.S. HOUK. 1986. Inductively coupled plasma mass spectrometric detection for multielement flow injection analysis and elemental speciation by reversed-phase liquid chromatography. *Anal. Chem.* 58:2541.
- VAUGHAN, M.A. & G. HORLICK. 1986. Oxide, hydroxide, and doubly charged analyte species in inductively coupled plasma/mass spectrometry. *Appl. Spectrosc.* 40:434.
- GARBARINO, J.R. & H.E. TAYLOR. 1987. Stable isotope dilution analysis of hydrologic samples by inductively coupled plasma mass spectrometry. *Anal. Chem.* 59:1568.
- BEAUCHEMIN, D., J.W. McLAREN, A.P. MYKYTIUK & S.S. BERMAN. 1987. Determination of trace metals in a river water reference material by inductively coupled plasma mass spectrometry. *Anal. Chem.* 59:778.
- THOMPSON, J.J. & R.S. HOUK. 1987. A study of internal standardization in inductively coupled plasma-mass spectrometry. *Appl. Spectrosc.* 41:801.
- JARVIS, K.E., A.L. GRAY & R.S. HOUK. 1992. Inductively Coupled Plasma Mass Spectrometry. Blackie Academic & Professional, Chapman & Hall, New York, N.Y.
- TAYLOR, D.B., H.M. KINGSTON, D.J. NOGAY, D. KOLLER & R. HUTTON. 1996. On-line solid-phase chelation for the determination of eight metals in environmental waters by ICP-MS. *JAAS* 11:187.
- KINGSTON, H.M.S. & S. HASWELL, eds. 1997. Microwave Enhanced Chemistry: Fundamentals, Sample Preparation, and Applications. ACS Professional Reference Book Ser., American Chemical Soc., Washington, D.C.
- U.S. ENVIRONMENTAL PROTECTION AGENCY. 1998. Inductively coupled plasma-mass spectrometry, Method 6020. *In* Solid Waste Methods. SW846, Update 4, U.S. Environmental Protection Agency, Environmental Monitoring Systems Lab., Cincinnati, Ohio.

3130 METALS BY ANODIC STRIPPING VOLTAMMETRY*

3130 A. Introduction

Anodic stripping voltammetry (ASV) is one of the most sensitive metal analysis techniques; it is as much as 10 to 100 times more sensitive than electrothermal atomic absorption spectroscopy for some metals. This corresponds to detection limits in the nanogram-per-liter range. The technique requires no sample

extraction or preconcentration, it is nondestructive, and it allows simultaneous determination of four to six trace metals, utilizing inexpensive instrumentation. The disadvantages of ASV are that it is restricted to amalgam-forming metals, analysis time is longer than for spectroscopic methods, and interferences and high sensitivity can present severe limitations. The analysis should be performed only by analysts skilled in ASV methodology because of the interferences and potential for trace background contamination.

* Approved by Standard Methods Committee, 1997.
Joint Task Group: 20th Edition—Malgorzata Ciszowska (chair), Margaret M. Goldberg, Janet G. Osteryoung, Robert S. Rodgers, Marek Wojciechowski.

3130 B. Determination of Lead, Cadmium, and Zinc

1. General Discussion

a. Principle: Anodic stripping voltammetry is a two-step electroanalytical technique. In the preconcentration step, metal ions in the sample solution are reduced at negative potential and concentrated into a mercury electrode. The concentration of the metal in the mercury is 100 to 1000 times greater than that of the metal ion in the sample solution. The preconcentration step is followed by a stripping step applying a positive potential scan. The amalgamated metal is oxidized rapidly and the accompanying current is proportional to metal concentration.

b. Detection limits and working range: The limit of detection for metal determination using ASV depends on the metal determined, deposition time, stirring rate, solution pH, sample matrix, working electrode (hanging mercury drop electrode, HMDE, or thin mercury film electrode, TMFE), and mode of the stripping potential scan (square wave or differential pulse). Cadmium, lead, and zinc are concentrated efficiently during pre-electrolysis because of their high solubility in mercury and thus have low detection limits ($<1 \mu\text{g/L}$). Long deposition times and high stirring rates increase the concentration of metal preconcentrated in the mercury phase and reduce detection limits. The effects of solution pH and matrix are more complicated. In general, add a high concentration of inert electrolyte to samples to maintain a high, constant ionic strength. Acidify sample to a low pH or add a pH buffer. If the pH buffer or other component of the sample matrix complexes the metal (3130B.1c), detection limits often are increased.

The choice of working electrode is determined largely by the working range of concentration required. The HMDE is best suited for analysis from approximately $1 \mu\text{g/L}$ to 10 mg/L , while the TMFE is superior for detection below $1 \mu\text{g/L}$.

c. Interferences: Major interferences include intermetallic compound formation, overlapping stripping peaks, adsorption of organics, and complexation. Intermetallic compounds can form in the mercury phase when high concentrations of certain metals are present simultaneously. Zinc forms intermetallic compounds with cobalt and nickel, and both zinc and cadmium form intermetallic compounds with copper, silver, and gold. As a result, the stripping peak for the constituent metals may be severely depressed or shifted and additional peaks due to intermetallic compound stripping may be observed. Minimize or avoid intermetallic compound formation by use of a hanging mercury drop electrode instead of a thin film mercury electrode when metal concentrations are above $1 \mu\text{g/L}$, application of a preconcentration potential sufficiently negative to reduce the desired but not the interfering metal, and use of a relatively short preconcentration period followed by a relatively large pulse modulation (50 mV) during the stripping stage. In general, suspect formation of intermetallic compounds if metals are present in concentrations above 1 mg/L . If metals are present at concentrations above 10 mg/L , do not use anodic stripping voltammetry. Concentrations above 10 mg/L usually can be quantitated by methods such as those given in Sections 3111 and 3120.

Separate overlapping stripping peaks by various methods, including appropriate choice of buffer and electrolyte.¹⁻³ If only one of the metal peaks is of interest, eliminate interfering peaks

by selective complexation with a suitable ligand, such as EDTA. Judicious choice of preconcentration potential can result in the deposition of the selected metal but not the interfering metal in the mercury electrode. Selection of buffer/ligand also may help to distinguish metals during the preconcentration step. Alternatively, use "medium exchange," in which preconcentration is performed with the electrodes in the sample and stripping is performed in a different electrolyte solution. In this procedure, metals are deposited from the sample into the amalgam as usual, but they may be stripped into a medium that provides different stripping peak potentials for the overlapping metals.

Minimize interferences from adsorption of organic compounds and complexation by removal of the organic matter. Digest samples with high-purity acids as described in Section 3030. Make standard additions to determine if complexation or adsorption remains a problem. Analyze a metal-free solution with a matrix similar to that of the sample both before and after addition of known quantities of standard. Repeat procedure for sample. If the slope of the stripping current versus added metal is significantly different in the sample relative to the metal-free solution, digest sample further. The choice of stripping waveform also is important. While both square-wave and differential-pulse stripping attempt to minimize the contribution of adsorption currents to the total measured stripping current, square-wave stripping does this more effectively. Thus, use square-wave stripping instead of differential-pulse stripping when adsorption occurs.

2. Apparatus

a. Electrochemical analyzer: The basic electrochemical analyzer for ASV applications contains a three-electrode potentiostat, which very precisely controls potential applied to the working electrode relative to the reference electrode, and a sensitive current measuring device. It is capable of delivering potential pulses of various amplitudes and frequencies, and provides several scan rates and current ranges. More advanced ASV instruments offer automated timing, gas purge and stirring, and data processing routines including curve smoothing, baseline correction, and background subtraction.

Two variations of stripping waveforms are commonly used: differential pulse (DPASV) and square wave (SWASV) waveforms. The differential-pulse waveform consists of a series of pulses superimposed on a linear voltage ramp, while the square-wave waveform consists of a series of pulses superimposed on a staircase potential waveform. Square-wave stripping is significantly faster than differential-pulse stripping and is typically ten times more sensitive. Most commercially available ASV instruments perform both differential-pulse and square-wave stripping.

b. Electrodes and cell: Provide working, reference, and auxiliary electrodes. Working electrodes are either hanging mercury drop or thin mercury film electrodes. Hanging mercury drop electrodes must be capable of dispensing mercury in very precisely controlled drop sizes. Three types of electrodes meet this requirement: static mercury drop electrodes, controlled growth mercury drop electrodes, or Kemula-type electrodes. In any case,

use a drop knocker to remove an old drop before dispensing a fresh mercury drop.

When the lowest detection limits are required, a thin mercury film electrode is preferred. This electrode consists of a rotating glassy carbon disk plated with mercury in situ during preconcentration of the analyte. A high-precision, constant-speed rotator controls the rotation rate of the electrode and provides reproducible mass transport.

Reference electrodes may be either saturated calomel or silver/silver chloride electrodes. Use a platinum wire for the auxiliary electrode.

Use cells constructed of glass, or preferably fused silica or TFE, because they are more resistant to solution adsorption or leaching. Cover cell with a lid that provides reproducible placement of the electrodes and gas purging tubes. Provide an additional hole in the cell lid for addition of standards. Most commercially available mercury drop electrodes include electrolytic cells, reference and auxiliary electrodes, and gas purging tubes.

Use a constant-speed stirring mechanism to provide reproducible mass transport in samples and standards.

Locate the cell in an area where temperature is relatively constant. Alternatively, use a constant-temperature water bath and cell jacket.

c. Oxygen-removal apparatus: Oxygen interferes in electrochemical analyses; remove it from solution before preconcentration by purging with nitrogen or argon. Provide two gas inlet tubes through the cell lid: one extends into the solution and the second purges the space above the solution. A gas outlet hole in the lid provides for removal of oxygen and excess purging gas.

d. Recording device: If the electrochemical analyzer is not equipped with a digital data acquisition system, use an XY plotter to record stripping voltammograms.

e. Timer: If preconcentration and equilibration periods are not controlled by the instrument, use an accurate timing device.

f. Polishing wheel: To obtain the high polish required for a glassy carbon disk electrode, use a motorized polishing wheel.

3. Reagents

CAUTION: Follow proper practices for disposal of any solutions containing mercury.

a. Metal-free water: Use deionized water to prepare buffers, electrolytes, standards, etc. Use water with at least 18 megohm-cm resistivity (see Section 1080).

*b. Nitric acid, HNO₃, conc, high-purity.**

c. Nitric acid, HNO₃, 6N, 1.6N (10%), and 0.01N.

d. Purging gas (nitrogen or argon), high-purity. Remove traces of oxygen in nitrogen or argon gas before purging the solution. Pass gas through sequential scrubbing columns containing vanadous chloride in the first, deionized water in the second, and buffer (or electrolyte) solution in the third column.

e. Metal standards: Prepare stock solutions containing 1 mg metal/mL in polyethylene bottles. Purchase these solutions commercially or prepare as in Section 3111. Daily prepare dilutions of stock standards in a matrix similar to that of the samples to cover the concentration range desired.

f. Electrolyte/buffer: Use one of the following:

1) *Acetate buffer, pH 4.5:* Dissolve 16.4 g anhydrous sodium acetate, NaC₂H₃O₂, in 800 mL water. Adjust to pH 4.5 with high-purity glacial acetic acid.* Dilute to 1 L with water.

2) *Citrate buffer, pH 3:* Dissolve 42.5 g citric acid (monohydrate) in 700 mL water. Adjust to pH 3 with high-purity NH₄OH.* Dilute to 1 L with water.

3) *Phosphate buffer, pH 6.8:* Dissolve 24 g NaH₂PO₄ in 500 mL water. Adjust to pH 6.8 with 1N NaOH. Dilute to 1 L with water.

g. Mercury: Use commercially available triply distilled metallic mercury for hanging mercury drop electrodes. **CAUTION:** Mercury vapors are highly toxic. Use only in well-ventilated area.

h. Mercuric nitrate solution, Hg(NO₃)₂: For thin mercury film electrodes, dissolve 0.325 g Hg(NO₃)₂ in 100 mL 0.01N HNO₃.

i. Reference electrode filling solution: Available from electrode manufacturer.

j. Amalgamated zinc: Dissolve 2 g Hg(NO₃)₂ in 25 mL conc HNO₃; dilute to 250 mL with water. In a separate beaker, clean approximately 50 g mossy zinc by gently oxidizing the surface with 10% HNO₃ and rinse with water. Add Hg(NO₃)₂ solution to cleaned zinc and stir with a glass rod. If barely visible bubbles do not appear, add a small amount of 6N HNO₃. Zinc should rapidly acquire a shiny, metallic appearance. Decant solution and store for amalgamating future batches of zinc. Rinse amalgamated zinc copiously with water and transfer to a gas scrubbing column.

k. Hydrochloric acid, HCl, conc.

l. Vanadous chloride: Add 2 g ammonium metavanadate, NH₄VO₃, to 25 mL conc HCl and heat to boiling. Solution should turn blue-green. Dilute to 250 mL with water. Pour solution into gas scrubbing column packed with amalgamated zinc and bubble purging gas through it until the solution turns a clear violet color. When the violet color is replaced by a blue, green, or brown color, regenerate vanadous chloride by adding HCl.

m. Siliconizing solution: Preferably use commercially available solutions in sealed ampules for siliconizing capillaries used for hanging mercury drops. **CAUTION:** Most commercial siliconizing reagents contain CCl₄, a toxic and cancer-suspect agent. Handle with gloves and avoid breathing vapors.

n. Alumina suspensions, 1, 0.3, and 0.05 μm. Use commercially available alumina suspensions in water, or make a suspension by adding a small amount of water to the alumina.

o. Hydrofluoric acid, HF, 5%: Dilute 5 mL conc HF to 100 mL with water.

p. Methanol.

q. Sodium hydroxide, NaOH, 1N.

4. Procedure

a. Sample preparation and storage: Collect samples in pre-cleaned, acid-soaked polyethylene or TFE bottles. Add 2 mL conc HNO₃/L sample and mix well. Cap tightly and store in refrigerator or freezer until ready for analysis.

b. Cell preparation: Soak clean cell in 6N HNO₃ overnight and rinse well with water before use.

* Ultrex, Suprapur, Aristar, or equivalent.

c. *Electrode preparation:*

1) **HMDE**—Follow manufacturer's guidance for capillary cleaning. If not available, use the following procedure. Remove all mercury from the capillary. Aspirate the following through the capillary in the order listed: 6*N* HNO₃, water, 5% HF, water, methanol, and air. Dry capillary at 100°C for 1 h. Siliconize cooled capillary using a siliconizing solution. Between uses, fill capillary with clean mercury and immerse tip in clean mercury. If the capillary fails to suspend a drop of mercury, repeat cleaning.

2) **TMFE**—Polish glassy carbon disks used for thin mercury film electrodes to a high metallic sheen with alumina suspensions, progressively decreasing particle size from 1 μm to 0.05 μm. Use a motorized polishing wheel for best results. Completely rinse off all traces of alumina with water. Check disk frequently for etching or pitting; repolish as necessary to maintain reproducible mercury film deposition.

d. *Instrumental conditions:* Use the following conditions:

Variable	Value
Initial potential	−1.00 V (Pb, Cd); −1.20 V (Zn)
Final potential	0.00 V
Equilibration potential	−1.00 V (Pb, Cd); −1.20 V (Zn)
HMDE drop size	medium
TMFE rotation rate	2000 rpm
DPASV:	
Pulse amplitude	25 mV
Pulse period	0.5 s
Pulse width	50 ms
Sample width	17 ms
Scan rate	5 mV/s
SWASV:	
SW amplitude	25 mV
Step potential	4 mV
Frequency	100 Hz

e. *Deoxygenation:* Pipet 2 mL sample and 3 mL electrolyte/buffer into cell. If using a TMFE, add 10 μL Hg(NO₃)₂ solution. Place electrodes in cell and secure cell lid. Deoxygenate solution with purified purging gas for 10 min while stirring. When solution purge is completed, purge the space above the solution with purified purging gas. Continue head-space purge throughout analysis.

f. *Preconcentration:* If using a HMDE, dispense a new mercury drop. Start preconcentration, stirring and timing simultaneously. Precisely control and keep constant preconcentration times and stirring rates for solutions and standards. Generally use 120 s and a rotation rate of 2000 rpm for the TMFE.

After the metal is sufficiently concentrated in the amalgam, stop stirring or TMFE rotation for an equilibration period of precisely 30 s.

g. *Anodic stripping:* After equilibration period, begin anodic stripping without stirring and make potential at the working electrode progressively more positive as a function of time. Monitor stripping current and plot as a function of applied potential in stripping voltammograms. Use peak current to quantify metal concentration and peak potential to identify the metal.

h. Add 5 to 50 μL standard solution and repeat analysis, beginning with deoxygenation of sample. Adjust volume of added standard solution to obtain 30 to 70% increase of the stripping peak. If 50 μL addition is not sufficient, use standard solution with a higher concentration of metal. Shorten deoxygenation step to 1 min after initial gas purge.

5. Calculations

Calculate the concentration of metal in the original sample using the following equation:

$$C_o = \frac{C_s \times V_s}{V_o} \times \frac{i_o}{(i_s - i_o)}$$

where:

C_o = concentration of metal in sample, mg/L,

C_s = concentration of metal in standard solution, mg/L,

i_o = stripping peak height in original sample,

i_s = stripping peak height in sample with standard addition,

V_o = volume of sample, mL, and

V_s = volume of standard solution added, mL.

6. Quality Control

Follow quality control guidelines outlined in Section 3020 with respect to use of additions, duplicates, and blanks for best

TABLE 3130:I. PRECISION OF Cd, Pb, AND Zn ANALYSIS BY ASV

Sample	Electrode	ASV Mode	Metal Concentration μg/L			RSD %		
			Cd	Pb	Zn	Cd	Pb	Zn
Tap water #1 ⁴	HMDE	SW	0.068	0.57	—	4.2	4.8	—
Tap water #2 ⁴	HMDE	SW	—	2.50	—	—	5.1	—
Seawater #1 ⁵	TFME	DP	0.0121	0.0086	—	10.7	8.1	—
Seawater #2 ⁵	TFME	DP	0.032	0.032	—	6.3	6.3	—
Soil extract #1 ⁶	HMDE	SW	189	11.8	—	2.5	5.6	—
Soil extract #1 ⁶	HMDE	DP	186	11.9	—	2.5	4.0	—
Deionized water ⁴	HMDE	SW	0.13	0.79	—	5.5	2.2	—
Wastewater #1 ⁷	HMDE	DP	—	74	26	—	4.3	4.5
Wastewater #2 ⁷	HMDE	DP	—	47	86	—	5.2	6.3
Wastewater #3 ⁷	HMDE	DP	—	46	65	—	4.6	6.2
Wastewater #4 ⁸	TFME	SW	5.2	60	12	5.2	6.1	7.4

results. Blanks are critical because of the high sensitivity of the method.

7. Precision and Bias

Table 3130:I gives precision data for analyses of samples with various matrices.

8. References

1. WANG, J. 1985. *Stripping Analysis: Principles, Instrumentation, and Applications*. VCH Publishers, Inc., Deerfield Beach, Fla.
2. BRAININA, K.H. & E. NEYMAN. 1993. *Electroanalytical Stripping Methods*. John Wiley & Sons, New York, N.Y.
3. KISSINGER, P.T. & W.R. HEINEMAN. 1996. *Laboratory Techniques in Electroanalytical Chemistry*, 2nd ed. Marcel Dekker, Inc., New York, N.Y.
4. MARTIN-GOLDBERG, M. 1989. Unpublished data. ResearchTriangle Institute, Metals Analysis Facility. ResearchTriangle Park, N.C.
5. BRULAND, K.W., K.H. COALE & L. MART. 1985. Analysis of seawater for dissolved Cd, Cu, and Pb: An intercomparison of voltammetric and atomic absorption methods. *Mar. Chem.* 17:285.
6. OSTAPCZUK, P., P. VALENTA & H.W. NURNBERG. 1986. Square wave voltammetry—a rapid and reliable determination method of Zn, Cd, Pb, Ni, and Co in biological and environmental samples. *J. Electroanal. Chem.* 214:51.

7. CLARK, B.R., D.W. DEPAOLI, D.R. McTAGGART & B.D. PATTON. 1988. An on-line voltammetric analyzer for trace metals in wastewater. *Anal. Chim. Acta* 215:13.
8. BRETT, C.M.A., A.M. OLIVEIRA-BRETT & L. TUGULEA. 1996. Anodic stripping voltammetry of trace metals by batch injection analysis. *Anal. Chim. Acta* 322:151.

9. Bibliography

- E.G. & G. PRINCETON APPLIED RESEARCH CORP. 1980. Differential pulse anodic stripping voltammetry of water and wastewater. Application Note W-1. Princeton, N.J.
- PETERSON, W.M. & R.V. WONG. 1981. Fundamentals of stripping voltammetry. *Amer. Lab.* 13(11):116.
- E.G. & G. PRINCETON APPLIED RESEARCH CORP. 1982. Basics of voltammetry and polarography. Application Note P-2. Princeton, N.J.
- NURNBERG, H.W. 1984. The voltammetric approach in trace metal chemistry of natural waters and atmospheric precipitation. *Anal. Chim. Acta* 164:1.
- SHUMAN, M.S. & MARTIN-GOLDBERG, M. 1984. Electrochemical methods: Anodic stripping. In R. Minear & L. Keith, eds. *Water Analysis: Inorganic Species*. II:345. Academic Press, Orlando, Fla.
- FLORENCE, T.M. 1986. Electrochemical approaches to trace element speciation in waters: A review. *Analyst* 111:489.
- FLORENCE, T.M. 1992. Trace element speciation by anodic stripping voltammetry. *Analyst* 117:551.
- TERCIER, M.-L. & J. BUFFLE. 1993. In situ voltammetric measurements in natural waters: Future prospects and challenges. *Electroanalysis* 5:187.

3500-AI ALUMINUM*

3500-AI A. Introduction

1. Occurrence and Significance

Aluminum (Al) is the second element in Group IIIA of the periodic table; it has an atomic number of 13, an atomic weight of 26.98, and a valence of 3. The average abundance in the earth's crust is 8.1%; in soils it is 0.9 to 6.5%; in streams it is 400 $\mu\text{g/L}$; in U.S. drinking waters it is 54 $\mu\text{g/L}$, and in groundwater it is <0.1 $\mu\text{g/L}$. Aluminum occurs in the earth's crust in combination with silicon and oxygen to formfeldspars, micas, and clay minerals. The most important minerals are bauxite and corundum, which is used as an abrasive. Aluminum and its alloys are used for heat exchangers, aircraft parts, building materials, containers, etc. Aluminum potassium sulfate (alum) is used in water-treatment processes to flocculate suspended particles, but it may leave a residue of aluminum in the finished water.

Aluminum's occurrence in natural waters is controlled by pH and by very finely suspended mineral particles. The cation Al^{3+} predominates at pH less than 4. Above neutral pH, the predominant dissolved form is $\text{Al}(\text{OH})_4^-$. Aluminum is nonessential for plants and animals. Concentrations exceeding 1.5 mg/L constitute a toxicity hazard in the marine environment, and levels below 200 $\mu\text{g/L}$ present a minimal risk. The United Nations Food and Agriculture Organization's recommended maximum level for irrigation waters is 5 mg/L. The possibility of a link between elevated aluminum levels in brain tissues and Alzheimer's disease has been raised. The U.S. EPA secondary drinking water regulations list an optimal secondary maximum contaminant level (SMCL) of 0.05 mg/L and maximum SMCL of 0.2 mg/L.

2. Selection of Method

The atomic absorption spectrometric methods (3111D and E, and 3113B) and the inductively coupled plasma methods (3120 and 3125) are free from such common interferences as fluoride and phosphate, and are preferred. The Eriochrome cyanine R colorimetric method (B) provides a means for estimating aluminum with simpler instrumentation.

* Approved by Standard Methods Committee, 2001.

Joint Task Group: 20th Edition—Brian J. Condike (chair), Deanna K. Anderson, Anthony Bright, Richard A. Cahill, Alois F. Clary, C. Ellen Gonter, Peter M. Grohse, Daniel C. Hillman, Albert C. Holler, Amy Hughes, J. Charles Jennett, Roger A. Minear, Marlene O. Moore, Gregg L. Oelker, S. Kusum Perera, James G. Poff, Jeffrey G. Skousen, Michael D. Wichman, John L. Wuepper.

3500-AI B. Eriochrome Cyanine R Method

1. General Discussion

a. Principle: With Eriochrome cyanine R dye, dilute aluminum solutions buffered to a pH of 6.0 produce a red to pink complex that exhibits maximum absorption at 535 nm. The intensity of the developed color is influenced by the aluminum concentration, reaction time, temperature, pH, alkalinity, and concentration of other ions in the sample. To compensate for color and turbidity, the aluminum in one portion of sample is complexed with EDTA to provide a blank. The interference of iron and manganese, two elements commonly found in water when aluminum is present, is eliminated by adding ascorbic acid. The optimum aluminum range lies between 20 and 300 $\mu\text{g/L}$ but can be extended upward by sample dilution.

b. Interference: Negative errors are caused by both fluoride and polyphosphates. When the fluoride concentration is constant, the percentage error decreases with increasing amounts of aluminum. Because the fluoride concentration often is known or can be determined readily, fairly accurate results can be obtained by adding the known amount of fluoride to a set of standards. A simpler correction can be determined from the family of curves in Figure 3500-AI:1. A procedure is given for the removal of complex phosphate interference. Orthophosphate in concentrations under 10 mg/L does not interfere. The interference caused by even small amounts of alkalinity is removed by acidifying the sample just beyond the neutralization point of methyl orange. Sulfate does not interfere up to a concentration of 2000 mg/L.

c. Minimum detectable concentration: The minimum aluminum concentration detectable by this method in the absence of fluorides and complex phosphates is approximately 6 $\mu\text{g/L}$.

d. Sample handling: Collect samples in clean, acid-rinsed bottles, preferably plastic, and examine them as soon as possible after collection. If only soluble aluminum is to be determined, filter a portion of sample through a 0.45- μm membrane filter; discard first 50 mL of filtrate and use succeeding filtrate for the determination. Do not use filter paper, absorbent cotton, or glass wool for filtering any solution that is to be tested for aluminum, because they will remove most of the soluble aluminum.

2. Apparatus

a. Colorimetric equipment: One of the following is required:

- 1) *Spectrophotometer*, for use at 535 nm, with a light path of 1 cm or longer.
- 2) *Filter photometer*, providing a light path of 1 cm or longer and equipped with a green filter with maximum transmittance between 525 and 535 nm.
- 3) *Nessler tubes*, 50-mL, tall form, matched.

b. Glassware: Treat all glassware with warm 1 + 1 HCl and rinse with aluminum-free distilled water to avoid errors due to materials absorbed on the glass. Rinse sufficiently to remove all acid.

3. Reagents

Use reagents low in aluminum, and aluminum-free distilled water.

a. Stock aluminum solution: Use either the metal (1) or the salt (2) for preparing stock solution; 1.00 mL = 500 $\mu\text{g Al}$:

1) Dissolve 500.0 mg aluminum metal in 10 mL conc HCl by heating gently. Dilute to 1000 mL with water, or

2) Dissolve 8.791 g aluminum potassium sulfate (also called potassium alum), $\text{AlK}(\text{SO}_4)_2 \cdot 12\text{H}_2\text{O}$, in water and dilute to 1000 mL. Correct this weight by dividing by the decimal fraction of assayed $\text{AlK}(\text{SO}_4)_2 \cdot 12\text{H}_2\text{O}$ in the reagent used.

b. Standard aluminum solution: Dilute 10.00 mL stock aluminum solution to 1000 mL with water; 1.00 mL = 5.00 $\mu\text{g Al}$. Prepare daily.

c. Sulfuric acid, H_2SO_4 , 0.02N and 6N.

d. Ascorbic acid solution: Dissolve 0.1 g ascorbic acid in water and make up to 100 mL in a volumetric flask. Prepare fresh daily.

e. Buffer reagent: Dissolve 136 g sodium acetate, $\text{NaC}_2\text{H}_3\text{O}_2 \cdot 3\text{H}_2\text{O}$, in water, add 40 mL 1N acetic acid, and dilute to 1 L.

f. Stock dye solution: Use any of the following products:

1) *Solochrome cyanine R-200** or *Eriochrome cyanine*:[†] Dissolve 100 mg in water and dilute to 100 mL in a volumetric flask. This solution should have a pH of about 2.9.

2) *Eriochrome cyanine R*:[‡] Dissolve 300 mg dye in about 50 mL water. Adjust pH from about 9 to about 2.9 with 1 + 1 acetic acid (approximately 3 mL will be required). Dilute with water to 100 mL.

3) *Eriochrome cyanine R*:[§] Dissolve 150 mg in about 50 mL water. Adjust pH from about 9 to about 2.9 with 1 + 1 acetic acid (approximately 2 mL will be required). Dilute with water to 100 mL.

Stock solutions have excellent stability and can be kept for at least a year.

g. Working dye solution: Dilute 10.0 mL of selected stock dye solution to 100 mL in a volumetric flask with water. Working solutions are stable for at least 6 months.

h. Methyl orange indicator solution, or bromcresol green indicator solution specified in the total alkalinity determination (Section 2320B.3d).

i. EDTA (sodium salt of ethylenediamine-tetraacetic acid dihydrate), 0.01M: Dissolve 3.7 g in water, and dilute to 1 L.

j. Sodium hydroxide, NaOH, 1N and 0.1N.

4. Procedure

a. Preparation of calibration curve:

1) Prepare a series of aluminum standards from 0 to 7 μg (0 to 280 $\mu\text{g/L}$ based on a 25-mL sample) by accurately measuring the calculated volumes of standard aluminum solution into 50-mL volumetric flasks or nessler tubes. Add water to a total volume of approximately 25 mL.

2) Add 1 mL 0.02N H_2SO_4 to each standard and mix. Add 1 mL ascorbic acid solution and mix. Add 10 mL buffer solution

* Arnold Hoffman & Co., Providence, RI.

† K & K Laboratories, K & K Lab. Div., Life Sciences Group, Plainview, NY.

‡ Pfaltz & Bauer, Inc., Stamford, CT.

§ EM Science, Gibbstown, NJ.




Figure 3500-Al:1. Correction curves for estimation of aluminum in the presence of fluoride. Above the mg F⁻/L present, locate the point corresponding to the apparent mg Al/L measured. From this point interpolate between the curves shown, if the point does not fall directly on one of the curves, to read the true mg Al/L on the ordinate, which corresponds to 0.00 mg F⁻/L. For example, an apparent 0.20 mg Al/L in a sample containing 1.00 mg F⁻/L would actually be 0.30 mg Al/L if no fluoride were present to interfere.

and mix. With a volumetric pipet, add 5.00 mL working dye reagent and mix. Immediately make up to 50 mL with distilled water. Mix and let stand for 5 to 10 min. The color begins to fade after 15 min.

3) Read transmittance or absorbance on a spectrophotometer, using a wavelength of 535 nm or a green filter providing maximum transmittance between 525 and 535 nm. Adjust instrument to zero absorbance with the standard containing no aluminum.

Plot concentration of Al (micrograms Al in 50 mL final volume) against absorbance.

b. Sample treatment in absence of fluoride and complex phosphates: Place 25.0 mL sample, or a portion diluted to 25 mL, in a porcelain dish or flask, add a few drops of methyl orange indicator, and titrate with 0.02N H₂SO₄ to a faint pink color. Record reading and discard sample. To two similar samples at room temperature add the same amount of 0.02N H₂SO₄ used in the titration and 1 mL in excess.

To one sample add 1 mL EDTA solution. This will serve as a blank by complexing any aluminum present and compensating for color and turbidity. To both samples add 1 mL ascorbic acid,

10 mL buffer reagent, and 5.00 mL working dye reagent as prescribed in ¶ a2) above.

Set instrument to zero absorbance or 100% transmittance using the EDTA blank. After 5 to 10 min contact time, read transmittance or absorbance and determine aluminum concentration from the calibration curve previously prepared.

c. *Visual comparison:* If photometric equipment is not available, prepare and treat standards and a sample, as described above, in 50-mL nessler tubes. Make up to mark with water and compare sample color with the standards after 5 to 10 min contact time. A sample treated with EDTA is not needed when nessler tubes are used. If the sample contains turbidity or color, the use of nessler tubes may result in considerable error.

d. *Removal of phosphate interference:* Add 1.7 mL 6N H₂SO₄ to 100 mL sample in a 200-mL erlenmeyer flask. Heat on a hot plate for at least 90 min, keeping solution temperature just below the boiling point. At the end of the heating period solution volume should be about 25 mL. Add water if necessary to keep it at or above that volume.

After cooling, neutralize to a pH of 4.3 to 4.5 with NaOH, using 1N NaOH at the start and 0.1N for the final fine adjustment. Monitor with a pH meter. Make up to 100 mL with water, mix, and use a 25-mL portion for the aluminum test.

Run a blank in the same manner, using 100 mL distilled water and 1.7 mL 6N H₂SO₄. Subtract blank reading from sample reading or use it to set instrument to zero absorbance before reading the sample.

e. *Correction for samples containing fluoride:* Measure sample fluoride concentration by the SPADNS or electrode method. Either:

- 1) Add the same amount of fluoride as in the sample to each aluminum standard, or
- 2) Determine fluoride correction from the set of curves in Figure 3500-Al:1.

5. Calculation

$$\text{mg Al/L} = \frac{\mu\text{g Al (in 50 mL final volume)}}{\text{mL sample}}$$

3500-Sb ANTIMONY

Antimony (Sb) is the fourth element in Group VA in the periodic table; it has an atomic number of 51, an atomic weight of 121.75, and valences of 3 and 5. The average abundance of Sb in the earth's crust is 0.2 ppm; in soils it is 1 ppm; in streams it is 1 µg/L, and in groundwaters it is <0.1 mg/L. Antimony is sometimes found native, but more commonly in stibnite (Sb₂S₃). It is used in alloys of lead and in batteries, bullets, solder, pyrotechnics, and semiconductors.

*Joint Task Group: 20th Edition—See 3500-Al.

6. Precision and Bias

A synthetic sample containing 520 µg Al/L and no interference in distilled water was analyzed by the Eriochrome cyanine R method in 27 laboratories. Relative standard deviation was 34.4% and relative error 1.7%.

A second synthetic sample containing 50 µg Al/L, 500 µg Ba/L, and 5 µg Be/L in distilled water was analyzed in 35 laboratories. Relative standard deviation was 38.5% and relative error 22.0%.

A third synthetic sample containing 500 µg Al/L, 50 µg Cd/L, 110 µg Cr/L, 1000 µg Cu/L, 300 µg Fe/L, 70 µg Pb/L, 50 µg Mn/L, 150 µg Ag/L, and 650 µg Zn/L in distilled water was analyzed in 26 laboratories. Relative standard deviation was 28.8% and relative error 6.2%.

A fourth synthetic sample containing 540 µg Al/L and 2.5 mg polyphosphate/L in distilled water was analyzed in 16 laboratories that hydrolyzed the sample in the prescribed manner. Relative standard deviation was 44.3% and relative error 1.3%. In 12 laboratories that applied no corrective measures, the relative standard deviation was 49.2% and the relative error 8.9%.

A fifth synthetic sample containing 480 µg Al/L and 750 µg F/L in distilled water was analyzed in 16 laboratories that relied on the curve to correct for the fluoride content. Relative standard deviation was 25.5% and relative error 2.3%. The 17 laboratories that added fluoride to the aluminum standards showed a relative standard deviation of 22.5% and a relative error of 7.1%.

7. Bibliography

SHULL, K.E. & G.R. GUTHAN. 1967. Rapid modified Eriochrome cyanine R method for determination of aluminum in water. *J. Amer. Water Works Assoc.* 59:1456.

The common aqueous species are SbO₂⁻, HSbO₂, and complexes with carbonate and sulfate. Soluble salts of antimony are toxic. The U.S. EPA primary drinking water standard MCL is 6 µg/L.

The electrothermal atomic absorption spectrometric method (3113B) or the inductively coupled plasma/mass spectrometric method (3125) are the methods of choice because of their sensitivity. Alternatively use the flame atomic absorption spectrometric method (3111B) or the inductively coupled plasma method (3120) when high sensitivity is not required.

3500-As ARSENIC*

3500-As A. Introduction

1. Occurrence and Significance

Arsenic (As) is the third element in Group VA of the periodic table; it has an atomic number of 33, an atomic weight of 74.92, and valences of 3 and 5. The average abundance of As in the earth's crust is 1.8 ppm; in soils it is 5.5 to 13 ppm; in streams it is less than 2 $\mu\text{g/L}$, and in groundwater it is generally less than 100 $\mu\text{g/L}$. It occurs naturally in sulfide minerals such as pyrite. Arsenic is used in alloys with lead, in storage batteries, and in ammunition. Arsenic compounds are widely used in pesticides and in wood preservatives.

Arsenic is nonessential for plants but is an essential trace element in several animal species. The predominant form between pH 3 and pH 7 is H_2AsO_4^- , between pH 7 and pH 11 it is HAsO_4^{2-} , and under reducing conditions it is $\text{HAsO}_2(\text{aq})$ (or H_3AsO_3). Aqueous arsenic in the form of arsenite, arsenate, and organic arsenicals may result from mineral dissolution, industrial discharges, or the application of pesticides. The chemical form of arsenic depends on its source (inorganic arsenic from minerals, industrial discharges, and pesticides; organic arsenic from industrial discharges, pesticides, and biological action on inorganic arsenic).

Severe poisoning can arise from the ingestion of as little as 100 mg arsenic trioxide; chronic effects may result from the accumulation of arsenic compounds in the body at low intake levels. Carcinogenic properties also have been imputed to arsenic compounds. The toxicity of arsenic depends on its chemical form. Arsenite is many times more toxic than arsenate. For

the protection of aquatic life, the average concentration of As^{3+} in water should not exceed 72 $\mu\text{g/L}$ and the maximum should not exceed 140 $\mu\text{g/L}$. The United Nations Food and Agriculture Organization's recommended maximum level for irrigation waters is 100 $\mu\text{g/L}$. The U.S. EPA primary drinking water standard MCL is 0.05 mg/L.

2. Selection of Method

Methods are available to identify and determine total arsenic, arsenite, and arsenate. Unpolluted fresh water normally does not contain organic arsenic compounds, but may contain inorganic arsenic compounds in the form of arsenate and arsenite. The electrothermal atomic absorption spectrometric method (3113B) is the method of choice in the absence of overwhelming interferences. The hydride generation-atomic absorption method (3114B) is preferred when interferences are present that cannot be overcome by standard electrothermal techniques (e.g., matrix modifiers, background correction). The silver diethyldithiocarbamate method (B), in which arsine is generated by reaction with sodium borohydride in acidic solution, is applicable to the determination of total inorganic arsenic when interferences are absent and when the sample contains no methylarsenic compounds. This method also provides the advantage of being able to identify and quantify arsenate and arsenite separately by generating arsine at different pHs. The inductively coupled plasma (ICP) emission spectroscopy method (3120) is useful at higher concentrations (greater than 50 $\mu\text{g/L}$) while the ICP-mass spectrometric method (3125) is applicable at lower concentrations if chloride does not interfere. When measuring arsenic species, document that speciation does not change over time. No universal preservative for speciation measurements has been identified.

* Approved by Standard Methods Committee, 1997.
Joint Task Group: 20th Edition—See 3500-A1.

3500-As B. Silver Diethyldithiocarbamate Method

1. General Discussion

a. Principle: Arsenite, containing trivalent arsenic, is reduced selectively by aqueous sodium borohydride solution to arsine, AsH_3 , in an aqueous medium of pH 6. Arsenate, methylarsonic acid, and dimethylarsenic acid are not reduced under these conditions. The generated arsine is swept by a stream of oxygen-free nitrogen from the reduction vessel through a scrubber containing glass wool or cotton impregnated with lead acetate solution into an absorber tube containing silver diethyldithiocarbamate and morpholine dissolved in chloroform. The intensity of the red color that develops is measured at 520 nm. To determine total inorganic arsenic in the absence of methylarsenic compounds, a sample portion is reduced at a pH of about 1. Alternatively, arsenate is measured in a sample from which arsenite has been removed by reduction to arsine gas at pH 6 as above.

The sample is then acidified with hydrochloric acid and another portion of sodium borohydride solution is added. The arsine formed from arsenate is collected in fresh absorber solution.

b. Interferences: Although certain metals—chromium, cobalt, copper, mercury, molybdenum, nickel, platinum, silver, and selenium—influence the generation of arsine, their concentrations in water are seldom high enough to interfere, except in the instance of acid rock drainage. H_2S interferes, but the interference is removed with lead acetate. Antimony is reduced to stibine, which forms a colored complex with an absorption maximum at 510 nm and interferes with the arsenic determination. Methylarsenic compounds are reduced at pH 1 to methylarsines, which form colored complexes with the absorber solution. If methylarsenic compounds are present, measurements of total arsenic and arsenate are unreliable. The results for arsenite are not influenced by methylarsenic compounds.




Figure 3500-As:1. Arsenic generator and absorber assembly.

c. Minimum detectable quantity: 1 μg arsenic.

2. Apparatus

a. *Arsenic generator, scrubber, and absorption tube:* See Figure 3500-As:1. Use a 200-mL three-necked flask with a sidearm (19/22 or similar size female ground-glass joint) through which the inert gas delivery tube reaching almost to the bottom of the flask is inserted; a 24/40 female ground-glass joint to carry the scrubber; and a second side arm closed with a rubber septum, or preferably by a screw cap with a hole in its top for insertion of a TFE-faced silicone septum. Place a small magnetic stirring bar in the flask. Fit absorber tube (20 mL capacity) to the scrubber and fill with silver diethyldithiocarbamate solution. Do not use rubber or cork stoppers because they may absorb arsine. Clean glass equipment with concentrated nitric acid.

b. *Fume hood:* Use apparatus in a well-ventilated hood with flask secured on top of a magnetic stirrer.

c. *Photometric equipment:*

- 1) *Spectrophotometer*, for use at 520 nm.
- 2) *Filter photometer*, with green filter having a maximum transmittance in the 500- to 540-nm range.
- 3) *Cells*, for spectrophotometer or filter photometer, 1-cm, clean, dry, and each equipped with a tightly fitting cover (TFE stopper) to prevent chloroform evaporation.

3. Reagents

a. *Reagent water:* See Section 1080A.

b. *Acetate buffer, pH 5.5:* Mix 428 mL 0.2M sodium acetate, $\text{NaC}_2\text{H}_3\text{O}_2$, and 72 mL 0.2M acetic acid, CH_3COOH .

c. *Sodium acetate, 0.2M:* Dissolve 16.46 g anhydrous sodium acetate or 27.36 g sodium acetate trihydrate, $\text{NaC}_2\text{H}_3\text{O}_2 \cdot 3\text{H}_2\text{O}$, in water. Dilute to 1000 mL with water.

d. *Acetic acid, CH_3COOH , 0.2M:* Dissolve 11.5 mL glacial acetic acid in water. Dilute to 1000 mL.

e. *Sodium borohydride solution, 1%:* Dissolve 0.4 g sodium hydroxide, NaOH (4 pellets), in 400 mL water. Add 4.0 g sodium borohydride, NaBH_4 (check for absence of arsenic). Shake to dissolve and to mix. Prepare fresh every few days.

f. *Hydrochloric acid, HCl, 2M:* Dilute 165 mL conc HCl to 1000 mL with water.

g. *Lead acetate solution:* Dissolve 10.0 g $\text{Pb}(\text{CH}_3\text{COO})_2 \cdot 3\text{H}_2\text{O}$ in 100 mL water.

h. *Silver diethyldithiocarbamate solution:* Dissolve 1.0 mL morpholine (CAUTION: Corrosive—avoid contact with skin) in 70 mL chloroform, CHCl_3 . Add 0.30 g silver diethyldithiocarbamate, $\text{AgSCSN}(\text{C}_2\text{H}_5)_2$; shake in a stoppered flask until most is dissolved. Dilute to 100 mL with chloroform. Filter and store in a tightly closed brown bottle in a refrigerator.

i. *Standard arsenite solution:* Dissolve 0.1734 g NaAsO_2 in water and dilute to 1000 mL with water. CAUTION: Toxic—avoid contact with skin and do not ingest. Dilute 10.0 mL to 100 mL with water; dilute 10.0 mL of this intermediate solution to 100 mL with water; 1.00 mL = 1.00 μg As.

j. *Standard arsenate solution:* Dissolve 0.416 g $\text{Na}_2\text{HAsO}_4 \cdot 7\text{H}_2\text{O}$ in water and dilute to 1000 mL. Dilute 10.0 mL to 100 mL with water; dilute 10 mL of this intermediate solution to 100 mL; 1.00 mL = 1.00 μg As.

4. Procedure

a. *Arsenite:*

1) Preparation of scrubber and absorber—Dip glass wool into lead acetate solution; remove excess by squeezing glass wool. Press glass wool between pieces of filter paper, then fluff it. Alternatively, if cotton is used treat it similarly but dry in a desiccator and fluff thoroughly when dry. Place a plug of loose glass wool or cotton in scrubber tube. Add 4.00 mL silver diethyldithiocarbamate solution to absorber tube (5.00 mL may be used to provide enough volume to rinse spectrophotometer cell).

2) Loading of arsenic generator—Pipet not more than 70 mL sample containing not more than 20.0 μg As (arsenite) into the generator flask. Add 10 mL acetate buffer. If necessary, adjust total volume of liquid to 80 mL. Flush flask with nitrogen at the rate of 60 mL/min.

3) Arsenic generation and measurement—While nitrogen is passing through the system, use a 30-mL syringe to inject through the septum 15 mL 1% sodium borohydride solution within 2 min. Stir vigorously with magnetic stirrer. Pass nitrogen through system for an additional 15 min to flush arsenic into absorber solution. Pour absorber solution into a clean and dry spectrophotometric cell and measure absorbance at 520 nm against chloroform. Determine concentration from a calibration curve obtained with arsenite standards. If arsenate also is to be determined for this sample by using the same sample portion, save the liquid in the generator flask.

4) Preparation of standard curves—Treat standard arsenite solution containing 0.0, 1.0, 2.0, 5.0, 10.0, and 20.0 μg As

described in ¶s 1) through 3) above. Plot absorbance versus micrograms arsenic in the standard.

b. Arsenate: After removal of arsenite as arsine, treat sample to convert arsenate to arsine:

If the lead acetate-impregnated glass wool has become ineffective in removing hydrogen sulfide (if it has become gray to black) replace glass wool [see ¶ a1)]. Pass nitrogen through system at the rate of 60 mL/min. Cautiously add 10 mL 2.0N HCl. Generate arsine as directed in ¶ 4a3) and prepare standard curves with standard solutions of arsenate according to procedure of ¶ 4a4).

c. Total inorganic arsenic: Prepare scrubber and absorber as directed in ¶ 4a1) and load arsine generator as directed in ¶ 4a2) using 10 mL 2.0N HCl instead of acetate buffer. Generate arsine and measure as directed in ¶ 4a3). Prepare standard curves according to ¶ 4a4). Curves obtained with standard arsenite solution are almost identical to those obtained with arsenate standard solutions. Therefore, use either arsenite or arsenate standards.

5. Calculation

Calculate arsenite, arsenate, and total inorganic arsenic from readings and calibration curves obtained in 4a, b, and c, respectively, as follows:

$$\text{mg As/L} = \frac{\mu\text{g As (from calibration curve)}}{\text{mL sample in generator flask}}$$

6. Precision and Bias

Interlaboratory comparisons are not available. The relative standard deviation of results obtained with arsenite/arsenate mixtures containing approximately 10 μg arsenic was less than 10%.

7. Bibliography

- PEOPLES, S.A., J. LAKSO & T. LAIS. 1971. The simultaneous determination of methylarsonic acid and inorganic arsenic in urine. *Proc. West. Pharmacol. Soc.* 14:178.
- AGGETT, J. & A.C. ASPELL. 1976. Determination of arsenic (III) and total arsenic by the silver diethyldithiocarbamate method. *Analyst* 101: 912.
- HOWARD, A.G. & M.H. ARBAB-ZAVAR. 1980. Sequential spectrophotometric determination of inorganic arsenic (III) and arsenic (V) species. *Analyst* 105:338.
- PANDE, S.P. 1980. Morpholine as a substitute for pyridine in determination of arsenic in water. *J. Inst. Chem. (India)* 52:256.
- IRGOLIC, K.J. 1986. Arsenic in the environment. In A. V. Xavier, ed. *Frontiers in Bioinorganic Chemistry*. VCH Publishers, Weinheim, Germany.
- IRGOLIC, K.J. 1987. Analytical procedures for the determination of organic compounds of metals and metalloids in environmental samples. *Sci. Total Environ.* 64:61.

3500-Ba BARIUM

Barium (Ba) is the fifth element in Group IIA in the periodic table; it has an atomic number of 56, an atomic weight of 137.33, and a valence of 2. The average abundance of Ba in the earth's crust is 390 ppm; in soils it is 63 to 810 ppm; in streams it is 10 mg/L; in U.S. drinking waters it is 49 $\mu\text{g/L}$; and in groundwaters it is 0.05 to 1 mg/L. It is found chiefly in barite (BaSO_4) or in witherite (BaCO_3). Barium's main use is in mud slurries used in drilling oil and exploration wells, but it is also used in pigments, rat poisons, pyrotechnics, and in medicine.

Joint Task Group: 20th Edition—See 3500-Al.

The solubility of barium in natural waters is controlled by the solubility of BaSO_4 , and somewhat by its adsorption on hydroxides. High concentrations of barium occur in some brines. Concentrations exceeding 1 mg/L constitute a toxicity hazard in the marine environment. The U.S. EPA primary drinking water standard MCL is 1 mg/L.

Perform analyses by the atomic absorption spectrometric methods (3111D or E), the electrothermal atomic absorption method (3113B), or the inductively coupled plasma methods (3120 or 3125).

3500-Be BERYLLIUM

Beryllium (Be) is the first element in Group IIA of the periodic table; it has an atomic number of 4, an atomic weight of 9.01,

and a valence of 2. The average abundance of Be in the earth's crust is 2 ppm; in soils it is 0.8 to 1.3 ppm; in streams it is 0.2 $\mu\text{g/L}$; in U.S. drinking waters and in groundwaters it is typically <0.1 $\mu\text{g/L}$. Beryllium occurs in nature in deposits of beryls in granitic rocks. Beryllium is used in high-strength alloys of cop-

Joint Task Group: 20th Edition—See 3500-Al.

per and nickel, windows in X-ray tubes, and as a moderator in nuclear reactors.

Beryllium solubility is controlled in natural waters by the solubility of beryllium hydroxides. The solubility at pH 6.0 is approximately 0.1 $\mu\text{g/L}$. It is nonessential for plants and animals. Acute toxicity occurs at 130 $\mu\text{g/L}$, and chronic toxicity at 5 $\mu\text{g/L}$ in freshwater species. The United Nations Food and Agriculture Organization recommended maximum level for irrigation waters is 100 $\mu\text{g/L}$. The U.S. EPA primary drinking water standard MCL for beryllium is 4 $\mu\text{g/L}$.

3500-Bi

Bismuth (Bi) is the fifth element in Group VA in the periodic table; it has an atomic number of 83, an atomic weight of 208.98, and valences of 3 and 5. The average abundance of Bi in the earth's crust is 0.08 ppm; in streams it is <0.02 mg/L, and in groundwaters it is <0.1 mg/L. Bismuth occurs in association with lead and silver ores, and occasionally as the native element. ^{210}Bi , ^{212}Bi , and ^{214}Bi are naturally occurring radioisotopes produced in the decay of uranium and thorium. The metal is used in alloys of lead, tin, and cadmium, and in some pharmaceuticals.

Joint Task Group: 20th Edition—See 3500-Al.

3500-B

BORON

See Section 4500-B.

3500-Cd

CADMIUM

Cadmium (Cd) is the second element in Group IIB of the periodic table; it has an atomic number of 48, an atomic weight of 112.41, and a valence of 2. The average abundance of Cd in the earth's crust is 0.16 ppm; in soils it is 0.1 to 0.5 ppm; in streams it is 1 $\mu\text{g/L}$, and in groundwaters it is from 1 to 10 $\mu\text{g/L}$. Cadmium occurs in sulfide minerals that also contain zinc, lead, or copper. The metal is used in electroplating, batteries, paint pigments, and in alloys with various other metals. Cadmium is usually associated with zinc at a ratio of about 1 part cadmium to 500 parts zinc in most rocks and soils.

The solubility of cadmium is controlled in natural waters by carbonate equilibria. Guidelines for maximum cadmium concentrations in natural water are linked to the hardness or alkalinity of the water (i.e., the softer the water, the lower the permitted level of cadmium). It is nonessential for plants and animals.

Joint Task Group: 20th Edition—See 3500-AI.

The atomic absorption spectrometric methods (3111D and E, and 3113B) and the inductively coupled plasma (ICP) methods (3120 and 3125) are the methods of choice. If atomic absorption or ICP instrumentation is not available, the aluminon colorimetric method detailed in the 19th Edition of *Standard Methods* may be used. This method has poorer precision and bias than the methods of choice.

In natural water, Bi^{3+} ion will occur, and complex ions with nitrate and chloride also might be expected. The iodide and telluride compounds are toxic by ingestion or inhalation.

Perform analyses by the atomic absorption spectrometric method (3111B) or by the electrothermal atomic absorption method (3113B). The inductively coupled plasma mass spectrometric method (3125) also may be applied successfully in most cases (with lower detection levels), even though bismuth is not specifically listed as an analyte in the method.

Cadmium is extremely toxic and accumulates in the kidneys and liver, with prolonged intake at low levels sometimes leading to dysfunction of the kidneys. The United Nations Food and Agriculture Organization recommended maximum level for cadmium in irrigation waters is 10 $\mu\text{g/L}$. The U.S. EPA primary drinking water standard MCL is 10 $\mu\text{g/L}$.

The electrothermal atomic absorption spectrometric method (3113B) is preferred. The flame atomic absorption methods (3111B and C) and inductively coupled plasma methods (3120 and 3125) provide acceptable precision and bias, with higher detection limits. Anodic stripping voltammetry (3130B) can achieve superior detection limits, but is susceptible to interferences from copper, silver, gold, and organic compounds. When atomic absorption spectrometric or inductively coupled plasma apparatus is unavailable and the desired precision is not as great, the dithizone method detailed in the 19th Edition of *Standard Methods* is suitable.

3500-Ca CALCIUM*

3500-Ca A. Introduction

1. Occurrence and Significance

Calcium (Ca) is the third element in Group IIA of the periodic table; it has an atomic number of 20, an atomic weight of 40.08, and a valence of 2. The average abundance of Ca in the earth's crust is 4.9%; in soils it is 0.07 to 1.7%; in streams it is about 15 mg/L; and in groundwaters it is from 1 to >500 mg/L. The most common forms of calcium are calcium carbonate (calcite) and calcium-magnesium carbonate (dolomite). Calcium compounds are widely used in pharmaceuticals, photography, lime, de-icing salts, pigments, fertilizers, and plasters. Calcium carbonate solubility is controlled by pH and dissolved CO₂. The CO₂, HCO₃⁻, and CO₃²⁻ equilibrium is the major buffering mechanism in fresh waters. Hardness is based on the concentration of calcium and magnesium salts, and often is used as a measure of potable water quality.

NOTE: Calcium is necessary in plant and animal nutrition and is an essential component of bones, shells, and plant structures. The presence of calcium in water supplies results from passage over deposits of limestone, dolomite, gypsum, and gypsiferous shale. Small concentrations of calcium carbonate combat corro-

sion of metal pipes by laying down a protective coating. Because precipitation of calcite in pipes and in heat-exchangers can cause damage, the amount of calcium in domestic and industrial waters is often controlled by water softening (e.g., ion exchange, reverse osmosis). Calcium carbonate saturation and water hardness are discussed in Sections 2330 and 2340, respectively.

Calcium contributes to the total hardness of water. Chemical softening treatment, reverse osmosis, electrodialysis, or ion exchange is used to reduce calcium and the associated hardness.

2. Selection of Method

The atomic absorption methods (3111B, D, and E) and inductively coupled plasma method (3120) are accurate means of determining calcium. The EDTA titration method gives good results for control and routine applications, but for samples containing high P levels (>50 mg/L) only the atomic absorption or atomic emission methods are recommended because of interferences often encountered with EDTA indicators.

3. Storage of Samples

The customary precautions are sufficient if care is taken to redissolve any calcium carbonate that may precipitate on standing.

* Approved by Standard Methods Committee, 1997.
Joint Task Group: 20th Edition—See 3500-Al.

3500-Ca B. EDTA Titrimetric Method

1. General Discussion

a. Principle: When EDTA (ethylenediaminetetraacetic acid or its salts) is added to water containing both calcium and magnesium, it combines first with the calcium. Calcium can be determined directly, with EDTA, when the pH is made sufficiently high that the magnesium is largely precipitated as the hydroxide and an indicator is used that combines with calcium only. Several indicators give a color change when all of the calcium has been complexed by the EDTA at a pH of 12 to 13.

b. Interference: Under conditions of this test, the following concentrations of ions cause no interference with the calcium hardness determination: Cu²⁺, 2 mg/L; Fe²⁺, 20 mg/L; Fe³⁺, 20 mg/L; Mn²⁺, 10 mg/L; Zn²⁺, 5 mg/L; Pb²⁺, 5 mg/L; Al³⁺, 5 mg/L; and Sn⁴⁺, 5 mg/L. Orthophosphate precipitates calcium at the pH of the test. Strontium and barium give a positive interference and alkalinity in excess of 300 mg/L may cause an indistinct end point in hard waters.

2. Reagents

a. Sodium hydroxide, NaOH, 1N.

b. Indicators: Many indicators are available for the calcium titration. Some are described in the literature (see Bibliography); others are commercial preparations and also may be used. Murexide (ammonium purpurate) was the first indicator available for detecting the calcium end point; directions for its use are presented in this procedure. Individuals who have difficulty recognizing the murexide end point may find the indicator Eriochrome Blue Black R (color index number 202) or Solochrome Dark Blue an improvement because of the color change from red to pure blue. Eriochrome Blue Black R is sodium-1-(2-hydroxy-1-naphthylazo)-2-naphthol-4-sulfonic acid. Other indicators specifically designed for use as end-point detectors in EDTA titration of calcium may be used.

1) *Murexide (ammonium purpurate) indicator:* This indicator changes from pink to purple at the end point. Prepare by dissolving 150 mg dye in 100 g absolute ethylene glycol. Water solutions of the dye are not stable for longer than 1 d. A ground mixture of dye powder and sodium chloride (NaCl) provides a stable form of the indicator. Prepare by mixing 200 mg murexide with 100 g solid NaCl and grinding the mixture to 40 to 50 mesh. Titrate immediately after adding indicator because it is unstable under alkaline conditions. Facilitate end-point recognition by preparing a color comparison blank containing 2.0 mL NaOH solution, 0.2 g solid indicator mixture (or 1 to 2 drops if a

solution is used), and sufficient standard EDTA titrant (0.05 to 0.10 mL) to produce an unchanging color.

2) *Eriochrome Blue Black R indicator*: Prepare a stable form of the indicator by grinding together in a mortar 200 mg powdered dye and 100 g solid NaCl to 40 to 50 mesh. Store in a tightly stoppered bottle. Use 0.2 g of ground mixture for the titration in the same manner as murexide indicator. During titration the color changes from red through purple to bluish purple to a pure blue with no trace of reddish or purple tint. The pH of some (not all) waters must be raised to 14 (rather than 12 to 13) by the use of 8N NaOH to get a good color change.

c. *Standard EDTA titrant, 0.01M*: Prepare standard EDTA titrant and standardize against standard calcium solution as described in Section 2340C to obtain EDTA/CaCO₃ equivalence. Standard EDTA titrant, 0.0100M, is equivalent to 1.000 mg CaCO₃/1.00 mL; use titrated equivalent for *B* in the calculations in 4.

3. Procedure

a. *Pretreatment of water and wastewater samples*: Follow the procedure described in Section 3030E or I if samples require preliminary digestion.

b. *Sample preparation*: Because of the high pH used in this procedure, titrate immediately after adding alkali and indicator. Use 50.0 mL sample, or a smaller portion diluted to 50 mL so that the calcium content is about 5 to 10 mg. Analyze hard waters with alkalinity higher than 300 mg CaCO₃/L by taking a smaller portion and diluting to 50 mL. Alternatively, adjust sample pH into the acid range (pH <6), boil for 1 min to dispel CO₂, and cool before beginning titration.

c. *Titration*: Add 2.0 mL NaOH solution or a volume sufficient to produce a pH of 12 to 13. Stir. Add 0.1 to 0.2 g indicator mixture selected (or 1 to 2 drops if a solution is used). Add EDTA titrant slowly, with continuous stirring to the proper end point. When using murexide, check end point by adding 1 to 2 drops of titrant in excess to make certain that no further color change occurs.

4. Calculation

$$\text{mg Ca/L} = \frac{A \times B \times 400.8}{\text{mL sample}}$$

$$\text{Calcium hardness as mg CaCO}_3/\text{L} = \frac{A \times B \times 1000}{\text{mL sample}}$$

where:

A = mL titrant for sample and

B = mg CaCO₃ equivalent to 1.00 mL EDTA titrant at the calcium indicator end point.

5. Precision and Bias

A synthetic sample containing 108 mg Ca/L, 82 mg Mg/L, 3.1 mg K/L, 19.9 mg Na/L, 241 mg Cl⁻/L, 1.1 mg NO₃⁻-N/L, 0.25 mg NO₂⁻-N/L, 259 mg SO₄²⁻/L, and 42.5 mg total alkalinity/L (contributed by NaHCO₃) in distilled water was analyzed in 44 laboratories by the EDTA titrimetric method, with a relative standard deviation of 9.2% and a relative error of 1.9%.

6. Bibliography

- DIEHL, H. & J.L. ELLINGBOE. 1956. Indicator for titration of calcium in the presence of magnesium using disodium dihydrogen ethylenediamine tetraacetate. *Anal. Chem.* 28:882.
- HILDEBRAND, G.P. & C.N. REILLEY. 1957. New indicator for complexometric titration of calcium in the presence of magnesium. *Anal. Chem.* 29:258.
- PATTON, J. & W. REEDER. 1956. New indicator for titration of calcium with (ethylenedinitrilo) tetraacetate. *Anal. Chem.* 28:1026.
- SCHWARZENBACH, G. 1957. *Complexometric Titrations*. Interscience Publishers, New York, N.Y.
- FURMAN, N.H. 1962. *Standard Methods of Chemical Analysis*, 6th ed. D. Van Nostrand Co., Inc., Princeton, N.J.
- KATZ, H. & R. NAVONE. 1964. Method for simultaneous determination of calcium and magnesium. *J. Amer. Water Works Assoc.* 56:121.

3500-Cs CESIUM

Cesium (Cs) is the sixth element in Group IA of the periodic table; it has an atomic number of 55, an atomic weight of 132.90, and a valence of 1. The average abundance of Cs in the earth's crust is 2.6 ppm; in soils it is 1 to 5 ppm; in streams it is 0.02 mg/L; and in groundwaters it is generally <0.1 mg/L. Cesium is found in lepidolite and in the water of certain mineral springs. ¹³⁷Cs, with a 33-year half-life, is widely dispersed on the earth's

surface as a result of the radioactive fallout from the atmospheric testing of nuclear weapons. Cesium compounds are used in photoelectric cells, as a catalyst, and in brewing. Some cesium compounds are fire hazards.

Perform analyses by the flame atomic absorption spectrometric method (3111B). The inductively coupled plasma/mass spectrometric method (3125) also may be applied successfully in most cases (with lower detection levels), even though cesium is not specifically listed as an analyte in the method.

3500-Cr CHROMIUM*

3500-Cr A. Introduction

1. Occurrence and Significance

Chromium (Cr) is the first element in Group VIB in the periodic table; it has an atomic number of 24, an atomic weight of 51.99, and valences of 0 and 2 through 6. The average abundance of Cr in the earth's crust is 122 ppm; in soils Cr ranges from 11 to 22 ppm; in streams it averages about 1 $\mu\text{g/L}$, and in groundwaters it is generally 100 $\mu\text{g/L}$. Chromium is found chiefly in chrome-iron ore ($\text{FeO}\cdot\text{Cr}_2\text{O}_3$). Chromium is used in alloys, in electroplating, and in pigments. Chromate compounds frequently are added to cooling water for corrosion control.

In natural waters trivalent chromium exists as Cr^{3+} , $\text{Cr}(\text{OH})_2^+$, $\text{Cr}(\text{OH})_2^+$, and $\text{Cr}(\text{OH})_4^-$; in the hexavalent form chromium exists as CrO_4^{2-} and as $\text{Cr}_2\text{O}_7^{2-}$. Cr^{3+} would be expected to form strong complexes with amines, and would be adsorbed by clay minerals.

Chromium is considered nonessential for plants, but an essential trace element for animals. Hexavalent compounds have been shown to be carcinogenic by inhalation and are corrosive to tissue. The chromium guidelines for natural water are linked to the hardness or alkalinity of the water (i.e., the softer the water, the lower the permitted level for chromium). The United Nations Food and Agriculture Organization recommended maximum level for irrigation waters is 100 $\mu\text{g/L}$. The U.S. EPA primary drinking water standard MCL is 100 $\mu\text{g/L}$ for total chromium.

* Approved by Standard Methods Committee, 2001.

Joint Task Group: Rock J. Vitale (chair), C. Ellen Gonter, Timothy S. Oostdyk; 20th Edition—See 3500-A1.

3500-Cr B. Colorimetric Method

1. General Discussion

a. Principle: This procedure measures only hexavalent chromium, Cr(VI). For total chromium determination, acid-digest the sample (see Section 3030) and follow with a suitable instrumental analysis technique. The hexavalent chromium is determined colorimetrically by reaction with diphenylcarbazide in acid solution. A red-violet colored complex of unknown composition is produced. The reaction is very sensitive, the molar absorptivity based on chromium being about 40 000 $\text{L g}^{-1} \text{cm}^{-1}$ at 530 or 540 nm. NOTE: Validation data for Method 3500-Cr.B were developed at 540 nm. Validation data for Method 3500-Cr.C were developed at 530 nm.

b. Interferences: The reaction with diphenylcarbazide is nearly specific for chromium. Hexavalent molybdenum and mercury salts will react to form color with the reagent but the intensities are much lower than that for chromium at the specified pH. Concentrations of Mo or Hg as high as 200 mg/L can be tolerated. Vanadium interferes strongly but concentrations up to

2. Selection of Method

The colorimetric method (B) is useful for the determination of hexavalent chromium in a natural or treated water in the range from 100 to 1000 $\mu\text{g/L}$. This range can be extended by appropriate sample dilution or concentration and/or use of longer cell paths. The ion chromatographic method with photometric detection (C) is suitable for determining dissolved hexavalent chromium in drinking water, groundwater, and industrial wastewater effluents from 0.5 to 5000 $\mu\text{g/L}$. The electrothermal atomic absorption spectrometric method (3113B) is suitable for determining low levels of total chromium (< 50 $\mu\text{g/L}$) in water and wastewater, and the flame atomic absorption spectrometric methods (3111B and C) and the inductively coupled plasma methods (3120 and 3125) are appropriate for measuring total chromium concentrations up to milligram-per-liter levels.

3. Sample Handling

If only the dissolved chromium content is desired, filter sample immediately through a 0.45- μm membrane filter at time of collection, and acidify filtrate with conc nitric acid (HNO_3) to pH < 2. If only dissolved hexavalent chromium is desired, adjust pH of filtrate to 9 with 1N sodium hydroxide solution and refrigerate to $4 \pm 2^\circ\text{C}$. If the total chromium content is desired, acidify unfiltered sample at time of collection with conc HNO_3 to pH < 2. If total hexavalent chromium is desired, adjust pH of unfiltered sample to 9 with 1N sodium hydroxide and refrigerate to $4 \pm 2^\circ\text{C}$.

10 times that of chromium will not result in significant analytical error. Iron in concentrations greater than 1 mg/L may produce a yellow color but the ferric ion (Fe^{3+}) color is not strong and no difficulty is encountered normally if the absorbance is measured photometrically at the appropriate wavelength.

2. Apparatus

a. Colorimetric equipment: One of the following is required:

- 1) *Spectrophotometer*, for use at 530 or 540 nm, with a light path of 1 cm or longer.
- 2) *Filter photometer*, providing a light path of 1 cm or longer and equipped with a greenish yellow filter having maximum transmittance at 530 or 540 nm.

b. Laboratory ware: Soak all reusable items (glass, plastic, etc.), including sample containers, overnight in laboratory-grade detergent, rinse, and soak for 4 h in a mixture of nitric acid (1 part), hydrochloric acid (2 parts), and reagent water (9 parts). Rinse with tap water and reagent water. NOTE: Never use chromic acid cleaning solution.

3. Reagents

Use reagent water (see Section 1080) for reagent preparation and analytical procedure.

a. Stock chromium solution: Dissolve 141.4 mg $K_2Cr_2O_7$ in water and dilute to 100 mL; 1.00 mL = 500 μg Cr. CAUTION: Hexavalent chromium is toxic and a suspected carcinogen. Handle with care.

b. Standard chromium solution: Dilute 1.00 mL stock chromium solution to 100 mL; 1.00 mL = 5.00 μg Cr. Prepare calibration standards at time of analysis.

c. Nitric acid, HNO_3 , conc.

d. Sulfuric acid, H_2SO_4 , conc, 18N, and 6N.

e. Sulfuric acid, H_2SO_4 , 0.2N: Dilute 17 mL 6N H_2SO_4 to 500 mL with water.

f. Phosphoric acid, H_3PO_4 , conc.

g. Diphenylcarbazide solution: Dissolve 250 mg 1,5-diphenylcarbazide (1,5-diphenylcarbohydrazide) in 50 mL acetone. Store in a brown bottle. Prepare weekly. Discard if the solution becomes discolored.

h. Sodium hydroxide, 1N: Dissolve 40 g NaOH in 1 L water. Store in plastic bottle.

4. Procedure

a. Preparation of calibration curve: To compensate for possible slight losses of chromium during analytical operations, treat standards and samples with the same procedure. Accordingly, pipet measured volumes of standard chromium solution (5 $\mu\text{g}/\text{mL}$) ranging from 2.00 to 20.0 mL, to give standards for 10 to 100 μg Cr, into 250-mL beakers or conical flasks. Depending on pretreatment used in ¶ b below, proceed with subsequent treatment of standards as if they were samples.

Develop color as for samples, transfer a suitable portion of each colored solution to a 1-cm absorption cell, and measure absorbance at 540 nm, using reagent water as reference. Correct absorbance readings of standards by subtracting absorbance of a reagent blank carried through the method.

Construct a calibration curve by plotting corrected absorbance values against micrograms chromium in 102 mL final volume.

b. Treatment of sample: If sample has been filtered and/or only hexavalent chromium is desired, start analysis within 24 h of collection and proceed to ¶ 4c. NOTE: Recent evidence suggests that preserved samples can be held for 30 d without substantial changes to Cr (VI) concentrations.^{1,2}

c. Color development and measurement: Add 0.25 mL (5 drops) H_3PO_4 . Use 0.2N H_2SO_4 and a pH meter to adjust solution to pH 2.0 ± 0.5 . Transfer solution to a 100-mL volumetric flask, dilute to 100 mL, and mix. Add 2.0 mL diphenylcarbazide solution, mix, and let stand 5 to 10 min for full color development. Transfer an appropriate portion to a 1-cm absorption cell and measure its absorbance at 530 or 540 nm, using reagent water as reference. Correct absorbance reading of sample by subtracting absorbance of a blank carried through the method (see also note below). From the corrected absorbance, determine micrograms chromium present by reference to the calibration curve.

NOTE: If the solution is turbid after dilution to 100 mL in ¶ c above, take an absorbance reading before adding carbazide reagent and correct absorbance reading of final colored solution by subtracting the absorbance measured previously.

5. Calculation

$$\text{mg Cr/L} = \frac{\mu\text{g Cr (in 102 mL final volume)}}{A}$$

where:

A = mL original sample.

6. Precision and Bias

Collaborative test data from 16 laboratories were obtained on reagent water, tap water, 10% NaCl solution, treated water from synthetic organic industrial waste, EPA extraction leachate, process water, lake water, and effluent from a steel pickle liquor treatment plant.³ The test data yielded the following relationships:

Reagent water:

$$\begin{aligned} S_r &= 0.037x + 0.006 \\ S_o &= 0.022x + 0.004 \end{aligned}$$

Drinking or wastewater:

$$\begin{aligned} S_r &= 0.067x + 0.004 \\ S_o &= 0.037x + 0.002 \end{aligned}$$

Leachate:

$$\begin{aligned} S_r &= 0.032x + 0.007 \\ S_o &= 0.017x + 0.004 \end{aligned}$$

where:

S_r = overall precision,

S_o = single-operator precision, and

x = chromium concentration, mg/L.

7. References

1. U.S. ENVIRONMENTAL PROTECTION AGENCY. 1996. Determination of hexavalent chromium by ion chromatography. Method 1636. EPA 821-R-96-003, U.S. Environmental Protection Agency, Washington, D.C.
2. EATON, A., A. HAGHANI & L. RAMIREZ. 2001. The Erin Brockovich factor. In Proc. Water Quality Technology Conf. (Nashville, Tenn., November 11–15, 2001). American Water Works Assoc., Denver, Colo.
3. AMERICAN SOCIETY FOR TESTING AND MATERIALS. 1996. Chromium, total. D1687-92, Annual Book of ASTM Standards, Vol. 11.01. American Soc. Testing & Materials, Philadelphia, Pa.

8. Bibliography

- ROWLAND, G.P., JR. 1939. Photoelectric colorimetry—Optical study of permanganate ion and of chromium-diphenylcarbazide system. *Anal. Chem.* 11:442.
- URONE, P.F. 1955. Stability of colorimetric reagent for chromium, 5-diphenylcarbazide, in various solvents. *Anal. Chem.* 27:1354.
- ALLEN, T.L. 1958. Microdetermination of chromium with 1,5-diphenylcarbohydrazide. *Anal. Chem.* 30:447.
- ONISHI, H. 1979. Colorimetric Determination of Traces of Metals, 4th ed. Interscience Publishers, New York, N.Y.

3500-Cr C. Ion Chromatographic Method

1. General Discussion

a. Principle: This method is applicable to determination of dissolved hexavalent chromium in drinking water, groundwater, and industrial wastewater effluents. An aqueous sample is filtered and its pH adjusted to 9 to 9.5 with a concentrated buffer. This pH adjustment reduces the solubility of trivalent chromium and preserves the hexavalent chromium oxidation state. The sample is introduced into the instrument's eluent stream of ammonium sulfate and ammonium hydroxide. Trivalent chromium in solution is separated from the hexavalent chromium by the column. After separation, hexavalent chromium reacts with an azide dye to produce a chromogen that is measured at 530 or 540 nm. NOTE: Validation data for 3500-Cr.C were developed at 530 nm. Hexavalent chromium is identified on the basis of retention time.

Although this method was developed using specific commercial equipment, use of another manufacturer's equipment should be acceptable if appropriate adjustments are made.

b. Interferences: Interferences may come from several sources. Use analytical grade salts for the buffer because trace amounts of chromium may be included.

Several soluble species of trivalent chromium in the sample may be oxidized to the hexavalent form in an alkaline medium in the presence of such oxidants as hydrogen peroxide, ozone, and manganese dioxide. The hexavalent form can be reduced to the trivalent in the presence of reducing species in an acid medium.

High ionic concentration may cause column overload. Samples high in chloride and/or sulfate might show this phenomenon, which is characterized by a change in peak geometry.

Interfering organic compounds are removed by the guard column.

c. Minimum detectable concentrations: The method detection levels obtained in a single laboratory with a 250- μ L loop were as follows:

Reagent water	0.4 μ g/L
Drinking water	0.3 μ g/L
Groundwater	0.3 μ g/L
Primary wastewater effluent	0.3 μ g/L
Electroplating waste	0.3 μ g/L

d. Sample preservation and holding time: Filter sample through a 0.45- μ m filter. Use a portion of sample to rinse syringe filter unit and filter, then collect the required volume of filtrate. Adjust pH to 9 to 9.5 by adding buffer solution dropwise while checking pH with a pH meter.

Ship and store sample at 4°C. Bring to room temperature before analysis. Analyze samples within 24 h of collection.

2. Apparatus

a. Ion chromatograph equipped with a pump capable of precisely delivering a flow of 1 to 5 mL/min. The metallic parts of the pump must not contact sample, eluent, or reagent. Sample loops should be available or the instrument should be capable of delivering from 50 to 250- μ L injections of sample. The visible absorption cell should not contain metallic parts that contact the eluent-sample flow. The cell must be usable at 530 nm. Use

plastic pressurized containers to deliver eluent and post-column reagent. Use high-purity helium (99.995%) to pressurize the eluent and post-column reagent vessel.

b. Guard column, to be placed before the separator column, containing an adsorbent capable of adsorbing organic compounds and particulates that would damage or interfere with the analysis or equipment.*

c. Separator column, packed with a high-capacity anion-exchange resin capable of resolving chromate from other sample constituents.†

d. Recorder, integrator, or computer for receiving signals from the detector as a function of time.

e. Laboratory ware: Soak all reusable items (glass, plastic, etc.) including sample containers, overnight in laboratory-grade detergent, rinse, and soak for 4 h in a mixture of nitric acid (1 part), hydrochloric acid (2 parts), and reagent water (9 parts). Rinse with tap water and reagent water. NOTE: Never use chromic acid cleaning solution.

f. Syringe, equipped with male luer-type fitting and a capacity of at least 3 mL.

3. Reagents

a. Reagent water: Deionized or distilled water free from interferences at the minimum detection level of each constituent, filtered through a 0.2- μ m membrane filter and having a conductance of less than 0.1 μ S/cm. Use for preparing all reagents (see Section 1080).

b. Cr(VI) stock solution, 100 mg Cr⁶⁺/L: Prepare from primary standard grade potassium dichromate. Dissolve 0.1414 g K₂Cr₂O₇ in water and dilute to 500 mL in a volumetric flask. pH adjustment is not required. Store in plastic. CAUTION: Hexavalent chromium is toxic and a suspected carcinogen; handle with care.

c. Eluent: Dissolve 33 g ammonium sulfate (NH₄)₂SO₄ in 500 mL water and add 6.5 mL conc ammonium hydroxide, NH₄OH. Dilute to 1 L with water.

d. Post-column reagent: Dissolve 0.5 g 1,5-diphenylcarbazide in 100 mL HPLC-grade methanol. Add with stirring to 500 mL water containing 28 mL conc H₂SO₄. Dilute to 1 L with water. Reagent is stable for 4 or 5 d; prepare only as needed.

e. Buffer solution: Dissolve 33 g ammonium sulfate, (NH₄)₂SO₄, in 75 mL water and add 6.5 mL conc ammonium hydroxide, NH₄OH. Dilute to 100 mL with water.

4. Procedure

a. Instrument setup: Establish ion chromatograph operating conditions as indicated in Table 3500-Cr:I. Set flow rate of the eluent pump at 1.5 mL/min and adjust pressure of reagent delivery module so that the system flow rate, measured after the detector, is 2.0 mL/min. Measure system flow rate using a graduated cylinder and stopwatch. Allow approximately 30 min after adjustment before measuring flow.

* IonPac NG1, Dionex, 4700 Lakeside Drive, Sunnyvale, CA 94086, or equivalent.

† IonPac AS7, Dionex, or equivalent.

TABLE 3500-Cr:I. ION CHROMATOGRAPHIC CONDITIONS

Variable	Value
Guard column	Dionex IonPac NG1
Separator column	Dionex IonPac AS7
Eluent	250 mM (NH ₄) ₂ SO ₄ 100 mM NH ₄ OH
Eluent flow rate	1.5 mL/min
Post-column reagent	2 mM diphenylcarbohydrazide 10% v/v CH ₃ OH 1N H ₂ SO ₄
Post-column reagent flow rate	0.5 mL/min
Detector	Visible @ 530 nm
Retention time	3.8 min

Use an injection loop size based on required sensitivity. A 50- μ L loop is sufficient, although a 250- μ L loop was used to determine the method detection level.

b. Calibration: Before sample analysis, construct a calibration curve using a minimum of a blank and three standards that bracket the expected sample concentration range. Prepare calibration standards from the stock standard (3b) by appropriate dilution with reagent water in volumetric flasks. Adjust to pH 9 to 9.5 with buffer solution (3e) before final dilution. Injection volumes of standards should be about 10 times the injection loop volume to insure complete loop flushing.

c. Sample analysis: Bring chilled, pH-adjusted sample to ambient temperature. Fill a clean syringe with sample, attach a 0.45- μ m syringe filter, and inject 10 times the sample loop volume into the instrument. Dilute any sample that has a concentration greater than the highest calibration standard.

5. Calculation

Determine area or height of the Cr(VI) peak in the calibration standard chromatograms. Establish a calibration curve by performing a regression analysis of peak height or peak area against standard concentration in mg/L. If correlation coefficient is less than 0.995, do not use these data. Check analytical process.

For samples, measure area (or height) of Cr(VI) peak in sample chromatogram, as determined by retention time. Calculate Cr(VI) concentration by interpolating from the calibration line. Correct data for any dilutions made.

Currently available instrumentation automates the entire measurement process (peak measurement, calibration, and sample measurements and calculations). Ensure that enough quality control samples are analyzed to monitor the instrumental processes.

6. Quality Control

a. Initial demonstration of performance: Before sample analysis, set up instrument and analyze enough known samples to determine estimates for the method detection level and linear calibration range. Use the initial demonstration of performance to characterize instrument performance, i.e., method detection levels (MDLs) and linear calibration range.

b. Initial and continuing calibration performance: Initially, after every 10 samples, and after the final sample, analyze an independent check sample and a calibration blank. The concentration of the calibration check sample should be near the mid-cal-

TABLE 3500-Cr:II. SINGLE-LABORATORY PRECISION AND BIAS

Sample Type	Concentration*	Mean Recovery	
	μ g/L	%	RPD [†]
Reagent water	100	100	0.8
	1000	100	0.0
Drinking water	100	105	6.7
	1000	98	1.5
Groundwater	100	98	0.0
	1000	96	0.8
Primary wastewater effluent	100	100	0.7
	1000	104	2.7
Electroplating effluent	100	99	0.4
	1000	101	0.4

* Sample fortified at this concentration level.

[†] RPD - relative percent difference between fortified duplicates.

ibration range; prepare from a source independent of the calibration standards. Use acceptance criteria for check standard recovery and calibration standard concentration based on project goals for precision and accuracy. Typical values for recovery of the check standard range from 90 to 110%. The acceptance criteria for the calibration blank are typically set at \pm the nominal MDL.

c. Reagent blank analysis: Analyze one laboratory reagent blank with each batch of samples. Significant Cr(VI) detected in the reagent blank is a sign of contamination. Identify source and eliminate contamination.

d. Laboratory-fortified matrix (also known as matrix spike) analysis: To a portion of a sample, add a known quantity of Cr(VI). After analysis, calculate percent recovery of the known addition. If the recovery falls outside the control limits (typically 75 to 125%), the matrix may be interfering with the analysis. If matrix interferences are thought to be present, use the method of standard additions, if appropriate, to minimize the effect of interferences. NOTE: There are situations where low recoveries of analyte additions can be overcome, or where the non-detected analyte results can be further validated by the application of three-point method of standard addition. The decision to apply method of standard additions should be a function of achieving desired data quality objectives. Analyze fortified matrix samples as frequently as dictated by project goals and anticipated similarity of matrices in the sample set.

e. Laboratory control sample: Analyze a laboratory control sample (LCS) from an external source with every sample batch. Process LCS and samples identically, including filtering and pH adjustment. Base acceptance criteria for LCS recovery on project goals for precision and bias. Typical values for acceptable recovery range from 90 to 110%.

7. Precision and Bias

The instrument operating conditions and data from a single-laboratory test of the method are shown in Tables 3500-Cr:I and 3500-Cr:II, respectively.

Multilaboratory test data are shown in Table 3500-Cr:III.† Fifteen laboratories analyzed samples ranging from 1.2 to 960 μ g Cr/L.

† The multilaboratory precision and bias data cited in this method were the result of a collaborative study carried out jointly between U.S. EPA Environmental Monitoring Systems Laboratory (Cincinnati) and Committee D-19 of ASTM.

TABLE 3500-Cr:III. MULTILABORATORY DETERMINATION OF BIAS FOR HEXAVALENT CHROMIUM*

Water	Amount Added	Amount Found	Bias		
	µg/L	µg/L	S_r	S_o	%
Reagent	6.00	6.68	1.03	0.53	+11.3
	8.00	8.64	1.10		+8.0
	16.0	17.4	2.25	0.77	+8.8
	20.0	21.4	2.31		+7.0
	100	101	1.91	3.76	+1.0
	140	143	5.52		+2.1
	800	819	24.3	12.7	+2.4
	960	966	18.5		+7.3
Waste	6.0	5.63	1.17	0.55	-6.2
	8.0	7.31	1.91		-8.6
	16.0	15.1	2.70	1.85	-5.6
	20.0	19.8	1.01		-1.0
	100	98.9	4.36	3.31	-1.1
	140	138	8.39		-1.4
	800	796	60.6	27.1	-0.5
	960	944	72.1		-1.7

*Each Youden pair was used to calculate one laboratory data point (S_r).

For reagent water matrix:

$$S_o = 0.033x + 0.106$$

3500-Co COBALT

Cobalt (Co) is the second element in Group VIII in the periodic table; it has an atomic number of 27, an atomic weight of 58.93, and valences of 1, 2, and 3. The average abundance of Co in the earth's crust is 29 ppm; in soils it is 1.0 to 14 ppm; in streams it is 0.2 µg/L; and in groundwaters it is 1 to 10 µg/L. Cobalt occurs only sparingly in ores, usually as the sulfide or the arsenide. It is widely used in alloys of various steels, in electroplating, in fertilizers, and in porcelain and glass.

The solubility of cobalt is controlled by coprecipitation or adsorption by oxides or manganese and iron, by carbonate pre-

$$S_r = 0.050x + 0.559$$

$$\text{Mean recovery} = 1.04x + 0.183$$

where:

S_o = single-operator precision,

S_r = overall precision, and

x = added amount.

For wastewater matrix:

$$S_o = 0.041x + 0.039$$

$$S_r = 0.059x + 1.05$$

$$\text{Mean recovery} = 0.989x - 0.41$$

The eleven water samples consisted of a reagent water blank and five Youden pairs. The nine wastewater samples consisted of a wastewater blank and four Youden pairs.

8. Bibliography

U.S. ENVIRONMENTAL PROTECTION AGENCY. 1991. Methods for the Determination of Metals in Environmental Samples. Method 218.6. EPA-600/4-91-010, Environmental Monitoring Systems Lab., Cincinnati, Ohio.
 DIONEX. 1990. Technical Note No. 26. Dionex, Sunnyvale, Calif.
 EDGEL, K.W., J.A. LONGBOTTOM & R.A. JOYCE. 1994. Determination of dissolved hexavalent chromium in drinking water, ground water, and industrial wastewater effluents by ion chromatography: Collaborative study. *J. Assoc. Offic. Anal. Chem.* 77:994.

Joint Task Group: 20th Edition—See 3500-Al.

3500-Cu COPPER*

3500-Cu A. Introduction

1. Occurrence and Significance

Copper (Cu) is the first element in Group IB in the periodic table; it has an atomic number of 29, an atomic weight of 63.54,

and by the formation of complex ions. Cobalt dust is flammable and is toxic by inhalation. Cobalt is considered essential for algae and some bacteria, nonessential for higher plants, and an essential trace element for animals. The United Nations Food and Agriculture Organization recommended maximum level for irrigation waters is 100 µg/L.

Perform analyses by the flame atomic absorption spectrometric methods (3111B and C), by the electrothermal atomic absorption method (3113B), or by the inductively coupled plasma methods (3120 and 3125).

Large amounts of chromium and its compounds are used in the manufacture of stainless steel and other alloys. Chromium is also used in the manufacture of pigments and dyes. Chromium is a component of many pigments and dyes. Chromium is also used in the manufacture of pigments and dyes.

Cyanide, sulfide, and organic matter interference can be removed by digestion procedure (see Section 3010).

A minimum detection concentration of 0.01 µg/L is possible with this concentration corresponding to 0.01 µg/L.

and valences of 1 and 2. The average abundance of Cu in the earth's crust is 68 ppm; in soils it is 9 to 33 ppm; in streams it is 4 to 12 µg/L; and in groundwater it is <0.1 mg/L. Copper occurs in its native state, but is also found in many minerals, the most important of which are those containing sulfide compounds (e.g., chalcopyrite), but also those with oxides and carbonates. Copper is widely used in electrical wiring, roofing, various

* Approved by Standard Methods Committee, 1999.

Joint Task Group: 20th Edition—See 3500-Al.

alloys, pigments, cooking utensils, piping, and in the chemical industry. Copper salts are used in water supply systems to control biological growths in reservoirs and distribution pipes and to catalyze the oxidation of manganese. Copper forms a number of complexes in natural waters with inorganic and organic ligands. Among the common aqueous species are Cu^{2+} , $\text{Cu}(\text{OH})_2$, and CuHCO_3^+ . Corrosion of copper-containing alloys in pipe fittings may introduce measurable amounts of copper into the water in a pipe system.

Copper is considered an essential trace element for plants and animals. Some compounds are toxic by ingestion or inhalation. The United Nations Food and Agriculture Organization recommended maximum level for irrigation waters is $200 \mu\text{g}/\text{L}$. Under the lead-copper rule, the U.S. EPA drinking water 90th percentile action level is $1.3 \text{ mg}/\text{L}$.

3500-Cu B. Neocuproine Method

1. General Discussion

a. Principle: Cuprous ion (Cu^+) in neutral or slightly acidic solution reacts with 2,9-dimethyl-1,10-phenanthroline (neocuproine) to form a complex in which 2 moles of neocuproine are bound by 1 mole of Cu^+ ion. The complex can be extracted by a number of organic solvents, including a chloroform-methanol ($\text{CHCl}_3\text{-CH}_3\text{OH}$) mixture, to give a yellow solution with a molar absorptivity of about 8000 at 457 nm. The reaction is virtually specific for copper; the color follows Beer's law up to a concentration of $0.2 \text{ mg Cu}/25 \text{ mL solvent}$; full color development is obtained when the pH of the aqueous solution is between 3 and 9; the color is stable in $\text{CHCl}_3\text{-CH}_3\text{OH}$ for several days.

The sample is treated with hydroxylamine-hydrochloride to reduce cupric ions to cuprous ions. Sodium citrate is used to complex metallic ions that might precipitate when the pH is raised. The pH is adjusted to 4 to 6 with NH_4OH , a solution of neocuproine in methanol is added, and the resultant complex is extracted into CHCl_3 . After dilution of the CHCl_3 to an exact volume with CH_3OH , the absorbance of the solution is measured at 457 nm.

b. Interference: Large amounts of chromium and tin may interfere. Avoid interference from chromium by adding sulfurous acid to reduce chromate and complex chromic ion. In the presence of much tin or excessive amounts of other oxidizing ions, use up to 20 mL additional hydroxylamine-hydrochloride solution.

Cyanide, sulfide, and organic matter interfere but can be removed by a digestion procedure (see Section 3030).

c. Minimum detectable concentration: The minimum detectable concentration, corresponding to 0.01 absorbance or 98% transmittance, is $3 \mu\text{g Cu}$ when a 1-cm cell is used and $0.6 \mu\text{g Cu}$ when a 5-cm cell is used.

2. Apparatus

a. Colorimetric equipment: One of the following is required:

1) *Spectrophotometer*, for use at 457 nm, providing a light path of 1 cm or longer.

2. Selection of Method

The atomic absorption spectrometric methods (3111B and C), the inductively coupled plasma methods (3120 and 3125), and the neocuproine method (B) are recommended because of their freedom from interferences. The electrothermal atomic absorption method (3113B) also may be used with success with an appropriate matrix modifier. The bathocuproine method (C) may be used for potable waters.

3. Sampling and Storage

Copper ion tends to be adsorbed on the surface of sample containers. Therefore, analyze samples as soon as possible after collection. If storage is necessary, use $0.5 \text{ mL } 1 + 1 \text{ HCl}/100 \text{ mL sample}$, or acidify to $\text{pH} < 2$ with HNO_3 , to prevent this adsorption.

2) *Filter photometer*, providing a light path of 1 cm or longer and equipped with a narrow-band violet filter having maximum transmittance in the range 450 to 460 nm.

b. Separatory funnels, 125-mL, Squibb form, with glass or TFE stopcock and stopper.

3. Reagents

a. Redistilled water, copper-free: Because most ordinary distilled water contains detectable amounts of copper, use redistilled water, prepared by distilling singly distilled water in a resistant-glass still, or distilled water passed through an ion-exchange unit, to prepare all reagents and dilutions.

b. Stock copper solution: To 200.0 mg polished electrolytic copper wire or foil in a 250-mL conical flask, add 10 mL water and 5 mL conc HNO_3 . After the reaction has slowed, warm gently to complete dissolution of the copper and boil to expel oxides of nitrogen, using precautions to avoid loss of copper. Cool, add about 50 mL water, transfer quantitatively to a 1-L volumetric flask, and dilute to the mark with water; $1 \text{ mL} = 200 \mu\text{g Cu}$.

c. Standard copper solution: Dilute 50.00 mL stock copper solution to 500 mL with water; $1.00 \text{ mL} = 20.0 \mu\text{g Cu}$.

d. Sulfuric acid, H_2SO_4 , conc.

e. Hydroxylamine-hydrochloride solution: Dissolve 50 g $\text{NH}_2\text{OH} \cdot \text{HCl}$ in 450 mL water.

f. Sodium citrate solution: Dissolve 150 g $\text{Na}_3\text{C}_6\text{H}_5\text{O}_7 \cdot 2\text{H}_2\text{O}$ in 400 mL water. Add 5 mL $\text{NH}_2\text{OH} \cdot \text{HCl}$ solution and 10 mL neocuproine reagent. Extract with 50 mL CHCl_3 to remove copper impurities and discard CHCl_3 layer.

g. Ammonium hydroxide, NH_4OH , 5N: Dilute 330 mL conc NH_4OH (28-29%) to 1000 mL with water. Store in a polyethylene bottle.

h. Congo red paper, or other pH test paper showing a color change in the pH range of 4 to 6.

i. Neocuproine reagent: Dissolve 100 mg 2,9-dimethyl-1,10-phenanthroline hemihydrate* in 100 mL methanol. This

* GFS Chemicals, Inc., Columbus, OH, or equivalent.

solution is stable under ordinary storage conditions for a month or more.

j. Chloroform, CHCl₃: Avoid or redistill material that comes in containers with metal-lined caps.

k. Methanol, CH₃OH, reagent grade.

l. Nitric acid, HNO₃, conc.

m. Hydrochloric acid, HCl, conc.

4. Procedure

a. Preparation of calibration curve: Pipet 50 mL water into a 125-mL separatory funnel for use as a reagent blank. Prepare standards by pipetting 1.00 to 10.00 mL (20.0 to 200 µg Cu) standard copper solution into a series of 125-mL separatory funnels, and dilute to 50 mL with water. Add 1 mL conc H₂SO₄ and use the extraction procedure given in ¶ 4b below.

Construct a calibration curve by plotting absorbance versus micrograms of copper.

To prepare a calibration curve for smaller amounts of copper, dilute 10.0 mL standard copper solution to 100 mL. Carry 1.00- to 10.00-mL volumes of this diluted standard through the previously described procedure, but use 5-cm cells to measure absorbance.

b. Treatment of sample: Transfer 100 mL sample to a 250-mL beaker, add 1 mL conc H₂SO₄ and 5 mL conc HNO₃. Add a few boiling chips and cautiously evaporate to dense white SO₃ fumes on a hot plate. If solution remains colored, cool, add another 5 mL conc HNO₃, and again evaporate to dense white fumes. Repeat, if necessary, until solution becomes colorless.

Cool, add about 80 mL water, and bring to a boil. Cool and filter into a 100-mL volumetric flask. Make up to 100 mL with water using mostly beaker and filter washings.

Pipet 50.0 mL or other suitable portion containing 4 to 200 µg Cu, from the solution obtained from preliminary treatment, into a 125-mL separatory funnel. Dilute, if necessary, to 50 mL with water. Add 5 mL NH₂OH · HCl solution and 10 mL sodium citrate solution, and mix thoroughly. Adjust pH to approximately

4 by adding 1-mL increments of NH₄OH until Congo red paper is just definitely red (or other suitable pH test paper indicates a value between 4 and 6).

Add 10 mL neocuproine reagent and 10 mL CHCl₃. Stopper and shake vigorously for 30 s or more to extract the copper-neocuproine complex into the CHCl₃. Let mixture separate into two layers and withdraw lower CHCl₃ layer into a 25-mL volumetric flask, taking care not to transfer any of the aqueous layer. Repeat extraction of the water layer with an additional 10 mL CHCl₃ and combine extracts. Dilute combined extracts to 25 mL with CH₃OH, stopper, and mix thoroughly.

Transfer an appropriate portion of extract to a suitable absorption cell (1 cm for 40 to 200 µg Cu; 5 cm for lesser amounts) and measure absorbance at 457 nm or with a 450- to 460-nm filter. Use a sample blank prepared by carrying 50 mL water through the complete digestion and analytical procedure.

Determine micrograms copper in final solution by reference to the appropriate calibration curve.

5. Calculation

$$\text{mg Cu/L} = \frac{\mu\text{g Cu (in 25 mL final volume)}}{\text{mL portion taken for extraction}}$$

6. Bibliography

- SMITH, G.F. & W.H. MCCURDY. 1952. 2,9-Dimethyl-1,10-phenanthroline: New specific in spectrophotometric determination of copper. *Anal. Chem.* 24:371.
- LUKE, C.L. & M.E. CAMPBELL. 1953. Determination of impurities in germanium and silicon. *Anal. Chem.* 25:1586.
- GAHLER, A.R. 1954. Colorimetric determination of copper with neocuproine. *Anal. Chem.* 26:577.
- FULTON, J.W. & J. HASTINGS. 1956. Photometric determinations of copper in aluminum and lead-tin solder with neocuproine. *Anal. Chem.* 28:174.
- FRANK, A.J., A.B. GOULSTON & A.A. DEACUTIS. 1957. Spectrophotometric determination of copper in titanium. *Anal. Chem.* 29:750.

3500-Cu C. Bathocuproine Method

1. General Discussion

a. Principle: Cuprous ion forms a water-soluble orange-colored chelate with bathocuproine disulfonate (2,9-dimethyl-4,7-diphenyl-1,10-phenanthrolinedisulfonic acid, disodium salt). While the color forms over the pH range 3.5 to 11.0, the recommended pH range is between 4 and 5. The sample is buffered at a pH of about 4.3 and reduced with hydroxylamine hydrochloride. The absorbance is measured at 484 nm. The method can be applied to copper concentrations up to at least 5 mg/L with a sensitivity of 20 µg/L.

b. Interference: The following substances can be tolerated with an error of less than ± 2%: Cyanide, thiocyanate, persulfate, and EDTA also can interfere.

c. Minimum detectable concentration: 20 µg/L with a 5-cm cell.

Substance	Concentration mg/L
Cations	
Aluminum	100
Beryllium	10
Cadmium	100
Calcium	1000
Chromium (III)	10
Cobalt (II)	5
Iron (II)	100
Iron (III)	100
Lithium	500
Magnesium	100
Manganese (II)	500
Nickel (II)	500
Sodium	1000

Substance	Concentration mg/L
Strontium	200
Thorium (IV)	100
Zinc	200
Anions	
Chlorate	1000
Chloride	1000
Fluoride	500
Nitrate	200
Nitrite	200
Orthophosphate	1000
Perchlorate	1000
Sulfate	1000
Compounds	
Residual chlorine	1
Linear alkylate sulfonate (LAS)	40

2. Apparatus

a. Colorimetric equipment: One of the following, with a light path of 1 to 5 cm (unless nessler tubes are used):

- 1) *Spectrophotometer*, for use at 484 nm.
- 2) *Filter photometer*, equipped with a blue-green filter exhibiting maximum light transmission near 484 nm.
- 3) *Nessler tubes*, matched, 100-mL, tall form.

b. Acid-washed glassware: Rinse all glassware with conc HCl and then with copper-free water.

3. Reagents

- a. Copper-free water:* See Method B, ¶ 3a.
- b. Stock copper solution:* Prepare as directed in Method B, ¶ 3b, but use 20.00 mg copper wire or foil; 1.00 mL = 20.00 µg Cu.
- c. Standard copper solution:* Dilute 250 mL stock copper solution to 1000 mL with water; 1.00 mL = 5.00 µg Cu. Prepare daily.
- d. Hydrochloric acid*, HCl, 1 + 1.
- e. Hydroxylamine hydrochloride solution:* See Method B, ¶ 3e.
- f. Sodium citrate solution:* Dissolve 300 g Na₃C₆H₅O₇ · 2H₂O in water and make up to 1000 mL.

g. Disodium bathocuproine disulfonate solution: Dissolve 1.000 g C₁₂H₄N₂(CH₃)₂(C₆H₄)₂(SO₃Na)₂ in water and make up to 1000 mL.

4. Procedure

Pipet 50.0 mL sample, or a suitable portion diluted to 50.0 mL, into a 250-mL erlenmeyer flask. In separate 250-mL erlenmeyer flasks, prepare a 50.0-mL water blank and a series of 50.0-mL copper standards containing 5.0, 10.0, 15.0, 20.0, and 25.0 µg Cu. To sample, blank, and standards add, mixing after each addition, 1.00 mL 1 + 1 HCl, 5.00 mL NH₂OH · HCl solution, 5.00 mL sodium citrate solution, and 5.00 mL disodium bathocuproine disulfonate solution. Transfer to cells and read sample absorbance against the blank at 484 nm. Plot absorbance against micrograms Cu in standards for the calibration curve. Estimate concentration from the calibration curve.

5. Calculation

$$\text{mg Cu/L} = \frac{\mu\text{g Cu (in 66 mL final volume)}}{\text{mL sample}}$$

6. Precision and Bias

A synthetic sample containing 1000 µg Cu/L, 500 µg Al/L, 50 µg Cd/L, 110 µg Cr/L, 300 µg Fe/L, 70 µg Pb/L, 50 µg Mn/L, 150 µg Ag/L, and 650 µg Zn/L was analyzed in 33 laboratories by the bathocuproine method, with a relative standard deviation of 4.1% and a relative error of 0.3%.

7. Bibliography

- SMITH, G.F. & D.H. WILKINS. 1953. New colorimetric reagent specific for copper. *Anal. Chem.* 25:510.
- BORCHARDT, L.G. & J.P. BUTLER. 1957. Determination of trace amounts of copper. *Anal. Chem.* 29:414.
- ZAK, B. 1958. Simple procedure for the single sample determination of serum copper and iron. *Clinica Chim. Acta* 3:328.
- BLAIR, D. & H. DIEHL. 1961. Bathophenanthrolinedisulfonic acid and bathocuproinedisulfonic acid, water soluble reagents for iron and copper. *Talanta* 7:163.

3500-Ga GALLIUM

Gallium (Ga) is the third element in Group IIIA in the periodic table; it has an atomic number of 31, an atomic weight of 67.72, and valences of 1, 2, and 3. The average abundance of Ga in the earth's crust is 19 ppm; in soils it is 1.9 to 29 ppm; in streams it is 0.09 µg/L; and in groundwaters it is <0.1 mg/L. Gallium occurs in many zinc ores, and nearly always in bauxite. Gallium compounds are used in semiconducting devices.

The element exists as Ga³⁺ in natural water, and its solubility is controlled by formation of the hydroxide. It is considered nonessential for plants and animals.

Perform analyses by the electrothermal atomic absorption method (3113B). The inductively coupled plasma/mass spectrometric method (3125) also may be applied successfully in most cases (with lower detection levels), even though gallium is not specifically listed as an analyte in the method.

Germanium (Ge) is the third element in Group IVA in the periodic table; it has an atomic number of 32, an atomic weight of 72.59, and valences of 2 and 4. The average abundance of Ge in the earth's crust is 1.5 ppm; in streams it is 0.03 to 0.1 $\mu\text{g/L}$; and in groundwaters it is <0.1 mg/L. Germanium is found in germanite, in certain zinc ores, and in elevated levels in certain hot spring waters. Germanium alloys are used in transistors, gold alloys, phosphors, and semiconducting devices.

Joint Task Group: 20th Edition—See 3500-A1.

3500-Au GOLD

Gold (Au) is the third element in Group IB in the periodic table; it has an atomic number of 79, an atomic weight of 196.97, and valences of 1 and 3. The average abundance of Au in the earth's crust is 0.004 ppm; in streams it is 2 $\mu\text{g/L}$; and in groundwater it is <0.1 mg/L. Gold occurs in the native form, and is associated with quartz or pyrite. The main uses of gold are in jewelry, dentistry, electronics, and the aerospace industry.

Gold solubility is restricted to acidic waters in the presence of oxidizing agents and chloride, or in alkaline solutions in the

Joint Task Group: 20th Edition—See 3500-A1.

3500-In INDIUM

Indium (In) is the fourth element in Group IIIA in the periodic table; it has an atomic number of 49, an atomic weight of 114.82, and valences of 1, 2, and 3. The average abundance of indium in the earth's crust is 0.19 ppm; in streams it is <0.01 $\mu\text{g/L}$; and in groundwaters it is <0.1 mg/L. Indium often occurs in combination with zinc ores, and sometimes with pyrites and siderite. Indium is used in alloys for bearings, brazing, solder, and in electrical devices.

Joint Task Group: 20th Edition—See 3500-A1.

3500-Ir IRIDIUM

Iridium (Ir) is the eighth element in Group VIII of the periodic table; it has an atomic number of 77, an atomic weight of 192.2, and valences of 1, 3, and 4. The average abundance of Ir in the earth's crust is probably <0.001 ppm, and in groundwaters it is <0.1 mg/L. Iridium occurs uncombined with platinum and other metals. It is used in alloys with platinum in catalysts, thermocouples, electrodes, and wires.

Joint Task Group: 20th Edition—See 3500-A1.

Germanium is present in natural waters in the tetravalent state, and its distribution in natural waters probably is controlled by adsorption on clay mineral surfaces. It is nonessential for plants and animals.

Perform analyses by the electrothermal atomic absorption method (3113B). The inductively coupled plasma/mass spectrometric method (3125) also may be applied successfully in most cases (with lower detection levels), even though germanium is not specifically listed as an analyte in the method.

presence of hydrogen sulfide. Its solubility may be influenced by natural organic acids. Compounds of gold containing thiosulfate and cyanide have some human toxicity.

Perform analyses by the atomic absorption spectrometric method (3111B) or by the electrothermal atomic absorption method (3113B). The inductively coupled plasma mass spectrometric method (3125) also may be applied successfully in most cases (with lower detection levels), even though gold is not specifically listed as an analyte in the method.

Indium exists as In^{3+} and as a number of complex ions. Its solubility is controlled by the formation of the insoluble hydroxide. The metal and its compounds are toxic by inhalation.

Perform analyses by the electrothermal atomic absorption method (3113B). The inductively coupled plasma mass spectrometric method (3125) also may be applied successfully in most cases (with lower detection levels), even though indium is not specifically listed as an analyte in the method.

The aqueous chemistry is controlled by complex compounds, although the solubility in natural waters is relatively unknown.

Perform analyses by the flame atomic absorption spectrometric method (3111B). The inductively coupled plasma/mass spectrometric method (3125) also may be applied successfully in most cases (with lower detection levels), even though iridium is not specifically listed as an analyte in the method.

3500-Fe IRON*

3500-Fe A. Introduction

1. Occurrence and Significance

Iron (Fe) is the first element in Group VIII of the periodic table; it has an atomic number of 26, an atomic weight of 55.85, and common valences of 2 and 3 (and occasionally valences of 1, 4, and 6). The average abundance of Fe in the earth's crust is 6.22%; in soils Fe ranges from 0.5 to 4.3%; in streams it averages about 0.7 mg/L; and in groundwater it is 0.1 to 10 mg/L. Iron occurs in the minerals hematite, magnetite, taconite, and pyrite. It is widely used in steel and in other alloys.

The solubility of ferrous ion (Fe^{2+}) is controlled by the carbonate concentration. Because groundwater is often anoxic, any soluble iron in groundwater is usually in the ferrous state. On exposure to air or addition of oxidants, ferrous iron is oxidized to the ferric state (Fe^{3+}) and may hydrolyze to form red, insoluble hydrated ferric oxide. In the absence of complex-forming ions, ferric iron is not significantly soluble unless the pH is very low.

Elevated iron levels in water can cause stains in plumbing, laundry, and cooking utensils, and can impart objectionable tastes and colors to foods. The United Nations Food and Agriculture Organization recommended level for irrigation waters is 5 mg/L. The U.S. EPA secondary drinking water standard MCL is 0.3 mg/L.

2. Selection of Method

Sensitivity and detection levels for the atomic absorption spectrometric methods (3111B and C), the inductively coupled plasma method (3120), and the phenanthroline colorimetric procedure described here (B) are similar and generally adequate for analysis of natural or treated waters. Lower detection levels can be achieved with electrothermal atomic absorption spectrometry (3113B) when an appropriate matrix modifier is used. The complexing reagents used in the colorimetric procedures are specific for ferrous iron but the atomic absorption procedures are not. However, because of the instability of ferrous iron, which is changed easily to the ferric form in solutions in contact with air, determination of ferrous iron requires special precautions and may need to be done in the field at the time of sample collection.

The procedure for determining ferrous iron using 1,10-phenanthroline (3500-Fe.B.4c) has a somewhat limited applicability; avoid long storage time or exposure of samples to light. A rigorous quantitative distinction between ferrous and ferric iron can be obtained with a special procedure using bathophenanthroline. Spectrophotometric methods using bathophenanthroline¹⁻⁶ and other organic complexing reagents such as ferrozine⁷ or TPTZ⁸ are capable of determining iron concentrations as low as 1 $\mu\text{g/L}$. A chemiluminescence procedure⁹ is stated to have a

detection limit of 5 ng/L. Additional procedures are described elsewhere.¹⁰⁻¹³

3. Sampling and Storage

Plan in advance the methods of collecting, storing, and pre-treating samples. Clean sample container with acid and rinse with reagent water. Equipment for membrane filtration of samples in the field may be required to determine iron in solution (dissolved iron). Dissolved iron, considered to be that passing through a 0.45- μm membrane filter, may include colloidal iron. The value of the determination depends greatly on the care taken to obtain a representative sample. Iron in well or tap water samples may vary in concentration and form with duration and degree of flushing before and during sampling. When taking a sample portion for determining iron in suspension, shake the sample bottle often and vigorously to obtain a uniform suspension of precipitated iron. Use particular care when colloidal iron adheres to the sample bottle. This problem can be acute with plastic bottles.

For a precise determination of total iron, use a separate container for sample collection. Treat with acid at the time of collection to place the iron in solution and prevent adsorption or deposition on the walls of the sample container. Take account of the added acid in measuring portions for analysis. The addition of acid to the sample may eliminate the need for adding acid before digestion (3500-Fe.B.4a).

4. References

- LEE, G.F. & W. STUMM. 1960. Determination of ferrous iron in the presence of ferric iron using bathophenanthroline. *J. Amer. Water Works Assoc.* 52:1567.
- GHOSH, M.M., J.T. O'CONNOR & R.S. ENGELBRECHT. 1967. Bathophenanthroline method for the determination of ferrous iron. *J. Amer. Water Works Assoc.* 59:897.
- BLAIR, D. & H. DIEHL. 1961. Bathophenanthroline-disulfonic acid and bathocuproine-disulfonic acid, water soluble reagents for iron and copper. *Talanta* 7:163.
- SHAPIRO, J. 1966. On the measurement of ferrous iron in natural waters. *Limnol. Oceanogr.* 11:293.
- MCMAHON, J.W. 1967. The influence of light and acid on the measurement of ferrous iron in lake water. *Limnol. Oceanogr.* 12:437.
- MCMAHON, J.W. 1969. An acid-free bathophenanthroline method for measuring dissolved ferrous iron in lake water. *Water Res.* 3:743.
- GIBBS, C. 1976. Characterization and application of ferrozine iron reagent as a ferrous iron indicator. *Anal. Chem.* 48:1197.
- DOUGAN, W.K. & A.L. WILSON. 1973. Absorbimetric determination of iron with TPTZ. *Water Treat. Exam.* 22:110.
- SEITZ, W.R. & D.M. HERCULES. 1972. Determination of trace amounts of iron (II) using chemiluminescence analysis. *Anal. Chem.* 44:2143.

* Approved by Standard Methods Committee, 1997.
Joint Task Group: 20th Edition—See 3500-A1.

10. MOSS, M.L. & M.G. MELLON. 1942. Colorimetric determination of iron with 2,2'-bipyridine and with 2,2',2''-tripyridine. *Ind. Eng. Chem., Anal. Ed.* 14:862.
11. WELCHER, F.J. 1947. *Organic Analytical Reagents*. D. Van Nostrand Co., Princeton, N.J., Vol. 3, pp. 100-104.

12. MORRIS, R.L. 1952. Determination of iron in water in the presence of heavy metals. *Anal. Chem.* 24:1376.
13. DOIG, M.T., III & D.F. MARTIN. 1971. Effect of humic acids on iron analyses in natural water. *Water Res.* 5:689.

3500-Fe B. Phenanthroline Method

1. General Discussion

a. Principle: Iron is brought into solution, reduced to the ferrous state by boiling with acid and hydroxylamine, and treated with 1,10-phenanthroline at pH 3.2 to 3.3. Three molecules of phenanthroline chelate each atom of ferrous iron to form an orange-red complex. The colored solution obeys Beer's law; its intensity is independent of pH from 3 to 9. A pH between 2.9 and 3.5 insures rapid color development in the presence of an excess of phenanthroline. Color standards are stable for at least 6 months.

b. Interference: Among the interfering substances are strong oxidizing agents, cyanide, nitrite, and phosphates (polyphosphates more so than orthophosphate), chromium, zinc in concentrations exceeding 10 times that of iron, cobalt and copper in excess of 5 mg/L, and nickel in excess of 2 mg/L. Bismuth, cadmium, mercury, molybdate, and silver precipitate phenanthroline. The initial boiling with acid converts polyphosphates to orthophosphate and removes cyanide and nitrite that otherwise would interfere. Adding excess hydroxylamine eliminates errors caused by excessive concentrations of strong oxidizing reagents. In the presence of interfering metal ions, use a larger excess of phenanthroline to replace that complexed by the interfering metals. Where excessive concentrations of interfering metal ions are present, the extraction method may be used.

If noticeable amounts of color or organic matter are present, it may be necessary to evaporate the sample, gently ash the residue, and redissolve in acid. The ashing may be carried out in silica, porcelain, or platinum crucibles that have been boiled for several hours in 6*N* HCl. The presence of excessive amounts of organic matter may necessitate digestion before use of the extraction procedure.

c. Minimum detectable concentration: Dissolved or total concentrations of iron as low as 10 µg/L can be determined with a spectrophotometer using cells with a 5 cm or longer light path. Carry a blank through the entire procedure to allow for correction.

2. Apparatus

a. Colorimetric equipment: One of the following is required:

- 1) *Spectrophotometer*, for use at 510 nm, providing a light path of 1 cm or longer.
- 2) *Filter photometer*, providing a light path of 1 cm or longer and equipped with a green filter having maximum transmittance near 510 nm.
- 3) *Nessler tubes*, matched, 100-mL, tall form.

b. Acid-washed glassware: Wash all glassware with conc hydrochloric acid (HCl) and rinse with reagent water before use to remove deposits of iron oxide.

c. Separatory funnels: 125-mL, Squibb form, with ground-glass or TFE stopcocks and stoppers.

3. Reagents

Use reagents low in iron. Use reagent water (see 1080 and 3111B.3c) in preparing standards and reagent solutions and in procedure. Store reagents in glass-stoppered bottles. The HCl and ammonium acetate solutions are stable indefinitely if tightly stoppered. The hydroxylamine, phenanthroline, and stock iron solutions are stable for several months. The standard iron solutions are not stable; prepare daily as needed by diluting the stock solution. Visual standards in nessler tubes are stable for several months if sealed and protected from light.

a. Hydrochloric acid, HCl, conc, containing less than 0.5 ppm iron.

b. Hydroxylamine solution: Dissolve 10 g NH₂OH·HCl in 100 mL water.

c. Ammonium acetate buffer solution: Dissolve 250 g NH₄C₂H₃O₂ in 150 mL water. Add 700 mL conc (glacial) acetic acid. Because even a good grade of NH₄C₂H₃O₂ contains a significant amount of iron, prepare new reference standards with each buffer preparation.

d. Sodium acetate solution: Dissolve 200 g NaC₂H₃O₂·3H₂O in 800 mL water.

e. Phenanthroline solution: Dissolve 100 mg 1,10-phenanthroline monohydrate, C₁₂H₈N₂·H₂O, in 100 mL water by stirring and heating to 80°C. Do not boil. Discard the solution if it darkens. Heating is unnecessary if 2 drops conc HCl are added to the water. (NOTE: One milliliter of this reagent is sufficient for no more than 100 µg Fe.)

f. Potassium permanganate, 0.02*M*: Dissolve 0.316 g KMnO₄ in reagent water and dilute to 100 mL.

g. Stock iron solution: Use metal (1) or salt (2) for preparing the stock solution.

1) Use electrolytic iron wire, or "iron wire for standardizing," to prepare the solution. If necessary, clean wire with fine sandpaper to remove any oxide coating and to produce a bright surface. Weigh 200.0 mg wire and place in a 1000-mL volumetric flask. Dissolve in 20 mL 6*N* sulfuric acid (H₂SO₄) and dilute to mark with water; 1.00 mL = 200 µg Fe.

2) If ferrous ammonium sulfate is preferred, slowly add 20 mL conc H₂SO₄ to 50 mL water and dissolve 1.404 g Fe(NH₄)₂(SO₄)₂·6H₂O. Slowly add potassium permanganate (¶ 1 *f*) until a faint pink color persists. Add the last few milliliters of the solution dropwise. Approximately 50 mL of the potassium permanganate will be required. Dilute to 1000 mL with water and mix; 1.00 mL = 200 µg Fe.

h. Standard iron solutions: Prepare daily for use.

1) Pipet 50.00 mL stock solution into a 1000-mL volumetric flask and dilute to mark with water; 1.00 mL = 10.0 μg Fe.

2) Pipet 5.00 mL stock solution into a 1000-mL volumetric flask and dilute to mark with water; 1.00 mL = 1.00 μg Fe.

i. Diisopropyl or isopropyl ether. CAUTION: Ethers may form explosive peroxides; test before using.

4. Procedure

a. Total iron: Mix sample thoroughly and measure 50.0 mL into a 125-mL erlenmeyer flask. If this sample volume contains more than 200 μg iron use a smaller accurately measured portion and dilute to 50.0 mL. Add 2 mL conc HCl and 1 mL $\text{NH}_2\text{OH}\cdot\text{HCl}$ solution. Add a few glass beads and heat to boiling. To insure dissolution of all the iron, continue boiling until volume is reduced to 15 to 20 mL. (If the sample is ashed, take up residue in 2 mL conc HCl and 5 mL water.) Cool to room temperature and transfer to a 50- or 100-mL volumetric flask or nessler tube. Add 10 mL $\text{NH}_4\text{C}_2\text{H}_3\text{O}_2$ buffer solution and 4 mL phenanthroline solution, and dilute to mark with water. Mix thoroughly and allow a minimum of 10 min for maximum color development.

b. Dissolved iron: Immediately after collection filter sample through a 0.45- μm membrane filter into a vacuum flask containing 1 mL conc HCl/100 mL sample. Analyze filtrate for total dissolved iron (§ 4a) and/or dissolved ferrous iron (§ 4c). (This procedure also can be used in the laboratory if it is understood that normal sample exposure to air during shipment may result in precipitation of iron.)

Calculate suspended iron by subtracting dissolved from total iron.

c. Ferrous iron: Determine ferrous iron at sampling site because of the possibility of change in the ferrous-ferric ratio with time in acid solutions. To determine ferrous iron only, acidify a separate sample with 2 mL conc HCl/100 mL sample at time of collection. Fill bottle directly from sampling source and stopper. Immediately withdraw a 50-mL portion of acidified sample and add 20 mL phenanthroline solution and 10 mL $\text{NH}_4\text{C}_2\text{H}_3\text{O}_2$ solution with vigorous stirring. Dilute to 100 mL and measure color intensity within 5 to 10 min. Do not expose to sunlight. (Color development is rapid in the presence of excess phenanthroline. The phenanthroline volume given is suitable for less than 50 μg total iron; if larger amounts are present, use a correspondingly larger volume of phenanthroline or a more concentrated reagent.)

Calculate ferric iron by subtracting ferrous from total iron.

d. Color measurement: Prepare a series of standards by accurately pipetting calculated volumes of standard iron solutions [use solution described in § 3h2) to measure 1- to 10- μg portions] into 125-mL erlenmeyer flasks and diluting to 50 mL by adding measured volumes of water. Add 2 mL conc HCl and 1 mL $\text{NH}_2\text{OH}\cdot\text{HCl}$ solution. Carry out the steps in § 4a beginning with transfer to a 100-mL volumetric flask or nessler tube.

For visual comparison, prepare a set of at least 10 standards, ranging from 1 to 100 μg Fe in the final 100-mL volume. Compare colors in 100-mL tall-form nessler tubes.

For photometric measurement, use Table 3500-Fe:I as a rough guide for selecting proper light path at 510 nm. Read standards

TABLE 3500-Fe:I. SELECTION OF LIGHT PATH LENGTH FOR VARIOUS IRON CONCENTRATIONS

Fe μg		Light Path cm
50-mL Final Volume	100-mL Final Volume	
50-200	100-400	1
25-100	50-200	2
10-40	20-80	5
5-20	10-40	10

against water set at zero absorbance and plot a calibration curve, including a blank (see § 3c and General Introduction).

If samples are colored or turbid, carry a second set of samples through all steps of the procedure without adding phenanthroline. Instead of water, use the prepared blanks to set photometer to zero absorbance and read each sample developed with phenanthroline against the corresponding blank without phenanthroline. Translate observed photometer readings into iron values by means of the calibration curve. This procedure does not compensate for interfering ions.

e. Samples containing organic interferences: Digest samples containing substantial amounts of organic substances according to the directions given in Sections 3030G or H.

1) If a digested sample has been prepared according to the directions given in Section 3030G or H, pipet 10.0 mL or other suitable portion containing 20 to 500 μg Fe into a 125-mL separatory funnel. If the volume taken is less than 10 mL, add water to make up to 10 mL. To the separatory funnel add 15 mL conc HCl for a 10-mL aqueous volume; or, if the portion taken was greater than 10.0 mL, add 1.5 mL conc HCl/mL sample. Mix, cool, and proceed with 4e3) below.

2) To prepare a sample solely for determining iron, measure a suitable volume containing 20 to 500 μg Fe and carry it through the digestion procedure described in either Section 3030G or H. However, use only 5 mL H_2SO_4 or HClO_4 and omit H_2O_2 . When digestion is complete, cool, dilute with 10 mL water, heat almost to boiling to dissolve slowly soluble salts, and, if the sample is still cloudy, filter through a glass-fiber, sintered-glass, or porcelain filter, washing with 2 to 3 mL water. Quantitatively transfer filtrate or clear solution to a 25-mL volumetric flask and make up to 25 mL with water. Empty flask into a 125-mL separatory funnel, rinse flask with 5 mL conc HCl and add to the funnel. Add 25 mL conc HCl measured with the same flask. Mix and cool to room temperature.

3) Extract the iron from the HCl solution in the separatory funnel by shaking for 30 s with 25 mL isopropyl ether (CAUTION). Draw off lower acid layer into a second separatory funnel. Extract acid solution again with 25 mL isopropyl ether, drain acid layer into a suitable clean vessel, and add ether layer to the ether in the first funnel. Pour acid layer back into second separatory funnel and re-extract with 25 mL isopropyl ether. Withdraw and discard acid layer and add ether layer to first funnel. Persistence of a yellow color in the HCl solution after three extractions does not signify incomplete separation of iron because copper, which is not extracted, gives a similar yellow color.

Shake combined ether extracts with 25 mL water to return iron to aqueous phase and transfer lower aqueous layer to a 100-mL volumetric flask. Repeat extraction with a second 25-mL portion of water, adding this to the first aqueous extract. Discard ether layer.

4) Add 1 mL $\text{NH}_2\text{OH}\cdot\text{HCl}$ solution, 10 mL phenanthroline solution, and 10 mL $\text{NaC}_2\text{H}_3\text{O}_2$ solution. Dilute to 100 mL with water, mix thoroughly, and let stand for a minimum of 10 min. Measure absorbance at 510 nm using a 5-cm absorption cell for amounts of iron less than 100 μg or 1-cm cell for quantities from 100 to 500 μg . As reference, use either water or a sample blank prepared by carrying the specified quantities of acids through the entire analytical procedure. If water is used as reference, correct sample absorbance by subtracting absorbance of a sample blank.

Determine micrograms of iron in the sample from the absorbance (corrected, if necessary) by reference to the calibration curve prepared by using a suitable range of iron standards containing the same amounts of phenanthroline, hydroxylamine, and sodium acetate as the sample.

5. Calculation

When the sample has been treated according to 4a, b, c, or 4e2):

$$\text{mg Fe/L} = \frac{\mu\text{g Fe (in 100 mL final volume)}}{\text{mL sample}}$$

When the sample has been treated according to 4e1):

$$\text{mg Fe/L} = \frac{\mu\text{g Fe (in 100 mL final volume)}}{\text{mL sample}} \times \frac{100}{\text{mL portion}}$$

Report details of sample collection, storage, and pretreatment if they are pertinent to interpretation of results.

6. Precision and Bias

Precision and bias depend on the method of sample collection and storage, the method of color measurement, the iron concentration, and the presence of interfering color, turbidity, and

foreign ions. In general, optimum reliability of visual comparison in nessler tubes is not better than 5% and often only 10%, whereas, under optimum conditions, photometric measurement may be reliable to 3% or 3 μg , whichever is greater. The sensitivity limit for visual observation in nessler tubes is approximately 1 μg Fe. Sample variability and instability may affect precision and bias of this determination more than will the errors of analysis. Serious divergences have been found in reports of different laboratories because of variations in methods of collecting and treating samples.

A synthetic sample containing 300 μg Fe/L, 500 μg Al/L, 50 μg Cd/L, 110 μg Cr/L, 470 μg Cu/L, 70 μg Pb/L, 120 μg Mn/L, 150 μg Ag/L, and 650 μg Zn/L in distilled water was analyzed in 44 laboratories by the phenanthroline method, with a relative standard deviation of 25.5% and a relative error of 13.3%.

7. Bibliography

- CHRONHEIM, G. & W. WINK. 1942. Determination of divalent iron (by o-nitrosophenol). *Ind. Eng. Chem., Anal. Ed.* 14:447.
- MEHLIG, R.P. & R.H. HULETT. 1942. Spectrophotometric determination of iron with o-phenanthroline and with nitro-o-phenanthroline. *Ind. Eng. Chem., Anal. Ed.* 14:869.
- CALDWELL, D.H. & R.B. ADAMS. 1946. Colorimetric determination of iron in water with o-phenanthroline. *J. Amer. Water Works Assoc.* 38:727.
- WELCHER, F.J. 1947. Organic Analytical Reagents. D. Van Nostrand Co., Princeton, N.J., Vol. 3, pp. 85-93.
- KOLTHOFF, I.M., T.S. LEE & D.L. LEUSSING. 1948. Equilibrium and kinetic studies on the formation and dissociation of ferroin and ferrin. *Anal. Chem.* 20:985.
- RYAN, J.A. & G.H. BOTHAM. 1949. Iron in aluminum alloys: Colorimetric determination using 1,10-phenanthroline. *Anal. Chem.* 21:1521.
- REITZ, L.K., A.S. O'BRIEN & T.L. DAVIS. 1950. Evaluation of three iron methods using a factorial experiment. *Anal. Chem.* 22:1470.
- SANDELL, E.B. 1959. Chapter 22 in *Colorimetric Determination of Traces of Metals*, 3rd ed. Interscience Publishers, New York, N.Y.
- SKOUGSTAD, M.W., M.J. FISHMAN, L.C. FRIEDMAN, D.E. ERDMANN & S.S. DUNCAN. 1979. Methods for determination of inorganic substances in water and fluvial sediment. Chapter A1 in Book 5, *Techniques of Water Resources Investigations of the United States Geological Survey*. U.S. Geological Surv., Washington, D.C.

3500-Pb LEAD*

3500-Pb A. Introduction

1. Occurrence and Significance

Lead (Pb) is the fifth element in Group IVA in the periodic table; it has an atomic number of 82, an atomic weight of 207.19, and valences of 2 and 4. The average abundance of Pb in the

earth's crust is 13 ppm; in soils it ranges from 2.6 to 25 ppm; in streams it is 3 $\mu\text{g}/\text{L}$, and in groundwaters it is generally <0.1 mg/L. Lead is obtained chiefly from galena (PbS). It is used in batteries, ammunition, solder, piping, pigments, insecticides, and alloys. Lead also was used in gasoline for many years as an anti-knock agent in the form of tetraethyl lead.

The common aqueous species are Pb^{2+} and hydroxide and carbonate complexes. Lead in a water supply may come from industrial, mine, and smelter discharges or from the dissolu-

*Approved by Standard Methods Committee, 1997.
Joint Task Group: 20th Edition—See 3500-AI.

tion of plumbing and plumbing fixtures. Tap waters that are inherently noncorrosive or not suitably treated may contain lead resulting from an attack on lead service pipes, lead interior plumbing, brass fixtures and fittings, or solder pipe joints.

Lead is nonessential for plants and animals. It is toxic by ingestion and is a cumulative poison. The Food and Drug Administration regulates lead content in food and in house paints. Under the lead-copper rule, the U.S. EPA drinking water 90th percentile action level is 15 $\mu\text{g/L}$.

2. Selection of Method

The atomic absorption spectrometric method (3111B) has a relatively high detection limit in the flame mode and requires

an extraction procedure (3111C) for the low concentrations common in potable water. The electrothermal atomic absorption (AA) method (3113B) is more sensitive for low concentrations and does not require extraction. The inductively coupled plasma/mass spectrometric method (3125) is even more sensitive than the electrothermal AA method. The inductively coupled plasma method (3120) has a sensitivity similar to that of the flame atomic absorption method. Anodic stripping voltammetry (3130B) can achieve superior detection levels, but is susceptible to interferences from copper, silver, gold, and organic compounds. The dithizone method (B) is sensitive and specific as a colorimetric procedure.

3500-Pb B. Dithizone Method

1. General Discussion

a. Principle: An acidified sample containing microgram quantities of lead is mixed with ammoniacal citrate-cyanide reducing solution and extracted with dithizone in chloroform (CHCl_3) to form a cherry-red dithizonate. The color of the mixed color solution is measured photometrically.^{1,2} Sample volume taken for analysis may be 2 L when digestion is used.

b. Interference: In a weakly ammoniacal cyanide solution (pH 8.5 to 9.5) dithizone forms colored complexes with bismuth, stannous tin, and monovalent thallium. In strongly ammoniacal citrate-cyanide solution (pH 10 to 11.5) the dithizonates of these ions are unstable and are extracted only partially.³ This method uses a high pH, mixed color, single dithizone extraction. Interference from stannous tin and monovalent thallium is reduced further when these ions are oxidized during preliminary digestion. A modification of the method allows detection and elimination of bismuth interference. Excessive quantities of bismuth, thallium, and tin may be removed.⁴

Dithizone in CHCl_3 absorbs at 510 nm; control its interference by using nearly equal concentrations of excess dithizone in samples, standards, and blank.

The method is without interference for the determination of 0.0 to 30.0 μg Pb in the presence of 20 μg Tl^+ , 100 μg Sn^{2+} , 200 μg In^{3+} , and 1000 μg each of Ba^{2+} , Cd^{2+} , Co^{2+} , Cu^{2+} , Mg^{2+} , Mn^{2+} , Hg^{2+} , Sr^{2+} , Zn^{2+} , Al^{3+} , Sb^{3+} , As^{3+} , Cr^{3+} , Fe^{3+} , Y^{3+} , PO_4^{3-} , and SO_4^{2-} . Gram quantities of alkali metals do not interfere. A modification is provided to avoid interference from excessive quantities of bismuth or tin.

c. Preliminary sample treatment: At time of collection acidify with conc HNO_3 to pH < 2 but avoid excess HNO_3 . Add 5 mL 0.1N iodine solution to avoid losses of volatile organo-lead compounds during handling and digesting of samples. Prepare a blank of lead-free water and carry through the procedure.

d. Digestion of samples: Unless digestion is shown to be unnecessary, digest all samples for dissolved or total lead as described in 3030H or K.

e. Minimum detectable concentration: 1.0 μg Pb/10 mL dithizone solution.

2. Apparatus

a. Spectrophotometer for use at 510 nm, providing a light path of 1 cm or longer.

b. pH meter.

c. Separatory funnels: 250-mL Squibb type. Clean all glassware, including sample bottles, with 1 + 1 HNO_3 . Rinse thoroughly with reagent water.

d. Automatic dispensing burets: Use for all reagents to minimize indeterminate contamination errors.

3. Reagents

Prepare all reagents in lead-free water.

a. Stock lead solution: Dissolve 0.1599 g lead nitrate, $\text{Pb}(\text{NO}_3)_2$ (minimum purity 99.5%), in approximately 200 mL water. Add 10 mL conc HNO_3 and dilute to 1000 mL with water. Alternatively, dissolve 0.1000 g pure Pb metal in 20 mL 1 + 1 HNO_3 and dilute to 1000 mL with water; 1.00 mL = 100 μg Pb.

b. Working lead solution: Dilute 2.0 mL stock solution to 100 mL with water; 1 mL = 2.00 μg Pb.

c. Nitric acid, HNO_3 , 1 + 4: Dilute 200 mL conc HNO_3 to 1 L with water.

d. Ammonium hydroxide, NH_4OH , 1 + 9: Dilute 10 mL conc NH_4OH to 100 mL with water.

e. Citrate-cyanide reducing solution: Dissolve 400 g dibasic ammonium citrate, $(\text{NH}_4)_2\text{HC}_6\text{H}_5\text{O}_7$, 20 g anhydrous sodium sulfite, Na_2SO_3 , 10 g hydroxylamine hydrochloride, $\text{NH}_2\text{OH} \cdot \text{HCl}$, and 40 g potassium cyanide, KCN (CAUTION: *Poison*) in water and dilute to 1 L. Mix this solution with 2 L conc NH_4OH . Do not pipet by mouth. Prepare solution in a fume hood.

f. Stock dithizone solution: The dithizone concentration in the stock dithizone solutions is based on having a 100% pure dithizone reagent. Some commercial grades of dithizone are contam-

inated with the oxidation product diphenylthiocarbodiazone or with metals. Purify dithizone as directed below. For dithizone solutions not stronger than 0.001% (w/v), calculate the exact concentration by dividing the absorbance of the solution in a 1.00-cm cell at 606 nm by 40.6×10^3 , the molar absorptivity.

In a fume hood, dissolve 100 mg dithizone in 50 mL CHCl_3 in a 150-mL beaker and filter through a 7-cm-diam paper.* Receive filtrate in a 500-mL separatory funnel or in a 125-mL erlenmeyer flask under slight vacuum; use a filtering device designed to handle the CHCl_3 vapor. Wash beaker with two 5-mL portions CHCl_3 , and filter. Wash the paper with three 5-mL portions CHCl_3 , adding final portion dropwise to edge of paper. If filtrate is in flask, transfer with CHCl_3 to a 500-mL separatory funnel.

Add 100 mL 1 + 99 NH_4OH to separatory funnel and shake moderately for 1 min; excessive agitation produces slowly breaking emulsions. Let layers separate, swirling funnel gently to submerge CHCl_3 droplets held on surface of aqueous layer. Transfer CHCl_3 layer to 250-mL separatory funnel, retaining the orange-red aqueous layer in the 500-mL funnel. Repeat extraction, receiving CHCl_3 layer in another 250-mL separatory funnel and transferring aqueous layer, using 1 + 99 NH_4OH , to the 500-mL funnel holding the first extract. Repeat extraction, transferring the aqueous layer to 500-mL funnel. Discard CHCl_3 layer.

To combined extracts in the 500-mL separatory funnel add 1 + 1 HCl in 2-mL portions, mixing after each addition, until dithizone precipitates and solution is no longer orange-red. Extract precipitated dithizone with three 25-mL portions CHCl_3 . Dilute combined extracts to 1000 mL with CHCl_3 ; 1.00 mL = 100 μg dithizone.

g. *Dithizone working solution:* Dilute 100 mL stock dithizone solution to 250 mL with CHCl_3 ; 1 mL = 40 μg dithizone.

h. *Special dithizone solution:* Dissolve 250 mg dithizone in 250 mL CHCl_3 . This solution may be prepared without purification because all extracts using it are discarded.

i. *Sodium sulfite solution:* Dissolve 5 g anhydrous Na_2SO_3 in 100 mL water.

j. *Iodine solution:* Dissolve 40 g KI in 25 mL water, add 12.7 g resublimed iodine, and dilute to 1000 mL.

4. Procedure

a. *With sample digestion:* CAUTION: Perform the following procedure (excluding use of spectrophotometer) in a fume hood.

To a digested sample containing not more than 1 mL conc acid add 20 mL 1 + 4 HNO_3 and filter through lead-free filter paper† and filter funnel directly into a 250-mL separatory funnel. Rinse digestion beaker with 50 mL water and add to filter. Add 50 mL ammoniacal citrate-cyanide solution, mix, and cool to room temperature. Add 10 mL dithizone working solution, shake stoppered funnel vigorously for 30 s, and let layers separate. Insert lead-free cotton in stem of separatory funnel and draw off lower layer. Discard 1 to 2 mL CHCl_3 layer, then fill absorption

cell. Measure absorbance of extract at 510 nm, using dithizone working solution, ¶ 3g, to zero spectrophotometer.

b. *Without sample digestion:* To 100 mL acidified sample (pH 2) in a 250-mL separatory funnel add 20 mL 1 + 4 HNO_3 and 50 mL citrate-cyanide reducing solution; mix. Add 10 mL dithizone working solution and proceed as in ¶ 4a.

c. *Calibration curve:* Plot concentration of at least five standards and a blank against absorbance. Determine concentration of lead in extract from curve. All concentrations are μg Pb/10 mL final extract.

d. *Removal of excess interferences:* The dithizonates of bismuth, tin, and thallium differ from lead dithizonate in maximum absorbance. Detect their presence by measuring sample absorbance at 510 nm and at 465 nm. Calculate corrected absorbance of sample at each wavelength by subtracting absorbance of blank at same wavelength. Calculate ratio of corrected absorbance at 510 nm to corrected absorbance at 465 nm. The ratio of corrected absorbances for lead dithizonate is 2.08 and for bismuth dithizonate is 1.07. If the ratio for the sample indicates interference, i.e., is markedly less than 2.08, proceed as follows with a new 100-mL sample: If the sample has not been digested, add 5 mL Na_2SO_3 solution to reduce iodine preservative. Adjust sample to pH 2.5 using a pH meter and 1 + 4 HNO_3 or 1 + 9 NH_4OH as required. Transfer sample to 250-mL separatory funnel, extract with a minimum of three 10-mL portions special dithizone solution, or until the CHCl_3 layer is distinctly green. Extract with 20-mL portions CHCl_3 to remove dithizone (absence of green). Add 20 mL 1 + 4 HNO_3 , 50 mL citrate-cyanide reducing solution, and 10 mL dithizone working solution. Extract as in ¶ 4a and measure absorbance.

5. Calculation

$$\text{mg Pb/L} = \frac{\mu\text{g Pb (in 10 mL, from calibration curve)}}{\text{mL sample}}$$

6. Precision and Bias

Single-operator precision in recovering 0.0104 mg Pb/L from Mississippi River water was 6.8% relative standard deviation and -1.4% relative error. At the level of 0.026 mg Pb/L, recovery was made with 4.8% relative standard deviation and 15% relative error.

7. References

1. SNYDER, L.J. 1947. Improved dithizone method for determination of lead—mixed color method at high pH. *Anal. Chem.* 19:684.
2. SANDELL, E.B. 1959. *Colorimetric Determination of Traces of Metals*, 3rd ed. Interscience, New York, N.Y.
3. WICHMANN, H.J. 1939. Isolation and determination of trace metals—the dithizone system. *Ind. Eng. Chem., Anal. Ed.* 11:66.
4. AMERICAN SOCIETY FOR TESTING AND MATERIALS. 1977. *Annual Book of ASTM Standards*. Part 26, Method D3112-77, American Soc. Testing & Materials, Philadelphia, Pa.

* Whatman No. 42 or equivalent.

† Whatman No. 541 or equivalent.

3500-Li LITHIUM*

3500-Li A. Introduction

1. Occurrence and Significance

Lithium (Li) is the second element in Group IA of the periodic table; it has an atomic number of 3, an atomic weight of 6.94, and a valence of 1. The average abundance of Li in the earth's crust is 18 ppm; in soils it is 14 to 32 ppm; in streams it is 3 µg/L, and in groundwaters it is <0.1 mg/L. The more important minerals containing lithium are lepidolite, spodumene, petalite, and amblygonite. Lithium compounds are used in pharmaceuticals, soaps, batteries, welding flux, ceramics, reducing agents (e.g., lithium aluminum hydride), and cosmetics.

Many lithium salts are only slightly soluble, and the metal's concentration in water is controlled by incorporation in clay

* Approved by Standard Methods Committee, 1997.
Joint Task Group: 20th Edition—See 3500-A1.

3500-Li B. Flame Emission Photometric Method

1. General Discussion

a. Principle: Lithium can be determined in trace amounts by flame photometric methods at a wavelength of 670.8 nm.

b. Interference: A molecular band of strontium hydroxide with an absorption maximum at 671.0 nm interferes in the flame photometric determination of lithium. Ionization of lithium can be significant in both the air-acetylene and nitrous oxide-acetylene flames and can be suppressed by adding potassium. See Section 3500-Na.B.1*b* for additional information on minimizing interferences in flame photometry.

c. Minimum detectable concentration: The minimum lithium concentration detectable is approximately 0.1 µg/L for reagent water analyzed on an atomic absorption spectrophotometer in the emission mode with an air-acetylene flame, or 0.03 µg/L with a nitrous oxide-acetylene flame.

d. Sampling and storage: Preferably collect sample in a polyethylene bottle, although borosilicate glass containers also may be used. At time of collection adjust sample to pH <2 with nitric acid (HNO₃).

2. Apparatus

Flame photometer: A flame photometer or an atomic absorption spectrometer operating in the emission mode using a lean air-acetylene flame is recommended.

Introduction

minerals of soils. Lithium is considered nonessential for plants and animals, but it is essential for some microorganisms. Some lithium salts are toxic by ingestion. The United Nations Food and Agriculture Organization recommended maximum level for lithium in irrigation waters is 2.5 mg/L.

2. Selection of Method

The atomic absorption spectrometric method (3111B) and the inductively coupled plasma method (3120) are preferred. The flame emission photometric method (B) also is available for laboratories not equipped to use preferred methods. The inductively coupled plasma/mass spectrometric method (3125) may be applied successfully in most cases (with lower detection levels), even though lithium is not specifically listed as an analyte in the method.

3. Reagents

Use reagent water (see 3111B.3c) in reagent preparation and analysis.

a. Potassium ionization suppressant: Dissolve 95.35 g KCl dried at 110°C and dilute to 1000 mL with water; 1.00 mL = 50 mg K.

b. Stock lithium solution: Dissolve 152.7 mg anhydrous lithium chloride, LiCl, in water and dilute to 250 mL; 1.00 mL = 100 µg Li. Dry salt overnight in an oven at 105°C. Cool in a desiccator and weigh immediately after removal from desiccator. Alternatively, purchase prepared stock from a reputable supplier.

c. Standard lithium solution: Dilute 10.00 mL stock LiCl solution to 500 mL with water; 1.00 mL = 2.0 µg Li.

4. Procedure

a. Pretreatment of polluted water and wastewater samples: Choose digestion method appropriate to matrix (see Section 3030).

b. Suppressing ionization: If necessary, filter sample through medium-porosity paper, add 1.0 mL potassium ionization suppressant to 50 mL volumetric flask, and dilute with sample for flame photometric determination. Sample solution will be in a 0.1% K matrix.

c. Treatment of standard solutions: Prepare dilutions of the Li standard solution to bracket sample concentration or to establish at least three points on a calibration curve of emission intensity

against Li concentration. Prepare standards by adding appropriate volumes of standard lithium solution to 25 mL water + 1.0 mL potassium ionization suppressant reagent in a 50-mL volumetric flask. Dilute to 50.0 mL and mix. Both samples and standards will be in a 0.1% K matrix to suppress ionization of lithium.

d. *Flame photometric measurement*: Determine lithium concentration by direct intensity measurements at a wavelength of 670.8 nm. The bracketing method (Section 3500-Na.B.4d) can be used with some photometric instruments, while the construction of a calibration curve is necessary with others. Run sample, water, and lithium standard as nearly simultaneously as possible. For best results, average several readings on each solution.

Follow the manufacturer's instructions for instrument operation.

5. Calculation

$$\mu\text{g Li/L} = (\mu\text{g Li/L in portion analyzed}) \times D$$

where:

D = dilution ratio

$$D = \frac{\text{mL sample} + \text{mL water}}{\text{mL sample}}$$

3500-Mg MAGNESIUM*

3500-Mg A. Introduction

1. Occurrence and Significance

Magnesium (Mg) is the second element in Group IIA of the periodic table; it has an atomic number of 12, an atomic weight of 24.30, and a valence of 2. The average abundance of Mg in the earth's crust is 2.1%; in soils it is 0.03 to 0.84%; in streams it is 4 mg/L, and in groundwaters it is >5 mg/L. Magnesium occurs commonly in the minerals magnesite and dolomite. Magnesium is used in alloys, pyrotechnics, flash photography, drying agents, refractories, fertilizers, pharmaceuticals, and foods.

The common aqueous species is Mg^{2+} . The carbonate equilibrium reactions for magnesium are more complicated than for calcium, and conditions for direct precipitation of dolomite in natural waters are not common. Important contributors to the hardness of a water, magnesium salts break down when heated, forming scale in boilers. Chemical soft-

6. Quality Control

Process a QC standard through entire analytical protocol as a way of determining systematic bias. The control limits for precision of duplicate determinations at concentrations (in water) of 4.0 $\mu\text{g/L}$ and 10.0 $\mu\text{g/L}$ were $4.09 \pm 0.056 \mu\text{g/L}$ and $9.96 \pm 0.094 \mu\text{g/L}$, respectively. The single-operator RSD was 1.38% for a lithium solution containing 10 $\mu\text{g/L}$.

7. Bibliography

- FISHMAN, M.J. 1962. Flame photometric determination of lithium in water. *J. Amer. Water Works Assoc.* 54:228.
- PICKETT, E.E. & S.R. KOIRTYOHANN. 1968. The nitrous oxide-acetylene flame in emission analysis-I. General characteristics. *Spectrochem. Acta.* 23B:235.
- KOIRTYOHANN, S.R. & E.E. PICKETT. 1968. The nitrous oxide-acetylene flame in emission analysis-II. Lithium and the alkaline earths. *Spectrochem. Acta.* 23B:673.
- URE, A.M. & R.L. MITCHELL. 1975. Lithium, sodium, potassium, rubidium, and cesium. In J.A. Dean & T.C. Rains, eds. *Flame Emission and Atomic Absorption Spectrometry*. Dekker, New York, N.Y.
- THOMPSON, K.C. & R.J. REYNOLDS. 1978. *Atomic Absorption Fluorescence, and Flame Spectroscopy—A Practical Approach*, 2nd ed. John Wiley & Sons, New York, N.Y.
- WILLARD, H.H., L.L. MERRIT, J.A. DEAN & F.A. SETTLE, JR. 1981. *Instrumental Methods of Analysis*, 6th ed. Wadsworth Publishing Co., Belmont, Calif.

ening, reverse osmosis, or ion exchange reduces magnesium and associated hardness to acceptable levels.

Magnesium is an essential element in chlorophyll and in red blood cells. Some salts of magnesium are toxic by ingestion or inhalation. Concentrations greater than 125 mg/L also can have a cathartic and diuretic effect.

2. Selection of Method

The methods presented are applicable to waters and wastewaters. Direct determinations can be made with the atomic absorption spectrometric method (3111B) and inductively coupled plasma method (3120). The inductively coupled plasma mass spectrometric method (3125) also may be applied successfully in most cases (with lower detection levels), even though magnesium is not specifically listed as an analyte in the method. These methods can be applied to most concentrations encountered, although sample dilution may be required. Choice of method is largely a matter of personal preference and analyst experience. A calculation method (B) also is available.

* Approved by Standard Methods Committee, 1997.
Joint Task Group: 20th Edition—See 3500-Al.

3500-Mg B. Calculation Method

Magnesium may be estimated as the difference between hardness and calcium as CaCO_3 if interfering metals are present in noninterfering concentrations in the calcium titration (Section 3500-Ca.B) and suitable inhibitors are used in the hardness titration (Section 2340C).

$$\text{mg Mg/L} = [\text{total hardness (as mg CaCO}_3\text{/L)}]$$

$$- \text{calcium hardness (as mg CaCO}_3\text{/L)}] \times 0.243$$

3500-Mn MANGANESE*

3500-Mn A. Introduction

1. Occurrence and Significance

Manganese (Mn) is the first element in Group VIIB in the periodic table; it has an atomic number of 25, an atomic weight of 54.94, and common valences of 2, 4, and 7 (and more rarely, valences of 1, 3, 5, and 6). The average abundance of Mn in the earth's crust is 1060 ppm; in soils it is 61 to 1010 ppm; in streams it is 7 $\mu\text{g/L}$, and in groundwaters it is <0.1 mg/L. Manganese is associated with iron minerals, and occurs in nodules in ocean, fresh waters, and soils. The common ores are pyrolusite (MnO_2) and psilomelane. Manganese is used in steel alloys, batteries, and food additives.

The common aqueous species are the reduced Mn^{2+} and the oxidized Mn^{4+} . The aqueous chemistry of manganese is similar to that of iron. Since groundwater is often anoxic, any soluble manganese in groundwater is usually in the reduced state (Mn^{2+}). Upon exposure to air or other oxidants, groundwater containing manganese usually will precipitate black MnO_2 . Elevated manganese levels therefore can cause stains in plumbing/laundry, and cooking utensils. It is considered an essential trace element for plants and animals. The United Nations Food and

Agriculture Organization recommended maximum level for manganese in irrigation waters is 0.2 mg/L. The U.S. EPA secondary drinking water standard MCL is 50 $\mu\text{g/L}$.

2. Selection of Method

The atomic absorption spectrometric methods (3111B and C), the electrothermal atomic absorption method (3113B), and the inductively coupled plasma methods (3120 and 3125) permit direct determination with acceptable sensitivity and are the methods of choice. Of the various colorimetric methods, the persulfate method (B) is preferred because the use of mercuric ion can control interference from a limited chloride ion concentration.

3. Sampling and Storage

Manganese may exist in a soluble form in a neutral water when first collected, but it oxidizes to a higher oxidation state and precipitates or becomes adsorbed on the container walls. Determine manganese very soon after sample collection. When delay is unavoidable, total manganese can be determined if the sample is acidified at the time of collection with HNO_3 to pH <2. See Section 3010B.

* Approved by Standard Methods Committee, 1999.
Joint Task Group: 20th Edition—See 3500-A1.

3500-Mn B. Persulfate Method

1. General Discussion

a. Principle: Persulfate oxidation of soluble manganous compounds to form permanganate is carried out in the presence of silver nitrate. The resulting color is stable for at least 24 h if excess persulfate is present and organic matter is absent.

b. Interference: As much as 0.1 g chloride (Cl^-) in a 50-mL sample can be prevented from interfering by adding 1 g mercuric sulfate (HgSO_4) to form slightly dissociated complexes. Bro-

midate and iodide still will interfere and only trace amounts may be present. The persulfate procedure can be used for potable water with trace to small amounts of organic matter if the period of heating is increased after more persulfate has been added.

For wastewaters containing organic matter, use preliminary digestion with nitric and sulfuric acids (HNO_3 and H_2SO_4) (see Section 3030G). If large amounts of Cl^- also are present, boiling with HNO_3 helps remove it. Interfering traces of Cl^- are eliminated by HgSO_4 in the special reagent.

Colored solutions from other inorganic ions are compensated for in the final colorimetric step.

Samples that have been exposed to air may give low results due to precipitation of manganese dioxide (MnO₂). Add 1 drop 30% hydrogen peroxide (H₂O₂) to the sample, after adding the special reagent, to redissolve precipitated manganese.

c. Minimum detectable concentration: The molar absorptivity of permanganate ion is about 2300 L g⁻¹ cm⁻¹. This corresponds to a minimum detectable concentration (98% transmittance) of 210 µg Mn/L when a 1-cm cell is used or 42 µg Mn/L when a 5-cm cell is used.

2. Apparatus

Colorimetric equipment: One of the following is required:

a. Spectrophotometer, for use at 525 nm, providing a light path of 1 cm or longer.

b. Filter photometer, providing a light path of 1 cm or longer and equipped with a green filter having maximum transmittance near 525 nm.

c. Nessler tubes, matched, 100-mL, tall form.

3. Reagents

a. Special reagent: Dissolve 75 g HgSO₄ in 400 mL conc HNO₃ and 200 mL distilled water. Add 200 mL 85% phosphoric acid (H₃PO₄), and 35 mg silver nitrate (AgNO₃). Dilute the cooled solution to 1 L.

b. Ammonium persulfate, (NH₄)₂S₂O₈, solid.

c. Standard manganese solution: Prepare a 0.1N potassium permanganate (KMnO₄) solution by dissolving 3.2 g KMnO₄ in distilled water and making up to 1 L. Age for several weeks in sunlight or heat for several hours near the boiling point, then filter through a fine fritted-glass filter crucible and standardize against sodium oxalate as follows:

Weigh several 100- to 200-mg samples of Na₂C₂O₄ to 0.1 mg and transfer to 400-mL beakers. To each beaker, add 100 mL distilled water and stir to dissolve. Add 10 mL 1 + 1 H₂SO₄ and heat rapidly to 90 to 95°C. Titrate rapidly with the KMnO₄ solution to be standardized, while stirring, to a slight pink end-point color that persists for at least 1 min. Do not let temperature fall below 85°C. If necessary, warm beaker contents during titration; 100 mg Na₂C₂O₄ will consume about 15 mL permanganate solution. Run a blank on distilled water and H₂SO₄.

$$\text{Normality of KMnO}_4 = \frac{\text{g Na}_2\text{C}_2\text{O}_4}{(A - B) \times 0.06701}$$

where:

A = mL titrant for sample and

B = mL titrant for blank.

Average results of several titrations. Calculate volume of this solution necessary to prepare 1 L of solution so that 1.00 mL = 50.0 µg Mn, as follows:

$$\text{mL KMnO}_4 = \frac{4.55}{\text{normality KMnO}_4}$$

To this volume add 2 to 3 mL conc H₂SO₄ and NaHSO₃ solution dropwise, with stirring, until the permanganate color disappears. Boil to remove excess SO₂, cool, and dilute to 1000 mL with distilled water. Dilute this solution further to measure small amounts of manganese.

d. Standard manganese solution (alternate): Dissolve 1.000 g manganese metal (99.8% min.) in 10 mL redistilled HNO₃. Dilute to 1000 mL with 1% (v/v) HCl; 1 mL = 1.000 mg Mn. Dilute 10 mL to 200 mL with distilled water; 1 mL = 0.05 mg Mn. Prepare dilute solution daily.

e. Hydrogen peroxide, H₂O₂, 30%.

f. Nitric acid, HNO₃, conc.

g. Sulfuric acid, H₂SO₄, conc.

h. Sodium nitrite solution: Dissolve 5.0 g NaNO₂ in 95 mL distilled water.

i. Sodium oxalate, Na₂C₂O₄, primary standard.

j. Sodium bisulfite: Dissolve 10 g NaHSO₃ in 100 mL distilled water.

4. Procedure

a. Treatment of sample: If a digested sample has been prepared according to directions for reducing organic matter and/or excessive chlorides in Section 3030G, pipet a portion containing 0.05 to 2.0 mg Mn into a 250-mL conical flask. Add distilled water, if necessary, to 90 mL and proceed as in ¶ b.

b. To a suitable sample portion add 5 mL special reagent and 1 drop H₂O₂. Concentrate to 90 mL by boiling or dilute to 90 mL. Add 1 g (NH₄)₂S₂O₈, bring to a boil, and boil for 1 min. Do not heat on a water bath. Remove from heat source, let stand 1 min, then cool under the tap. (Boiling too long results in decomposition of excess persulfate and subsequent loss of permanganate color; cooling too slowly has the same effect.) Dilute to 100 mL with distilled water free from reducing substances and mix. Prepare standards containing 0, 5.00, . . . 1500 µg Mn by treating various amounts of standard Mn solution in the same way.

c. Nessler tube comparison: Use standards prepared as in ¶ 4b and containing 5 to 100 µg Mn/100 mL final volume. Compare samples and standards visually.

d. Photometric determination: Use a series of standards from 0 to 1500 µg Mn/100 mL final volume. Make photometric measurements against a distilled water blank. The following table shows light path length appropriate for various amounts of manganese in 100 mL final volume:

Mn Range µg	Light Path cm
5-200	15
20-400	5
50-1000	2
100-1500	1

Prepare a calibration curve of manganese concentration vs. absorbance from the standards and determine Mn in the samples from the curve. If turbidity or interfering color is present, make corrections as in ¶ 4e.

e. Correction for turbidity or interfering color: Avoid filtration because of possible retention of some permanganate on the filter paper. If visual comparison is used, the effect of turbidity

only can be estimated and no correction can be made for interfering colored ions. When photometric measurements are made, use the following "bleaching" method, which also corrects for interfering color: As soon as the photometer reading has been made, add 0.05 mL H₂O₂ solution directly to the sample in the optical cell. Mix and, as soon as the permanganate color has faded completely and no bubbles remain, read again. Deduct absorbance of bleached solution from initial absorbance to obtain absorbance due to Mn.

5. Calculation

a. When all of the original sample is taken for analysis:

$$\text{mg Mn/L} = \frac{\mu\text{g Mn}/100 \text{ mL}}{\text{mL sample}} \times \frac{100}{\text{mL portion}}$$

b. When a portion of the digested sample (100 mL final volume) is taken for analysis:

$$\text{mg Mn/L} = \frac{\mu\text{g Mn (in 100 mL final volume)}}{\text{mL sample}}$$

3500-Hg MERCURY

Mercury (Hg) is the third element in Group IIB in the periodic table; it has an atomic number of 80, an atomic weight of 200.59, and valences of 1 and 2. The average abundance of Hg in the earth's crust is 0.09 ppm; in soils it is 30 to 160 ppb; in streams it is 0.07 $\mu\text{g/L}$, and in groundwaters it is 0.5 to 1 $\mu\text{g/L}$. Mercury occurs free in nature, but the chief source is cinnabar (HgS). Mercury is used in amalgams, mirror coatings, vapor lamps, paints, measuring devices (thermometers, barometers, manometers), pharmaceuticals, pesticides, and fungicides. It is often used in paper mills as a mold retardant for paper.

The common aqueous species are Hg²⁺, Hg(OH)₂⁰, Hg⁰, and stable complexes with organic ligands. Inorganic mercury can be methylated in sediments when sulfides are present to form dimethyl mercury, (CH₃)₂Hg, which is very toxic and can concentrate in the aquatic food chain. Mercury poisoning occurred in Japan in the 1950s as the result of consumption of shellfish that

Joint Task Group: 20th Edition—See 3500-AI.

3500-Mo MOLYBDENUM

Molybdenum (Mo) is the second element in Group VIB in the periodic table; it has an atomic number of 42, and atomic weight of 95.95, and valences of 2, 3, 4, 5, and 6. The average abundance of Mo in the earth's crust is 1.2 ppm; in soils it is 2.5 ppm; in streams it is 1 $\mu\text{g/L}$, and in groundwaters it is <0.1 mg/L. Molybdenum occurs naturally as molybdenite (MoS₂) and wulfenite (PbMoO₄). It is used in alloys, ink pigments, catalysts, and lubricants.

Joint Task Group: 20th Edition—See 3500-AI.

6. Precision and Bias

A synthetic sample containing 120 $\mu\text{g Mn/L}$, 500 $\mu\text{g Al/L}$, 50 $\mu\text{g Cd/L}$, 110 $\mu\text{g Cr/L}$, 470 $\mu\text{g Cu/L}$, 300 $\mu\text{g Fe/L}$, 70 $\mu\text{g Pb/L}$, 150 $\mu\text{g Ag/L}$, and 650 $\mu\text{g Zn/L}$ in distilled water was analyzed in 33 laboratories by the persulfate method, with a relative standard deviation of 26.3% and a relative error of 0%.

A second synthetic sample, similar in all respects except for 50 $\mu\text{g Mn/L}$ and 1000 $\mu\text{g Cu/L}$, was analyzed in 17 laboratories by the persulfate method, with a relative standard deviation of 50.3% and a relative error of 7.2%.

7. Bibliography

- RICHARDS, M.D. 1930. Colorimetric determination of manganese in biological material. *Analyst* 55:554.
- NYDAHL, F. 1949. Determination of manganese by the persulfate method. *Anal. Chem. Acta.* 3:144.
- MILLS, S.M. 1950. Elusive manganese. *Water Sewage Works* 97:92.
- SANDELL, E.B. 1959. Colorimetric Determination of Traces of Metals, 3rd ed. Interscience Publishers, New York, N.Y., Chapter 26.
- DELFINO, J.J. & G.F. LEE. 1969. Colorimetric determination of manganese in lake waters. *Environ. Sci. Technol.* 3:761.

had accumulated mercury. In times past, mercury was used in the haberdashery industry to block hats (the cause of the "mad hatter" syndrome).

Mercury is considered nonessential for plants and animals. The U.S. EPA primary drinking water standard MCL is 2 $\mu\text{g/L}$.

The cold-vapor atomic absorption method (3112B) is the method of choice for all samples. The inductively coupled plasma mass spectrometric method (3125) also may be applied successfully in some cases, even though mercury is not specifically listed as an analyte in the method. The dithizone method detailed in the 19th edition of *Standard Methods* can be used for determining high levels of mercury (>2 $\mu\text{g/L}$) in potable waters.

Because mercury can be lost readily from samples, preserve them by treating with HNO₃ to reduce the pH to <2 (see Section 1060). Glass storage containers are preferred to plastic, because they can extend the holding time to 30 d, rather than only the 14 d allowed in plastic containers.

The common aqueous species are HMoO₄⁻, MoO₄²⁻, and organic complexes. It is considered an essential trace element for plants and animals. The United Nations Food and Agriculture Organization recommended maximum level for irrigation waters is 0.01 mg/L.

Use one of the flame atomic absorption spectrometric methods (3111D or E), the electrothermal atomic absorption spectrometric method (3113B), or one of the inductively coupled plasma methods (3120 or 3125).

mg NO₃⁻/mL, 1.1 mg NO₂⁻/mL, 259 mg SO₄²⁻/mL, 3500-Ni NICKEL

Nickel (Ni) is the third element in Group VIII in the periodic table; it has an atomic number of 28, an atomic weight of 58.69, and a common valence of 2 and less commonly 1, 3, or 4. The average abundance of Ni in the earth's crust is 1.2 ppm; in soils it is 2.5 ppm; in streams it is 1 µg/L, and in groundwaters it is <0.1 mg/L. Nickel is obtained chiefly from pyrrhotite and garnierite. Nickel is used in alloys, magnets, protective coatings, catalysts, and batteries.

The common aqueous species is Ni²⁺. In reducing conditions insoluble sulfides can form, while in aerobic conditions nickel

Joint Task Group: 20th Edition—See 3500-AI.

3500-Os OSMIUM

Osmium (Os) is the seventh element in Group VIII in the periodic table; it has an atomic number of 76, an atomic weight of 190.2, and valences of 3, 4, and 6, and less commonly 1, 2, 5, 7, and 8. The average abundance of Os in the earth's crust is probably <0.005 ppm, and in groundwaters it is <0.01 mg/L.

Joint Task Group: 20th Edition—See 3500-AI.

3500-Pd PALLADIUM

Palladium (Pd) is the sixth element in Group VIII of the periodic table; it has an atomic number of 46, an atomic weight of 106.42, and valences of 2 and 4. Palladium occurs with platinum in nature. It is used in alloys to make electrical relays, catalysts, in the making of "white gold," and in protective coatings.

Joint Task Group: 20th Edition—See 3500-AI.

3500-Pt PLATINUM

Platinum (Pt) is the ninth element in Group VIII of the periodic table; it has an atomic number of 78, an atomic weight of 195.1, and valences of 2 and 4. The average abundance of Pt in the earth's crust is probably <0.01 ppm, and in groundwaters it is <0.1 mg/L. Platinum is usually found in its native state, but also may be found as sperrylite (PtAs₂). Platinum is used as a catalyst and in laboratory ware, jewelry, and surgical wire.

Joint Task Group: 20th Edition—See 3500-AI.

complexes with hydroxide, carbonates, and organic ligands can form. It is suspected to be an essential trace element for some plants and animals. The United Nations Food and Agriculture Organization recommended maximum level for irrigation waters is 200 µg/L. The U.S. EPA primary drinking water standard MCL is 0.1 mg/L.

The atomic absorption spectrometric methods (3111B and C), the inductively coupled plasma methods (3120 and 3125), and the electrothermal atomic absorption spectrometric method (3113B) are the methods of choice for all samples.

Osmium occurs in iridosime and in platinum-bearing river sands. Osmium is used as a hardener with iridium and as a catalyst with platinum.

The aqueous chemistry is controlled by complex compounds, although the solubility in natural waters is relatively unknown.

Analyze by flame atomic absorption methods (3111D and E).

Palladium has no known toxic effects. The United Nations Food and Agriculture Organization recommended maximum level for irrigation waters is 5 mg/L.

Preferably analyze by flame atomic absorption method (3111B). The inductively coupled plasma mass spectrometric method (3125) also may be applied successfully in most cases (with lower detection levels), even though palladium is not specifically listed as an analyte in the method.

The aqueous chemistry in natural waters is relatively unknown, although its solubility is probably controlled by complex compounds. In powder form, platinum can be flammable, and its soluble salts are toxic by inhalation.

Preferably analyze by flame atomic absorption method (3111B). The inductively coupled plasma mass spectrometric method (3125) also may be applied successfully in most cases (with lower detection levels), even though platinum is not specifically listed as an analyte in the method.

3500-K POTASSIUM*

3500-K A. Introduction

1. Occurrence and Significance

Potassium (K) is the fourth element in Group IA of the periodic table; it has an atomic number of 19, an atomic weight of 39.10, and a valence of 1. The average abundance of K in the earth's crust is 1.84%; in soils it has a range of 0.1 to 2.6%; in streams it is 2.3 mg/L, and in groundwaters it has a range of 0.5 to 10 mg/L. Potassium is commonly associated with aluminosilicate minerals such as feldspars. ^{40}K is a naturally occurring radioactive isotope with a half-life of 1.3×10^9 years. Potassium compounds are used in glass, fertilizers, baking powder, soft drinks, explosives, electroplating, and pigments. Potassium is an essential element in both plant and human nutrition, and occurs in groundwaters as a result of mineral dissolution, from decomposing plant material, and from agricultural runoff.

The common aqueous species is K^+ . Unlike sodium, it does not remain in solution, but is assimilated by plants and is incorporated into a number of clay-mineral structures.

* Approved by Standard Methods Committee, 1997.
Joint Task Group: 20th Edition—See 3500-A1.

3500-K B. Flame Photometric Method

1. General Discussion

a. Principle: Trace amounts of potassium can be determined in either a direct-reading or internal-standard type of flame photometer at a wavelength of 766.5 nm. Because much of the information pertaining to sodium applies equally to the potassium determination, carefully study the entire discussion dealing with the flame photometric determination of sodium (Section 3500-Na.B) before making a potassium determination.

b. Interference: Interference in the internal-standard method may occur at sodium-to-potassium ratios of 5:1 or greater. Calcium may interfere if the calcium-to-potassium ratio is 10:1 or more. Magnesium begins to interfere when the magnesium-to-potassium ratio exceeds 100:1.

c. Minimum detectable concentration: Potassium levels of approximately 0.1 mg/L can be determined.

2. Apparatus

See Section 3500-Na.B.2.

3. Reagents

To minimize potassium pickup, store all solutions in plastic bottles. Shake each container thoroughly to dissolve accumulated salts from walls before pouring.

2. Selection of Method

Methods for the determination of potassium include flame atomic absorption (3111B), inductively coupled plasma (3120), flame photometry (B), and selective ion electrode (C). The inductively coupled plasma/mass spectrometric method (3125) usually may be applied successfully (with lower detection levels), even though potassium is not specifically listed as an analyte in the method. The preferred methods are rapid, sensitive, and accurate; selection depends on instrument availability and analyst choice.

3. Storage of Samples

Do not store samples in soft-glass bottles because of the possibility of contamination from leaching of the glass. Use acid-washed polyethylene or borosilicate glass bottles. Adjust sample to $\text{pH} < 2$ with nitric acid. This will dissolve potassium salts and reduce adsorption on vessel walls.

a. Reagent water: See Section 1080. Use this water for preparing all reagents and calibration standards, and as dilution water.

b. Stock potassium solution: Dissolve 1.907 g KCl dried at 110°C and dilute to 1000 mL with water; 1 mL = 1.00 mg K.

c. Intermediate potassium solution: Dilute 10.0 mL stock potassium solution with water to 100 mL; 1.00 mL = 0.100 mg K.

Use this solution to prepare calibration curve in potassium range of 1 to 10 mg/L.

d. Standard potassium solution: Dilute 10.0 mL intermediate potassium solution with water to 100 mL; 1.00 mL = 0.010 mg K. Use this solution to prepare calibration curve in potassium range of 0.1 to 1.0 mg/L.

4. Procedure

Make determination as described in Section 3500-Na.B.4, but measure emission intensity at 766.5 nm.

5. Calculation

See Section 3500-Na.B.5.

6. Precision and Bias

A synthetic sample containing 3.1 mg K^+ /L, 108 mg Ca^{2+} /L, 82 mg Mg^{2+} /L, 19.9 mg Na^+ /L, 241 mg Cl^- /L, 0.25

mg NO₂⁻-N/L, 1.1 mg NO₃⁻-N/L, 259 mg SO₄²⁻/L, and 42.5 mg total alkalinity/L (contributed by NaHCO₃) was analyzed in 33 laboratories by the flame photometric method, with a relative standard deviation of 15.5% and a relative error of 2.3%.

3500-K C. Potassium-Selective Electrode Method

1. General Discussion

a. Principle: Potassium ion is measured potentiometrically by using a potassium ion-selective electrode and a double-junction, sleeve-type reference electrode. The analysis is performed with either a pH meter having an expanded millivolt scale capable of being read to the nearest 0.1 mV or a specific ion meter having a direct concentration scale for potassium.

Before measurement, an ionic strength adjustor reagent is added to both standards and samples to maintain a constant ionic strength. The electrode response is measured in standard solutions with potassium concentrations spanning the range of interest using a calibration line derived either by the instrument meter or manually. The electrode response in sample solutions is measured following the same procedure and potassium concentration determined from the calibration line or instrument direct readout.

b. Interferences: Although most sensitive to potassium, the potassium electrode will respond to other cations at high concentrations; this can result in a positive bias. Table 3500-K:I lists the concentration of common cations causing a 10% error at various concentrations of potassium chloride with a background ionic strength of 0.12*N* sodium chloride. Of the cations listed, ammonium ion is most often present in samples at concentrations high enough to result in a significant bias. It can be converted to gaseous ammonia by adjusting to pH > 10.

An electrode exposed to interfering cations tends to drift and respond sluggishly. To restore normal performance soak electrode for 1 h in distilled water and then for several hours in a standard potassium solution.

TABLE 3500-K:I. CONCENTRATION OF CATIONS INTERFERING AT VARIOUS CONCENTRATIONS OF POTASSIUM

Cation	Concentration Causing 10% Error mg/L		
	K conc = 1 mg/L	K conc = 10 mg/L	K conc = 100 mg/L
Cs ⁺	1.0	10	100
NH ₄ ⁺	2.7	27	270
Tl ⁺	31.4	314	3 140
Ag ⁺	2 765	27 650	276 500
Tris ⁺	3 105	31 050	310 500
Li ⁺	356	3 560	35 600
Na ⁺	1 179	11 790	117 900
H ⁺	3.6*	2.6*	1.6*

* pH.

7. Bibliography

MEHLICH, A. & R.J. MONROE. 1952. Report on potassium analyses by means of flame photometer methods. *J. Assoc. Offic. Agr. Chem.* 35:588. Also see 3500-Na.B.7.

c. Detection limits: Samples containing from 0.1 to 1000 mg K⁺/L may be analyzed. To measure higher concentrations dilute the sample.

2. Apparatus

- Expanded-scale or digital pH meter or ion-selective meter.
- Potassium ion-selective electrode.
- Sleeve-type double-junction reference electrode: Fill outer sleeve with reference electrode filling solution (see ¶ 3b). Fill inner sleeve with inner filling solution provided with the electrode.
- pH electrode.
- Mixer, magnetic, with a TFE-coated stirring bar.

3. Reagents

- Ionic strength adjustor (ISA): Dissolve 29.22 g NaCl in reagent water and dilute to 100 mL.
- Reference electrode outer sleeve filling solution: Dilute 2 mL ISA solution to 100 mL with reagent water.
- Stock potassium solution: See 3500-K.B.3b.
- Sodium hydroxide, NaOH, 6*N*.
- Reagent water: See Section 1080.

4. Procedure

a. Preparation of standards: Prepare a series of standards containing 100.0, 10.0, 1.0, and 0.1 mg K⁺/L by making serial dilutions of the stock potassium solution as in Section 3500-K.B.3c and 3d.

b. Instrument calibration: Fill reference electrode according to the manufacturer's instructions using reference electrode filling solution. Transfer 100 mL 0.1 mg K⁺/L standard into a 150-mL beaker and add 2 mL ISA. Raise pH to about 11. Stir gently with magnetic mixer. Immerse electrodes, wait approximately 2 min for potential stabilization and record meter reading. Thoroughly rinse electrodes and blot dry. Repeat for each standard solution in order of increasing concentration. Prepare calibration curve on semilogarithmic graph paper by plotting observed potential in millivolts (linear scale) against concentration (log scale). Alternatively, calculate calibration line by regression analysis.

c. Analysis of samples: Transfer 100 mL sample into a 150-mL beaker and follow procedure applied to standards in ¶ 4b above. From the measured response, calculate K⁺ concentration from calibration curve.

5. Precision

Reproducibility of potential measured, over the method's range, can be expected to be ± 0.4 mV, corresponding to about $\pm 2.5\%$ in concentration.

6. Quality Assurance

The slope of the calibration line should be -56 mV/10-fold concentration change. If the slope is outside the range of -56 ± 3 mV, the electrode may require maintenance (replace filling solutions). If the proper electrode response cannot be obtained, replace electrode.

Analyze an independent check standard with a mid-range potassium concentration throughout analysis of a series, initially, every

ten samples, and after final sample. If the value has changed by more than 5%, recalibrate electrode. Analyze a reagent blank at the same frequency. Readings must represent a lower concentration than the lowest concentration standard (0.1 mg/L).

7. Bibliography

- PIODA, L., V. STANKOVA & W. SIMON. 1969. Highly selective potassium ion responsive liquid membrane electrode. *Anal. Lett.* 2 (12): 665.
- MIDGLEY, D. & K. TORRANCE. 1978. *Potentiometric Water Analysis*. John Wiley & Sons, New York, N.Y.
- BAILEY, P.L. 1980. *Analysis with Ion-Selective Electrodes*. Heyden & Son Ltd., Philadelphia, Pa.

3500-Re RHENIUM

Rhenium (Re) is the third element in Group VIIB in the periodic table; it has an atomic number of 75, an atomic weight of 186.21, and valences of 1 through 7, with 7 being the most stable. The average abundance of Re in the earth's crust is 7 ppm, and in groundwaters it is <0.1 mg/L. Rhenium is found in columbite, tantalite, and wolframite, as well as in molybdenum

ore concentrates. It is used in tungsten-molybdenum-based alloys, thermocouples, filaments, and flash bulbs. Rhenium in the powder form can be flammable.

For analysis methods, see flame atomic absorption methods (3111D and E). The inductively coupled plasma mass spectrometric method (3125) also may be applied successfully in most cases (with lower detection levels), even though rhenium is not specifically listed as an analyte in the method.

Joint Task Group: 20th Edition—See 3500-A1.

3500-Rh RHODIUM

Rhodium (Rh) is the fifth element in Group VIII in the periodic table; it has an atomic number of 45, an atomic weight of 102.91, and valences of 1 through 6, the most common being 1 and 3. Rhodium is found in its native state in platinum-bearing sands. It is used in platinum alloys for thermocouples, electrical contacts, and jewelry.

The aqueous chemistry in natural waters is relatively unknown. The metal is flammable in the powder form, and its salts are toxic by inhalation.

For analysis see flame atomic absorption method (3111B). The inductively coupled plasma/mass spectrometric method (3125) also may be applied successfully in most cases (with lower detection levels), even though rhodium is not specifically listed as an analyte in the method.

Joint Task Group: 20th Edition—See 3500-A1.

3500-Ru RUTHENIUM

Ruthenium (Ru) is the fourth element in Group VIII in the periodic table; it has an atomic number of 44, an atomic weight of 101.07, and valences of 1 through 7, the most common being 2, 3, and 4. The average abundance of Ru in the earth's crust is probably <0.01 ppm, and in groundwaters it is <0.1 mg/L. It occurs in its native state in platinum-bearing river sands. It is

used in jewelry with platinum, in electrical contacts, and as a catalyst.

The aqueous chemistry in natural waters is relatively unknown. Ruthenium has no known toxic effects.

For analysis see flame atomic absorption method (3111B). The inductively coupled plasma mass spectrometric method (3125) also may be applied successfully in most cases (with lower detection levels), even though ruthenium is not specifically listed as an analyte in the method.

Joint Task Group: 20th Edition—See 3500-A1.

3500-Se A. Introduction

1. Occurrence and Significance

Selenium (Se) is the third element in Group VIA in the periodic table; it has an atomic number of 34, an atomic weight of 78.96, and valences of 2, 4, or 6. The average abundance of Se in the earth's crust is 0.2 ppm; in soils it is 0.27 to 0.74 ppm; in streams it is 0.2 $\mu\text{g/L}$, and in groundwaters it is <0.1 mg/L. Selenium is used in electronics, ceramics, and shampoos.

The inorganic fraction of dissolved selenium consists predominantly of selenium as the selenate ion (SeO_4^{2-}), designated here as Se(VI), and selenium as the selenite ion (SeO_3^{2-}), Se(IV). Other common aqueous species include Se^{2-} , HSe^- , and Se^0 . Selenium is considered a nonessential trace element for most plants, but is an essential trace nutrient for most animals, and selenium deficiency diseases are well known in veterinary medicine. Above trace levels, ingested selenium is toxic to animals and may be toxic to humans. While the selenium concentration of most natural waters is low, the pore water in seleniferous soils in semiarid areas may contain up to hundreds or thousands of micrograms dissolved selenium per liter. Certain plants that grow in such areas accumulate large concentrations of selenium and may poison livestock that graze on them. Water drained from such soil may cause severe environmental pollution and wildlife toxicity. Selenopolysulfide ions (SSe^{2-}) may occur in the presence of hydrogen sulfide in waterlogged, anoxic soils. Selenium derived from microbial degradation of seleniferous organic matter includes selenite, selenate, and the volatile organic compounds dimethylselenide and dimethyldiselenide. Nonvolatile organic selenium compounds may be released to water by microbial processes. Soluble selenium may be leached from coal ash and fly ash at electric power plants that burn seleniferous coal.


The United Nations Food and Agriculture Organization recommended maximum level for selenium in irrigation waters is 20 $\mu\text{g/L}$. The U.S. EPA primary drinking water standard MCL is 50 $\mu\text{g/L}$.

2. Selection of Method

The selenium methods using hydride generation atomic absorption (3114B and C), electrothermal atomic absorption (3113B), and derivatization colorimetry (C) are the most sensitive currently available. For determination of selenium at higher concentrations, the inductively coupled plasma methods (3120 and 3125) may be used.

By using suitable preparatory steps to convert other chemical species to Se(IV), it is possible to distinguish the chemical species in the sample. In drinking water and most surface and ground waters, Se(IV), Se(VI), and particulate selenium frequently are the only significant species. However, when speciation is important, for example, when a new matrix is being analyzed, the general analytical scheme shown in Figure 3500-Se:1 may be carried out as follows: Determine volatile selenium by stripping sample with nitrogen or air and collecting selenium

in alkaline hydrogen peroxide (see Method D). To obtain an estimate of selenium in suspended particles, determine total selenium, filter sample, and make a second determination of total selenium. In any case, filter sample. Occasionally, a filtered sample may have the odor of hydrogen sulfide and a yellow color; such a sample may contain selenopolysulfides, which may be estimated by comparing results of total selenium analyses before and after acidification, stripping with nitrogen, settling for 10 min, and refiltration. Determine selenite, Se(IV), by analyzing filtered water sample directly by Methods 3114B or C, or by 3500-Se.C or E. In principle, sample digestion with HCl will convert Se(VI) to Se(IV), and the value determined will equal the sum of the two species. In practice, samples frequently contain an unknown masking agent that produces an unduly low result. Test for this effect by analyzing samples with known additions of both species. If recovery is good, the HCl digestion followed by analyses will yield reliable results. If recovery is poor and organic selenium is to be determined subsequently, attempt to remove the interference by sample pretreatment with resin (B.1). Interference also can be eliminated by digestion with



* Approved by Standard Methods Committee, 1999.
Joint Task Group: 20th Edition—See 3500-A1.

Figure 3500-Se:1. General scheme for speciation of selenium in water.

an oxidizing agent (B.2, 3, and 4), but these procedures prevent distinguishing of Se(VI) and organic selenium and also oxidize many organic selenium compounds. To measure nonvolatile organic selenium compounds, use Method E.

The choice of digestion method for oxidizing interferences and organic selenium depends on sample matrix. The methods described in B.2, 3, and 4, in order of increasing complexity and digestive ability, use ammonium (or potassium) persulfate, hydrogen peroxide, and potassium permanganate. Ammonium persulfate digestion is adequate for most filtered ground, drinking, and surface water. Hydrogen peroxide digestion may be required if organic selenium compounds are present, and potassium permanganate digestion may be needed with unfiltered samples or those containing refractory organic selenium compounds. Confirm results obtained with one digestion method by using a more rigorous method when characterizing a new matrix.

3. Interferences

Interferences are found in certain reagents, as well as in samples. Recognition of the presence of an interferant is critical, especially when unknown sample matrices are being analyzed. Routinely add Se(IV) and Se(VI) to test for interference. If present, characterize the interference and correct by the method of standard additions. A slope less than one indicates interference. In cases of mild interference (recoveries reduced by 25% or less), the standard additions method will largely correct determined values.

Because the hydride atomic absorption method is extremely sensitive, samples frequently need to be diluted to bring them within the linear range of the instrument. Diluting a filtered water sample frequently will eliminate sample-related interferences.

3500-Se B. Sample Preparation

1. Removal of Organic and Iron Interference by Resin Pretreatment

Interferences are common in selenium analysis, particularly when chemical speciation is attempted. Routine pretreatment of water samples as described here is not necessary. The methods in this section should be tried when poor recovery of standard additions indicates a problem.

Many waters contain iron and/or dissolved organic matter (humic acid) in quantities sufficient to interfere. Reduction of Se(VI) to Se(IV) usually is nonquantitative. When Se(VI) standard additions show poor recovery, treat the sample before analysis. To remove dissolved organic compounds, pass an acidified sample through a resin. Because dissolved organic selenium compounds also may be removed by this treatment, also determine total selenium in the untreated water sample (see 2, 3, or 4 below). To remove iron use a strong base ion exchange resin.* Iron is removed as the

anionic chloro complex. In this treatment the acidity and ion exchanger do not alter speciation; complete speciation of selenium is possible.

4. Bibliography

- U.S. NATIONAL ACADEMY OF SCIENCES. 1976. Selenium: Medical and Biological Effects of Environmental Pollutants. National Academy of Sciences, Washington, D.C.
- CUTTER, G.A. 1978. Species determination of selenium in natural waters. *Anal. Chim. Acta* 98:59.
- ROBBERECHT, H. & R. VAN GRIEKEN. 1982. Selenium in environmental waters: Determination, speciation and concentration levels. *Talanta* 29:823.
- KUBOTA, J. & E.E. CARY. 1982. Cobalt, molybdenum and selenium. In A.L. Page et al., eds. *Methods of soil analysis, Part 2*, 2nd ed. *Agronomy* 9:485.
- RAPTIS, S. et al. 1983. A survey of selenium in the environment and a critical review of its determination at trace levels. *Fresenius Z. Anal. Chem.* 316:105.
- CUTTER, G. 1983. Elimination of nitrite interference in the determination of selenium by hydride generation. *Anal. Chim. Acta* 149:391.
- CAMPBELL, A. 1984. Critical evaluation of analytical methods for the determination of trace elements in various matrices. Part 1. Determination of selenium in biological materials and water. *Anal. Chem.* 56:645.
- LEMLY, A.D. 1985. Ecological basis for regulating aquatic emissions from the power industry: The case with selenium. *Regul. Toxic. Pharm.* 5:465.
- OHLENDORF, H.M., D.J. HOFFMAN, M.K. SAIKI & T.W. ALDRICH. 1986. Embryonic mortality and abnormalities of aquatic birds: Apparent impacts of selenium from irrigation drainwater. *Sci. Total Environ.* 52:49.

anionic chloro complex. In this treatment the acidity and ion exchanger do not alter speciation; complete speciation of selenium is possible.

a. Apparatus:

- 1) *Chromatography column* for organics removal, glass, about 0.8 cm ID × 30 cm long, with fluorocarbon metering valve.
- 2) *Chromatography column* for ion exchange, disposable polyethylene.†
- 3) *pH meter*.

b. Reagents:

- 1) *Organics-removal resin*: Thoroughly rinse 16 to 50 mesh resin‡ with deionized water and remove resin fines by decanting. Rinse three times with pH 12 solution. Store resin in pH 12 solution and refrigerate to prevent bacterial growth.
- 2) *Anion exchange resin*: Add 100 to 200 mesh anion exchange resin* to a beaker and thoroughly rinse with deionized

* Bio-Rad AG1-X8 or equivalent.

† Bio-Rad Econo-Columns or equivalent.

‡ Amberlite XAD-8, Supelco, or equivalent.

water. Cover resin with 4*N* HCl, stir, and let settle. Decant and repeat acid rinse twice more. Store resin in 4*N* HCl.

3) *Hydrochloric acid*, conc: Before use, bubble helium through the acid for 3 h at rate of 100 mL/min. (CAUTION: Use a fume hood.)

4) *pH 1.6 solution*: Adjust pH of deionized water to 1.6 with HCl.

5) *pH 12 solution*: Adjust pH of deionized water to 12 with KOH.

c. Procedure:

1) *Organic removal*—Place 5 cm washed resin in a 0.8-cm-ID column. Precondition column, at 1 mL/min, with 30 mL pH 12 solution and 20 mL pH 1.6 solution. Using HCl and a pH meter adjust sample to pH 1.6 to 1.8. Pass sample through preconditioned column at rate of 1 mL/min. Discard first 10 mL and use next 11 to 50 mL collected for Se(IV) determinations by Methods 3114B or C, or 3500-Se.C preceded, if Se(VI) also is to be determined, by preparatory step B.5. If more than 50 mL sample are needed, use another column or use a column with twice as much resin.

2) *Iron removal*—Place 4 cm prepared resin in a small chromatographic column (add resin to column filled with 4*N* HCl to avoid air bubbles). Rinse column with 10 mL 4*N* HCl at flow rate < 6 mL/min. Let solution drain to top of resin, but do not let the column run dry. Adjust sample to 4*N* HCl and pour into column. Discard first 10 mL and collect the next 11 to 100 mL for Se(IV) analysis by Methods 3114B or C, or 3500-Se.C preceded, if necessary, by preparatory step B.2 and if Se(VI) also is to be determined, by preparatory step B.5 below.

2. Removal of Interference by Persulfate Digestion

The combination of this procedure with step B.5 below and Methods 3114B or C, or 3500-Se.C is, in most cases, the preferred method for determining total selenium in filtered water. A small amount of ammonium or potassium persulfate is added to the mixture of sample and HCl to remove interference from reducing agents and to oxidize relatively labile organic selenium compounds such as selenoamino acids and methaneseleninic acid.

If the sample contains hydrogen sulfide or a large concentration of organic matter or is otherwise suspect or to confirm method accuracy, reanalyze sample using digestion procedure 3 or 4 below.

Prepare 2% potassium or ammonium persulfate solution by dissolving 2.0 g in 100 mL deionized water (prepare weekly). Add 0.2 mL persulfate solution to the mixed sample and HCl of ¶ B.5c before heating and proceeding with pretreatment and analysis.

After completing analysis multiply concentration of selenium determined in the acidified sample by 2.04 to obtain total selenium in original sample.

3. Removal of Interference by Alkaline Hydrogen Peroxide Digestion

Occasionally, digestion with persulfate gives incomplete recovery of total selenium. In this case, digestion with hydrogen peroxide is used to remove all reducing agents that might interfere and to fully oxidize organic selenium to Se(VI). The result-

ing solution can be analyzed for total selenium after pretreatment according to step B.5 below.

This method is suitable for determining total selenium in unfiltered water samples, where particulate selenium is present. When working with a new matrix, confirm results obtained by reanalyzing the sample using digestion procedure 4, below.

a. Apparatus:

1) *Beakers*, 150-mL.

2) *Watch glasses*.

3) *Hot plate*.

4) *Pipettor*, 1-mL, and tips.

5) *Graduated cylinder*, 25-mL.

b. Reagents:

1) *Hydrogen peroxide*, H₂O₂, 30%. Keep refrigerated.

2) *Sodium hydroxide*, NaOH, 1*N*.

3) *Hydrochloric acid*, HCl, 1.5*N*: Dilute 125 mL conc HCl to 1 L with deionized water.

c. Procedure: Add 2 mL 30% H₂O₂ and 1 mL 1 *N* NaOH to 25 mL sample in a beaker. Cover beaker to control spattering and simmer on hot plate until fine bubbles characteristic of H₂O₂ decomposition subside and are replaced by ordinary boiling. Add 1 mL 1.5*N* HCl to redissolve any precipitate that may have formed, let cool, and pour into graduated cylinder. Rinse beaker with deionized water into graduated cylinder and make volume up to 25 mL. Proceed to B.5 and chosen analytical method.

4. Removal of Interference by Permanganate Digestion

This digestion method utilizes potassium permanganate to oxidize selenium and remove interfering organic compounds. Excess KMnO₄ and MnO₂ are removed by reaction with hydroxylamine. HCl digestion is included here, because it is conveniently performed in the same reaction vial. Selenite may then be determined directly by Methods 3114B or C, or 3500-Se.C.

Verify recovery for the given matrix. This method gives good recovery even with heavily contaminated water samples that contain organic selenium compounds, dissolved organic matter, and visible suspended material.

Permanganate may oxidize chloride ion to free chlorine. Part of the chlorine (which interferes with hydride analysis) is removed by reaction with hydroxylamine, but the best way to eliminate free chlorine is by prolonged heating in an open vial during the digestion step. Excess hydroxylamine may reduce recovery by reducing selenium to Se(0).

a. Apparatus:

1) *Oven with thermostat*, for continuous operation at 110 ± 5°C.

2) *Digestion vials*, 40-mL, with fluorocarbon-lined screw caps.

3) *Metal support rack* to hold 40 digestion vials.

b. Reagents:

1) *Hydrochloric acid*, HCl, conc. [See B.1b3].

2) *Hydrochloric acid*, HCl, 10*N*: Dilute 1000 mL conc HCl to 1200 mL with deionized water.

3) *Sulfuric acid*, H₂SO₄, conc and 25%. NOTE: Many brands of H₂SO₄ are contaminated with selenium. Run reagent blanks when starting a new bottle. Make 25% v/v solution by adding 250 mL conc H₂SO₄ to 500 mL deionized water, and diluting to 1 L.

4) *Potassium permanganate solution*, KMnO₄, 5% (w/v): Dissolve 50 g KMnO₄ in 1000 mL deionized water.

5) *Hydroxylamine hydrochloride solution*: Dissolve 100 g $\text{NH}_2\text{OH} \cdot \text{HCl}$ in 1000 mL deionized water.

c. Procedure: Pipet 5 mL sample into digestion vial, add 5 mL 25% H_2SO_4 and 1 mL KMnO_4 solution. Lightly screw cap on and place in preheated oven at 110°C for 1 h. CAUTION: Excessive pressure may build up in tightly capped vials. Carefully remove tray with vials from the oven (CAUTION: potential for acid gas vapor release) and cool to room temperature. Open vial, carefully add a few drops hydroxylamine hydrochloride solution, mix, and wait until sample is decolorized and residual manganese dioxide is dissolved. Avoid excess hydroxylamine solution, which can cause a low reading. Add 10 mL conc HCl to the sample and heat vial 60 min at 95°C without cap. Let cool to room temperature. Transfer sample to a 25-mL volumetric flask or graduated cylinder, rinse vial into flask, dilute to mark, and mix well. Proceed to analyze by Methods 3114B or C, or 3500-Se.C. If Method 3114 B or C is used, multiply spectrometer readings by the dilution factor as follows:

$$\text{Concentration, } \mu\text{g/L} = \frac{\text{final volume}}{\text{volume of sample}} \times \text{reading}$$

5. Reduction of Se(VI) to Se(IV) by Hydrochloric Acid Digestion

Se(VI) is reduced to Se(IV) by digestion with HCl. Determine Se(IV) + Se(VI) by either hydride generation atomic absorption spectrometer (Method 3114B or C) or as an organic derivative (Method C).

Test any given sample matrix to ensure recovery of added Se(VI). If recovery is poor, try to remove interference by the procedure of ¶ B.1, above, or if at least 75% recovery is achieved, use the method of standard additions. The method described here is of limited utility for direct analysis of water samples, but is useful as a step in determining total selenium in a sample where selenium has been oxidized to Se(VI).

NOTE: Available literature does not provide definitive criteria for collection and preservation of samples for which speciation is desired. If speciation is needed, fully evaluate collection and preservation as part of the data quality objectives.

a. Apparatus:

- 1) *Dispenser*, bottle type, 5-mL, suitable for dispensing concentrated HCl.
- 2) *Pipettors*, 0.2- and 5-mL.
- 3) *Screw-cap culture tubes*, borosilicate glass, 25- × 150-mm.
- 4) *Boiling water bath*, suitable for heating culture tubes; a 1-L beaker on a hot plate is suitable.

b. Reagents:

1) *Sodium selenate additions solution*: Dilute 1000 mg/L stock selenate solution with deionized water to prepare a solution of 1 to 10 mg/L, such that the concentration of the additions solution will be approximately 50 times greater than anticipated total selenium in the sample to be analyzed.

2) *Hydrochloric acid*, HCl, conc: See B.1b3).

c. Procedure: Calibrate acid dispenser using water. Preheat water bath. Pipet 5 mL filtered sample into a culture tube. Add 5 mL conc HCl. Loosely cap tube (do not tighten) and place in boiling water bath for 20 min. Let tube cool and tighten cap. Determine total Se(IV) by Methods 3114B or C, or 3500-Se.C.

Add 0.200 mL additions solution with a microliter pipet to sample and proceed as above. Analyze a deionized water blank and a blank with the addition to ensure absence of contamination and to determine the true value of the addition.

Multiply the concentration of selenium determined in the acidified sample by 2.00 to obtain total concentration of Se(IV) + Se(VI). Multiply reading obtained for sample with addition by 2.04.

6. Bibliography

- JANGHORBANI, M., B. TING, A. NAHAPETLAN & R. YOUNG. 1982. Conversion of urinary selenium to selenium IV by wet oxidation. *Anal. Chem.* 54:1188.
- ADELOJU, S.B. & A.M. BOND. 1984. Critical evaluation of some wet digestion methods for the stripping voltammetric determination of selenium in biological materials. *Anal. Chem.* 56:2397.
- BRIMMER, S.P., W.R. FAWCETT & K.A. KULHAVY. 1987. Quantitative reduction of selenate ion to selenite in aqueous samples. *Anal. Chem.* 59:1470.

3500-Se C. Colorimetric Method

1. General Discussion

a. Principle: This method is specific to determining selenite ion in aqueous solution. Selenite ion reacts with 2,3-diaminonaphthalene to produce a brightly colored and strongly fluorescent piaszelenol compound, which is extracted in cyclohexane and measured colorimetrically.

The optimum pH for formation of the piaszelenol complex is approximately 1.5 but should not be above 2.5 because above pH 2, the rate of formation of the colored compound is critically dependent on pH. When indicators are used to adjust pH, results frequently are erratic; results can be improved when pH is monitored electrochemically.

b. Interference: No inorganic compounds are known to give a positive interference. Colored organic compounds extractable by cyclohexane may be encountered, but usually they are absent or can be removed by oxidizing the sample (see B.2, 3, or 4) or by treating it to remove dissolved organics (see B.1). Negative interference results from compounds that reduce the concentration of diaminonaphthalene by oxidizing it. Addition of EDTA eliminates negative interference from at least 2.5 mg Fe^{2+} .

c. Minimum detectable quantity: 10 μg Se/L.

2. Apparatus

a. Colorimetric equipment: A spectrophotometer, for use at 480 nm, providing a light path of 1 cm or longer.

- b. *Separatory funnel*, 250-mL, preferably with a fluorocarbon stopcock.
- c. *Thermostatically controlled water bath* (50°C) with cover.
- d. *pH meter*.
- e. *Centrifuge*, with rotor for 50-mL tubes (optional).
- f. *Centrifuge bottles*, 60-mL, screw-capped, fluorocarbon.
- g. *Shaker*, suitable for separatory funnel (optional).

3. Reagents

Use reagent water (see Section 1080) in preparing reagents.

a. *Selenium standard reference solution*: Dissolve 2.190 g sodium selenite, Na_2SeO_3 , in water containing 10 mL HCl and dilute to 1 L. 1.00 mL = 1.00 mg Se(IV).

b. *Working standard selenium solutions*: Dilute selenium reference standard solution with water or suitable background solution to produce a series of working standards spanning the concentration range of interest.

c. *Hydrochloric acid*: HCl, conc and 0.1N.

d. *Ammonium hydroxide*, NH_4OH , 50% v/v.

e. *Cyclohexane*, C_6H_{12} .

f. *2,3-Diaminonaphthalene (DAN) solution*: Dissolve 200 mg DAN in 200 mL 0.1N HCl. Shake 5 min. Extract three times with 25-mL portions of cyclohexane, retaining aqueous phase and discarding organic portions. Filter into opaque container* and store in cool, dark place for no longer than 8 h. CAUTION: *Toxic, handle with extreme care.*

g. *Hydroxylamine-EDTA solution (HA-EDTA)*: Dissolve 4.5 g Na_2EDTA in approximately 450 mL water. Add 12.5 g hydroxylamine hydrochloride ($\text{NH}_2\text{OH} \cdot \text{HCl}$) and adjust volume to 500 mL.

4. Procedure

a. *Formation of piaszelenol*: Add 2 mL HA-EDTA solution to 10 mL sample in 60-mL centrifuge bottle (filtered if Se(IV) is to be determined; oxidized using Method B.2, 3, or 4, then reduced using Method B.5 for total Se). Adjust to pH 1.5 ± 0.3 with 0.1N HCl and 50% NH_4OH , using a pH meter. Add 5 mL DAN solution and heat in a covered water bath at 50°C for 30 min.

b. *Extraction of piaszelenol*: Cool and add 2.0 mL cyclohexane. Cap container securely and shake vigorously for 5 min. Let solution stand for 5 min or until cyclohexane layer becomes well separated. If separation is slow, centrifuge for 5 min at 2000 rpm. Place bottle in a clamp on a ringstand at a 45° angle to the vertical. Remove aqueous phase using a disposable pipet attached to a vacuum line. Transfer organic phase to a small capped container using a clean disposable pipet, or to the spectrophotometer cuvette if absorbance is to be read immediately.

* Use Whatman No. 42 filter paper, or equivalent.

c. *Determination of absorbance*: Read absorbance at 480 nm using a zero standard. The piaszelenol color is very stable but evaporation of the cyclohexane concentrates the color unless the container is capped. CAUTION: *Avoid inhaling cyclohexane vapors.* Beer's Law is obeyed up to 2 mg/L.

5. Calculation

Construct calibration curve using at least a three-point standard curve to bracket the expected sample concentration. Plot absorbance vs. concentration. Correct for digestion blank and any reagent blank.

6. Precision and Bias

Three standard reference materials (wheat flour, water, and a commercial standard) were used to evaluate Se recovery.¹ The wheat flour sample was digested using HNO_3 and HClO_4 to convert total selenium to Se(VI), digested with HCl to convert Se(VI) to Se(IV) and finally, the colorimetric method was used. Results were as follows:

Standard	Selenium Concentration $\mu\text{g Se/L}$	
	Expected	Recovered*
NBS, SRM 1567, wheat flour†	1097 ± 197	1113 ± 8
NBS, SRM 1543ib, water	9.7 ± 0.5	8.7 ± 0
Fisher Certified AAS Standard	1002 ± 8	1002 ± 0

* Analyses in triplicate.

† Dry weight basis.

7. Reference

- HOLTZCLAW, K.M., R.H. NEAL, G. SPOSITO & S.J. TRAINA. 1987. A sensitive colorimetric method for the quantitation of selenite in soil solutions and natural waters. *Soil Sci. Soc. Amer. J.* 51:75.

8. Bibliography

- HOSTE, J. & J. GILLIS. 1955. Spectrophotometric determination of traces of selenium with 3,3'-diaminobenzidine. *Anal. Chim. Acta.* 12:158.
- CHENG, K. 1956. Determination of traces of selenium. *Anal. Chem.* 28:1738.
- MAGIN, G.B. et al. 1960. Suggested modified method for colorimetric determination of selenium in natural waters. *J. Amer. Water Works Assoc.* 52:1199.
- ROSSUM, J.R. & P.A. VILLARRUZ. 1962. Suggested methods for determining selenium in water. *J. Amer. Water Works Assoc.* 54:746.
- OLSEN, O.E. 1973. Simplified spectrophotometric analysis of plants for selenium. *J. Assoc. Offic. Anal. Chem.* 56:1073.

3500-Se D. Determination of Volatile Selenium

1. General Discussion

Dimethylselenide and dimethyldiselenide are low boiling, extremely malodorous organic compounds sparingly soluble in water. They are produced by microbial processes in seleniferous soil and decaying seleniferous organic matter, and occasionally are present in natural waters. They are readily air stripped from a water sample and can be collected with high efficiency in an alkaline solution of hydrogen peroxide, which oxidizes them quantitatively to Se(VI). Total selenium is determined by digestion with HCl and analysis by the hydride atomic absorption (3114B or C) or colorimetric (3500-Se.C) methods.

Either nitrogen or air may be used to strip the sample. Preferably use nitrogen if air-sensitive compounds (e.g., selenopolysulfides) are suspected.

Because volatile selenium can be lost in the course of sample collection and handling, preferably air-strip the sample in the field immediately after it is collected. After boiling to decompose H_2O_2 , return the alkaline peroxide solution to the laboratory for analysis.

2. Apparatus

All apparatus required for selenate reduction (B.5) and Methods 3114B or C, or 3500-Se.C, plus:

a. *Gas washing bottles*, borosilicate glass, 250-mL, with coarse porous glass gas dispersion frit. Mark 100-mL level on side of bottle.

b. *Rotameter*, to measure 3 L/min air flow.

c. *Gas flow regulator*

d. *Hot plate*.

e. *Graduated cylinder*, 100-mL.

f. *Beakers*, 250-mL.

g. *Rubber tubing*, to interconnect gas washing bottles and other gas equipment.

h. *Rubber gloves*.

3. Reagents

All reagents required for selenate reduction (B.5) and Methods 3114B or C, or 3500-Se.C, plus:

a. *Hydrogen peroxide*, H_2O_2 , 30%. Refrigerate.

b. *Sodium hydroxide solution*, NaOH, 1N.

c. *Compressed air or nitrogen*.

4. Procedure

Set up air-flow train in this order: Regulated air supply → rotameter → gas washing bottle 1 → gas washing bottle 2.

Prepare alkaline peroxide solution immediately before use by pouring 20 mL 30% H_2O_2 into a 100-mL graduated cylinder, adding 50 to 60 mL deionized water and 5 mL 1N NaOH, and making up to 100 mL. CAUTION: *Alkaline H_2O_2 is unstable. Do not keep in glass bottle; hold at about 0 °C in oversized plastic bottle. Solution is corrosive; protect eyes and skin.* Pour into gas washing bottle 2. Pour approximately 100 mL freshly collected sample into gas washing bottle 1. Do not attempt to measure sample volume accurately before volatile selenium determination, as unnecessary handling may cause volatile selenium to be lost.

Connect and check all air lines, turn on air and adjust flow to 3 L/min. Strip for 30 min or more. After 30 min, turn off air, disconnect gas washing bottle 2, and place it on the hot plate. Adjust heat to produce a gentle simmering of oxygen bubbles from decomposition of H_2O_2 . Continue heating until the characteristic effervescence of oxygen subsides and is replaced by ordinary boiling. Remove from hot plate and let cool. Pour solution into beaker. (Volume will be very near to 100 mL, and correction will usually be unnecessary.) Analyze for total selenium using HCl digestion (B.5) and Methods 3114B or C, or 3500-Se.C. Once boiled, this solution may be safely stored and transported in plastic bottles. Measure volume of sample in gas washing bottle 1.

5. Calculation

The concentration of volatile selenium compounds in the original water sample can be calculated as:

$$C = \frac{100}{\text{volume of original sample}} \times \text{conc of Se in solution}$$

6. Precision and Bias

Approximately 90% of dimethylselenide in samples will be recovered with 30 min air stripping. The recovery of dimethyldiselenide is not known. Loss of gases to the atmosphere during sampling and handling that precede analysis may cause a significant negative error.

3500-Se E. Determination of Nonvolatile Organic Selenium Compounds

1. General Discussion

In principle, the total amount of dissolved organic selenium plus polysulfidic selenium may be estimated by comparing "total Se," determined by oxidation and HCl digestion (B.2 or 3 and B.5, or B.4), followed by Methods 3114B or C, or 3500-Se.C,

with Se(IV) + Se(VI) determined by HCl digestion (B.5) and Method 3114C, or 3500-Se.C. In practice, this will give a meaningful estimate only if a known addition of Se(VI) is fully recovered. Even if recovery is good, this estimate may be unreliable, because it is the difference of two (frequently larger) numbers determined by slightly different methods. Comparing

total Se before and after treatment with resin [B.1c1]) gives a similarly unreliable estimate of nonvolatile organic Se.

It is preferable to separate and directly determine nonvolatile organic selenium. One method involves adsorption of dissolved organic matter onto a C-18 reverse phase HPLC resin, elution with an organic solvent, and determination of selenium in this fraction. While this technique is relatively simple, it is affected by pH and small organic molecules (e.g., individual selenium containing amino acids) are not retained by the resins. Adjust sample pH to 1.5 to 2.0 before using the column, but because the latter problem cannot be solved easily, the use of organic adsorbents provides only an estimate of organic selenium concentration.

Alternatively, isolate specific compounds and determine their selenium content. In some natural waters selenium may be associated with dissolved polypeptides or small proteins, and even small amounts of free selenoamino acids may be present. Because selenoamino acids are the most toxic form of the element, a direct determination is sometimes desirable.

To determine selenium in dissolved peptides, hydrolyze with acid and isolate the free amino acids via ligand exchange chromatography. Elute the selenoamino acids from the column and determine selenium. Selenoamino acids are unstable during acid hydrolysis and even using nonoxidizing methyl sulfonic acid and nitrogen-purged glass ampules, selenoamino acid recoveries are only 50 to 80%. This method is good only for estimating protein-bound selenium. A somewhat more reliable estimate of free selenoamino acids and selenium associated with small oligopeptides is obtained by performing a similar procedure without the hydrolysis step.

While these methods are too intricate for routine use, semi-quantitative, and sensitive only to certain classes of organic compounds, at present they are the only ones available with any practical experience.

Imperfect separation of organic selenium compounds from inorganic forms of selenium may cause interference. In parallel with the actual determination, always perform the procedure using a solution compounded to resemble the actual matrix and containing a similar amount of selenium, but in the form of Se(IV) and Se(VI) to determine degree of interference.

2. Apparatus

- Rotary evaporator*, with temperature-control bath and 30-mL pear-shaped flasks.
- Glass ampule sealing apparatus*, or oxygen-gas torch.
- Heating block*, 100°C, or pressure cooker.
- Glass chromatography columns*, 15 cm long, 0.7 cm ID.*
- Glass syringe*, 50-mL.
- Glass ampules*, 10- or 20- mL. Clean by heating in a muffle furnace at 400°C for 24 h.
- pH meter*.

3. Reagents

- Hydrochloric acid*, HCl, 1*N*.
- Methyl sulfonic acid*, conc.
- Ammonium hydroxide*, NH₄OH, 1.5*N*: Dilute 100 mL conc NH₄OH to 1 L with deionized water.

* Bio-Rad glass Econo-Columns or equivalent.

d. Sodium hydroxide, NaOH pellets and 1*N* solution.

e. pH 1.6 solution: Adjust pH of deionized water to 1.6 using HCl.

f. pH 9.0 solution: Adjust pH of deionized water to 9.0 using NaOH.

g. Copper sulfate solution, CuSO₄, 1*M*: Dissolve 25 g CuSO₄ · 5H₂O in deionized water and dilute to 100 mL.

h. Methanol, purified.

i. C-18 cartridges:[†] Using a glass syringe, pass the following sequence of reagents through the cartridge to clean the resin (6 mL/min or less): 10 mL deionized water; 20 mL 1*N* HCl; 10 mL deionized water; 20 mL methanol; 10 mL deionized water; and 20 mL pH 1.6 solution. Refrigerate but do not freeze cleaned cartridges.

j. Ligand-exchange chromatographic resin, 100/200 mesh: Rinse resin[‡] with deionized water to remove fines. Then rinse resin three times in the following sequence: 1*N* HCl; 1.5*N* NH₄OH; and deionized water. Store resin wet.

k. Copper-treated ligand-exchange chromatographic resin, 100/200 mesh: Rinse resin[‡] with deionized water to remove fines. Then rinse three times with 1*N* HCl, followed by deionized water. Using NaOH adjust pH of supernatant above the resin to about 7. Add CuSO₄ solution to the resin and stir. After settling, decant supernatant and add more CuSO₄ solution. Decant CuSO₄ solution and rinse with deionized water until no copper is noticeable in the supernatant. Rinse resin three times with 1.5*N* NH₄OH and three times with deionized water. Store resin wet.

4. Procedures

a. Extractable organic selenium: Adjust sample (5 to 50 mL) to pH 1.5 to 2.0 using HCl and place in a clean glass syringe with an attached cleaned C-18 cartridge. Push sample through the cartridge at a rate of 6 mL/min. After removing the cartridge, draw 2 mL pH 1.6 solution into syringe as a rinse, reattach cartridge, and push the rinse through the cartridge. Repeat two additional times. The cartridge can be refrigerated for storage. To elute organic selenium, push 10 mL methanol through the cartridge at rate of 2 mL/min and collect eluate in a 30-mL pear-shaped flask. Remove methanol by rotary evaporation, with the water bath temperature less than 40°C. Use deionized water to solubilize and transfer the residue into the vessel used for total selenium digestions. Determine total dissolved selenium by digestion with persulfate (B.2) or peroxide (B.3), reduction of Se(VI) (B.5), and analysis by Methods 3114B or C, or 3500-Se.C.

b. Hydrolysis of protein-bound selenium: Place filtered sample in a 10- or 20-mL glass ampule (depending on desired volume), and add conc methyl sulfonic acid to adjust concentration to 4*M*. Purge acidified sample with nitrogen for 10 min and seal top with a torch. Heat sealed vial at 100°C for 24 h in a heating block or pressure cooker. Transfer cooled hydrolysis solution with deionized water rinses to a 50-mL beaker and place in an ice bath. Using NaOH pellets and 1*N* NaOH, adjust to pH 9.0, taking care not to allow solution to heat to boiling.

c. Determination of selenoamino acids: If sample is not hydrolyzed as in ¶ 4b, filter, and adjust pH to 9.0 using 1*N* NaOH.

Fill an empty chromatographic column with deionized water and add ammonium-form resin to a depth of 2 cm. Add copper-

[†] Sep-Pak, Waters Associates, or equivalent.

[‡] Chelax 100 (ammonium form) Bio Rad, or equivalent.

treated resin to form a 12-cm length of resin. (The ammonium-form resin removes any copper that bleeds from the copper-treated resin above it). Rinse with deionized water until the pH of the effluent is 9.0; maintain flow through the column by gravity. Pass sample through column, rinse sample beaker with 5 mL pH 9 solution, and place the rinse on column (after the last of the sample reaches the top of the resin). Rinse beaker twice more. Discard flow through column.

Place clean beaker under column and add 20 mL 1.5N NH₄OH to the column. Neutralize NH₄OH eluate with 2.5 mL conc HCl. Determine total dissolved selenium by digestion with persulfate (B.2) or peroxide (B.3), reduction of Se(VI) (B.5) and analysis by Methods 3114B or C, or 3500-Se.C.

3500-Ag SILVER

Silver (Ag) is the second element in Group IB of the periodic table; it has an atomic number of 47, an atomic weight of 107.87, and valences of 1 and 2. The average abundance of Ag in the earth's crust is 0.08 ppm; in soils it is <0.01 to 0.5 ppm; in streams it is 0.3 µg/L; in U.S. drinking waters it is 0.23 µg/L, and in groundwater it is <0.1 µg/L. Silver occurs in its native state and in combination with many nonmetallic elements such as argentite (Ag₂S) and horn silver (AgCl). Lead and copper ores also may yield considerable silver. Silver is widely used in photography, silverware, jewelry, mirrors, and batteries. Silver iodide has been used in the seeding of clouds, and silver oxide to a limited extent is used as a disinfectant for water.

In acidic water Ag⁺ would predominate, and in high-chloride water a series of complexes would be expected. Silver is nonessential for plants and animals. Silver can cause argyria, a permanent, blue-gray discoloration of the skin and eyes that imparts a ghostly appearance. Concentrations in the range of 0.4 to 1 mg/L have caused pathological changes in the kidneys, liver, and spleen of rats. Toxic effects on fish in fresh water have been observed at concentrations as low as

Joint Task Group: 20th Edition—See 3500-AI.

3500-Na SODIUM*

3500-Na A. Introduction

1. Occurrence and Significance

Sodium (Na) is the third element in Group IA of the periodic table; it has an atomic number of 11, an atomic weight of 22.99,

* Approved by Standard Methods Committee, 1997.

Joint Task Group: 20th Edition—See 3500-AI.

5. Precision and Bias

These procedures are only semiquantitative. Typical relative standard deviation is 12% for the C-18 isolation of dissolved organic selenium, and 15% for protein-bound selenium.

6. Bibliography

- CUTTER, G. 1982. Selenium in reducing waters. *Science* 217:829.
 COOKE, T.D. & K.W. BRULAND. 1987. Aquatic chemistry of selenium: evidence of biomethylation. *Environ. Sci. Technol.* 21:1214.

0.17 µg/L. For freshwater aquatic life, total recoverable silver should not exceed 1.2 mg/L.

The atomic absorption spectrometric methods (3111B and C) and the inductively coupled plasma methods (3120 and 3125) are preferred. The electrothermal atomization method (3113B) is the most sensitive for determining silver in natural waters. The dithizone method detailed in the 19th edition of *Standard Methods* can be used when an atomic absorption spectrometer is unavailable. A method suitable for analysis of silver in industrial or other wastewaters at levels above 1 mg/L is available.¹

If total silver is to be determined, acidify sample with conc nitric acid (HNO₃) to pH <2 at time of collection. If sample contains particulate matter and only the "dissolved" metal content is to be determined, filter through a 0.45-µm membrane filter at time of collection. After filtration, acidify filtrate with HNO₃ to pH <2. Complete analysis as soon after collection as possible. Some samples may require special storage and digestion; see Section 3030D.

1. Reference

1. U.S. ENVIRONMENTAL PROTECTION AGENCY. 1994. Approved Inorganic Test Procedures. *Federal Register* 59(20):4504.

and a valence of 1. The average abundance of Na in the earth's crust is 2.5%; in soils it is 0.02 to 0.62%; in streams it is 6.3 mg/L, and in groundwaters it is generally >5 mg/L. Sodium occurs with silicates and with salt deposits. Sodium compounds are used in many applications, including caustic soda, salt, fertilizers, and water treatment chemicals.

Sodium is very soluble, and its monovalent ion Na⁺ can reach concentrations as high as 15 000 mg/L in equilibrium with sodium

bicarbonate. The ratio of sodium to total cations is important in agriculture and human physiology. Soil permeability can be harmed by a high sodium ratio. In large concentrations it may affect persons with cardiac difficulties. A limiting concentration of 2 to 3 mg/L is recommended in feedwaters destined for high-pressure boilers. When necessary, sodium can be removed by the hydrogen-exchange process or by distillation. The U.S. EPA advisory limit for sodium in drinking water is 20 mg/L.

2. Selection of Method

Method 3111B uses an atomic absorption spectrometer in the flame absorption mode. Method 3120B uses inductively coupled plasma; this method is not as sensitive as the other methods, but usually this is not important. Method 3500-Na.B uses either a flame

3500-Na B. Flame Emission Photometric Method

1. General Discussion

a. Principle: Trace amounts of sodium can be determined by flame emission photometry at 589 nm. Sample is nebulized into a gas flame under carefully controlled, reproducible excitation conditions. The sodium resonant spectral line at 589 nm is isolated by interference filters or by light-dispersing devices such as prisms or gratings. Emission light intensity is measured by a phototube, photomultiplier, or photodiode. The light intensity at 589 nm is approximately proportional to the sodium concentration. Alignment of the wavelength dispersing device and wavelength readout may not be precise. The appropriate wavelength setting, which may be slightly more or less than 589 nm, can be determined from the maximum emission intensity when aspirating a sodium standard solution, and then used for emission measurements. The calibration curve may be linear but has a tendency to level off or even reverse at higher concentrations. Work in the linear to near-linear range.

b. Interferences: Minimize interference by incorporation of one or more of the following:

- 1) Operate at the lowest practical concentration range.
- 2) Add releasing agents, such as strontium or lanthanum at 1000 mg/L, to suppress ionization and anion interference. Among common anions capable of causing interference are Cl^- , SO_4^{2-} and HCO_3^- in relatively large amounts.
- 3) Matrix-match standards and samples by adding identical amounts of interfering substances present in the sample to calibration standards.
- 4) Apply an experimentally determined correction in those instances where the sample contains a single important interference.

5) Remove interfering ions.

6) Remove burner-clogging particulate matter from the sample by filtration through a filter paper of medium retentiveness.

7) Use the standard addition technique as described in the flame photometric method for strontium (3500-Sr.B). The method involves preparing a calibration curve using the sample matrix as a diluent, and determining the sample concentration

photometer or an atomic absorption spectrometer in the flame emission mode. The inductively coupled plasma/mass spectrometric method (3125) also may be applied successfully in most cases (with lower detection limits), even though sodium is not specifically listed as an analyte in the method. When all of these instruments are available, the choice will depend on factors including relative quality of the instruments, precision and sensitivity required, number of samples and analytes per sample, matrix effects, and relative ease of instrument operation. If an atomic absorption spectrometer is used, operation in the emission mode is preferred.

3. Storage of Sample

Store alkaline samples or samples containing low sodium concentrations in polyethylene bottles to eliminate the possibility of sample contamination due to leaching of the glass container.

either mathematically or graphically.

8) Use the internal standard technique. Potassium and calcium interfere with sodium determination by the internal-standard method if the potassium-to-sodium ratio is $\geq 5:1$ and the calcium-to-sodium ratio is $\geq 10:1$. When these ratios are exceeded, determine calcium and potassium concentrations and matrix-match sodium calibration standards by addition of approximately equivalent concentrations of interfering ions. Interference from magnesium is not significant until the magnesium-to-sodium ratio exceeds 100, a rare occurrence.

c. Minimum detectable concentration: Better flame photometers or atomic absorption spectrometers operating in the emission mode can be used to determine sodium levels approximating 5 $\mu\text{g/L}$.

2. Apparatus

a. Flame photometer (either direct-reading or internal-standard type) or atomic absorption spectrometer operating in the flame emission mode.

b. Glassware: Rinse all glassware with 1 + 15 HNO_3 followed by several portions of reagent water (¶ 3a).

3. Reagents

To minimize sodium contamination, store all solutions in plastic bottles. Use small containers to reduce the amount of dry element that may be picked up from the bottle walls when the solution is poured. Shake each container vigorously to wash accumulated salts from walls before pouring solution.

a. Reagent water: See Section 1080. Use reagent water to prepare all reagents and calibration standards, and as dilution water.

b. Stock sodium solution: Dissolve 2.542 g NaCl dried at 140°C to constant weight and dilute to 1000 mL with water; 1.00 mL = 1.00 mg Na.

c. Intermediate sodium solution: Dilute 10.00 mL stock sodium solution with water to 100.0 mL; 1.00 mL = 0.10 mg Na

(1.00 mL = 100 μg Na). Use this intermediate solution to prepare calibration curve in sodium range of 1 to 10 mg/L.

d. Standard sodium solution: Dilute 10.00 mL intermediate sodium solution with water to 100 mL; 1.00 mL = 10.0 μg Na. Use this solution to prepare calibration curve in sodium range of 0.1 to 1.0 mg/L.

4. Procedure

a. Pretreatment of polluted water and wastewater samples: Follow the procedures described in Section 3030.

b. Instrument operation: Because of differences between makes and models of instruments, it is impossible to formulate detailed operating instructions. Follow manufacturer's recommendation for selecting proper photocell and wavelength, adjusting slit width and sensitivity, appropriate fuel and oxidant gas pressures, and the steps for warm-up, correcting for interferences and flame background, rinsing of burner, igniting flame, and measuring emission intensity.

c. Direct-intensity measurement: Prepare a blank and sodium calibration standards in stepped amounts in any of the following applicable ranges: 0 to 1.0, 0 to 10, or 0 to 100 mg/L. Determine emission intensity at 589 nm. Aspirate calibration standards and samples enough times to secure a reliable average reading for each. Construct a calibration curve from the sodium standards. Determine sodium concentration of sample from the calibration curve. Where a large number of samples must be run routinely, the calibration curve provides sufficient accuracy. If greater precision and less bias are desired and time is available, use the bracketing approach described in ¶ 4d below.

d. Bracketing approach: From the calibration curve, select and prepare sodium standards that immediately bracket the emission intensity of the sample. Determine emission intensities of the bracketing standards (one sodium standard slightly less and the other slightly greater than the sample) and the sample as nearly simultaneously as possible. Repeat the determination on bracketing standards and sample. Calculate the sodium concentration by the equation in ¶ 5b and average the findings.

5. Calculation

a. For direct reference to the calibration curve:

$$\text{mg Na/L} = (\text{mg Na/L in portion}) \times D$$

b. For the bracketing approach:

$$\text{mg Na/L} = \left[\frac{(B - A)(s - a)}{(b - a)} + A \right] D$$

where:

B = mg Na/L in upper bracketing standard,

A = mg Na/L in lower bracketing standard,

b = emission intensity of upper bracketing standard,

a = emission intensity of lower bracketing standard,

s = emission intensity of sample, and

D = dilution ratio

$$D = \frac{\text{mL sample} + \text{mL water}}{\text{mL sample}}$$

6. Precision and Bias

A synthetic sample containing 19.9 mg Na^+/L , 108 mg Ca^{2+}/L , 82 mg Mg^{2+}/L , 3.1 mg K^+/L , 241 mg Cl^-/L , 0.25 mg NO_2^-/L , 1.1 mg NO_3^-/L , 259 mg $\text{SO}_4^{2-}/\text{L}$, and 42.5 mg total alkalinity/L (as CaCO_3) was analyzed in 35 laboratories by the flame photometric method, with a relative standard deviation of 17.3% and a relative error of 4.0%.

7. Bibliography

- WEST, P.W., P. FOLSE & D. MONTGOMERY. 1950. Application of flame spectrophotometry to water analysis. *Anal. Chem.* 22:667.
- COLLINS, C.G. & H. POLKINHORNE. 1952. An investigation of anionic interference in the determination of small quantities of potassium and sodium with a new flame photometer. *Analyst* 77:430.
- MELOCHE, V.W. 1956. Flame photometry. *Anal. Chem.* 28:1844.
- BURRIEL-MARTI, F. & J. RAMIREZ-MUNOZ. 1957. *Flame Photometry: A Manual of Methods and Applications*. D. Van Nostrand Co., Princeton, N.J.
- DEAN, J.A. 1960. *Flame Photometry*. McGraw-Hill Publishing Co., New York, N.Y.
- URE, A.M. & R.L. MITCHELL. 1975. Lithium, sodium, potassium, rubidium, and cesium. In J.A. Dean & T.C. Rains, eds. *Flame Emission and Atomic Absorption Spectrometry*. Dekker, New York, N.Y.
- THOMPSON, K.C. & REYNOLDS, R.J. 1978. *Atomic Absorption, Fluorescence, and Flame Spectroscopy—A Practical Approach*, 2nd ed. John Wiley & Sons, New York, N.Y.
- WILLARD, H.H., L.L. MERRIT, JR., J.A. DEAN & F.A. SETTLE, JR. 1981. *Instrumental Methods of Analysis*, 6th ed. Wadsworth Publishing Co., Belmont, Calif.
- AMERICAN SOCIETY FOR TESTING AND MATERIALS. 1988. Method D 1428-82: Standard test methods for sodium and potassium in water and water-formed deposits by flame photometry. *Annual Book of ASTM Standards*, Vol. 11.01. American Soc. Testing & Materials, Philadelphia, Pa.

3500-Sr STRONTIUM*

3500-Sr A. Introduction

1. Occurrence and Significance

Strontium (Sr) is the fourth element in Group IIA of the periodic table; it has an atomic number of 38, an atomic weight of 87.62, and a valence of 2. The average abundance of Sr in the earth's crust is 384 ppm; in soils Sr ranges from 3.6 to 160 ppm; in streams it averages 50 $\mu\text{g/L}$, and in groundwaters it ranges from 0.01 to 10 mg/L. Strontium is found chiefly in celestite (SrSO_4) and in strontianite (SrCO_3). Strontium compounds are used in pigments, pyrotechnics, ceramics, and flares. ^{90}Sr is a fission product of nuclear reactor fuels, and was widely distributed on the earth's surface as a result of fallout from nuclear weapons testing.

The common aqueous species is Sr^{2+} . The solubility of strontium is controlled by carbonate and sulfate. Some compounds are toxic by ingestion and inhalation. Although there is no U.S. EPA drinking water standard MCL for concentra-

tion of strontium, strontium-90 measurements are required when the gross beta activity of a water sample is greater than 50 pCi/L. The U.S. EPA primary drinking water standard MCL for ^{90}Sr is 8 pCi/L.

A method for determination of ^{90}Sr is found in Section 7500-Sr.

2. Selection of Method

The atomic absorption spectrometric method (3111B) and inductively coupled plasma methods (3120 and 3125) are preferred. The flame emission photometric method (B) also is available for those laboratories that do not have the equipment needed for one of the preferred methods.

3. Sampling and Storage

Polyethylene bottles are preferable for sample storage, although borosilicate glass containers also may be used. At time of collection adjust sample to pH <2 with nitric acid (HNO_3).

* Approved by Standard Methods Committee, 1997.
Joint Task Group: 20th Edition—See 3500-A1.

3500-Sr B. Flame Emission Photometric Method

1. General Discussion

a. Principle: The flame photometric method can be used for the determination of strontium in the concentration range prevalent in natural waters. The strontium emission is measured at a wavelength of 460.7 nm, while the background intensity is measured at a wavelength of 466 nm. The difference in readings obtained at these two wavelengths measures the light intensity emitted by strontium.

b. Interference: Emission intensity is a linear function of strontium concentration and concentration of other constituents. The standard addition technique distributes the same ions throughout the standards and the sample, thereby equalizing the radiation effect of possible interfering substances. A very low pH (<1) could produce an interference, but sample dilution should eliminate this interference.

c. Minimum detectable concentration: Strontium levels of about 0.2 mg/L can be detected by the flame photometric method without prior sample concentration.

2. Apparatus

Spectrophotometer, equipped with photomultiplier tube and flame accessories; or an atomic absorption spectrophotometer capable of operation in flame emission mode.

3. Reagents

a. Stock strontium solution: Dissolve 2.415 g strontium nitrate, $\text{Sr}(\text{NO}_3)_2$, dried to constant weight at 140°C, in 1000 mL 1% (v/v) HNO_3 ; 1.00 mL = 1.00 mg Sr.

b. Standard strontium solution: Dilute 25.00 mL stock strontium solution to 1000 mL with water; 1.00 mL = 25.0 μg Sr. Use this solution for preparing Sr standards in the 0.2- to 25-mg/L range.

c. Nitric acid, HNO_3 , conc.

4. Procedure

a. Pretreatment of polluted water and wastewater samples: Select an appropriate procedure from Section 3030.

b. Preparation of strontium standards: Dilute samples, if necessary, to contain less than 400 mg Ca or Ba/L and less than 40 mg Sr/L. Add 25.0 mL sample (or a lesser but consistent volume to keep all standards in the linear range of the instrument) to 25.0 mL of each of a series of four or more strontium standards containing from 0 mg/L to a concentration exceeding that of the sample. For most natural waters 0, 2.0, 5.0, and 10.0 mg Sr/L standards are sufficient. A broader range curve might be preferable for brines. Dilute the brine sufficiently to eliminate burner splatter and clogging.

c. Concentration of low-level strontium samples: Concentrate samples containing less than 2 mg Sr/L. Polluted water or

wastewater samples can be concentrated during digestion by starting with a larger volume (see Section 3030D). For other samples, add 3 to 5 drops conc HNO₃ to 250 mL sample and evaporate to about 25 mL. Cool and make up to 50.0 mL with distilled water. Proceed as in ¶ b. The HNO₃ concentration in the sample prepared for atomization can approach 0.4 mL/50 mL without producing interference.

d. *Flame photometric measurement:* Measure emission intensity of prepared samples (standards plus sample) at wavelengths of 460.7 and 466 nm. Follow manufacturer's instructions for correct instrument operation. Use a fuel-rich nitrous oxide-acetylene flame, if possible.

5. Calculation

a. Using a calculator or computer with linear regression capability, enter the net intensity (reading at 460.7 nm minus reading at 466 nm) versus concentration added to the sample and solve the equation for zero emissions. The negative of this number multiplied by any dilution factor is the sample concentration.

b. Plot net intensity (reading at 460.7 nm minus reading at 466 nm) against strontium concentration added to the sample. Because the plot forms a straight calibration line that intersects the ordinate, strontium concentration can be calculated from the equation:

$$\text{mg Sr/L} = \frac{A - B}{C} \times \frac{D}{E}$$

where:

- A = sample emission-intensity reading of sample plus 0 mg/L at 460.7 nm,
- B = background radiation reading at 466 nm, and
- C = slope of calibration line.

Use the ratio D/E only when E mL of sample are concentrated to a final volume D mL (typically 50 mL).

c. *Graphical method:* Strontium concentration also can be evaluated by the graphical method illustrated in Figure 3500-Sr:1. Plot net intensity against strontium concentration added to sample. If the line intersects the ordinate at Y emissions, the strontium concentration is where the abscissa value of the point on the calibration line has an ordinate value of 2Y emissions due to the two-fold dilution with standards (if sample and standards are mixed in equal volumes). The calibration line in the example intersects the ordinate at 12. Thus, Y = 12 and 2Y = 24. The strontium concentration of the sample is the abscissa value of the point on the calibration line having an ordinate value of 24. In the example, the strontium concentration is 9.0 mg/L.

d. Report strontium concentrations below 10 mg/L to the nearest 0.1 mg/L and above 10 mg/L to the nearest whole number.




Figure 3500-Sr:1. Graphical method of computing strontium concentration.

6. Quality Control

See Part 1000 and Section 3020 for specific quality control procedures to be followed during sample preparation and analysis.

7. Precision and Bias

Strontium concentrations in the range 12.0 to 16.0 mg/L can be determined with an accuracy within ±1 to 2 mg/L.

8. Bibliography

CHOW, T.J. & T.G. THOMPSON. 1955. Flame photometric determination of strontium in sea water. *Anal. Chem.* 27:18.
 NICHOLS, M.S. & D.R. McNALL. 1957. Strontium content of Wisconsin municipal waters. *J. Amer. Water Works Assoc.* 49:1493.
 HERR, C.A. 1959. A survey of analytical methods for the determination of strontium in natural water. U.S. Geol. Surv. Water Supply Pap. No. 1496A.

3500-Te TELLURIUM

Tellurium (Te) is the fourth element in Group VIA in the periodic table; it has an atomic number of 52, an atomic weight of 127.60, and valences of 2, 4, and 6. The average abundance of Te in the earth's crust is 0.002 ppm; in soils it is 0.001 to 0.01 ppm; and in groundwaters it is <0.1 mg/L. Tellurium is found in its native state and as the telluride of gold and other metals. It is used in alloys, catalysts, batteries, and as a coloring agent in glass and ceramics.

Joint Task Group: 20th Edition—See 3500-A1.

3500-Tl THALLIUM

Thallium (Tl) is the fifth element in Group IIIA in the periodic table; it has an atomic number of 81, an atomic weight of 204.38, and valences of 1 and 3. The average abundance in the earth's crust is 0.07 ppm, and in groundwaters it is <0.1 mg/L. The metal occurs chiefly in pyrites. Thallium is used in the production of glasses and rodenticides, in photoelectric applications, and in electrodes for dissolved oxygen meters.

Joint Task Group: 20th Edition—See 3500-A1.

3500-Th THORIUM

Thorium (Th) is the first element in the actinium series of the periodic table; it has an atomic number of 90, an atomic weight of 232.04, and a valence of 4. The average abundance in the earth's crust is 8.1 ppm; in soils it is 13 ppm; in streams it is 0.1 $\mu\text{g/L}$, and in groundwaters it is <0.1 mg/L. Thorium is a radioactive element, with ^{232}Th having a half-life of 1.4×10^{10} years. It is widely distributed in the earth, with the principal mineral being monazite. Thorium is used in sun lamps, photoelectric cells, incandescent lighting, and gas mantles.

Joint Task Group: 20th Edition—See 3500-A1.

3500-Sn TIN

Tin (Sn) is the fourth element in Group IVA in the periodic table; it has an atomic number of 50, an atomic weight of 118.69, and valences of 2 and 4. The average abundance in the earth's crust is 2.1 ppm; in soils it is 10 ppm; in streams it is 0.1 $\mu\text{g/L}$, and in groundwaters it is <0.1 mg/L. Tin is found mostly in the mineral cassiterite (SnO_2), in association with granitic rocks. Tin is used in reducing agents, solder, bronze, pewter, and coatings for various metals.

The common aqueous species are Sn^{4+} , $\text{Sn}(\text{OH})_4$, $\text{SnO}(\text{OH})_2$, and $\text{SnO}(\text{OH})_3^-$. Tin is adsorbed to suspended

Joint Task Group: 20th Edition—See 3500-A1.

The common aqueous species is TeO_3^{2-} . The metal and its compounds are toxic by inhalation.

Perform analyses by the electrothermal atomic absorption method (3113B). The inductively coupled plasma mass spectrometric method (3125) also may be applied successfully in most cases (with lower detection limits), even though tellurium is not specifically listed as an analyte in the method.

The common aqueous species is Tl^+ . It is nonessential for plants and animals. Compounds of thallium are toxic on contact with moisture, and by inhalation. The U.S. EPA primary drinking water standard MCL is 2 $\mu\text{g/L}$.

For analysis, use one of the atomic absorption spectrometric methods (3111B or 3113B), or one of the inductively coupled plasma methods (3120 or 3125), depending upon sensitivity requirements.

The aqueous chemistry of thorium is controlled by the Th^{4+} ion, which forms a set of complex species with hydroxides. Thorium's radioactive decay isotopes are dangerous when inhaled or ingested as thorium dust particles.

Either of the flame atomic absorption spectrometric methods (3111D or E) may be used for analysis. The inductively coupled plasma mass spectrometric method (3125) also may be applied successfully in most cases (with lower detection limits), even though thorium is not specifically listed as an analyte in the method.

solids, sulfides, and hydroxides. Tin can be methylated in sediments. Tributyl tin undergoes biodegradation quickly. Organo-tin compounds are toxic. Tin is considered nonessential for plants and animals.

Either the flame atomic absorption method (3111B) or the electrothermal atomic absorption method (3113B) may be used successfully for analyses, depending upon the sensitivity desired. The inductively coupled plasma/mass spectrometric method (3125) also may be applied successfully in most cases (with lower detection limits), even though tin is not specifically listed as an analyte in the method.

3500-Ti TITANIUM

Titanium (Ti) is the first element in Group IVB in the periodic table; it has an atomic number of 22, an atomic weight of 47.88, and valences of 2, 3, and 4. The average abundance of Ti in the earth's crust is 0.6%; in soils it is 1700 to 6600 ppm; in streams it is 3 $\mu\text{g/L}$, and in groundwaters it is <0.1 mg/L. The element is commonly associated with iron minerals. Titanium is used in alloys for aircraft, marine, and food-handling equipment. Compounds of the metal are used in pigments and as a reducing agent.

Joint Task Group: 20th Edition—See 3500-A1.

Titanium species are usually insoluble in natural waters, with the Ti^{4+} species being the most common ion when found. Some compounds are toxic by ingestion and the pure metal is flammable.

Either of the flame atomic absorption spectrometric methods (3111D or E) may be used. The inductively coupled plasma mass spectrometric method (3125) also may be applied successfully in most cases (with lower detection limits), even though titanium is not specifically listed as an analyte in the method.

3500-U URANIUM

Uranium (U) is the third element in the actinide series of the periodic table; it has an atomic number of 92, an atomic weight of 238.04, and valences of 3, 4, and 6. The average abundance of U in the earth's crust is 2.3 ppm, and in soils it is 1.8 ppm. Concentrations of uranium in drinking waters usually are expressed in terms of picocuries per liter, but that is now being replaced by Becquerel per liter (Bq/L). The approximate conversion factor, assuming equilibrium between ^{234}U and ^{238}U , is 1 μg uranium equals 0.67 pCi. The mean concentration of uranium in drinking water is 1.8 pCi/L. The chief ore is uraninite, or pitchblende, uranous uranate [$\text{U}(\text{UO}_4)_2$]. Uranium is known mainly for its use in the nuclear industry, but has also been used in glass, ceramics, and photography.

Uranium compounds are radioactive and are thereby toxic by inhalation and ingestion. There are three natural radioiso-

topes of uranium. Uranium-238 has a half-life of 4.5×10^9 years, represents 99% of uranium's natural abundance, and is not fissionable, but can be used to form plutonium-239, which is fissionable. Uranium-235 has a half-life of 7.1×10^8 years, represents 0.75% of uranium's natural abundance, is readily fissionable, and was the energy source in the original atomic bombs. Uranium-234 has a half-life of 2.5×10^5 years and represents only 0.006% of uranium's natural abundance.

The common forms in natural water are U^{4+} and UO_2^{2+} . In natural waters below pH 5, UO_2^{2+} would dominate; in the pH range of 5 to 10, soluble carbonate complexes predominate. Although there is no U.S. EPA drinking water standard MCL for uranium, an analysis for uranium is required if the gross alpha activity of a water sample is greater than 15 pCi/L.

Perform analyses by the inductively coupled plasma/mass spectrometry method (3125) or by one of the methods in 7500-U (for regulatory compliance purposes).

Joint Task Group: 20th Edition—See 3500-A1.

3500-V VANADIUM*

3500-V A. Introduction

1. Occurrence and Significance

Vanadium (V) is the first element in Group VB in the periodic table; it has an atomic number of 23, an atomic weight of 50.94, and valences of 2, 3, 4, and 5. The average abundance of V in the earth's crust is 136 ppm; in soils it ranges from 15 to 110 ppm; in streams it averages about 0.9 $\mu\text{g/L}$, and in groundwaters it is generally <0.1 mg/L. Though relatively rare, vanadium is found in a variety of minerals; most important among these are vanadinite [$\text{Pb}_5(\text{VO}_4)_3\text{Cl}$], and patronite (possibly VS_4), occurring chiefly in Peru. Vanadium complexes have been noted in coal

and petroleum deposits. Vanadium is used in steel alloys and as a catalyst in the production of sulfuric acid and synthetic rubber.

The dominant form in natural waters is V^{5+} . It is associated with organic complexes and is insoluble in reducing environments. It is considered nonessential for most higher plants and animals, although it may be an essential trace element for some algae and microorganisms. Laboratory and epidemiological evidence suggests that vanadium may play a beneficial role in the prevention of heart disease. In water supplies in New Mexico, which has a low incidence of heart disease, vanadium has been found in concentrations of 20 to 150 $\mu\text{g/L}$. In a state where incidence of heart disease is high, vanadium was not found in water supplies. However, vanadium pentoxide dust causes gastrointestinal and respiratory disturbances. The United Nations

* Approved by Standard Methods Committee, 1997.

Joint Task Group: 20th Edition—See 3500-A1.

Food and Agriculture Organization recommended maximum level for irrigation waters is 0.1 mg/L.

2. Selection of Method

The atomic absorption spectrometric methods (3111D and E), the electrothermal atomic absorption method (3113B), the in-

ductively coupled plasma methods (3120 and 3125), and gallic acid method (3500-V.B) are suitable for potable water samples. The atomic absorption spectrometric and inductively coupled plasma methods are preferred for polluted samples. The electrothermal atomic absorption method also may be used successfully with an appropriate matrix modifier.

3500-V B. Gallic Acid Method

1. General Discussion

a. Principle: The concentration of trace amounts of vanadium in water is determined by measuring the catalytic effect it exerts on the rate of oxidation of gallic acid by persulfate in acid solution. Under the given conditions of concentrations of reactants, temperature, and reaction time, the extent of oxidation of gallic acid is proportional to the concentration of vanadium. Vanadium is determined by measuring the absorbance of the sample at 415 nm and comparing it with that of standard solutions treated identically.

b. Interference: The substances listed in Table 3500-V:1 will interfere in the determination of vanadium if the specified concentrations are exceeded. This is not a serious problem for Cr^{6+} , Co^{2+} , Mo^{6+} , Ni^{2+} , Ag^+ , and U^{6+} because the tolerable concentration is greater than that commonly encountered in fresh water. However, in some samples the tolerable concentration of Cu^{2+} , Fe^{2+} , and Fe^{3+} may be exceeded. Because of the high sensitivity of the method, interfering substances in concentrations only slightly above tolerance limits can be rendered harmless by dilution.

Traces of Br^- and I^- interfere seriously and dilution alone will not always reduce the concentration below tolerance limits. Mercuric ion may be added to complex these halides and minimize their interference; however, mercuric ion itself interferes if in excess. Adding 350 μg mercuric nitrate, $\text{Hg}(\text{NO}_3)_2$, per sample permits determination of vanadium in the presence of up to 100 mg Cl^-/L , 250 μg Br^-/L , and 250 μg I^-/L . Dilute samples containing high concentrations of these ions to concentrations below the values given above and add $\text{Hg}(\text{NO}_3)_2$.

TABLE 3500-V:1. CONCENTRATION AT WHICH VARIOUS IONS INTERFERE IN THE DETERMINATION OF VANADIUM

Ion	Concentration mg/L
Cr^{6+}	1.0
Co^{2+}	1.0
Cu^{2+}	0.05
Fe^{2+}	0.3
Fe^{3+}	0.5
Mo^{6+}	0.1
Ni^{2+}	3.0
Ag^+	2.0
U^{6+}	3.0
Br^-	0.1
Cl^-	100.0
I^-	0.001

c. Minimum detectable concentration: 0.025 μg V in approximately 13 mL final volume or approximately 2 μg V/L.

2. Apparatus

- a. Water bath,* capable of being operated at $25 \pm 0.5^\circ\text{C}$.
- b. Colorimetric equipment:* One of the following is required:
 - 1) *Spectrophotometer,* for measurements at 415 nm, with a light path of 1 to 5 cm.
 - 2) *Filter photometer,* providing a light path of 1 to 5 cm and equipped with a violet filter with maximum transmittance near 415 nm.

3. Reagents

Use reagent water (see Section 1080) in preparation of reagents, for dilutions, and as blanks.

a. Stock vanadium solution: Dissolve 229.6 mg ammonium metavanadate, NH_4VO_3 , in a volumetric flask containing approximately 800 mL water and 15 mL 1 + 1 nitric acid (HNO_3). Dilute to 1000 mL; 1.00 mL = 100 μg V.

b. Intermediate vanadium solution: Dilute 1.00 mL stock vanadium solution with water to 100 mL; 1.00 mL = 1.00 μg V.

c. Standard vanadium solution: Dilute 1.00 mL intermediate vanadium solution with water to 100 mL; 1.00 mL = 0.010 μg V.

d. Mercuric nitrate solution: Dissolve 350 mg $\text{Hg}(\text{NO}_3)_2 \cdot \text{H}_2\text{O}$ in 1000 mL water.

e. Ammonium persulfate-phosphoric acid reagent: Dissolve 2.5 g $(\text{NH}_4)_2\text{S}_2\text{O}_8$ in 25 mL water. Bring just to a boil, remove from heat, and add 25 mL conc H_3PO_4 . Let stand approximately 24 h before use. Discard after 48 h.

f. Gallic acid solution: Dissolve 2 g $\text{H}_6\text{C}_7\text{O}_5$ in 100 mL warm water, heat to a temperature just below boiling, and filter through filter paper.* Prepare a fresh solution for each set of samples.

4. Procedure

a. Preparation of standards and sample: Prepare both blank and sufficient standards by diluting 0- to 8.0-mL portions (0 to 0.08 μg V) of standard vanadium solution to 10 mL with water. Pipet sample (10.00 mL maximum) containing less than 0.08 μg V into a suitable container and adjust volume to 10.0 mL with water. Filter colored or turbid samples. Add 1.0 mL $\text{Hg}(\text{NO}_3)_2$

* Whatman No. 42 or equivalent.

solution to each blank, standard, and sample. Place containers in a water bath regulated to $25 \pm 0.5^\circ\text{C}$ and allow 30 to 45 min for samples to come to the bath temperature.

b. Color development and measurement: Add 1.0 mL ammonium persulfate-phosphoric acid reagent (temperature equilibrated), swirl to mix thoroughly, and return to water bath. Add 1.0 mL gallic acid solution (temperature equilibrated), swirl to mix thoroughly, and return to water bath. Add gallic acid to successive samples at intervals of 30 s or longer to permit accurate control of reaction time. Exactly 60 min after adding gallic acid, remove sample from water bath and measure its absorbance at 415 nm, using water as a reference. Subtract absorbance of blank from absorbance of each standard and sample. Construct a calibration curve by plotting absorbance values of standards versus micrograms vanadium. Determine amount of vanadium in a sample by referring to the corresponding absorbance on the calibration curve. Prepare a calibration curve with each set of samples.

5. Calculation

$$\text{mg V/L} = \frac{\mu\text{g V (in 13 mL final volume)}}{\text{original sample volume, mL}}$$

6. Precision and Bias

In a synthetic sample containing $6 \mu\text{g V/L}$, $40 \mu\text{g As/L}$, $250 \mu\text{g Be/L}$, $240 \mu\text{g B/L}$, and $20 \mu\text{g Se/L}$ in distilled water, vanadium was measured in 22 laboratories with a relative standard deviation of 20% and no relative error.

7. Bibliography

FISHMAN, M.J. & M.V. SKOUGSTAD. 1964. Catalytic determination of vanadium in water. *Anal. Chem.* 36:1643.

3500-Zn ZINC*

3500-Zn A. Introduction

1. Occurrence and Significance

Zinc (Zn) is the first element in Group IIB in the periodic table; it has an atomic number of 30, an atomic weight of 65.38, and a valence of 2. The average abundance of Zn in the earth's crust is 76 ppm; in soils it is 25 to 68 ppm; in streams it is $20 \mu\text{g/L}$, and in groundwaters it is $<0.1 \text{ mg/L}$. The solubility of zinc is controlled in natural waters by adsorption on mineral surfaces, carbonate equilibrium, and organic complexes. Zinc is used in a number of alloys such as brass and bronze, and in batteries, fungicides, and pigments. Zinc is an essential growth element for plants and animals but at elevated levels it is toxic to some species of aquatic life. The United Nations Food and Agriculture Organization recommended level for zinc in irrigation waters is 2 mg/L . The U.S. EPA secondary drinking water

standard MCL is 5 mg/L . Concentrations above 5 mg/L can cause a bitter astringent taste and an opalescence in alkaline waters. Zinc most commonly enters the domestic water supply from deterioration of galvanized iron and dezincification of brass. In such cases lead and cadmium also may be present because they are impurities of the zinc used in galvanizing. Zinc in water also may result from industrial waste pollution.

2. Selection of Method

The atomic absorption spectrometric methods (3111B and C) and inductively coupled plasma methods (3120 and 3125) are preferred. The zincon method (B), suitable for analysis of both potable and polluted waters, may be used if instrumentation for the preferred methods is not available.

3. Sampling and Storage

See Section 3010B.2 for sample handling and storage.

3500-Zn B. Zincon Method

1. General Discussion

a. Principle: Zinc forms a blue complex with 2-carboxy-2'-hydroxy-5'-sulfoformazyl benzene (zincon) in a solution buffered to pH 9.0. Other heavy metals likewise form colored complexes with zincon. Cyanide is added to complex zinc and heavy metals. Cyclohexanone is added to free zinc selectively from its cyanide complex so that it can be complexed with zincon to form a blue color. Sodium ascorbate reduces manganese interference. The developed color is stable except in the presence of copper (see table below).

b. Interferences: The following ions interfere at concentrations exceeding those listed:

Ion	mg/L	Ion	mg/L
Cd ²⁺	1	Cr ³⁺	10
Al ³⁺	5	Ni ²⁺	20
Mn ²⁺	5	Cu ²⁺	30
Fe ³⁺	7	Co ²⁺	30
Fe ²⁺	9	CrO ₄ ²⁻	50

c. Minimum detectable concentration: 0.02 mg Zn/L .

* Approved by Standard Methods Committee, 1997.
Joint Task Group: 20th Edition—See 3500-A1.

2. Apparatus

a. *Colorimetric equipment:* One of the following is required:

1) *Spectrophotometer,* for measurements at 620 nm, providing a light path of 1 cm or longer.

2) *Filter photometer,* providing a light path of 1 cm or longer and equipped with a red filter having maximum transmittance near 620 nm. Deviation from Beer's Law occurs when the filter band pass exceeds 20 nm.

b. *Graduated cylinders,* 50-mL, with ground-glass stoppers, Class B or better.

c. *Erlenmeyer flasks,* 50-mL.

d. *Filtration apparatus:* 0.45- μ m filters and filter holders.

3. Reagents

a. *Metal-free water:* See Section 3111B.3c. Use water for rinsing apparatus and preparing solutions and dilutions.

b. *Stock zinc solution:* Dissolve 1000 mg (1.000 g) zinc metal in 10 mL 1 + 1 HNO₃. Dilute and boil to expel oxides of nitrogen. Dilute to 1000 mL; 1.00 mL = 1.00 mg Zn.

c. *Standard zinc solution:* Dilute 10.00 mL stock zinc solution to 1000 mL; 1.00 mL = 10.00 μ g Zn.

d. *Sodium ascorbate,* fine granular powder, USP.

e. *Potassium cyanide solution:* Dissolve 1.00 g KCN in approximately 50 mL water and dilute to 100 mL. CAUTION: Potassium cyanide is a deadly poison. Avoid skin contact or inhalation of vapors. Do not pipet by mouth or bring in contact with acids.

f. *Buffer solution,* pH 9.0: Dissolve 8.4 g NaOH pellets in about 500 mL water. Add 31.0 g H₃BO₃ and swirl or stir to dissolve. Dilute to 1000 mL with water and mix thoroughly.

g. *Zincon reagent:* Dissolve 100 mg zincon (2-carboxy-2'-hydroxy-5'-sulfoformazyl benzene) in 100 mL methanol. Because zincon dissolves slowly, stir and/or let stand overnight.

h. *Cyclohexanone,* purified.

i. *Hydrochloric acid,* HCl, conc and 1N.

j. *Sodium hydroxide,* NaOH, 6N and 1N.

4. Procedure

a. *Preparation of colorimetric standards:* Accurately deliver 0, 0.5, 1.0, 3.0, 5.0, 10.0, and 14.0 mL standard zinc solution to a series of 50-mL graduated mixing cylinders. Dilute each to 20.0 mL to yield solutions containing 0, 0.25, 0.5, 1.5, 2.5, 5.0, and 7.0 mg Zn/L, respectively. (Lower-range standards may be prepared to extend the quantitation range. Longer optical path cells can be used. Verify linearity of response in this lower concentration range.) Add the following to each solution in sequence, mixing thoroughly after each addition: 0.5 g sodium

ascorbate, 5.0 mL buffer solution, 2.0 mL KCN solution, and 3.0 mL zincon solution. Pipet 20.0 mL of the solution into a clean 50-mL erlenmeyer flask. Reserve remaining solution to zero the instrument. Add 1.0 mL cyclohexanone to the erlenmeyer flask. Swirl for 10 s and note time. Transfer portions of both solutions to clean sample cells. Use solution without cyclohexanone to zero colorimeter. Read and record absorbance for solution with cyclohexanone after 1 min. The calibration curve does not pass through zero because of the color enhancement effect of cyclohexanone on zincon.

b. *Treatment of samples:* To determine readily acid-extractable total zinc, add 1 mL conc HCl to 50 mL sample and mix thoroughly. Filter and adjust to pH 7. To determine dissolved zinc, filter sample through a 0.45- μ m membrane filter. Adjust to pH 7 with 1N NaOH or 1N HCl if necessary after filtering.

c. *Sample analysis:* Cool samples to less than 30°C if necessary. Analyze 20.0 mL of prepared sample as described in ¶ 4a above, beginning with "Add the following to each solution . . ." If the zinc concentration exceeds 7 mg Zn/L prepare a sample dilution and analyze a 20.0-mL portion.

5. Calculation

Read zinc concentration (in milligrams per liter) directly from the calibration curve.

6. Precision and Bias

A synthetic sample containing 650 μ g Zn/L, 500 μ g Al/L, 50 μ g Cd/L, 110 μ g Cr/L, 470 μ g Cu/L, 300 μ g Fe/L, 70 μ g Pb/L, 120 μ g Mn/L, and 150 μ g Ag/L in doubly demineralized water was analyzed in a single laboratory. A series of 10 replicates gave a relative standard deviation of 0.96% and a relative error of 0.15%. A wastewater sample from an industry in Standard Industrial Classification (SIC) No. 3333, primary smelting and refining of zinc, was analyzed by 10 different persons. The mean zinc concentration was 3.36 mg Zn/L and the relative standard deviation was 1.7%. The relative error compared to results from an atomic absorption analysis of the same sample was -1.0%.

7. Bibliography

- PLATTE, J.A. & V.M. MARCY. 1959. Photometric determination of zinc with zincon. *Anal. Chem.* 31:1226.
- RUSH, R.M. & J.H. YOE. 1954. Colorimetric determination of zinc and copper with 2-carboxy-2'-hydroxy-5'-sulfoformazyl-benzene. *Anal. Chem.* 26:1345.
- MILLER, D.G. 1979. Colorimetric determination of zinc with zincon and cyclohexanone. *J. Water Pollut. Control Fed.* 51:2402.
- PANDE, S.P. 1980. Study on the determination of zinc in drinking water. *J. IWWA XII* (3):275.

